



Peace Airshed Zone Association

Ambient Air Monitoring Network Summary

**Ambient Air Quality Monitoring Program
Monthly Report
October 2018**

November 30, 2018

Alberta Environment and Parks

11th Floor, Oxbridge Place
9820-106 Street
Edmonton Alberta T5K 2J6

RE: Peace Airshed Zone Association (PAZA) – October 2018 Ambient Air Report

Enclosed is the PAZA Ambient Monitoring Network Report for the month of October 2018.

The Person Responsible is Elizabeth Bell, Executive Director.

This report is submitted by PAZA on behalf of the industrial member companies to satisfy the requirements of the following facility Operating Approvals:

Company	Facility	LSD	EPEA Approval Number
Advantage Oil & Gas Ltd.	Glacier	05-02-076-13-W6	262479-00-00
Alberta Power (2000) Ltd. (an ATCO company)	Sturgeon	SW-06-069-21-W5	10283-02-02
ATCO Power Canada	Poplar Hill	11-19-073-08-W6	67774-01-01
ATCO Power Canada	Valleyview	SW-06-069-21-W5	147709-01-01
AltaGas Ltd.	Pouce Coupe	03-03-081-13-W6	247673-00-00
	Ante Creek	02-26-068-25-W5	266694-00-00
	Gordondale	16-31-78-11-W6M	287474-00-00
Apache Canada Ltd.	House Mountain	01-08-070-10-W5	10137-02-02
Birchcliff Energy Ltd.	Pouce Coupe	03-22-078-12-W6	252529-00-00
Canadian Natural Resources Limited	Bonanza	11-25-081-11-W6	00000029-01-00
	Progress/Gordondale	01-01-077-10-W6	00010036-02-00
	Gold Creek	13-26-067-05-W6	00010446-02-00

Company	Facility	LSD	EPEA Approval Number
	Teepee Creek	SE-2-074-04-W6	00001635-02-00
	Sturgeon/Valleyview	02-02-069-22-W5	1633-02-00
Canfor Forest Products	Grande Prairie	SW-23-071-06-W6	152645-01-00
Conocophillips Canada Energy Partnership	Wembley	06-19-073-08-W6	00000212-01-00
Devon Canada	NW Belloy (Dunvegan)	16-36-079-03-W6	00009810-02-00
	Eaglesham (South)	02-14-077-25-W5	00047669-01-00
	North Normanville	03-36-079-23-W5	00047455-01-00
	West Culp	05-34-078-25-W6	00136284-00-00
	Cecil	08-15-084-08-W6	00010032-02-00
Encana Corporation	Sexsmith	04-08-075-07-W6	00010002-01-00
Enerplus Resources	Pouce Coupe	SW-06-069-21-W5	1464-02-03
Exshaw Oil Corporation	Spirit River	03-10-077-07-W6	344521-00-00
Grande Prairie Generation Inc.	Northern Prairie Power Project	04-19-073-08-W6	00238762-00-00
Long Run Exploration	Eaglesham	01-25-076-01-W6	00241532-00-00
	Kakut	14-12-075-03-W6	00248469-00-00
	Donnelly	06-01-077-21-W5	00000087-02-00
	Puskwaskau	03-26-074-01-W6	00017524-01-00
Longview Oil Corp.	Sunset House	06-22-070-20-W5	138884-01-00
Penn West Petroleum Ltd.	Tangent	13-29-080-23-W5	00001746-02-00

Company	Facility	LSD	EPEA Approval Number
	Pouce Coupe	16-07-078-11-W6	00000614-01-00
Petrus Resources	Rycroft	08-25-077-06-W6	11351-02-00
	Spirit River	08-34-077-06-W6	11096-02-00
Spectra Energy Midstream Corporation	Fourth Creek	16-11-082-09-W6	00000263-01-00
	Gordondale	11-26-079-09-W6	00011495-01-01
	Pouce Coupe/Bonanza	3-23-080-13-W6	00070203-01-01
Suncor Energy Inc.	Progress	07-22-078-09-W6	00011428-02-00
TAQA North Ltd.	Valhalla	13-21-076-09-W6	00017620-01-00
Veresen Energy	Hythe Brainard	11-18-074-12-W6	00010910-02-00
Weyerhaeuser Canada	Grande Prairie Pulp and Wood Plant	01-14-070-05-W6	00000113-02-00

Included in this report is a summary of the monthly passive monitoring and continuous monitoring programs, detailed hourly average reports and multipoint calibration reports of all instruments. Operational summaries can be found on the "Monthly Continuous Data Summary" and "Continuous Network Equipment Summary" pages of the report.

Continuous Monitoring: Six (6) Stations including Henry Pirker (Grande Prairie), Evergreen Park, Smoky Heights, Beaverlodge, Valleyview, and Rycroft-Portable.

During the month of October, the following events were noted:

Beaverlodge Station:

- ◆ The measured ambient air quality was within the AAAQO for the Beaverlodge station.
- ◆ The RH sensor malfunctioned, returning an uptime of 35.4%. Faulty wiring is suspected and is under investigation. All other analyzers and meteorological sensors at the Beaverlodge station had an operational uptime greater than 90% for the month of October.

Evergreen Park Station:

- ◆ The measured ambient air quality was within the AAAQO for the Evergreen Park station.
- ◆ All analyzers and meteorological sensors at the Evergreen Park station had an operational uptime greater than 90% for the month of October.

Henry Pirker Station:

- ◆ The measured ambient air quality was within the AAAQO for the Henry Pirker station.
- ◆ One (1) 1-hour exceedance of the AAAQO for PM_{2.5} was recorded in October.

Station Name	Date	Start Time	End Time	PM _{2.5} Concentration	Reference Number
		(MST)	(MST)	(µg/m ³)	
Henry Pirker	22-Oct-18	19:00	20:00	82.3	342093

Smoky Heights Station:

- ◆ The measured ambient air quality was within the AAAQO for the Smoky Heights station.
- ◆ All analyzers and meteorological sensors at the Smoky Heights station had an operational uptime greater than 90% for the month of October.

Valleyview Station:

- ◆ The measured ambient air quality was within the AAAQO for the Valleyview station.
- ◆ All analyzers and meteorological sensors at the Valleyview station had an operational uptime greater than 90% for the month of October.

Donnelly Station:

- ◆ The measured ambient air quality was within the AAAQO for the Donnelly station.
- ◆ All analyzers and meteorological sensors at the Donnelly station had an operational uptime greater than 90% for the month of October.

Rycroft-Portable Station:

- ◆ The measured ambient air quality was within the AAAQO for the Rycroft-Portable station.
- ◆ All analyzers and meteorological sensors at the Rycroft-Portable station had an operational uptime greater than 90% for the month of October.
- ◆ The station was removed from service on October 27th and relocated to the Wembley area. This report marks the end of monitoring at this site.

Passive Monitoring - 27 Stations throughout the PAZA zone:

There were five duplicate sites sampled in the month of October: Eaglesham and Little Smoky (SO₂), Bay Tree (O₃), Wanham and Little Smoky (NO₂), and Griouxville 3 (H₂S). The passive sample analyses were performed by MAXXAM Analytics Inc.

A summary of the passive data collected are reported as follows:

- Monthly average concentrations for SO₂ passives ranged from 0.1 ppb to 0.4 ppb, with a mean of 0.2 ppb.
- Monthly average concentrations for NO₂ passives ranged from 0.2 ppb to 8.3 ppb, with a mean of 1.9 ppb.
- Monthly average concentrations for O₃ passives ranged from 25.6 ppb to 41.2 ppb, with a mean of 30.9 ppb.
- Monthly average concentrations for H₂S passives ranged from 0.1 to 0.2 ppb, with a mean of 0.2 ppb.


If you have any questions or concerns, please contact Mandeep Dhaliwal at 403.608.9697 or the PAZA office at 780.833.4343.

On Behalf of the Peace Airshed Zone Association



Mandeep Dhaliwal, B.Sc., P.Chem.
Technical Manager

Location of PAZA Continuous Monitoring Stations



PAZA Monthly Continuous Data Summary

Oct-2018 Peace Airshed Zone Association									Maximum Recorded Values					
Parameter (units)	Objectives			Station	Monthly Average	Exceedence			1-hr		24-hr / 8-hr		Operational Time (%)	Calibration Date
	1-hr	24-hr	30-day			1-hr	24-hr	30-day	Conc	Day	Conc	Day		
SO ₂ (ppb)	172	48	8	Beaverlodge	0.1	0	0	0	2.3	Oct-03 13:00	0.4	Oct-22	100.0	Oct-14
NO (ppb)				Beaverlodge	0.4	0	0	-	8.0	Oct-31 10:00	1.5	Oct-23	100.0	Oct-13
NO ₂ (ppb)	159	106		Beaverlodge	4.0	0	0	-	22.9	Oct-24 04:00	13.1	Oct-23	100.0	Oct-13
NO _x (ppb)				Beaverlodge	4.4	0	0	-	25.7	Oct-23 09:00	14.6	Oct-23	100.0	Oct-13
O ₃ (ppb)	82			Beaverlodge	24.6	0	0	-	52.4	Oct-17 04:00	44.8	Oct-17	100.0	Oct-13
O ₃ (ppb) - 8-hr				Beaverlodge		0					47.0	Oct-17		-
PM _{2.5} (µg/m3)	80	30		Beaverlodge	4.8	0	0	-	14.2	Oct-24 00:00	9.3	Oct-24	100.0	Aug-08
RH (%)				Beaverlodge	56.4	-	-	-	94.8	Oct-06 20:00	67.6	Oct-25	35.4	-
Temp (°C)				Beaverlodge	4.0	-	-	-	20.7	Oct-16 15:00	14.5	Oct-17	100.0	-
WSPD s (km/hr)				Beaverlodge	10.6	-	-	-	49.9	Oct-20 12:00	26.5	Oct-20	100.0	01/06/2017
WSPD v (km/hr)				Beaverlodge	3.2	-	-	-	49.9	Oct-20 12:00	25.6	Oct-20	100.0	01/06/2017
WDIR				Beaverlodge	W	-	-	-	-	-	-	-	100.0	01/06/2017

Parameter	Analyzer		Notes
	Make	Model	
SO ₂	TECO	43I	
NO/NO ₂ /NO _x	TECO	42I	
O ₃	TECO	49I	
PM _{2.5}	SHARP	5030	
ET/RH			RH probe failed to return valid data, faulty wiring is suspected
WS/WD	MetOne	50.5	

Oct-2018 Peace Airshed Zone Association									Maximum Recorded Values					
Parameter (units)	Objectives			Station	Monthly Average	Exceedence			1-hr		24-hr / 8-hr		Operational Time (%)	Calibration Date
	1-hr	24-hr	30-day			1-hr	24-hr	30-day	Conc	Day	Conc	Day		
SO ₂ (ppb)	172	48	8	Evergreen Park	0.2	0	0	0	5.2	Oct-22 13:00	0.7	Oct-22	100.0	Oct-19
TRS (ppb)				Evergreen Park	0.4	-	-	-	1.9	Oct-11 17:00	0.7	Oct-22	100.0	Oct-19
PM _{2.5} (µg/m3)	80	30		Evergreen Park	5.0	0	0	-	56.5	Oct-20 10:00	15.4	Oct-20	100.0	Oct-19
RH (%)				Evergreen Park	68.4	-	-	-	98.8	Oct-12 07:00	95.3	Oct-07	100.0	-
Temp (°C)				Evergreen Park	3.1	-	-	-	21.4	Oct-16 15:00	12.0	Oct-15	100.0	-
WSPD s (km/hr)				Evergreen Park	9.6	-	-	-	55.0	Oct-20 13:00	26.7	Oct-20	100.0	-
WSPD v (km/hr)				Evergreen Park	4.7	-	-	-	54.3	Oct-20 13:00	26.0	Oct-20	100.0	-
WDIR				Evergreen Park	WSW	-	-	-	-	-	-	-	100.0	-

Parameter	Analyzer		Notes
	Make	Model	
SO ₂	TECO	43I	
TRS	TECO	43I	
PM _{2.5}	SHARP	5030	
ET/RH			
WS/WD	Gil	MetPak	

Oct-2018 Peace Airshed Zone Association									Maximum Recorded Values					
Parameter (units)	Objectives			Station	Monthly Average	Exceedence			1-hr		24-hr / 8-hr		Operational Time (%)	Calibration Date
	1-hr	24-hr	30-day			1-hr	24-hr	30-day	Conc	Day	Conc	Day		
SO ₂ (ppb)	172	48	8	Henry Pirker	0.3	0	0	0	2.4	Oct-22 15:00	0.7	Oct-22	100.0	Oct-05
TRS (ppb)				Henry Pirker	0.2	-	-	-	1.6	Oct-17 10:00	0.6	Oct-22	100.0	Oct-09
NO (ppb)				Henry Pirker	8.1	0	0	-	152.7	Oct-17 09:00	34.5	Oct-22	100.0	Oct-05
NO ₂ (ppb)	159	106		Henry Pirker	10.9	0	0	-	60.1	Oct-17 09:00	22.6	Oct-22	100.0	Oct-05
NO _x (ppb)				Henry Pirker	19.2	0	0	-	213.5	Oct-17 09:00	57.5	Oct-22	100.0	Oct-05
O ₃ (ppb)	82			Henry Pirker	17.5	0	0	-	49.1	Oct-16 14:00	28.2	Oct-18	100.0	Oct-05
O ₃ (ppb) - 8-hr				Henry Pirker		0					40.1	Oct-15		-
CO (ppm)	13			Henry Pirker	0.21	0	0	-	1.4	Oct-17 09:00	0.4	Oct-24	100.0	Oct-09
CO (ppm) - 8-hr	5			Henry Pirker		0					0.7	Oct-17		-
THC (ppm)				Henry Pirker	2.0	-	-	-	3.1	Oct-23 05:00	2.4	Oct-23	100.0	Oct-05
CH ₄ (ppm)				Henry Pirker	2.0	-	-	-	2.9	Oct-22 05:00	2.3	Oct-23	100.0	Oct-05
NMHC (ppm)				Henry Pirker	0.0	-	-	-	0.4	Oct-23 04:00	0.1	Oct-23	100.0	Oct-05
PM _{2.5} (µg/m3)	80	30		Henry Pirker	6.6	1	0	-	82.3	Oct-22 20:00	16.0	Oct-22	100.0	Jun-11
RH (%)				Henry Pirker	62.3	-	-	-	89.3	Oct-12 06:00	84.2	Oct-12	100.0	-
SR (W/m ²)				Henry Pirker	71.0	-	-	-	549.0	Oct-02 13:00	132.9	Oct-02	100.0	-
Temp (°C)				Henry Pirker	3.8	-	-	-	21.3	Oct-16 15:00	11.9	Oct-17	100.0	-
WSPD s (km/hr)				Henry Pirker	6.4	-	-	-	33.8	Oct-20 13:00	17.5	Oct-20	100.0	01/07/2017
WSPD v (km/hr)				Henry Pirker	2.7	-	-	-	33.8	Oct-20 13:00	17.0	Oct-20	100.0	01/07/2017
WDIR				Henry Pirker	W	-	-	-	-	-	-	-	100.0	01/07/2017

Parameter	Analyzer		Notes
	Make	Model	
SO ₂	TECO	43I	
TRS	TECO	450I	
NO/NO ₂ /NO _x	TECO	42I	
O ₃	TECO	49I	
CO	TECO	48I-TLE	
THC/CH ₄ /NMHC	TECO	55I	
PM _{2.5}	SHARP	5030	
ET/RH			
WS/WD	MetOne	50.5	

Oct-2018Peace Airshed Zone Association									Maximum Recorded Values					
									1-hr		24-hr / 8-hr			
Parameter (units)	Objectives			Station	Monthly Average	Exceedence			Conc	Day	Conc	Day	Operational Time (%)	Calibration Date
	1-hr	24-hr	30-day			1-hr	24-hr	30-day						
SO ₂ (ppb)	172	48	8	Smoky Heights	0.3	0	0	0	3.4	Oct-03 00:00	0.9	Oct-16	100.0	Oct-19
TRS (ppb)				Smoky Heights	0.5	-	-	-	2.3	Oct-13 19:00	0.9	Oct-29	100.0	Oct-19
PM _{2.5} (µg/m3)	80	30		Smoky Heights	4.8	0	0	-	44.5	Oct-22 20:00	11.3	Oct-24	100.0	Aug-29
Temp (°C)				Smoky Heights	2.9	-	-	-	20.4	Oct-16 16:00	11.7	Oct-17	100.0	-
WSPD s (km/hr)				Smoky Heights	11.8	-	-	-	50.6	Oct-20 12:00	25.4	Oct-20	100.0	01/04/2017
WSPD v (km/hr)				Smoky Heights	5.9	-	-	-	50.6	Oct-20 12:00	24.9	Oct-20	100.0	01/04/2017
WDIR				Smoky Heights	WSW	-	-	-	-	-	-	-	100.0	01/04/2017

Parameter	Analyzer		Notes
	Make	Model	
SO ₂	TECO	43I	
TRS	TECO	43I	
PM _{2.5}	SHARP	5030	
ET/RH			
WS/WD	MetOne	50.5	

Oct-2018Peace Airshed Zone Association									Maximum Recorded Values					
									1-hr			24-hr / 8-hr		
Parameter (units)	Objectives			Station	Monthly Average	Exceedence			Conc	Day	Conc	Day	Operational Time (%)	Calibration Date
	1-hr	24-hr	30-day			1-hr	24-hr	30-day						
SO ₂ (ppb)	172	48	8	Valleyview	0.2	0	0	0	8.6	Oct-30 11:00	1.4	Oct-30	100.0	Oct-25
H ₂ S (ppb)	10	3		Valleyview	0.1	0	0	-	1.3	Oct-04 02:00	0.3	Oct-04	100.0	Oct-25
RH (%)				Valleyview	73.7	-	-	-	100.0	Oct-07 02:00	98.4	Oct-29	100.0	-
Temp (°C)				Valleyview	6.1	-	-	-	24.2	Oct-17 15:00	13.5	Oct-15	100.0	-
WSPD s (km/hr)				Valleyview	4.1	-	-	-	24.5	Oct-20 13:00	11.4	Oct-20	100.0	-
WSPD v (km/hr)				Valleyview	1.3	-	-	-	24.5	Oct-20 13:00	9.8	Oct-20	100.0	-
WDIR				Valleyview	W	-	-	-	-	-	-	-	100.0	-

Parameter	Analyzer		Notes
	Make	Model	
SO ₂	TECO	43I	
H ₂ S	TECO	450I	
ET/RH			
WS/WD	Gil	MetPak	

Oct-2018Peace Airshed Zone Association									Maximum Recorded Values					
									1-hr		24-hr / 8-hr			
Parameter (units)	Objectives			Station	Monthly Average	Exceedence			Conc	Day	Conc	Day	Operational Time (%)	Calibration Date
	1-hr	24-hr	30-day			1-hr	24-hr	30-day						
SO ₂ (ppb)	172	48	8	Donnelly	0.1	0	0	0	1.5	Oct-17 11:00	0.4	Oct-16	100.0	Oct-24
H ₂ S (ppb)	10	3		Donnelly	0.2	0	0	-	1.9	Oct-12 06:00	0.5	Oct-13	100.0	Oct-24
Temp (°C)				Donnelly	3.4	-	-	-	20.3	Oct-17 14:00	13.0	Oct-17	100.0	-
WSPD s (km/hr)				Donnelly	11.1	-	-	-	45.0	Oct-20 14:00	25.3	Oct-20	100.0	01/06/2017
WSPD v (km/hr)				Donnelly	10.9	-	-	-	44.3	Oct-20 14:00	25.1	Oct-20	100.0	01/06/2017
WDIR				Donnelly	SW	-	-	-	-	-	-	-	100.0	01/06/2017

Parameter	Analyzer		Notes
	Make	Model	
SO ₂	TECO	43I	
H ₂ S	TECO	450I	
ET/RH			
WS/WD	RMYoung		

Oct-2018 Peace Airshed Zone Association									Maximum Recorded Values					
									1-hr		24-hr / 8-hr			
Parameter (units)	Objectives			Station	Monthly Average	Exceedence			Conc	Day	Conc	Day	Operational Time (%)	Calibration Date
	1-hr	24-hr	30-day			1-hr	24-hr	30-day						
SO ₂ (ppb)	172	48	8	Rycroft-Portable	0.2	0	0	0	1.8	Oct-04 11:00	0.5	Oct-25	100.0	Oct-23
TRS (ppb)				Rycroft-Portable	0.2	-	-	-	3.7	Oct-04 02:00	0.8	Oct-13	100.0	Oct-17
NO (ppb)				Rycroft-Portable	0.5	0	0	-	22.7	Oct-17 21:00	2.1	Oct-23	100.0	Oct-23
NO ₂ (ppb)	159	106		Rycroft-Portable	2.5	0	0	-	14.4	Oct-17 21:00	8.0	Oct-23	100.0	Oct-23
NO _x (ppb)				Rycroft-Portable	3.0	0	0	-	37.2	Oct-17 21:00	10.2	Oct-23	100.0	Oct-23
O ₃ (ppb)	82			Rycroft-Portable	23.2	0	0	-	48.0	Oct-16 14:00	41.4	Oct-17	100.0	Oct-23
O ₃ (ppb) - 8-hr				Rycroft-Portable		0					45.5	Oct-17		-
THC (ppm)				Rycroft-Portable	1.9	-	-	-	2.4	Oct-01 07:00	2.1	Oct-24	98.0	Oct-23
CH ₄ (ppm)				Rycroft-Portable	1.9	-	-	-	2.4	Oct-01 07:00	2.1	Oct-24	98.0	Oct-23
NMHC (ppm)				Rycroft-Portable	0.0	-	-	-	0.0	Oct-04 12:00	0.0	Oct-04	98.0	Oct-23
PM _{2.5} (µg/m3)	80	30		Rycroft-Portable	2.8	0	0	-	19.4	Oct-23 21:00	7.8	Oct-24	100.0	Oct-23
Temp (°C)				Rycroft-Portable	3.6	-	-	-	21.2	Oct-16 15:00	14.5	Oct-17	100.0	-
WSPD s (km/hr)				Rycroft-Portable	9.6	-	-	-	42.9	Oct-20 15:00	19.3	Oct-20	99.5	01/03/2017
WSPD v (km/hr)				Rycroft-Portable	9.4	-	-	-	42.9	Oct-20 15:00	19.2	Oct-20	99.5	01/03/2017
WDIR				Rycroft-Portable	S	-	-	-	-	-	-	-	99.5	01/03/2017

Parameter	Analyzer		Notes
	Make	Model	
SO ₂	TECO	43I	Station was removed from service and relocated October 27
TRS	TECO	43I	
NO/NO ₂ /NO _x	TECO	42I	
O ₃	TECO	49I	
THC/CH ₄ /NMHC	TECO	55I	Hydrogen tank expired Oct 16, replaced that day. Nitrogen tank expired Oct 23, replaced and analyzer calibrated
PM _{2.5}		TEOM	
ET/RH			
WS/WD	MetOne	50.5	Wind system returned invalid values Oct 12

PAZA

Beaverlodge Station

Monthly Summary Tables, Graphs and
Roses

Hourly Averages


Sulphur Dioxide (SO₂) - ppb

Beaverlodge - October 2018

Number of Exceedences (AAAQO): Maximum Value: 2.3 ppb on Oct 3 13:00																			1-hr: 0 24-hr: 0 Maximum Daily Average: 0.4 ppb on Oct 22					Hours in Service: 744 Hours of Data: 710 Hours of Missing Data: 34 Hours of Calibration: 34 Percent Operational Time: 100.0				
Minimum Value: 0 ppb on Oct 20 15:00 Maximum Diurnal Average: 0.2 ppb at hour 13 Monthly Average: 0.14 ppb																			Minimum Daily Average: 0.0 ppb on Oct 31 Minimum Diurnal Average: 0.1 ppb at hour 8 Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.1 Q ₃ = 0.2 P ₉₀ = 0.3 P ₉₉ = 0.9									
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum		
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24				
1-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
2-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3		
3-Oct	0	0	0	0	A	0	0	0	0	0	0	1	2	0	0	0	0	0	0	1	0	0	0	0	0.3	2.3		
4-Oct	1	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0.2	0.7		
5-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	1	0	0	1	0	0	0	0	0	0	0	0.3	0.9		
6-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0.3	0.9		
7-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1		
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
9-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
10-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
11-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4		
12-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
13-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3		
14-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	C	C	C	0	0	0	0	0	0	0	0	0.1	0.2		
15-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0.2	1.1		
16-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1		
17-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1		
18-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.5		
19-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0.2	0.5		
20-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
21-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	1	0.2	1.3		
22-Oct	1	0	0	0	A	0	0	0	0	0	0	0	0	0	1	1	0	0	1	1	1	1	0	0	0.4	1.1		
23-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5		
24-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4		
25-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3		
26-Oct	0	0	0	0	A	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0.2	0.6		
27-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1		
28-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3		
29-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
30-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1		
31-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1		
																								Diurnal Average				
																								Diurnal Maximum				
C - Calibration A - Automated Daily Zero Span																												
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 172 ppb 24-hr 48 ppb 30-day 11 ppb																												

Hourly Averages

Sulphur Dioxide (SO₂) - ppb
Beaverlodge - October 2018



Hourly Maximums


Sulphur Dioxide (SO₂) - ppb

Beaverlodge - October 2018

Maximum Value: 6.6 ppb on Oct 3 13:00																			Maximum Daily Average: 0.8 ppb on Oct 3						Hours in Service: 744	
Minimum Value: 0 ppb on Oct 20 17:00																			Minimum Daily Average: 0.1 ppb on Oct 31						Hours of Data: 710	
Maximum Diurnal Average: 0.5 ppb at hour 13																			Minimum Diurnal Average: 0.2 ppb at hour 7						Hours of Missing Data: 34	
Monthly Average: 0.28 ppb																			Percentiles: P ₁ = 0.1 P ₁₀ = 0.1 Q ₁ = 0.1 Median = 0.2 Q ₃ = 0.3 P ₉₀ = 0.5 P ₉₉ = 1.4						Hours of Calibration: 34	
																			Percent Operational Time: 100.0							
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4
2-Oct	0	0	0	0	A	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.6
3-Oct	0	0	0	0	A	0	0	0	0	0	0	4	7	0	0	0	0	0	1	1	0	0	1	1	0.8	6.6
4-Oct	1	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	1	0	0	0.4	1.3
5-Oct	0	0	0	0	A	0	0	0	0	0	0	1	1	1	1	0	4	1	0	0	0	0	0	0	0.5	3.6
6-Oct	0	0	0	0	A	0	0	0	0	0	1	1	0	0	1	1	1	0	0	0	0	0	0	0	0.4	1.4
7-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.2
9-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
10-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
11-Oct	0	0	0	0	A	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0.3	0.5
12-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
13-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4
14-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	C	C	C	0	0	0	0	0	0	0	0	0.2	0.3
15-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	1	0	0.3	1.3
16-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
17-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
18-Oct	1	1	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	1.0
19-Oct	0	0	0	0	A	0	0	0	0	0	0	0	1	1	1	0	1	1	0	0	0	1	1	0	0.4	1.2
20-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4
21-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	3	3	0.5	2.9
22-Oct	2	1	0	0	A	0	0	0	0	0	0	0	0	0	1	1	0	0	1	1	1	2	1	0	0.7	2.1
23-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	1	1	1	1	1	1	1	0	0.4	1.2
24-Oct	0	0	0	0	A	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0.3	0.6
25-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4
26-Oct	0	0	0	0	A	0	0	0	0	0	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0.4	1.2
27-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.2
28-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4
29-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
30-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
31-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
0.3 0.2 0.2 0.2 -- 0.2 0.2 0.2 0.2 0.3 0.3 0.4 0.5 0.3 0.3 0.3 0.4 0.3 0.3 0.3 0.3 0.3 0.4 0.3																									Diurnal Average	
2.1 1.0 0.5 0.5 -- 0.3 0.4 0.4 0.4 0.6 0.6 4.0 6.6 0.7 1.1 1.4 3.6 1.2 0.9 1.3 1.3 1.7 2.8 2.9																									Diurnal Maximum	
C - Calibration					A - Automated Daily Zero Span																					


Hourly Maximums

Sulphur Dioxide (SO₂) - ppb
Beaverlodge - October 2018




Pollutant Rose

Sulphur Dioxide (SO₂) - ppb
Beaverlodge - October 2018




Pollutant Classes (ppb)



Span Responses

Sulphur Dioxide (SO₂)
Beaverlodge - October 2018



Hourly Averages


Nitrogen Dioxide (NO₂) - ppb

Beaverlodge - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 22.9 ppb on Oct 24 04:00 Maximum Daily Average: 13.1 ppb on Oct 23																				Hours in Service: 744 Hours of Data: 707 Hours of Missing Data: 37 Hours of Calibration: 37 Percent Operational Time: 100.0						
Minimum Value: 0 ppb on Oct 15 01:00 Maximum Diurnal Average: 6.1 ppb at hour 6 Monthly Average: 4.02 ppb										Minimum Daily Average: 1.2 ppb on Oct 2 Minimum Diurnal Average: 2.3 ppb at hour 14 Percentiles: P ₁ = 0.4 P ₁₀ = 0.8 Q ₁ = 1.5 Median = 2.9 Q ₃ = 5.3 P ₉₀ = 8.4 P ₉₉ = 16.8																
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	2	2	2	3	A	7	7	6	2	2	2	1	1	1	2	2	2	3	1	1	1	1	1	2.2	6.6	
2-Oct	0	1	0	1	A	4	2	2	2	2	1	1	1	1	0	0	1	1	1	2	1	1	1	2	1.2	3.9
3-Oct	2	3	3	3	A	6	5	7	4	2	2	1	1	0	0	1	1	4	5	4	4	9	6	6	3.6	9.4
4-Oct	6	9	7	5	A	8	7	7	6	5	3	2	1	1	1	1	1	2	4	4	7	4	4	4	4.1	8.7
5-Oct	4	3	5	11	A	9	7	8	5	5	4	1	2	2	2	3	3	4	6	6	7	6	11	11	5.5	11.2
6-Oct	9	8	5	4	A	7	6	6	7	6	3	2	1	0	3	2	3	6	6	4	4	3	2	1	4.3	9.5
7-Oct	2	2	2	2	A	5	3	2	2	2	2	2	2	3	3	3	4	3	5	4	3	2	2	2	2.6	5.3
8-Oct	2	2	2	2	A	4	2	2	2	2	2	2	2	2	1	1	2	1	2	2	1	1	1	1	1.8	3.7
9-Oct	1	2	2	2	A	4	3	3	4	4	3	3	3	3	3	2	2	3	3	3	3	3	3	4	3.0	4.3
10-Oct	4	4	4	5	A	6	6	5	6	5	5	5	5	3	4	4	5	3	4	6	6	5	7	7	5.0	6.5
11-Oct	6	6	6	7	A	7	8	7	5	4	4	3	1	1	1	1	1	2	2	4	3	1	1	1	3.5	7.8
12-Oct	3	3	3	4	A	9	5	3	2	3	2	1	1	1	1	1	1	2	5	6	3	5	11	5	3.5	11.3
13-Oct	5	2	1	1	A	3	3	4	3	2	C	C	C	C	C	C	1	3	3	2	2	2	2	2	--	5.5
14-Oct	2	2	1	1	A	5	4	3	2	2	2	1	1	0	1	1	1	1	1	1	0	0	1	1	1.4	5.0
15-Oct	0	1	1	1	A	3	1	2	2	1	1	1	1	1	1	1	1	1	2	2	4	5	3	2	1.6	5.1
16-Oct	1	1	1	1	A	4	3	4	3	2	3	1	1	1	1	1	1	1	2	4	4	2	1	1	1.9	4.2
17-Oct	2	1	1	1	A	4	5	5	3	2	3	2	1	2	2	3	1	2	2	3	2	2	1	1	2.2	5.1
18-Oct	2	1	1	1	A	5	3	11	5	2	1	1	1	1	1	1	1	3	2	5	6	5	3	3	2.6	10.8
19-Oct	2	3	2	1	A	7	4	12	10	9	8	6	3	3	4	4	6	5	4	4	3	4	5	6	5.0	12.0
20-Oct	7	6	3	0	A	3	2	1	1	1	4	14	11	6	2	2	1	0	1	1	1	1	3	1	3.1	13.9
21-Oct	1	1	2	2	A	4	4	3	3	5	4	3	1	1	1	1	2	3	5	4	5	5	6	7	3.2	6.8
22-Oct	8	7	8	8	A	13	10	16	11	6	4	3	3	3	5	5	8	13	16	14	9	12	13	17	9.2	16.8
23-Oct	18	13	14	12	A	16	16	14	19	11	8	8	7	6	7	5	8	15	17	14	16	15	20	23	13.1	22.8
24-Oct	19	16	20	23	A	16	8	11	11	8	8	13	10	3	3	2	5	7	7	5	4	7	5	4	9.4	22.9
25-Oct	3	4	5	3	A	6	8	11	11	9	5	4	3	3	2	1	2	1	2	1	2	3	2	2	3.9	10.9
26-Oct	1	1	1	1	A	4	2	5	5	7	14	13	15	8	11	10	6	8	9	9	2	0	0	0	5.8	15.3
27-Oct	0	0	0	0	A	2	1	1	1	1	1	1	1	1	0	1	2	4	3	6	3	2	1	2	1.5	5.5
28-Oct	2	2	4	4	A	4	4	3	5	6	6	6	6	6	5	6	6	7	7	7	10	10	8	6	5.7	10.2
29-Oct	5	4	4	4	A	6	5	5	5	6	4	2	1	1	1	1	2	7	7	3	3	2	3	3	3.7	6.9
30-Oct	3	4	3	3	A	3	2	2	3	2	1	1	1	1	1	1	1	1	1	2	5	4	3	4	2.2	5.1
31-Oct	3	3	3	2	A	5	5	10	11	14	9	4	6	6	5	4	6	8	8	7	10	5	4	2	6.1	13.7
4.1 3.7 3.7 3.9 -- 6.1 4.9 5.9 5.1 4.5 4.0 3.6 3.1 2.3 2.4 2.4 2.8 4.0 4.5 4.3 4.2 4.3 4.3 4.2 Diurnal Average																										
18.7 16.5 20.3 22.9 -- 16.4 16.4 15.7 19.0 13.7 14.4 13.9 15.3 7.8 11.3 10.3 8.3 14.9 16.7 13.8 15.7 14.7 19.5 22.8 Diurnal Maximum																										
C - Calibration A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 159 ppb 24-hr 106 ppb																										

Hourly Averages

Nitrogen Dioxide (NO₂) - ppb
Beaverlodge - October 2018



Hourly Maximums


Nitrogen Dioxide (NO₂) - ppb

Beaverlodge - October 2018

Maximum Value: 47.7 ppb on Oct 23 09:00																				Maximum Daily Average: 19.4 ppb on Oct 23					Hours in Service: 744	
Minimum Value: 1 ppb on Oct 15 01:00																				Minimum Daily Average: 2.3 ppb on Oct 8					Hours of Data: 707	
Maximum Diurnal Average: 9.6 ppb at hour 8																				Minimum Diurnal Average: 4.9 ppb at hour 17					Hours of Missing Data: 37	
Monthly Average: 6.77 ppb																				Percentiles: P ₁ = 0.8 P ₁₀ = 1.6 Q ₁ = 2.6 Median = 5.0 Q ₃ = 8.7 P ₉₀ = 14.7 P ₉₉ = 27.1					Hours of Calibration: 37	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	2	3	2	4	A	8	10	8	4	5	4	2	2	6	17	3	3	4	2	1	1	1	1	1	4.1	17.1
2-Oct	1	1	1	2	A	6	3	3	3	2	1	1	2	1	1	1	1	3	5	4	2	2	4	4	2.3	6.3
3-Oct	4	4	4	4	A	8	6	11	11	4	3	2	2	1	1	1	4	9	7	5	8	15	8	8	5.7	15.4
4-Oct	9	11	12	6	A	15	16	9	10	7	36	2	2	1	2	1	1	2	3	4	7	11	7	4	7.8	35.8
5-Oct	5	4	8	24	A	18	13	14	11	9	8	3	2	4	3	4	4	5	10	9	9	14	13	12	8.9	23.9
6-Oct	13	9	6	6	A	10	10	8	10	9	4	17	8	1	8	6	8	9	9	6	6	3	3	2	7.5	17.1
7-Oct	2	3	3	3	A	8	4	3	3	3	4	2	2	3	4	4	6	4	7	5	4	2	2	3	3.6	7.7
8-Oct	2	2	3	2	A	6	2	2	2	3	4	3	2	2	2	2	2	2	2	2	2	2	1	1	2.3	6.1
9-Oct	2	3	3	2	A	6	4	4	4	4	5	3	6	3	3	3	3	5	6	4	3	5	4	5	3.9	6.4
10-Oct	5	5	5	6	A	8	8	6	9	6	6	5	13	4	4	5	8	5	9	10	7	6	9	9	6.9	13.5
11-Oct	7	7	6	7	A	9	9	9	8	7	5	11	2	2	3	3	2	2	4	3	10	10	1	1	5.6	10.6
12-Oct	4	4	6	6	A	12	8	5	3	5	2	2	1	1	2	2	2	3	9	12	4	12	14	8	5.5	14.4
13-Oct	16	5	2	2	A	5	5	11	5	4	C	C	C	C	C	C	3	6	7	4	5	4	3	2	--	15.5
14-Oct	4	4	1	2	A	6	5	3	3	3	4	3	1	1	2	2	2	3	1	2	2	1	1	1	2.5	6.2
15-Oct	1	1	2	2	A	5	2	3	3	2	3	2	2	2	2	3	2	3	4	5	7	10	6	6	3.3	9.5
16-Oct	2	1	1	9	A	8	5	27	9	3	7	5	2	2	2	14	4	3	3	4	8	7	6	4	6.0	27.4
17-Oct	4	2	1	1	A	12	11	12	6	3	11	6	2	6	6	27	5	4	5	6	4	4	5	2	6.4	27.1
18-Oct	2	1	1	3	A	8	7	15	12	3	2	2	26	11	2	1	2	8	12	18	11	6	5	4	7.1	25.7
19-Oct	5	5	2	2	A	17	6	18	20	12	10	8	7	5	6	6	9	8	6	4	3	7	7	7	7.8	20.2
20-Oct	9	7	5	1	A	5	3	1	2	3	23	31	28	25	9	3	1	1	2	3	2	2	5	3	7.7	31.3
21-Oct	2	3	4	3	A	7	7	5	6	10	6	5	3	1	1	2	5	5	7	5	6	6	7	9	5.0	10.3
22-Oct	9	8	10	9	A	19	17	19	17	10	9	4	4	5	6	6	11	17	27	15	13	17	15	21	12.5	27.1
23-Oct	24	16	20	16	A	21	21	25	48	15	11	9	8	8	11	7	15	20	26	23	22	31	24	27	19.4	47.7
24-Oct	23	20	29	25	A	25	10	18	15	13	10	23	16	12	5	5	8	9	12	7	17	12	9	4	14.2	29.4
25-Oct	4	6	6	6	A	8	17	17	20	14	8	5	6	10	2	2	3	3	8	2	8	6	3	3	7.3	20.2
26-Oct	2	3	2	2	A	5	3	12	7	17	19	14	21	10	20	20	8	9	13	12	8	1	1	1	9.1	21.2
27-Oct	1	1	1	2	A	4	2	2	2	3	1	1	1	2	1	2	6	7	6	14	8	2	2	2	3.2	14.5
28-Oct	3	5	7	7	A	5	5	5	7	8	8	9	7	7	7	8	7	9	13	10	19	15	12	8	8.2	18.6
29-Oct	6	5	4	5	A	9	6	7	6	11	5	4	2	26	2	2	5	11	13	7	4	3	4	4	6.6	26.4
30-Oct	3	5	3	5	A	5	5	4	5	7	2	1	2	2	2	2	2	4	3	5	10	7	4	10	4.3	10.1
31-Oct	7	3	3	5	A	8	8	14	14	20	15	7	11	8	11	9	10	10	10	10	15	10	5	4	9.5	20.2
5.8 5.0 5.4 5.8 -- 9.6 7.6 9.6 9.2 7.3 7.9 6.4 6.5 5.7 4.9 5.2 4.9 6.2 8.1 7.1 7.6 7.5 6.2 5.9																									Diurnal Average	
24.1 20.1 29.4 24.8 -- 25.0 21.2 27.4 47.7 20.2 35.8 31.3 28.1 26.4 19.7 27.1 15.0 19.5 27.1 22.9 22.3 30.5 24.5 26.7																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										


Hourly Maximums

Nitrogen Dioxide (NO₂) - ppb
Beaverlodge - October 2018




Pollutant Rose

Nitrogen Dioxide (NO₂) - ppb
Beaverlodge - October 2018




Pollutant Classes (ppb)



Span Responses

Nitrogen Dioxide (NO₂)
Beaverlodge - October 2018




Hourly Averages

Nitrogen Oxide (NO) - ppb Beaverlodge - October 2018

Maximum Value: 8.0 ppb on Oct 31 10:00																				Maximum Daily Average: 1.5 ppb on Oct 23										Hours in Service: 744	
Minimum Value: 0 ppb on Oct 7 01:00																				Minimum Daily Average: 0.1 ppb on Oct 15										Hours of Data: 707	
Maximum Diurnal Average: 1.8 ppb at hour 10																				Minimum Diurnal Average: 0.0 ppb at hour 2										Hours of Missing Data: 37	
Monthly Average: 0.39 ppb																				Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.1 Q ₃ = 0.3 P ₉₀ = 1.0 P ₉₉ = 4.2										Hours of Calibration: 37	
																				Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3					
2-Oct	0	0	0	0	A	0	0	0	0	1	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0.1	0.6					
3-Oct	0	0	0	0	A	0	0	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0.3	1.2					
4-Oct	0	0	0	0	A	0	0	1	4	3	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	4.0					
5-Oct	0	0	0	1	A	0	0	1	2	3	2	1	1	1	1	1	1	0	0	0	0	0	0	0	0.6	2.7					
6-Oct	0	0	0	0	A	0	0	0	1	2	1	0	0	1	0	0	0	0	0	0	0	0	0	0	0.3	2.0					
7-Oct	0	0	0	0	A	0	0	0	0	0	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0.3	1.0					
8-Oct	0	0	0	0	A	0	0	0	0	0	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0.2	1.0					
9-Oct	0	0	0	0	A	0	0	0	0	0	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0.3	0.9					
10-Oct	0	0	0	0	A	0	0	0	1	1	1	2	1	1	1	1	0	0	0	0	0	0	0	0	0.4	1.7					
11-Oct	0	0	0	0	A	0	2	1	1	2	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0.4	2.2					
12-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0.2	0.8					
13-Oct	0	0	0	0	A	0	0	0	0	1	C	C	C	C	C	C	0	0	0	0	0	0	0	0	--	0.7					
14-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.5					
15-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1					
16-Oct	0	0	0	0	A	0	0	1	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.9					
17-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3					
18-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.4					
19-Oct	0	0	0	0	A	0	0	3	4	6	4	2	1	1	1	1	1	0	0	0	0	0	0	0	1.1	5.6					
20-Oct	0	0	0	0	A	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0.2	0.9					
21-Oct	0	0	0	0	A	0	0	0	0	2	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0.2	1.6					
22-Oct	0	0	0	0	A	1	0	0	3	2	2	1	1	1	1	1	1	1	1	0	0	0	0	0	0.8	2.6					
23-Oct	0	0	0	0	A	0	0	0	7	8	4	4	3	2	2	1	1	1	0	0	0	0	0	0	1.5	7.7					
24-Oct	1	1	1	1	A	0	0	0	2	3	3	6	2	0	0	0	0	0	0	0	0	0	0	0	0.9	6.3					
25-Oct	0	0	0	0	A	0	1	1	1	3	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0.5	2.7					
26-Oct	0	0	0	0	A	0	0	0	0	2	4	3	4	1	2	3	1	0	0	0	0	0	0	0	0.9	4.3					
27-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2					
28-Oct	0	0	0	0	A	0	0	0	0	0	0	1	1	1	1	1	1	0	0	0	1	0	0	0	0.4	1.0					
29-Oct	0	0	0	0	A	0	0	0	0	6	2	1	0	0	0	0	0	0	0	0	0	0	0	0	0.5	5.6					
30-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2					
31-Oct	0	0	0	0	A	0	0	0	1	8	2	0	1	1	1	0	0	0	0	0	0	0	0	0	0.7	8.0					
0.1 0.0 0.0 0.1 -- 0.1 0.2 0.4 1.0 1.8 1.3 1.0 0.8 0.5 0.5 0.4 0.3 0.2 0.1 0.0 0.1 0.1 0.1 0.1																									Diurnal Average						
0.6 0.6 0.6 1.2 -- 0.5 1.8 3.3 6.8 8.0 4.1 6.3 4.3 2.1 2.5 2.7 1.2 0.8 1.1 0.2 0.9 0.4 0.8 0.4																									Diurnal Maximum						
C - Calibration					A - Automated Daily Zero Span																										

Hourly Averages

Nitrogen Oxide (NO) - ppb
Beaverlodge - October 2018




Hourly Maximums

Nitrogen Oxide (NO) - ppb Beaverlodge - October 2018

Maximum Value: 37.5 ppb on Oct 23 09:00																				Maximum Daily Average: 4.6 ppb on Oct 23					Hours in Service: 744		
Minimum Value: 0 ppb on Oct 8 23:00																				Minimum Daily Average: 0.2 ppb on Oct 15					Hours of Data: 707		
Maximum Diurnal Average: 5.4 ppb at hour 10																				Minimum Diurnal Average: 0.3 ppb at hour 2					Hours of Missing Data: 37		
Monthly Average: 1.42 ppb																				Percentiles: P ₁ = 0.1 P ₁₀ = 0.1 Q ₁ = 0.1 Median = 0.4 Q ₃ = 1.1 P ₉₀ = 3.1 P ₉₉ = 18.0					Hours of Calibration: 37		
																				Percent Operational Time: 100.0							
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	0	0	0	0	A	0	0	1	0	1	1	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0.4	0.8
2-Oct	0	0	0	0	A	0	0	0	1	1	0	1	1	0	0	0	1	1	1	0	0	0	0	0	0	0.4	1.0
3-Oct	0	0	0	0	A	0	0	3	3	1	2	1	1	0	0	0	1	1	0	0	0	0	0	0	0	0.8	3.1
4-Oct	0	0	0	0	A	0	0	1	20	4	3	1	0	0	1	0	1	0	0	0	0	1	0	0	0	1.5	20.2
5-Oct	0	0	0	7	A	1	1	2	5	5	5	1	1	1	2	2	1	1	1	0	0	1	1	0	0	1.7	7.1
6-Oct	0	0	0	0	A	0	1	1	3	4	1	3	2	28	4	0	1	1	1	0	1	0	0	0	0	2.4	27.7
7-Oct	0	0	0	0	A	0	0	0	0	1	2	1	1	2	1	1	1	0	0	0	0	0	0	0	0	0.6	1.5
8-Oct	0	0	0	0	A	0	0	0	0	1	3	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0.5	2.8
9-Oct	0	0	0	0	A	0	0	1	0	1	2	1	3	1	1	1	1	1	0	0	0	0	0	0	0	0.7	2.9
10-Oct	0	0	0	0	A	0	1	0	2	2	2	14	4	1	1	1	1	1	0	0	0	0	1	1	0	1.4	13.7
11-Oct	0	0	0	0	A	1	4	3	2	4	3	4	1	0	1	1	0	0	0	0	0	0	0	0	0	1.1	4.3
12-Oct	0	0	1	0	A	1	0	1	1	1	1	0	0	0	0	0	0	0	0	0	1	1	2	0	0	0.5	2.2
13-Oct	3	0	0	0	A	0	0	0	1	2	C	C	C	C	C	C	0	0	0	0	1	0	0	0	0	--	2.7
14-Oct	0	0	0	0	A	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	1.1
15-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0.2	0.7
16-Oct	0	0	0	1	A	1	0	17	1	1	2	1	0	1	0	13	1	0	0	0	0	0	0	0	0	1.7	17.0
17-Oct	0	0	0	0	A	0	0	0	0	0	3	1	4	1	11	2	0	0	0	0	0	0	0	0	0	1.0	10.8
18-Oct	0	0	0	0	A	0	0	1	1	1	0	4	0	13	0	0	0	0	4	0	0	0	0	0	0	1.1	13.2
19-Oct	0	0	0	0	A	1	0	11	25	19	8	4	4	3	2	2	1	1	0	0	0	0	0	0	0	3.5	25.3
20-Oct	0	0	0	0	A	0	0	0	0	0	3	2	2	1	1	0	0	0	0	0	0	0	0	0	0	0.6	2.6
21-Oct	0	0	0	0	A	0	0	0	1	6	2	2	1	0	0	0	1	0	0	0	0	0	0	0	0	0.6	5.8
22-Oct	0	0	0	0	A	4	5	1	12	7	6	2	2	1	2	1	2	2	13	0	0	0	0	5	0	2.9	12.8
23-Oct	0	0	0	0	A	0	0	7	37	18	8	5	4	3	3	2	2	2	0	0	1	6	2	1	0	4.6	37.5
24-Oct	3	5	4	12	A	0	0	3	6	8	6	16	6	3	1	1	1	0	0	0	1	0	0	0	0	3.4	16.1
25-Oct	0	0	0	0	A	0	18	7	4	8	3	2	1	3	0	0	0	0	0	8	1	0	0	0	0	2.4	18.0
26-Oct	0	0	0	0	A	0	0	2	1	10	13	5	11	2	10	10	2	1	1	0	0	1	0	0	0	3.0	12.6
27-Oct	0	0	0	1	A	1	0	0	0	1	0	0	0	1	0	0	1	0	0	0	0	0	0	0	0	0.4	1.0
28-Oct	0	0	0	0	A	0	1	1	1	1	1	2	2	1	1	1	1	1	1	0	7	2	0	0	0	1.1	7.2
29-Oct	0	0	0	0	A	1	1	2	1	32	4	3	1	1	1	1	0	0	0	0	0	0	0	0	0	2.1	32.3
30-Oct	0	0	0	0	A	0	0	0	0	1	4	1	3	1	0	1	0	1	0	0	0	0	0	0	0	0.6	3.6
31-Oct	0	0	0	0	A	0	0	1	2	27	8	1	3	2	2	1	1	1	0	0	1	0	0	0	0	2.2	26.6
0.3 0.3 0.3 0.8 -- 0.5 1.2 2.2 4.3 5.4 3.1 2.6 2.1 2.4 1.6 1.5 0.8 0.6 0.8 0.4 0.5 0.5 0.3 0.4																									Diurnal Average		
2.7 5.3 4.5 11.8 -- 3.7 18.0 17.0 37.5 32.3 12.6 16.1 10.9 27.7 10.8 13.4 2.0 2.3 12.8 7.6 7.2 6.0 2.3 4.7																									Diurnal Maximum		
C - Calibration A - Automated Daily Zero Span																											

Hourly Maximums

Nitrogen Oxide (NO) - ppb
Beaverlodge - October 2018



Hourly Averages


Oxides of Nitrogen (NO_x) - ppb

Beaverlodge - October 2018

Maximum Value: 25.7 ppb on Oct 23 09:00																				Maximum Daily Average: 14.6 ppb on Oct 23					Hours in Service: 744	
Minimum Value: 0 ppb on Oct 15 01:00																				Minimum Daily Average: 1.3 ppb on Oct 2					Hours of Data: 707	
Maximum Diurnal Average: 6.3 ppb at hour 10																				Minimum Diurnal Average: 2.8 ppb at hour 16					Hours of Missing Data: 37	
Monthly Average: 4.41 ppb																				Percentiles: P ₁ = 0.5 P ₁₀ = 0.9 Q ₁ = 1.6 Median = 3.2 Q ₃ = 5.9 P ₉₀ = 9.5 P ₉₉ = 19.4					Hours of Calibration: 37	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	2	2	2	3	A	7	7	6	3	3	2	1	2	2	2	2	3	1	1	1	1	1	1	1	2.4	6.7
2-Oct	0	1	0	1	A	4	2	2	3	2	1	1	2	1	1	0	1	1	2	2	1	1	1	2	1.3	4.0
3-Oct	2	3	3	3	A	7	5	8	5	3	3	2	1	1	1	1	2	5	5	4	4	9	6	6	3.9	9.5
4-Oct	6	9	7	5	A	9	7	8	10	7	4	2	1	1	1	1	1	1	2	4	4	7	4	4	4.5	10.4
5-Oct	4	3	5	12	A	10	7	9	7	8	6	2	3	3	4	4	4	4	6	6	7	6	11	11	6.1	12.4
6-Oct	10	8	5	4	A	7	6	6	8	8	4	3	1	1	3	2	3	6	6	4	4	3	2	1	4.6	9.5
7-Oct	2	2	2	2	A	5	3	2	2	2	3	2	2	4	4	3	4	3	5	4	3	2	2	2	2.9	5.4
8-Oct	2	2	2	2	A	4	2	2	2	3	3	3	3	2	2	2	2	2	2	2	1	1	1	1	2.1	3.8
9-Oct	2	2	2	2	A	4	4	4	4	4	4	4	4	4	4	3	3	3	4	3	3	3	4	4	3.3	4.4
10-Oct	4	4	4	5	A	6	7	5	7	6	6	6	6	4	4	5	5	3	4	6	6	5	7	7	5.4	6.7
11-Oct	6	6	6	7	A	7	10	8	6	6	5	4	1	1	1	1	1	1	2	2	4	3	1	1	3.9	9.7
12-Oct	3	3	3	4	A	9	5	3	2	3	2	2	1	1	1	1	1	2	5	6	3	5	12	5	3.7	12.1
13-Oct	6	2	2	1	A	3	3	4	3	3	C	C	C	C	C	C	2	3	3	2	2	2	2	2	--	5.7
14-Oct	2	2	1	1	A	5	4	3	2	2	2	1	1	0	1	1	1	1	1	1	1	0	0	1	1.5	5.0
15-Oct	0	1	1	1	A	3	2	2	2	1	1	1	1	1	1	1	1	1	2	2	4	5	3	2	1.6	5.1
16-Oct	1	1	1	1	A	4	3	5	3	2	3	1	1	1	1	2	1	1	1	2	4	4	2	1	2.0	5.0
17-Oct	2	1	1	1	A	4	5	5	3	2	3	3	1	2	2	3	1	2	2	3	2	2	1	1	2.3	5.2
18-Oct	2	1	1	1	A	5	3	11	5	2	1	1	1	1	1	1	1	3	2	5	6	5	3	3	2.7	10.9
19-Oct	2	3	2	1	A	7	4	15	14	14	11	8	4	4	5	5	6	5	4	4	3	4	5	6	6.0	15.3
20-Oct	7	6	3	1	A	3	2	1	1	1	5	15	12	6	3	2	1	1	1	1	1	1	3	1	3.3	14.7
21-Oct	1	1	2	2	A	4	4	3	4	7	6	4	2	1	1	1	2	3	5	4	5	5	6	7	3.4	6.8
22-Oct	8	7	8	8	A	14	10	16	14	8	7	4	4	4	6	6	9	13	17	14	9	12	13	17	10.0	17.3
23-Oct	18	13	14	12	A	16	16	15	26	18	12	12	10	8	9	6	9	16	17	14	16	15	20	23	14.6	25.7
24-Oct	19	17	21	24	A	16	8	12	12	11	11	19	12	3	3	3	5	7	7	5	5	7	6	4	10.3	24.0
25-Oct	3	4	5	3	A	6	9	11	12	12	7	5	4	4	2	1	2	1	2	1	2	3	2	2	4.4	12.0
26-Oct	1	1	2	1	A	4	2	6	5	9	18	15	20	9	14	13	7	8	9	9	3	1	1	0	6.8	19.5
27-Oct	0	0	0	0	A	3	2	1	1	1	1	1	1	1	1	1	2	4	4	6	3	2	1	2	1.6	5.6
28-Oct	2	3	4	4	A	4	4	4	5	6	6	7	7	7	6	7	7	8	8	7	11	11	8	6	6.1	10.7
29-Oct	5	4	4	4	A	6	5	5	5	12	6	3	1	1	1	1	2	7	7	3	3	2	3	3	4.2	11.7
30-Oct	3	4	3	3	A	3	2	2	3	2	1	1	1	1	1	1	1	1	1	2	5	4	3	4	2.3	5.1
31-Oct	3	3	3	2	A	5	5	10	11	22	12	5	7	7	5	5	7	8	8	7	10	5	4	3	6.8	21.7
4.2 3.8 3.8 4.0 -- 6.2 5.0 6.2 6.1 6.3 5.2 4.7 3.9 2.8 2.9 2.8 3.1 4.2 4.6 4.3 4.3 4.4 4.4 4.3																									Diurnal Average	
19.3 17.1 20.8 24.0 -- 16.5 16.3 16.2 25.7 21.7 17.9 19.4 19.5 9.1 13.7 13.0 9.3 15.6 17.3 13.8 15.8 14.8 19.6 23.1																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										

Hourly Averages

Oxides of Nitrogen (NO_x) - ppb
Beaverlodge - October 2018



Hourly Maximums


Oxides of Nitrogen (NO_x) - ppb

Beaverlodge - October 2018

Maximum Value: 78.6 ppb on Oct 23 09:00																				Maximum Daily Average: 23.2 ppb on Oct 23										Hours in Service: 744	
Minimum Value: 1 ppb on Oct 15 01:00																				Minimum Daily Average: 2.7 ppb on Oct 2										Hours of Data: 707	
Maximum Diurnal Average: 13.1 ppb at hour 9																				Minimum Diurnal Average: 5.1 ppb at hour 2										Hours of Missing Data: 37	
Monthly Average: 7.90 ppb																				Percentiles: P ₁ = 0.9 P ₁₀ = 1.9 Q ₁ = 2.9 Median = 5.3 Q ₃ = 9.7 P ₉₀ = 16.9 P ₉₉ = 36.8										Hours of Calibration: 37	
																														Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	2	3	2	4	A	8	10	8	4	6	4	2	3	6	17	3	4	4	2	1	1	1	1	1	4.3	17.1					
2-Oct	1	1	1	2	A	6	3	3	4	3	2	1	3	1	1	1	2	4	5	4	3	2	4	5	2.7	6.3					
3-Oct	4	4	4	4	A	9	6	14	14	5	5	3	3	1	1	1	4	10	7	5	8	15	8	8	6.3	15.5					
4-Oct	9	11	12	6	A	15	16	10	29	11	37	3	2	1	3	1	2	2	3	4	7	11	7	5	9.1	36.9					
5-Oct	5	4	8	31	A	19	14	16	16	14	14	4	3	5	4	5	4	5	10	9	9	14	14	13	10.4	31.0					
6-Oct	13	9	6	6	A	10	11	9	13	14	5	17	10	1	10	6	10	9	10	6	7	3	3	3	8.3	17.5					
7-Oct	3	3	3	3	A	8	4	3	4	3	5	3	3	5	5	5	7	4	8	5	4	2	3	3	4.1	7.8					
8-Oct	2	2	3	2	A	6	3	2	3	3	7	4	3	3	2	2	3	2	2	2	2	2	1	1	2.8	6.8					
9-Oct	2	3	3	3	A	6	4	5	4	4	7	4	9	4	5	4	5	5	6	4	3	5	5	5	4.5	8.6					
10-Oct	5	5	5	6	A	8	9	6	11	8	8	14	18	5	5	6	9	5	10	10	7	6	9	10	8.0	17.8					
11-Oct	7	7	6	7	A	9	12	12	10	11	8	13	3	2	4	3	2	3	4	3	10	10	2	2	6.5	13.3					
12-Oct	4	4	6	6	A	13	9	5	3	6	3	2	1	1	2	2	2	3	9	12	5	13	16	8	5.9	16.4					
13-Oct	18	5	2	2	A	5	5	11	5	5	C	C	C	C	C	C	3	7	7	4	6	4	3	2	--	18.2					
14-Oct	4	3	1	2	A	6	5	4	3	4	5	3	1	1	3	2	2	3	1	2	2	1	1	1	2.7	6.1					
15-Oct	1	1	2	2	A	5	2	3	3	2	3	2	2	2	2	4	2	3	4	5	7	10	6	6	3.4	9.6					
16-Oct	2	1	1	10	A	9	5	44	10	4	9	5	3	2	2	16	4	3	3	4	8	7	6	4	7.1	43.7					
17-Oct	4	2	1	1	A	12	11	12	6	3	13	7	2	6	7	29	5	4	5	6	4	4	5	2	6.7	28.8					
18-Oct	2	1	1	3	A	8	6	16	12	4	2	3	26	12	2	1	2	8	16	19	11	6	5	4	7.4	25.8					
19-Oct	5	5	3	2	A	17	6	27	44	29	17	11	11	7	8	7	10	9	6	4	3	7	7	7	10.9	44.2					
20-Oct	9	7	5	1	A	5	4	2	2	3	26	33	30	26	10	4	2	1	3	3	2	2	5	3	8.2	33.1					
21-Oct	2	3	4	3	A	7	7	5	8	16	8	6	4	1	1	3	6	5	7	5	6	6	7	9	5.6	16.1					
22-Oct	9	8	10	9	A	19	22	19	29	17	15	6	5	6	7	8	12	17	40	15	13	17	15	26	15.0	39.9					
23-Oct	24	16	20	16	A	21	21	28	79	31	19	14	12	11	14	9	17	21	26	23	23	36	27	27	23.2	78.6					
24-Oct	24	25	31	36	A	25	10	21	20	21	15	39	22	14	6	5	9	10	13	7	17	12	9	4	17.2	39.3					
25-Oct	4	6	6	6	A	9	34	24	24	19	10	6	7	12	3	2	3	3	8	2	9	6	3	3	9.1	34.3					
26-Oct	2	3	3	2	A	5	3	13	8	28	32	18	32	12	29	30	9	10	14	12	8	2	1	1	12.0	32.1					
27-Oct	1	1	2	3	A	5	2	2	2	3	2	2	1	2	1	2	6	7	6	15	8	3	2	2	3.5	14.5					
28-Oct	3	5	7	7	A	5	5	6	8	9	8	11	9	9	8	9	8	9	13	10	24	17	12	8	9.0	23.6					
29-Oct	6	5	4	5	A	9	7	8	7	43	9	7	3	27	3	2	5	11	13	7	4	3	4	4	8.6	43.1					
30-Oct	3	4	3	6	A	5	5	4	5	9	3	2	3	2	2	3	2	5	3	5	10	7	4	10	4.7	10.4					
31-Oct	7	3	3	4	A	8	8	14	16	41	23	8	14	10	13	10	12	10	10	10	15	10	5	4	11.3	41.4					
6.0 5.1 5.5 6.4 -- 9.8 8.7 11.5 13.1 12.3 10.8 8.5 8.3 6.7 6.1 6.2 5.6 6.6 8.8 7.2 8.0 7.9 6.4 6.1																									Diurnal Average						
24.4 24.9 30.6 35.5 -- 25.0 34.3 43.7 78.6 43.1 36.9 39.3 32.1 27.0 29.3 30.5 16.9 21.3 39.9 22.9 23.6 35.7 26.6 27.5																									Diurnal Maximum						
C - Calibration A - Automated Daily Zero Span																															


Hourly Maximums

Oxides of Nitrogen (NO_x) - ppb
Beaverlodge - October 2018



Span Responses

Oxides of Nitrogen (NO_x)
Beaverlodge - October 2018



Hourly Averages


Ozone (O₃) - ppb

Beaverlodge - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 52.4 ppb on Oct 17 04:00 Maximum Daily Average: 44.8 ppb on Oct 17																								Hours in Service: 744 Hours of Data: 710 Hours of Missing Data: 34 Hours of Calibration: 34 Percent Operational Time: 100.0			
Minimum Value: 3 ppb on Oct 11 07:00 Maximum Diurnal Average: 30.7 ppb at hour 15 Monthly Average: 24.63 ppb												Minimum Daily Average: 12.1 ppb on Oct 7 Minimum Diurnal Average: 19.0 ppb at hour 8 Percentiles: P ₁ = 5.6 P ₁₀ = 11.9 Q ₁ = 16.3 Median = 25.0 Q ₃ = 32.3 P ₉₀ = 37.1 P ₉₉ = 47.7															
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	27	25	24	20	A	17	14	16	25	26	28	30	30	31	30	30	27	27	30	29	27	27	26		25.9	31.2	
2-Oct	26	26	26	27	A	28	28	28	25	29	33	33	32	34	34	34	34	32	28	26	25	24	24	21		28.5	34.2
3-Oct	21	20	19	19	A	16	15	12	17	20	22	26	32	34	36	37	37	30	27	30	30	24	27	26		25.0	37.2
4-Oct	24	16	17	18	A	20	15	13	11	15	22	28	33	35	35	35	34	34	31	26	26	22	24	22		24.3	35.2
5-Oct	20	20	18	12	A	13	14	11	13	16	20	25	25	26	27	27	27	26	22	21	20	19	14	13		19.6	27.2
6-Oct	13	15	16	16	A	15	12	11	12	15	28	30	35	35	31	30	25	20	16	17	15	19	18	19		20.2	34.7
7-Oct	17	15	13	12	A	11	10	10	10	10	10	11	11	11	12	12	11	11	10	11	13	15	15	16		12.1	17.0
8-Oct	15	15	14	14	A	15	15	14	13	12	11	11	15	18	21	24	22	21	21	20	20	20	20	20		17.0	24.5
9-Oct	19	17	17	17	A	16	16	16	16	16	17	17	17	18	19	19	19	18	17	16	17	17	16	16		17.1	18.8
10-Oct	15	15	15	14	A	13	9	10	9	11	12	12	13	15	15	14	14	16	15	10	8	9	8	8		12.2	16.1
11-Oct	8	8	8	5	A	6	3	5	9	10	15	19	31	31	32	33	32	31	27	27	25	25	23	22		18.9	32.6
12-Oct	19	17	11	11	A	14	13	14	15	17	18	20	21	23	25	26	25	21	15	11	16	12	5	12		16.5	25.6
13-Oct	12	18	20	22	A	23	21	19	18	19	23	24	28	C	C	C	28	26	22	21	20	19	20	19		21.1	28.4
14-Oct	17	19	21	22	A	23	23	23	25	24	34	39	41	42	41	37	39	40	39	37	34	32	31	31		31.1	42.4
15-Oct	32	31	30	31	A	38	35	38	40	42	41	41	41	40	40	39	38	36	34	32	32	30	32	33		35.9	41.7
16-Oct	35	37	37	36	A	35	36	35	36	37	38	50	50	49	45	41	39	38	37	38	35	35	40	42		39.0	50.0
17-Oct	44	48	50	52	A	48	45	42	44	46	45	46	48	46	44	43	46	45	45	42	42	41	41	40		44.8	52.4
18-Oct	38	39	44	45	A	41	39	26	32	35	38	37	37	37	37	37	36	32	34	31	26	27	31	29		35.2	45.4
19-Oct	29	25	28	28	A	18	23	8	13	11	17	25	30	32	32	34	33	31	33	33	34	33	30	27		26.4	34.4
20-Oct	22	22	28	32	A	35	34	36	38	38	36	31	32	35	36	37	37	37	36	35	35	36	30	30		33.4	38.3
21-Oct	32	31	31	33	A	33	27	28	26	23	24	30	37	39	39	38	35	33	33	34	32	32	30	27		31.6	38.7
22-Oct	25	23	20	21	A	12	14	7	12	17	20	25	26	29	30	29	23	17	13	17	29	22	16	8		19.8	30.2
23-Oct	9	16	13	14	A	12	10	12	10	11	18	20	24	28	30	29	25	17	15	20	16	15	11	5		16.6	30.2
24-Oct	7	4	6	4	A	11	18	13	11	14	16	12	23	35	36	34	29	28	28	29	30	20	26	28		20.1	35.6
25-Oct	26	26	23	24	A	22	17	12	14	19	25	27	29	29	35	36	35	35	34	35	32	30	30	30		27.2	35.7
26-Oct	30	33	33	32	A	30	32	25	28	22	13	16	14	22	17	17	21	18	18	20	29	35	33	34		25.0	35.1
27-Oct	35	35	35	35	A	35	34	35	34	33	34	35	35	34	35	34	32	29	29	27	29	29	30	29		32.6	34.9
28-Oct	30	30	27	27	A	29	27	25	23	18	19	18	18	17	17	16	14	14	16	17	12	11	16	16		19.9	30.1
29-Oct	17	16	14	10	A	8	7	8	10	10	14	18	28	33	33	33	32	27	25	27	28	28	25	23		20.7	33.4
30-Oct	24	23	19	18	A	25	24	25	27	28	31	32	33	33	35	35	33	33	32	30	26	27	27	26		28.2	35.1
31-Oct	24	24	22	24	A	23	19	13	12	8	15	19	18	17	22	24	16	13	14	15	9	14	14	14		17.2	24.3
23.0 22.9 22.5 22.4 -- 22.0 21.0 19.0 20.3 21.1 23.7 26.1 28.6 30.3 30.7 30.5 29.1 26.9 25.6 25.4 24.9 24.1 23.7 23.0																								Diurnal Average			
43.7 48.0 49.7 52.4 -- 47.9 45.1 42.2 43.5 45.9 44.7 49.6 50.0 48.6 44.6 43.2 45.8 44.6 44.6 42.0 41.7 40.6 40.8 42.4																								Diurnal Maximum			
C - Calibration A - Automated Daily Zero Span Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 82 ppb 24-hr na																											

Hourly Averages

Ozone (O₃) - ppb
Beaverlodge - October 2018



Hourly Maximums


Ozone (O₃) - ppb

Beaverlodge - October 2018

Maximum Value: 54.3 ppb on Oct 17 04:00																								Hours in Service: 744					
Maximum Daily Average: 47.2 ppb on Oct 17																								Hours of Data: 710					
Minimum Value: 5 ppb on Oct 11 07:00																								Hours of Missing Data: 34					
Maximum Diurnal Average: 32.6 ppb at hour 15																								Hours of Calibration: 34					
Monthly Average: 27.19 ppb																								Percent Operational Time: 100.0					
Minimum Daily Average: 13.3 ppb on Oct 7																													
Minimum Diurnal Average: 23.2 ppb at hour 8																													
Percentiles: P ₁ = 8.7 P ₁₀ = 15.3 Q ₁ = 19.2 Median = 27.3 Q ₃ = 34.5 P ₉₀ = 38.9 P ₉₉ = 49.3																													
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum			
1-Oct	29	27	25	24	A	18	17	25	26	28	30	32	33	33	35	35	35	35	33	31	27	29	31	30	29	27	27	27.9	32.9
2-Oct	27	27	27	29	A	29	29	29	27	32	34	33	33	33	35	35	35	35	33	31	27	27	27	26	26	23	30.0	35.2	
3-Oct	22	21	20	20	A	19	16	16	20	21	25	29	35	36	38	38	38	35	30	32	31	27	28	30	30	27.2	38.3		
4-Oct	29	20	22	20	A	23	19	15	14	19	27	31	36	36	36	36	35	35	34	28	27	25	26	24	24	26.8	36.2		
5-Oct	21	21	20	19	A	18	18	15	17	21	24	26	26	27	28	28	28	28	24	23	22	22	16	14	14	22.0	28.1		
6-Oct	14	17	19	19	A	24	24	16	24	29	30	35	36	36	35	32	29	25	18	18	17	20	19	20	20	24.2	35.9		
7-Oct	18	17	16	14	A	12	11	11	11	11	11	12	12	12	13	13	12	12	12	12	15	16	16	16	16	13.3	18.2		
8-Oct	16	16	15	15	A	15	15	15	14	13	12	12	17	19	25	26	23	22	22	22	21	20	21	21	21	18.2	26.1		
9-Oct	20	18	17	17	A	17	17	17	17	17	17	20	18	19	19	19	20	19	18	17	18	17	17	17	17	17.9	19.8		
10-Oct	17	16	16	15	A	14	12	12	12	13	13	13	15	16	16	15	16	17	18	11	11	11	10	10	10	13.8	17.7		
11-Oct	9	9	9	7	A	8	5	8	11	16	17	22	34	32	34	34	33	33	30	28	28	27	26	23	23	20.8	33.6		
12-Oct	23	19	17	12	A	18	18	15	18	18	20	21	22	25	27	27	27	22	19	15	18	18	11	15	15	19.4	27.2		
13-Oct	16	21	21	23	A	24	23	20	20	21	25	26	31	C	C	C	30	29	25	22	22	21	22	21	21	23.1	30.5		
14-Oct	19	21	23	23	A	23	24	25	28	28	39	40	43	44	42	41	41	41	40	38	36	33	32	33	33	32.9	43.5		
15-Oct	33	32	31	33	A	38	37	40	42	43	42	42	42	41	41	40	39	37	36	34	33	33	34	34	34	37.2	42.6		
16-Oct	37	38	38	41	A	36	38	40	40	41	42	52	52	51	47	44	40	39	38	39	38	38	45	45	45	41.6	52.1		
17-Oct	47	49	52	54	A	52	49	45	46	49	49	49	49	48	47	47	47	46	46	44	43	42	42	41	41	47.2	54.3		
18-Oct	39	43	46	46	A	44	40	40	37	38	39	39	39	38	38	38	37	35	36	35	33	31	34	33	33	38.0	46.4		
19-Oct	33	28	30	30	A	26	26	26	22	19	23	29	34	33	34	37	37	33	34	34	35	34	33	29	29	30.4	37.1		
20-Oct	25	26	33	34	A	36	36	37	40	39	38	38	38	38	37	38	38	38	38	36	36	37	36	35	35	35.9	39.5		
21-Oct	34	34	33	34	A	35	33	33	27	26	30	32	41	40	40	40	39	36	36	35	34	32	32	31	31	34.1	40.9		
22-Oct	27	24	22	23	A	18	19	13	18	20	25	26	28	31	32	32	29	21	17	22	33	33	30	17	17	24.4	33.5		
23-Oct	19	22	22	22	A	16	14	15	15	17	21	22	25	32	32	32	30	20	17	24	21	17	16	10	10	21.0	32.1		
24-Oct	16	10	10	9	A	20	20	19	15	16	19	16	29	39	37	36	36	32	31	31	38	26	30	32	32	24.6	38.6		
25-Oct	29	28	25	27	A	23	24	24	19	25	27	30	31	35	37	38	37	36	36	36	36	32	32	34	34	30.5	38.0		
26-Oct	35	36	36	34	A	33	32	30	30	29	18	21	23	25	24	23	25	23	23	23	35	36	35	35	35	28.9	36.2		
27-Oct	35	35	35	36	A	35	35	37	35	34	35	35	35	35	35	35	34	32	31	33	33	32	33	30	30	34.2	36.8		
28-Oct	31	32	29	30	A	31	28	28	25	20	20	20	19	18	19	18	17	17	19	19	19	16	19	19	19	22.4	31.7		
29-Oct	19	19	15	13	A	8	8	10	12	13	16	22	33	35	34	34	34	30	30	32	30	29	29	25	25	23.0	34.5		
30-Oct	25	25	25	21	A	27	26	26	29	30	32	34	34	34	36	36	35	34	33	47	31	29	30	30	30	30.8	47.0		
31-Oct	27	26	23	28	A	25	22	17	16	20	20	21	22	21	27	29	24	15	17	17	12	16	15	16	16	20.8	28.6		
25.5 25.1 24.8 24.8 -- 24.7 23.8 23.2 23.4 24.7 26.5 28.4 31.1 32.1 32.6 32.4 31.5 29.2 28.0 28.0 27.8 26.6 26.5 25.5																										Diurnal Average			
47.1 49.3 51.8 54.3 -- 52.0 49.3 45.3 46.2 49.0 48.5 52.1 51.7 50.6 46.7 47.1 47.5 46.2 46.1 47.0 43.2 42.0 44.9 44.8																										Diurnal Maximum			
C - Calibration A - Automated Daily Zero Span																													


Hourly Maximums

Ozone (O₃) - ppb
Beaverlodge - October 2018



Pollutant Rose

Ozone (O₃) - ppb
Beaverlodge - October 2018



Eight Hour Running Averages


Ozone (O₃) - ppb

Beaverlodge - October 2018

Maximum Value: 47.0 ppb on Oct 17 07:00																								Hours in Service: 744	
Minimum Value: 6.2 ppb on Oct 11 08:00																								Hours of Data: 738	
																								Hours of Missing Data: 6	
Percentiles: P ₁ = 7.7 P ₁₀ = 13.4 Q ₁ = 16.9 Median = 24.0 Q ₃ = 31.6 P ₉₀ = 35.9 P ₉₉ = 45.4																								Hours of Calibration: 6	
																								Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																							Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23		24
1-Oct	28	27	27	26	25	24	22	20	20	20	21	22	23	25	27	29	30	30	30	30	29	29	28	28	29.7
2-Oct	27	27	27	27	26	26	27	27	27	27	28	29	29	30	31	32	33	33	33	32	31	30	28	27	33.2
3-Oct	25	24	22	22	21	20	19	17	17	17	17	18	20	22	25	28	31	32	32	33	33	31	30	29	33.0
4-Oct	27	25	24	23	22	21	19	18	16	16	17	18	20	22	24	27	30	32	33	33	32	30	29	28	33.2
5-Oct	26	24	22	21	20	19	17	16	15	14	14	16	17	19	20	22	24	25	26	25	24	24	22	20	25.8
6-Oct	18	17	16	16	15	14	14	14	14	14	16	18	20	22	25	27	29	29	28	26	24	22	20	19	29.3
7-Oct	18	17	17	16	16	15	14	13	12	11	10	10	10	10	11	11	11	11	11	11	12	12	12	13	17.6
8-Oct	13	14	14	15	15	15	15	15	14	14	13	13	13	14	14	16	17	18	19	20	21	21	21	20	21.2
9-Oct	20	20	19	19	18	18	17	17	17	16	16	16	17	17	17	17	18	18	18	18	18	18	17	17	20.0
10-Oct	17	16	16	16	16	15	14	13	12	12	11	11	11	11	12	13	13	14	14	14	13	13	12	11	16.6
11-Oct	10	9	8	8	8	7	6	6	6	7	8	10	12	15	19	23	25	28	29	30	30	29	28	26	30.4
12-Oct	25	23	21	19	18	17	15	14	13	13	14	16	16	18	19	21	22	22	22	21	20	19	16	15	24.7
13-Oct	13	13	13	15	15	16	18	19	20	20	21	21	22	22	22	N	N	N	N	N	N	23	22	22	22.8
14-Oct	21	20	20	20	20	20	21	21	22	23	25	27	29	32	34	36	37	39	40	40	39	37	36	35	39.9
15-Oct	35	33	32	32	31	32	33	34	35	36	38	39	40	40	40	40	40	39	38	37	36	35	34	33	40.5
16-Oct	33	33	33	34	34	35	35	36	36	36	36	38	39	41	42	43	43	44	43	42	40	38	38	38	43.5
17-Oct	38	40	41	43	44	46	47	47	47	47	46	45	45	45	45	45	45	45	45	45	44	43	43	43	47.0
18-Oct	42	41	41	41	41	41	41	39	38	38	37	36	36	35	35	36	37	36	36	35	34	33	32	31	41.6
19-Oct	30	29	28	28	28	27	26	23	20	18	17	16	18	20	21	24	27	29	31	32	33	33	33	32	32.9
20-Oct	30	29	29	28	28	28	29	30	32	34	36	35	35	35	35	35	35	35	35	36	36	36	35	34	36.1
21-Oct	34	33	32	32	32	32	31	31	30	29	28	27	28	29	31	32	33	34	35	36	35	34	33	32	35.9
22-Oct	31	29	28	26	25	22	20	17	16	15	15	15	17	19	21	24	25	25	24	23	24	23	21	18	30.7
23-Oct	16	16	16	16	14	12	12	12	12	12	12	13	15	17	19	21	23	24	24	24	23	21	19	16	24.1
24-Oct	13	12	10	8	7	7	8	9	10	11	12	14	15	18	20	23	25	27	28	30	31	29	28	27	31.1
25-Oct	27	27	26	25	25	25	24	21	20	19	19	19	21	21	24	27	29	31	32	33	34	34	33	33	33.9
26-Oct	32	32	32	31	31	31	32	31	31	29	26	24	23	22	20	19	18	17	18	18	20	22	24	26	32.1
27-Oct	28	30	32	34	35	34	35	35	35	34	34	34	34	34	34	34	34	33	33	32	31	30	30	29	34.7
28-Oct	29	29	29	29	29	29	28	28	27	25	24	23	22	21	20	18	17	17	16	16	15	15	14	14	29.2
29-Oct	15	15	15	14	14	14	13	11	10	10	10	11	13	16	19	22	25	27	29	30	30	29	28	27	29.9
30-Oct	26	26	25	24	23	23	22	23	23	24	26	28	28	29	31	32	33	33	33	33	32	31	30	29	33.3
31-Oct	28	27	26	25	25	24	23	21	20	17	16	16	16	15	16	17	17	18	18	17	16	16	15	14	28.1
Diurnal Maximums																									
N - Not Valid																									

Span Responses

Ozone (O₃)
Beaverlodge - October 2018



Hourly Averages


Particulate Matter 2.5 (PM_{2.5}) - µg/m³

Beaverlodge - October 2018

Number of Exceedences: 1-hr: 0 24-hr: 0 Maximum Value: 14.2 µg/m³ on Oct 24 00:00 Maximum Daily Average: 9.3 µg/m³ on Oct 24																								Hours in Service: 744 Hours of Data: 744 Hours of Missing Data: 0 Hours of Calibration: 0 Percent Operational Time: 100.0			
Minimum Value: 1 µg/m³ on Oct 25 16:00 Maximum Diurnal Average: 6.8 µg/m³ at hour 2 Monthly Average: 4.76 µg/m³ Minimum Daily Average: 2.2 µg/m³ on Oct 14 Minimum Diurnal Average: 2.5 µg/m³ at hour 16 Percentiles: P ₁ = 1.4 P ₁₀ = 1.8 Q ₁ = 2.5 Median = 4.2 Q ₃ = 6.4 P ₉₀ = 8.5 P ₉₉ = 13.2																											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	3	4	4	4	4	4	4	4	3	3	3	2	2	2	2	2	3	4	4	4	5	4	4	4	3.5	4.5	
2-Oct	4	4	4	3	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	3	4	4	4	2.4	3.8	
3-Oct	3	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	4	5	6	6	6	2.7	5.8	
4-Oct	5	7	6	7	7	6	8	10	9	9	7	6	4	3	2	2	3	5	5	7	8	9	9	9	6.4	9.6	
5-Oct	8	8	8	7	7	8	7	6	5	4	3	2	2	2	2	2	2	2	3	5	6	7	8	8	5.1	8.4	
6-Oct	8	9	8	7	7	7	6	6	6	6	5	3	2	2	2	2	2	2	3	4	5	6	7	6	5.1	8.5	
7-Oct	7	8	9	9	9	8	8	7	8	7	6	6	6	6	6	6	5	4	4	4	4	4	4	4	6.2	9.1	
8-Oct	4	5	4	4	4	4	4	4	3	2	2	2	2	2	2	2	2	2	3	3	3	3	3	4	3.0	4.5	
9-Oct	6	8	9	9	9	9	8	8	9	9	9	8	7	6	5	5	4	4	4	4	5	5	5	5	6.6	9.4	
10-Oct	6	7	7	7	7	7	8	7	7	7	7	8	8	7	7	7	7	6	6	7	7	7	8	8	7.1	8.0	
11-Oct	9	10	9	9	8	7	6	6	5	5	4	4	3	2	1	1	1	1	2	2	3	3	3	3	4.5	9.5	
12-Oct	4	5	5	6	5	5	5	4	5	6	6	6	5	5	4	4	5	6	5	5	5	5	5	4	5.0	5.9	
13-Oct	6	6	6	5	4	4	4	3	4	3	2	2	2	2	2	2	2	2	2	3	4	5	5	5	3.5	6.1	
14-Oct	4	3	3	3	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2.2	3.8	
15-Oct	3	4	4	5	5	4	4	4	4	4	4	4	3	3	2	2	3	4	3	5	7	8	9	9	4.5	9.2	
16-Oct	9	10	10	9	10	9	8	8	7	5	4	2	2	2	2	2	2	2	2	4	6	8	9	9	5.9	9.6	
17-Oct	11	12	11	9	8	7	7	6	7	8	7	6	5	4	4	3	2	2	2	3	4	5	6	7	6.1	11.8	
18-Oct	8	8	8	7	7	7	7	8	7	6	4	4	3	3	2	2	2	3	3	4	5	5	6	6	5.2	8.4	
19-Oct	5	6	6	6	6	6	6	6	7	6	5	3	2	2	2	2	2	2	3	4	5	5	6	6	4.5	6.7	
20-Oct	8	8	8	7	5	4	4	4	4	3	3	2	2	2	2	2	2	2	3	4	5	6	6	6	4.2	8.3	
21-Oct	6	6	5	5	4	4	4	4	4	3	2	2	2	2	2	2	2	2	2	4	5	5	5	5	3.6	6.0	
22-Oct	7	8	9	9	9	9	8	9	8	6	4	3	2	2	2	2	2	3	6	7	8	10	11	13	6.5	13.1	
23-Oct	14	13	13	13	13	12	12	10	10	8	5	4	3	2	2	2	2	4	7	9	11	12	14	14	8.7	14.2	
24-Oct	13	14	14	14	13	12	10	10	11	10	9	7	6	6	6	5	5	5	6	8	8	10	10	9	9.3	14.1	
25-Oct	9	8	8	8	8	8	8	7	6	5	3	2	1	1	1	1	1	2	2	3	3	4	4	4	4.7	8.8	
26-Oct	5	6	6	6	5	5	5	4	5	4	4	4	3	3	3	3	2	2	3	4	4	3	3	3	3.9	6.2	
27-Oct	2	2	2	2	2	2	2	2	2	3	3	2	2	2	2	2	2	2	4	5	6	6	6	6	2.9	6.3	
28-Oct	5	4	4	3	3	3	2	2	2	3	3	3	4	4	3	3	4	5	5	6	6	6	6	6	4.1	6.3	
29-Oct	7	7	6	6	6	5	5	5	4	3	2	2	2	2	2	2	2	2	2	3	3	4	3	3	3.6	6.7	
30-Oct	4	4	4	4	3	3	3	3	4	4	3	2	2	2	2	2	2	3	4	4	4	4	4	4	3.2	4.4	
31-Oct	4	4	3	4	3	3	3	3	4	4	4	3	3	3	3	3	3	4	4	4	4	4	4	4	3.5	4.4	
6.4 6.8 6.7 6.4 6.1 5.8 5.5 5.4 5.3 4.9 4.2 3.5 3.0 2.7 2.6 2.5 2.6 3.0 3.6 4.5 5.2 5.7 5.9 6.0																								Diurnal Average			
13.7 14.1 14.0 13.7 13.1 12.3 11.7 10.5 10.9 10.0 8.7 8.0 7.8 7.0 7.0 6.6 6.8 6.4 7.2 9.2 11.3 12.5 13.8 14.2																								Diurnal Maximum			
Alberta Ambient Air Quality Guideline (AAAQG): 1-hr 80 µg/m³ Alberta Ambient Air Quality Objective (AAAQO): 24-hr 30 µg/m³																											

Hourly Averages

Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Beaverlodge - October 2018



Hourly Maximums


Particulate Matter 2.5 (PM_{2.5}) - µg/m³

Beaverlodge - October 2018

Maximum Value: 15.1 μg/m ³ on Oct 24 00:00																				Maximum Daily Average: 10.0 μg/m ³ on Oct 24					Hours in Service: 744	
Minimum Value: 1 μg/m ³ on Oct 25 16:00																				Minimum Daily Average: 2.3 μg/m ³ on Oct 14					Hours of Data: 744	
Maximum Diurnal Average: 7.1 μg/m ³ at hour 1																				Minimum Diurnal Average: 2.7 μg/m ³ at hour 16					Hours of Missing Data: 0	
Monthly Average: 5.21 μg/m ³																				Percentiles: P ₁ = 1.6 P ₁₀ = 1.9 Q ₁ = 2.9 Median = 4.6 Q ₃ = 7.0 P ₉₀ = 9.0 P ₉₉ = 14.1					Hours of Calibration: 0	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	4	4	4	5	4	4	4	4	4	4	3	3	3	3	2	2	3	4	4	4	5	5	4	4	3.8	4.9
2-Oct	4	4	4	3	3	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	3	4	4	4	2.6	3.9
3-Oct	4	3	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	4	5	7	7	6	3.0	6.6
4-Oct	7	8	7	7	7	7	9	13	10	10	8	7	6	3	3	3	4	5	7	7	9	10	9	9	7.2	12.7
5-Oct	9	8	8	8	8	8	8	7	6	4	3	2	2	2	2	2	2	2	4	6	7	8	8	9	5.5	8.5
6-Oct	9	9	9	8	7	7	7	6	6	6	5	4	3	2	2	3	3	3	4	5	6	7	7	7	5.6	8.8
7-Oct	8	8	9	9	9	9	9	8	8	7	8	6	6	6	6	6	4	4	4	4	4	4	4	4	6.7	9.5
8-Oct	5	5	5	4	4	4	4	4	4	2	2	2	2	2	2	2	2	3	3	3	3	3	4	4	3.3	4.8
9-Oct	8	9	9	9	9	10	9	9	10	10	9	8	7	6	6	5	5	4	4	5	5	5	6	6	7.1	9.8
10-Oct	7	7	8	8	8	8	8	8	7	8	8	9	8	8	8	7	7	7	7	8	8	8	8	8	7.6	8.8
11-Oct	10	10	10	10	9	8	7	7	6	5	5	4	3	2	1	1	1	1	2	2	3	3	3	3	4.8	9.8
12-Oct	5	5	5	6	6	6	5	5	6	6	6	6	5	5	5	4	6	6	6	5	5	5	5	5	5.4	6.1
13-Oct	6	6	6	5	5	4	4	4	4	4	3	2	2	2	2	2	2	2	2	4	5	5	6	6	3.8	6.3
14-Oct	5	3	3	3	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2.3	5.3
15-Oct	4	5	5	5	5	4	5	4	5	5	5	4	4	3	3	2	4	4	4	7	8	10	10	11	5.1	11.0
16-Oct	10	10	10	10	10	10	9	8	7	6	5	3	2	2	2	2	4	2	3	5	8	9	9	11	6.5	10.7
17-Oct	11	12	12	10	8	7	7	7	8	8	7	7	6	5	5	4	3	2	3	4	5	6	7	8	6.7	12.4
18-Oct	8	9	9	8	7	7	7	8	8	6	5	4	4	3	3	2	3	3	4	5	5	6	6	6	5.7	9.3
19-Oct	6	7	6	6	6	7	6	7	8	7	6	4	2	2	2	2	2	3	4	5	5	6	6	7	5.0	8.0
20-Oct	8	9	9	7	6	5	4	4	6	5	3	3	5	2	2	2	2	2	4	5	6	6	7	7	4.9	9.0
21-Oct	6	6	6	5	5	4	4	4	4	4	2	2	2	2	2	2	2	2	3	4	6	6	5	6	4.0	6.3
22-Oct	8	9	9	9	9	10	9	10	9	7	5	3	2	2	2	2	2	4	7	8	9	11	12	14	7.2	14.2
23-Oct	15	13	13	14	13	13	12	11	12	9	7	5	3	2	2	2	3	6	9	10	12	13	15	15	9.6	15.1
24-Oct	14	15	15	15	14	13	11	11	11	11	9	8	7	6	6	6	6	6	7	8	10	10	11	10	10.0	15.1
25-Oct	9	9	9	9	9	8	8	8	7	6	4	3	2	2	1	1	1	2	3	3	4	4	4	4	5.0	9.3
26-Oct	7	7	6	6	6	5	5	5	5	5	4	4	6	3	3	3	2	3	4	4	4	4	3	3	4.4	6.6
27-Oct	3	2	2	2	2	2	2	2	3	3	3	2	2	2	2	2	2	3	5	6	7	7	6	6	3.1	7.0
28-Oct	6	4	4	3	4	3	3	2	3	3	4	4	4	4	4	3	5	6	6	6	7	7	7	7	4.5	7.0
29-Oct	7	7	7	7	6	6	5	5	4	4	3	2	2	2	2	2	2	2	2	3	4	4	4	3	3.9	7.1
30-Oct	4	4	4	4	4	3	3	3	4	4	4	2	2	2	2	2	3	4	4	4	5	5	5	4	3.6	4.6
31-Oct	4	4	4	4	3	3	4	4	4	5	4	4	3	3	3	3	3	4	4	5	5	4	4	5	3.9	4.9
7.1 7.1 7.1 6.8 6.5 6.2 5.9 5.9 6.0 5.5 4.7 3.9 3.5 3.0 2.8 2.7 3.1 3.4 4.1 5.0 5.7 6.2 6.3 6.5																									Diurnal Average	
14.7 14.9 15.1 14.6 13.7 13.2 12.4 12.7 12.2 11.0 9.5 8.8 8.3 7.5 7.5 7.3 7.3 7.1 8.8 10.0 12.4 13.2 14.9 15.1																									Diurnal Maximum	


Hourly Maximums

Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Beaverlodge - October 2018



Pollutant Rose

Particulate Matter 2.5 (PM_{2.5}) - µg/m³
Beaverlodge - October 2018



Hourly Averages

External Temperature (ET) - °C


Beaverlodge - October 2018

Maximum Value: 20.7 °C on Oct 16 15:00																				Maximum Daily Average: 14.5 °C on Oct 17					Hours in Service: 744		
Minimum Value: -7 °C on Oct 3 07:00																				Minimum Daily Average: -3.4 °C on Oct 2					Hours of Data: 744		
Maximum Diurnal Average: 8.1 °C at hour 15																				Minimum Diurnal Average: 0.9 °C at hour 8					Hours of Missing Data: 0		
Monthly Average: 3.99 °C																				Percentiles: P ₁ = -6.6 P ₁₀ = -2.1 Q ₁ = -0.2 Median = 2.7 Q ₃ = 8.3 P ₉₀ = 12.0 P ₉₉ = 18.2					Hours of Calibration: 0		
																				Percent Operational Time: 100.0							
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	0	-1	-1	-1	-1	-1	-1	-2	-1	-1	0	0	-1	0	-1	0	0	-1	-2	-2	-3	-3	-3	-4	-1.3	0.0	
2-Oct	-4	-4	-5	-4	-4	-4	-4	-4	-4	-4	-3	-2	-2	-1	0	0	0	-1	-3	-4	-5	-5	-6	-7	-3.4	-0.2	
3-Oct	-7	-7	-7	-7	-7	-7	-7	-7	-6	-4	-1	1	4	5	6	6	6	5	3	2	2	2	2	1	-0.9	6.2	
4-Oct	1	-1	-2	-2	-2	-2	-3	-4	-2	1	3	5	7	9	9	9	9	8	6	3	2	1	0	0	2.3	9.4	
5-Oct	-1	-2	-2	-3	-4	-4	-4	-4	-4	-2	-2	0	2	4	5	5	5	4	2	0	0	0	-1	-1	-0.4	4.9	
6-Oct	-1	-1	-1	0	0	1	0	1	2	4	6	8	9	10	9	8	8	7	5	4	3	2	2	2	3.7	9.6	
7-Oct	1	1	0	0	0	0	0	0	0	0	0	0	-1	-1	-1	-1	-1	-1	-2	-2	-2	-2	-2	-2	-0.5	1.1	
8-Oct	-2	-2	-2	-2	-2	-2	-2	-2	-2	-2	-1	-1	0	1	1	1	0	0	0	0	-1	-1	-1	-1	-2	-1.1	1.3
9-Oct	-2	-2	-2	-2	-2	-2	-2	-2	-2	-2	-2	-1	-1	-1	0	0	0	0	-1	-1	-1	-2	-2	-2	-2	-1.4	0.0
10-Oct	-2	-2	-2	-2	-2	-2	-2	-2	-2	-2	-1	-1	-1	0	0	0	0	0	0	-1	-1	-1	-1	-2	-1.2	0.0	
11-Oct	-3	-3	-3	-4	-4	-3	-3	-3	-2	-1	1	3	6	7	8	9	9	8	6	6	6	6	6	6	2.5	9.3	
12-Oct	6	5	5	5	5	5	5	2	0	0	1	0	0	0	0	1	1	1	1	0	0	1	0	1	1.9	6.1	
13-Oct	1	1	1	2	1	1	1	2	2	4	5	7	8	8	9	9	9	8	5	4	2	1	1	0	3.9	9.2	
14-Oct	0	0	0	0	0	2	2	3	5	8	10	12	13	13	13	13	13	12	12	12	12	11	11	11	7.8	12.9	
15-Oct	11	11	11	11	11	12	11	12	12	13	14	14	15	16	16	16	15	15	13	12	11	11	11	9	12.6	16.2	
16-Oct	9	9	8	7	6	5	5	6	8	9	14	18	19	21	21	20	19	17	15	14	13	12	13	13	12.5	20.7	
17-Oct	11	12	12	13	13	12	11	11	13	15	16	18	19	18	18	18	18	18	16	14	14	13	13	12	14.5	18.9	
18-Oct	12	12	13	13	11	10	10	7	8	11	12	12	12	12	13	13	12	10	9	8	6	6	6	5	10.2	13.3	
19-Oct	4	2	2	2	1	-1	1	-1	0	2	5	7	10	11	12	13	13	11	10	10	10	9	8	7	6.1	12.6	
20-Oct	5	5	10	12	11	9	8	9	9	11	11	12	12	12	13	12	12	11	9	7	7	7	4	4	9.3	12.6	
21-Oct	4	3	4	4	4	3	0	0	1	3	6	8	9	11	11	11	9	8	7	6	5	5	5	4	5.4	10.9	
22-Oct	3	1	0	1	-1	-2	-2	-3	-2	2	3	5	7	10	12	12	10	9	7	8	8	5	4	1	4.1	11.9	
23-Oct	2	2	0	0	0	0	0	0	0	2	4	7	9	12	14	14	13	11	9	8	7	5	5	3	5.2	14.2	
24-Oct	2	-1	1	1	0	1	1	0	1	4	9	10	12	14	14	14	14	12	10	9	9	4	6	5	6.3	14.1	
25-Oct	5	4	3	2	2	3	3	3	3	7	8	10	12	13	13	10	10	9	8	7	5	3	3	3	6.2	12.9	
26-Oct	3	4	3	2	1	1	2	0	1	1	0	2	3	4	5	5	6	4	4	4	5	6	5	6	3.2	6.4	
27-Oct	6	6	6	5	5	5	5	5	5	6	7	8	9	9	9	9	8	6	4	3	2	2	1	1	5.6	9.2	
28-Oct	1	2	2	2	2	3	3	2	2	0	1	0	0	0	1	1	2	2	2	1	0	-1	-1	-1	1.0	3.0	
29-Oct	-1	-1	-2	-3	-3	-2	-2	-2	-1	0	3	5	9	10	10	9	7	6	5	4	3	3	3	2	2.6	10.0	
30-Oct	3	3	3	3	3	4	3	2	3	4	7	9	9	9	10	9	7	6	6	5	4	3	3	3	5.1	9.7	
31-Oct	2	2	2	3	2	2	1	1	1	1	2	3	4	4	4	4	3	2	2	2	2	2	1	1	2.1	3.8	
2.2 1.9 1.9 1.9 1.6 1.5 1.2 0.9 1.5 2.9 4.4 5.8 7.0 7.6 8.1 8.1 7.7 6.6 5.4 4.6 4.1 3.5 3.1 2.6																									Diurnal Average		
11.7 12.3 13.2 13.3 13.2 11.9 11.5 11.8 12.7 14.8 15.9 18.3 19.3 20.5 20.7 20.0 18.8 17.7 15.7 14.2 14.3 13.2 13.2 12.8																									Diurnal Maximum		

Hourly Averages

External Temperature (ET) - °C

Beaverlodge - October 2018




Hourly Averages

Relative Humidity (RH) - %
Beaverlodge - October 2018

Maximum Value: 94.8 % on Oct 6 20:00																								Maximum Daily Average: 67.6 % on Oct 25										Hours in Service: 744	
Minimum Value: 27 % on Oct 23 15:00																								Minimum Daily Average: 43.6 % on Oct 20										Hours of Data: 263	
Maximum Diurnal Average: 67.8 % at hour 5																								Minimum Diurnal Average: 41.1 % at hour 15										Hours of Missing Data: 481	
Monthly Average: 56.40 %																								Percentiles: P ₁ = 30.6 P ₁₀ = 36.8 Q ₁ = 45.5 Median = 56.4 Q ₃ = 66.8 P ₉₀ = 72.8 P ₉₉ = 89.0										Hours of Calibration: 0	
																								Percent Operational Time: 35.4											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum									
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24											
1-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
2-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
3-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
4-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
5-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
6-Oct	N	N	N	N	N	N	N	N	N	N	N	56	49	49	57	58	64	73	94	95	N	N	N	N	--	94.8									
7-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
8-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
9-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
10-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
11-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
12-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
13-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
14-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
15-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
16-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
17-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	--									
18-Oct	N	N	N	N	N	N	N	N	N	59	50	43	41	38	36	34	33	35	40	41	44	53	54	50	56	--	58.6								
19-Oct	58	65	64	67	73	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	72.8									
20-Oct	N	N	N	N	N	61	62	55	45	42	40	36	35	33	33	32	33	36	40	45	47	45	54	56	43.6	61.5									
21-Oct	55	58	54	51	50	51	63	62	62	58	54	45	35	31	30	32	37	40	37	39	43	45	45	47	46.8	62.6									
22-Oct	51	56	61	59	64	75	78	84	82	67	58	50	44	36	33	33	39	43	48	44	36	46	56	66	54.5	84.4									
23-Oct	65	62	67	68	67	70	70	69	69	63	52	44	38	31	27	30	33	39	43	42	46	52	54	63	52.6	69.9									
24-Oct	67	82	72	73	75	69	64	69	74	66	57	55	47	41	39	40	43	49	53	56	55	77	69	70	60.8	82.4									
25-Oct	74	79	85	89	85	83	87	89	86	67	57	52	48	46	45	53	52	54	56	58	67	70	72	69	67.6	89.0									
26-Oct	67	60	59	63	65	62	61	67	64	69	78	73	67	61	64	67	61	68	72	72	65	57	65	57	65.2	78.3									
27-Oct	52	51	51	50	51	48	48	50	51	50	45	40	38	36	35	36	38	45	50	53	56	58	61	63	48.2	62.8									
28-Oct	61	59	62	61	60	54	54	57	58	67	65	81	N	N	N	N	N	N	N	N	N	N	N	N	--	81.2									
29-Oct	N	N	N	N	N	N	N	N	N	N	N	N	57	48	45	46	53	57	63	65	68	69	71	75	--	75.4									
30-Oct	71	71	86	N	N	N	N	N	N	N	67	51	47	50	50	46	53	57	59	62	65	66	68	69	--	85.6									
31-Oct	72	72	75	72	87	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--	87.1									
63.1 65.1 66.9 65.3 67.8 63.7 65.1 66.8 64.9 59.9 56.0 51.9 45.2 41.4 41.1 42.1 45.2 50.0 54.7 56.3 54.7 58.1 60.4 62.7																								Diurnal Average											
74.5 82.4 85.6 89.0 87.1 83.1 86.5 89.0 86.1 68.7 78.3 81.2 67.4 60.8 64.3 66.8 63.8 72.9 93.8 94.8 68.1 77.2 72.2 75.4																								Diurnal Maximum											
N - Not Valid																																			

Hourly Averages

Relative Humidity (RH) - %
Beaverlodge - October 2018



Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Beaverlodge - October 2018

Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1 Spd	11	9	7	2	6	6	5	7	7	9	8	8	6	5	2	5	2	8	15	18	15	11	10	7	7.2	18.3
Dir	31	21	21	343	359	4	357	359	9	359	2	1	335	297	286	45	338	19	32	32	31	32	29	40	17	32
2 Spd	7	4	9	8	7	7	7	6	4	5	7	6	4	7	9	12	13	15	11	8	7	7	7	8	3.2	15.2
Dir	55	47	32	20	6	355	356	357	332	357	22	331	301	282	242	235	228	230	237	242	227	236	245	217	281	230
3 Spd	9	8	7	5	6	4	1	1	3	2	6	5	8	7	7	7	6	6	7	8	7	4	2	3	3.8	8.8
Dir	212	211	211	198	214	204	189	113	193	164	200	166	163	164	181	179	145	107	88	84	99	170	257	78	168	212
4 Spd	4	3	3	4	3	3	2	3	4	2	5	5	7	15	17	15	12	11	10	13	8	11	12	9	4.9	17.2
Dir	46	202	64	54	56	66	155	59	63	138	234	237	282	308	289	299	351	344	11	15	352	331	327	330	333	289
5 Spd	8	13	11	11	4	6	5	5	1	1	2	3	2	2	5	7	10	10	8	9	8	7	8	6	2.8	12.8
Dir	325	325	332	343	82	127	134	132	272	264	153	133	130	167	140	122	127	110	102	80	85	122	126	120	97	325
6 Spd	5	5	3	4	5	5	3	3	1	2	4	6	10	7	7	5	4	7	12	12	14	13	12	13	3.7	13.5
Dir	137	134	130	41	290	23	197	29	348	24	304	305	314	345	13	28	69	81	94	95	100	82	86	90	67	100
7 Spd	6	3	5	2	10	9	11	8	8	4	6	9	13	11	11	11	13	10	8	7	11	12	9	9	7.4	12.8
Dir	98	116	112	79	110	105	96	89	80	68	14	29	34	39	34	35	27	37	29	19	37	41	56	66	55	34
8 Spd	12	13	11	14	14	13	12	13	13	15	16	18	19	20	23	26	22	24	26	23	22	21	22	21	17.8	26.4
Dir	67	61	62	71	73	71	74	82	92	93	95	88	90	92	103	98	87	88	89	91	92	90	88	87	87	98
9 Spd	18	20	19	17	15	15	19	19	18	18	18	19	18	19	19	20	19	16	17	16	17	17	14	12	17.2	19.9
Dir	83	86	86	83	84	85	89	90	90	90	89	94	97	97	101	106	107	111	103	103	108	110	112	95	86	
10 Spd	11	8	9	3	5	5	4	2	2	4	5	6	4	4	3	3	3	5	3	3	1	3	4	3	1.0	10.8
Dir	115	102	81	111	104	86	35	29	205	246	301	304	294	326	355	323	329	346	267	247	105	145	139	108	64	115
11 Spd	3	4	3	3	3	3	4	4	3	4	8	7	16	19	20	21	21	15	9	5	5	4	8	8	5.1	21.3
Dir	123	108	107	122	59	149	143	136	160	148	175	201	240	244	258	264	261	269	246	259	251	292	332	352	246	261
12 Spd	6	6	4	4	5	5	9	22	16	13	19	19	19	19	18	15	11	9	5	1	5	4	3	4	8.4	21.9
Dir	357	43	136	107	54	42	9	0	339	324	322	324	327	330	323	327	323	318	293	315	39	300	288	302	337	0
13 Spd	8	10	10	12	12	9	4	5	8	9	11	12	12	11	12	11	9	9	10	10	11	7	7	4	8.3	12.3
Dir	292	291	294	308	305	306	277	299	302	300	324	309	324	320	304	288	283	283	264	267	246	223	219	200	291	309
14 Spd	3	2	2	2	3	2	1	1	6	5	14	24	23	33	33	37	31	27	26	22	27	22	19	18	14.6	37.2
Dir	96	127	112	162	70	85	181	135	258	203	265	265	273	274	264	264	260	263	274	275	270	270	273	277	267	264
15 Spd	24	20	17	17	12	20	17	25	28	31	34	31	27	26	28	29	24	21	14	10	14	9	5	1	19.7	33.5
Dir	280	284	276	275	288	275	271	260	262	259	264	269	282	288	292	291	292	287	292	286	271	272	287	253	277	264
16 Spd	5	5	1	0	4	5	5	2	1	4	6	27	25	28	29	29	24	21	10	12	6	7	13	10	9.5	29.4
Dir	289	287	38	182	60	53	51	65	66	51	264	265	253	267	274	259	253	254	260	255	200	231	250	239	261	274
17 Spd	4	13	11	16	10	6	8	4	8	6	7	17	20	16	14	13	22	10	13	12	20	11	9	8	10.6	21.9
Dir	216	244	257	242	239	215	211	190	241	252	226	245	247	251	248	228	263	280	286	269	271	278	325	311	254	263
18 Spd	6	16	22	27	17	9	1	3	11	14	25	22	20	20	18	19	10	8	9	5	3	1	3	2	10.7	27.3
Dir	305	300	280	264	275	282	21	164	250	258	268	285	305	302	305	311	329	355	354	323	324	350	297	338	291	264
19 Spd	1	5	6	2	1	2	7	3	4	3	4	4	11	14	19	20	17	12	12	14	12	5	0	2	6.3	19.7
Dir	270	54	51	81	30	39	35	283	54	75	81	153	136	121	113	114	114	99	97	105	104	103	64	307	102	114
20 Spd	1	1	21	23	25	19	20	30	33	43	45	50	49	50	45	39	32	29	26	11	11	9	8	3	25.6	50.0
Dir	306	308	276	262	257	246	246	250	249	261	257	256	257	262	265	266	261	261	262	274	267	272	245	319	260	256
21 Spd	2	2	7	6	6	6	2	2	2	2	2	5	4	2	3	5	7	7	7	7	9	9	8	5	1.2	8.8
Dir	263	240	295	308	304	315	252	57	130	141	187	285	314	160	220	130	101	95	73	82	89	90	90	160	86	90
22 Spd	4	2	1	3	1	2	3	1	4	3	3	4	4	2	3	5	8	5	6	5	3	6	0	2	1.7	7.6
Dir	198	145	174	54	202	281	66	114	49	104	154	170	173	169	175	179	170	132	80	70	33	26	197	214	127	170


Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Beaverlodge - October 2018


Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum			
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24					
23 Spd	1	2	1	0	1	2	3	2	1	2	2	2	3	2	2	4	5	5	4	1	5	3	13	8	0.6	12.7			
Dir	303	47	50	120	82	37	46	54	67	84	168	171	155	129	146	173	171	129	84	68	258	6	317	17	66	317			
24 Spd	2	3	4	1	0	2	3	4	5	4	1	1	6	13	10	9	6	4	3	3	2	3	6	2	1.1	13.5			
Dir	337	28	21	253	107	58	72	63	46	57	131	142	282	269	274	270	239	209	227	40	337	166	54	119	299	269			
25 Spd	5	2	3	4	4	3	3	1	2	2	2	4	4	3	19	27	22	23	15	21	14	6	5	8	5.7	27.4			
Dir	57	106	123	83	132	86	72	12	70	125	196	186	194	196	270	259	245	240	246	246	233	195	198	222	236	259			
26 Spd	7	10	4	2	13	3	3	3	3	3	3	1	6	6	1	5	5	6	2	6	23	35	27	30	6.4	35.3			
Dir	226	242	216	182	243	22	69	95	85	50	224	2	289	322	158	159	176	200	141	210	236	247	245	246	238	247			
27 Spd	33	30	32	29	30	31	32	29	30	31	37	37	29	23	22	16	10	11	10	3	1	4	3	2	20.2	36.8			
Dir	250	253	254	255	255	261	262	259	260	265	270	274	280	285	292	300	281	260	261	314	72	76	140	76	266	270			
28 Spd	5	8	7	6	13	27	24	21	17	8	1	5	7	4	3	4	3	2	5	6	4	9	7	2	3.2	26.8			
Dir	51	43	77	82	82	86	87	96	95	193	229	221	229	185	235	284	246	314	338	326	315	265	289	293	85	86			
29 Spd	4	2	3	1	3	3	2	5	2	3	3	9	17	22	26	26	16	8	1	3	3	2	4	3	3.8	26.1			
Dir	47	86	79	143	59	68	76	54	95	166	198	203	252	260	267	262	268	288	202	247	60	83	162	83	257	262			
30 Spd	3	2	2	7	13	18	14	12	12	14	28	32	35	31	38	32	22	21	15	8	9	6	4	6	15.4	38.0			
Dir	62	236	196	229	252	253	250	255	252	246	261	260	260	262	261	261	254	259	248	240	225	230	218	216	254	261			
31 Spd	4	6	4	3	1	2	4	3	3	2	3	4	2	3	5	7	7	3	6	5	5	11	15	17	1.9	17.5			
Dir	173	204	183	223	148	203	200	81	188	49	124	181	88	37	50	41	1	347	354	339	325	318	316	330	331	330			
Spd	0.6	1.3	1.7	2.3	1.7	0.8	0.2	0.5	1.9	3.1	5.8	7.8	8.2	8.3	8.7	7.7	5.7	3.9	1.8	1.2	1.1	1.3	1.6	0.8	Diurnal Average				
Dir	321	297	294	276	276	318	69	339	266	263	266	267	271	276	274	268	261	268	293	318	259	268	291	297	Diurnal Maximum				
Spd	33.4	29.9	32.0	28.7	30.0	31.2	32.0	29.8	33.5	43.0	45.0	50.0	48.9	49.6	44.9	38.7	31.6	29.0	26.2	23.1	27.1	35.3	26.5	30.3	Diurnal Maximum				
Dir	250	253	254	255	255	261	262	250	249	261	257	256	257	262	265	266	261	261	89	91	270	247	245	246					
Maximum Speed Value: 50 km/h on Oct 20 12:00																		Minimum Speed Value: 0 km/h on Oct 23 04:00										Hours in Service: 744	
Maximum Daily Speed Average: 25.6 km/h on Oct 20																		Minimum Daily Speed Average: 0.6 km/h on Oct 23										Hours of Data: 744	
Maximum Diurnal Speed Average: 8.7 km/h at hour 15																		Minimum Diurnal Speed Average: 0.2 km/h at hour 7										Hours of Missing Data: 0	
Monthly Average Velocity: 3.15 km/h 273.5 deg																		Speed Percentiles: P ₁ = 0.7 P ₁₀ = 2.1 Q ₁ = 3.5 Median = 7.0 Q ₃ = 13.7 P ₉₀ = 22.9 P ₉₉ = 37.6										Percent Operational Time: 100.0	
All monthly, daily, and diurnal averages have been calculated using vector methods																													
Frequency Distribution																													
Speed Range (km/h)																													
Direction	0 to 5		5 to 11		11 to 19		19 to 28		28 to 38		> 38		Total																
North	17		33		3		1		0		0		54																
NorthEast	53		27		11		0		0		0		91																
East	48		29		41		19		0		0		137																
SouthEast	45		12		2		2		0		0		61																
South	50		15		0		0		0		0		65																
SouthWest	22		37		14		8		2		0		83																
West	19		34		33		40		36		7		169																
NorthWest	27		26		28		3		0		0		84																
Total	281		213		132		73		38		7		744																

Wind Rose

Wind Speed (WS) (km/h)
Beaverlodge - October 2018



Wind Speed Classes (km/h)



Hourly Averages - Wind Speed (Scalar)

Wind Speed (km/h)

Beaverlodge - October 2018

Maximum Speed: 50 km/h on Oct 20 12:00																				Maximum Daily Speed Average: 26.5 km/h on Oct 20										Hours in Service: 744	
Minimum Speed: 1 km/h on Oct 5 09:00																				Minimum Daily Speed Average: 3.6 km/h on Oct 23										Hours of Data: 744	
Maximum Diurnal Speed Average: 15.9 km/h at hour 16																				Minimum Diurnal Speed Average: 7.8 km/h at hour 1										Hours of Missing Data: 0	
Monthly Average Speed: 10.56 km/h																				Percentiles: P ₁ = 1.8 P ₁₀ = 2.9 Q ₁ = 4.3 Median = 7.4 Q ₃ = 13.9 P ₉₀ = 23.0 P ₉₉ = 37.8										Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	11	9	7	3	6	6	5	7	7	9	8	8	7	6	5	5	4	10	15	18	15	11	10	7	8.3	18.4					
2-Oct	7	5	9	9	7	7	7	6	4	6	7	7	6	8	10	12	14	15	11	8	7	7	8	8	8.1	15.4					
3-Oct	9	8	7	5	6	4	2	2	3	2	6	6	9	8	8	8	7	7	7	8	7	4	2	3	5.8	8.8					
4-Oct	5	4	3	4	3	3	2	3	4	3	6	6	10	15	18	17	13	11	10	13	8	12	12	10	8.1	17.9					
5-Oct	9	13	12	11	5	6	5	5	1	2	4	4	4	4	6	8	11	10	8	9	8	7	8	6	6.9	12.8					
6-Oct	5	5	4	5	8	5	4	4	3	2	4	7	10	7	7	5	4	7	12	12	14	13	12	13	7.2	13.7					
7-Oct	7	4	5	3	10	10	11	8	8	5	8	10	13	12	11	11	13	10	8	7	11	12	9	9	9.0	12.9					
8-Oct	12	13	11	14	14	13	12	13	13	15	16	18	19	20	23	26	22	24	26	23	22	21	22	21	18.2	26.5					
9-Oct	18	20	19	17	15	16	19	19	18	18	18	19	18	19	19	20	19	16	17	16	17	17	14	12	17.5	20.0					
10-Oct	11	8	9	4	5	5	4	3	3	4	5	6	5	5	3	4	3	5	3	3	2	3	4	3	4.5	10.9					
11-Oct	3	4	3	3	3	3	4	4	4	4	8	8	16	19	20	21	22	16	9	5	6	4	8	8	8.5	21.6					
12-Oct	7	6	5	4	6	6	10	23	16	13	19	19	19	19	18	15	11	9	5	4	5	5	3	4	10.5	22.6					
13-Oct	8	10	10	12	12	9	5	5	8	9	11	13	13	12	12	12	10	9	11	11	11	7	7	5	9.7	13.0					
14-Oct	3	2	2	3	3	2	2	2	6	5	14	24	23	33	33	37	31	27	26	22	27	22	19	18	16.2	37.3					
15-Oct	24	20	17	17	12	20	17	26	28	31	34	31	27	27	29	29	24	21	14	10	14	9	5	2	20.3	33.6					
16-Oct	5	5	3	4	4	5	5	3	2	4	8	28	25	28	30	29	24	21	11	12	6	8	13	11	12.2	29.6					
17-Oct	5	13	12	16	10	7	8	4	9	8	9	18	21	16	15	14	22	10	13	12	20	11	9	8	12.1	22.0					
18-Oct	6	16	23	27	17	10	2	3	11	14	25	23	21	20	18	19	11	8	9	5	4	2	4	4	12.6	27.4					
19-Oct	4	5	6	2	2	5	7	4	5	4	4	5	11	14	19	20	17	12	12	14	12	5	2	3	8.1	19.8					
20-Oct	3	4	21	23	25	19	20	30	34	43	45	50	49	50	45	39	32	29	26	12	11	10	9	6	26.5	50.2					
21-Oct	3	4	7	7	6	6	5	4	3	2	2	5	5	4	4	5	7	7	7	8	9	9	9	5	5.5	8.9					
22-Oct	4	3	3	3	1	4	4	2	4	3	4	5	4	3	3	5	8	5	6	5	3	6	2	2	4.0	7.6					
23-Oct	2	2	2	2	2	2	3	2	2	2	3	2	3	2	2	4	5	6	5	3	6	4	13	10	3.6	12.9					
24-Oct	5	4	5	2	1	2	3	4	6	4	2	2	6	14	10	9	7	4	3	3	5	4	6	3	4.8	13.8					
25-Oct	5	3	3	4	4	3	5	4	3	3	3	4	5	4	19	28	22	23	15	21	14	7	6	8	9.0	27.7					
26-Oct	7	12	6	5	13	7	3	4	4	7	4	3	8	6	4	5	5	6	4	7	24	35	27	30	9.9	35.4					
27-Oct	33	30	32	29	30	31	32	29	30	31	37	37	29	24	22	16	10	11	10	5	2	4	3	3	21.7	37.0					
28-Oct	5	9	8	6	13	27	24	21	18	8	4	7	8	4	4	4	4	3	5	7	7	9	7	5	9.0	26.9					
29-Oct	4	2	3	4	4	4	3	5	3	4	4	9	18	22	26	26	16	8	3	5	3	2	4	3	7.7	26.5					
30-Oct	3	3	3	7	13	18	14	12	12	14	28	33	36	31	38	32	22	21	15	8	9	6	5	6	16.2	38.2					
31-Oct	5	7	5	6	2	4	5	3	4	3	4	4	3	3	5	7	7	3	6	5	6	12	15	18	5.8	17.6					
7.8 8.1 8.5 8.5 8.5 8.6 8.2 8.6 8.8 9.2 11.4 13.5 14.5 14.8 15.8 15.9 13.7 12.1 10.8 9.7 10.2 9.3 8.9 8.2																									Diurnal Average						
33.5 30.0 32.1 28.8 30.1 31.2 32.0 30.0 33.6 43.4 45.2 50.2 49.2 49.9 45.1 38.9 31.8 29.1 26.2 23.2 27.2 35.4 26.6 30.4																									Diurnal Maximum						

All monthly, daily, and diurnal averages have been calculated using scalar methods

Hourly Standard Deviations

Wind Direction (WD) - deg
Beaverlodge - October 2018

Maximum Value: 97.4 deg on Oct 25 08:00																			Hours in Service: 744						
Minimum Value: 2.4 deg on Oct 30 08:00																			Hours of Data: 744						
																			Hours of Missing Data: 0						
Percentiles: P ₁ = 3.5 P ₁₀ = 5.0 Q ₁ = 6.8 Median = 14.8 Q ₃ = 33.9 P ₉₀ = 60.0 P ₉₉ = 88.7																			Hours of Calibration: 0						
																			Percent Operational Time: 100.0						
Day	Hourly Period Ending At (MST)																								Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
1-Oct	5	4	5	43	17	10	10	10	11	10	10	18	29	43	76	30	77	35	4	4	4	6	5	6	77.2
2-Oct	11	28	6	7	7	6	5	7	18	35	17	22	51	43	28	14	18	7	4	6	7	8	18	6	50.7
3-Oct	4	6	24	25	13	20	46	55	21	43	31	26	24	28	29	28	26	17	6	5	17	42	57	58	58.1
4-Oct	64	49	58	11	16	27	30	35	15	42	23	29	59	17	17	24	14	12	10	4	18	16	5	28	63.9
5-Oct	52	4	5	10	41	18	11	29	63	68	74	56	76	73	51	29	14	10	10	4	5	18	10	7	75.8
6-Oct	5	9	41	52	49	29	38	66	69	37	17	22	15	21	10	15	28	15	10	13	8	8	9	9	68.8
7-Oct	50	47	25	51	6	22	13	13	8	26	39	13	7	8	10	9	7	8	9	8	12	7	8	9	51.1
8-Oct	7	8	10	6	4	4	6	6	8	6	6	7	7	7	8	5	7	6	4	5	6	6	5	5	9.8
9-Oct	6	6	6	5	5	6	5	4	4	4	5	7	7	7	7	7	6	7	7	4	5	5	6	6	7.4
10-Oct	6	9	12	59	11	10	26	30	57	25	24	18	36	24	32	28	29	15	51	27	70	23	9	22	69.8
11-Oct	13	13	19	21	16	25	13	13	38	25	9	21	10	7	6	10	8	13	8	17	16	28	8	13	37.7
12-Oct	27	16	39	24	34	35	29	15	6	9	7	7	8	9	7	11	8	8	10	79	11	28	8	18	79.4
13-Oct	7	5	5	6	6	6	27	13	9	9	10	14	20	19	14	20	14	6	15	9	10	5	14	50	49.7
14-Oct	42	23	39	59	32	53	74	58	32	16	11	7	9	6	5	5	6	8	7	8	5	4	5	4	74.0
15-Oct	5	5	6	9	6	5	5	6	5	5	4	6	6	7	9	6	6	6	4	8	3	9	15	95	94.5
16-Oct	19	9	74	84	25	6	43	49	66	15	83	7	11	11	7	5	7	5	19	6	30	18	23	23	84.1
17-Oct	29	5	17	11	13	25	16	58	22	54	71	10	6	7	10	18	5	9	6	17	4	6	16	14	71.3
18-Oct	22	12	8	5	9	24	86	48	16	15	8	8	9	9	8	10	10	7	8	33	56	69	45	55	85.8
19-Oct	88	15	7	37	75	93	30	69	36	33	15	27	13	11	6	6	5	6	5	4	8	43	80	87	93.2
20-Oct	92	77	11	7	8	4	15	6	5	7	5	6	6	7	7	7	5	4	5	10	7	13	45	60	92.2
21-Oct	65	81	24	16	13	13	66	61	46	43	49	25	49	68	59	28	9	15	5	9	6	5	13	13	81.0
22-Oct	34	53	64	43	66	81	71	65	40	32	21	22	30	37	23	18	7	23	6	11	18	18	79	56	81.1
23-Oct	71	34	69	91	53	73	59	49	70	35	25	37	32	34	38	15	8	30	15	96	14	47	11	31	95.6
24-Oct	76	44	60	77	81	73	45	26	13	20	59	54	41	12	11	6	18	15	47	22	63	55	8	35	80.8
25-Oct	21	44	27	21	28	17	68	97	43	38	70	19	27	43	18	8	5	6	12	5	17	45	52	18	97.4
26-Oct	42	45	51	69	12	67	24	30	32	76	64	70	68	30	81	17	18	13	55	18	10	4	3	4	80.5
27-Oct	4	4	4	4	4	4	4	3	4	4	5	6	8	8	12	8	6	5	12	50	92	26	28	39	91.7
28-Oct	13	7	17	6	7	5	7	7	30	16	89	60	48	23	35	22	20	37	10	14	64	7	13	75	89.4
29-Oct	12	38	24	71	49	39	26	30	35	39	59	16	20	12	10	8	12	6	70	75	38	29	32	45	74.9
30-Oct	17	69	66	16	5	3	2	2	5	3	5	5	7	5	6	7	5	5	3	18	10	29	61	21	69.4
31-Oct	23	22	25	67	49	77	30	20	38	57	42	23	53	21	16	12	20	22	5	20	20	8	6	8	76.9
92.2 81.0 74.2 90.8 80.8 93.2 85.8 97.4 70.0 76.1 89.4 70.3 75.8 73.2 80.5 30.4 77.2 36.9 70.4 95.6 91.7 69.3 79.6 94.5																									

PAZA
Evergreen Park Station
Monthly Summary Tables, Graphs and
Roses


Hourly Averages

Sulphur Dioxide (SO₂) - ppb Evergreen Park - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 5.2 ppb on Oct 22 13:00 Maximum Daily Average: 0.7 ppb on Oct 22																				Hours in Service: 744 Hours of Data: 710 Hours of Missing Data: 34 Hours of Calibration: 34 Percent Operational Time: 100.0						
Minimum Value: 0 ppb on Oct 1 01:00 Maximum Diurnal Average: 0.5 ppb at hour 13 Monthly Average: 0.21 ppb																				Minimum Daily Average: 0.0 ppb on Oct 7 Minimum Diurnal Average: 0.0 ppb at hour 4 Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.1 Q ₃ = 0.2 P ₉₀ = 0.5 P ₉₉ = 2.2						
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
3-Oct	1	2	1	0	0	0	0	0	0	A	1	1	2	1	1	0	0	0	0	0	0	0	0	0	0.4	1.7
4-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0.2	0.9
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.2
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0.1	0.7
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.0
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
10-Oct	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0.2	0.6
11-Oct	0	A	0	0	0	0	0	0	0	0	0	0	0	1	2	1	2	0	0	0	0	0	0	0	0.3	1.7
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0.1	0.2
13-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0.1	0.4
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	3	1	0	1	0	1	2	1	A	0	0	0.5	2.5
15-Oct	1	0	0	0	0	0	0	0	3	3	2	1	0	0	0	0	0	0	0	1	A	0	0	0	0.5	3.4
16-Oct	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	2	1	1	0	A	0	0	0	0	0.3	1.6
17-Oct	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0.2	0.9
18-Oct	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.2	0.9
19-Oct	0	0	0	0	0	0	0	0	0	0	0	1	0	C	C	0	A	0	0	0	0	0	0	0	0.1	0.5
20-Oct	0	0	0	0	0	0	0	0	0	1	0	0	1	1	1	A	1	0	0	0	0	0	0	0	0.2	1.1
21-Oct	0	0	0	0	0	0	0	0	0	0	0	0	1	0	A	0	0	0	0	0	0	0	0	0	0.1	0.5
22-Oct	0	0	0	0	0	0	0	0	0	0	0	2	5	A	3	2	2	1	0	0	0	0	0	0	0.7	5.2
23-Oct	0	0	0	0	0	0	0	1	0	0	0	0	A	1	1	1	0	0	0	0	0	0	0	0	0.3	1.1
24-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
25-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5
26-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0.1	0.6
27-Oct	0	0	0	0	0	0	0	0	A	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.9
28-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
29-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	1	0	1	1	0	0	0	0	0	0	0	0.3	1.0
30-Oct	0	0	0	0	0	A	0	0	0	0	0	1	1	1	1	1	1	0	0	0	0	0	0	0	0.3	1.3
31-Oct	0	0	0	0	A	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.8
																									Diurnal Average	
																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 172 ppb 24-hr 48 ppb 30-day 11 ppb																										

Hourly Averages

Sulphur Dioxide (SO₂) - ppb
Evergreen Park - October 2018




Hourly Maximums

Sulphur Dioxide (SO₂) - ppb Evergreen Park - October 2018

Maximum Value: 7.4 ppb on Oct 15 10:00																				Maximum Daily Average: 2.1 ppb on Oct 14										Hours in Service: 744									
Minimum Value: 0 ppb on Oct 14 06:00																				Minimum Daily Average: 0.5 ppb on Oct 7										Hours of Data: 710									
Maximum Diurnal Average: 1.5 ppb at hour 15																				Minimum Diurnal Average: 0.6 ppb at hour 5										Hours of Missing Data: 34									
Monthly Average: 0.94 ppb																				Percentiles: P ₁ = 0.3 P ₁₀ = 0.4 Q ₁ = 0.5 Median = 0.7 Q ₃ = 0.9 P ₉₀ = 1.7 P ₉₉ = 5.4										Hours of Calibration: 34									
																				Percent Operational Time: 100.0																			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum													
1-Oct	1	1	0	1	0	0	1	1	1	1	1	A	1	1	0	1	1	0	1	0	1	1	0	0	0.5	0.7													
2-Oct	0	1	1	1	1	1	1	1	1	1	A	1	1	1	1	3	1	1	1	0	1	1	1	1	2	0.8	2.6												
3-Oct	3	5	4	1	0	1	1	1	1	A	1	1	4	3	2	1	1	1	1	1	1	1	0	1	1.6	5.4													
4-Oct	1	0	1	1	1	1	1	1	A	1	1	2	2	1	2	2	2	1	1	1	1	1	1	1	1.0	2.4													
5-Oct	1	1	1	1	1	1	1	A	0	1	1	1	1	1	1	1	1	1	0	1	0	0	1	0	0.6	0.8													
6-Oct	0	0	0	0	1	1	A	1	1	1	1	1	1	2	1	1	1	1	1	1	0	0	1	0	0.6	1.9													
7-Oct	0	1	0	0	0	A	0	0	0	0	0	0	0	1	1	0	0	1	1	0	1	1	1	1	0.5	0.7													
8-Oct	0	0	1	0	A	0	1	1	1	1	0	1	0	0	1	0	1	1	1	1	1	1	1	1	0.5	0.8													
9-Oct	1	1	1	A	1	1	1	1	1	1	1	1	0	1	0	1	1	1	1	1	1	1	0	1	0.6	0.8													
10-Oct	1	1	A	1	1	0	1	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.4													
11-Oct	1	A	1	1	1	1	1	1	1	1	1	1	1	4	4	3	5	2	1	1	1	1	1	1	1.4	5.3													
12-Oct	A	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	1	1	1	1	A	0.6	1.0													
13-Oct	0	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	1	A	1	0.7	1.0													
14-Oct	1	1	0	1	0	0	1	1	0	1	1	1	3	5	6	1	5	1	6	7	6	A	1	2	2.1	7.0													
15-Oct	2	1	1	1	0	0	1	4	7	7	5	5	1	1	0	0	1	0	1	2	A	1	2	0	1.8	7.4													
16-Oct	0	1	0	1	0	0	0	1	0	1	1	1	2	3	3	5	4	5	1	A	0	1	1	1	1.4	5.0													
17-Oct	1	0	1	1	1	1	1	1	2	2	2	1	1	1	1	1	1	1	A	1	1	1	1	1	0.9	1.8													
18-Oct	1	1	1	1	1	1	2	2	4	5	1	3	1	1	1	1	1	A	1	1	1	1	1	1	1.2	4.8													
19-Oct	1	0	0	0	1	1	1	1	1	1	1	1	C	C	1	A	1	0	0	0	1	0	0	0	0.7	1.1													
20-Oct	0	0	0	0	0	1	1	2	2	3	2	1	1	2	2	A	2	2	1	0	0	0	0	0	1.1	3.1													
21-Oct	1	0	1	0	0	1	0	1	0	1	1	1	1	1	A	1	1	1	0	0	0	0	1	1	0.6	1.5													
22-Oct	0	0	1	0	0	0	1	1	1	1	1	4	7	A	3	3	3	2	1	1	1	1	1	1	1.5	6.9													
23-Oct	1	1	1	1	1	0	1	1	1	1	1	1	A	2	1	1	1	1	1	1	1	1	1	1	0.9	1.6													
24-Oct	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	0.7	0.9													
25-Oct	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.4													
26-Oct	1	1	0	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0.7	1.4													
27-Oct	1	0	0	0	0	2	2	1	A	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	2.3													
28-Oct	0	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.7	0.9													
29-Oct	1	1	1	1	0	1	A	1	1	1	1	1	1	1	1	2	3	2	1	2	1	1	1	1	1.0	2.6													
30-Oct	1	1	1	1	1	A	1	1	1	1	1	2	2	2	3	3	2	1	1	1	1	1	1	1	1.1	2.9													
31-Oct	1	1	1	1	A	1	1	1	1	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.8													
0.8		0.7	0.7	0.6	0.6	0.6	0.8	1.0	1.1	1.3	1.0	1.2	1.3	1.4	1.5	1.3	1.4	1.0	0.8	0.9	0.8	0.6	0.7	0.7	Diurnal Average														
2.8		5.4	4.1	1.5	0.8	1.7	2.3	4.3	6.6	7.4	5.2	4.7	6.9	5.1	5.6	4.7	5.3	5.0	5.8	7.0	5.6	1.0	1.7	2.5	Diurnal Maximum														
C - Calibration		A - Automated Daily Zero Span																																					


Hourly Maximums

Sulphur Dioxide (SO₂) - ppb
Evergreen Park - October 2018




Pollutant Rose

Sulphur Dioxide (SO₂) - ppb
Evergreen Park - October 2018




Pollutant Classes (ppb)



Span Responses

Sulphur Dioxide (SO₂)
Evergreen Park - October 2018



Hourly Averages


Total Reduced Sulphur (TRS) - ppb

Evergreen Park - October 2018

Maximum Value: 1.9 ppb on Oct 11 17:00																			Maximum Daily Average: 0.7 ppb on Oct 22						Hours in Service: 744	
Minimum Value: 0 ppb on Oct 19 05:00																			Minimum Daily Average: 0.2 ppb on Oct 2						Hours of Data: 709	
Maximum Diurnal Average: 0.5 ppb at hour 18																			Minimum Diurnal Average: 0.3 ppb at hour 5						Hours of Missing Data: 35	
Monthly Average: 0.44 ppb																			Percentiles: P ₁ = 0.1 P ₁₀ = 0.3 Q ₁ = 0.3 Median = 0.4 Q ₃ = 0.5 P ₉₀ = 0.6 P ₉₉ = 1.1						Hours of Calibration: 35	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	1	0	0	0	0	0	0	0.3	0.6
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
3-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.5
4-Oct	0	0	0	1	1	1	1	1	A	1	1	1	0	0	0	0	0	0	0	0	1	1	1	1	0.5	0.8
5-Oct	1	1	1	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.6
6-Oct	0	0	0	0	1	0	A	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.5
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.4
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.4
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.5
10-Oct	0	0	A	0	0	0	0	0	0	0	1	0	1	0	0	1	1	1	1	0	0	0	0	0	0.4	0.9
11-Oct	0	A	0	0	0	0	0	0	0	0	0	0	0	0	1	1	2	1	0	0	0	0	0	0	0.5	1.9
12-Oct	A	0	0	0	0	0	1	1	0	0	1	0	0	0	0	1	0	0	1	0	0	0	0	A	0.4	0.6
13-Oct	0	0	1	1	1	1	0	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0.5	0.8
14-Oct	0	0	0	0	0	0	0	0	0	1	1	0	1	1	1	1	1	1	1	1	1	A	0	1	0.5	1.3
15-Oct	0	0	1	0	0	1	0	1	1	1	1	0	0	0	0	0	0	0	0	1	A	0	0	0	0.5	0.6
16-Oct	0	0	0	0	0	0	0	0	0	1	1	1	0	1	0	0	0	0	0	A	0	0	0	0	0.4	0.8
17-Oct	0	0	0	0	0	0	0	1	1	1	1	0	0	0	0	0	0	0	A	0	0	0	0	0	0.5	1.1
18-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.4	0.5
19-Oct	0	0	0	0	0	0	0	0	0	0	0	C	C	C	1	1	A	0	0	0	0	0	0	1	0.3	0.8
20-Oct	1	1	1	1	0	0	0	0	1	1	1	1	1	1	1	A	1	0	0	0	0	0	0	0	0.5	0.7
21-Oct	1	0	1	0	0	1	1	1	1	1	1	1	1	1	A	1	0	1	1	0	0	0	0	0	0.5	0.7
22-Oct	0	0	0	0	0	0	0	1	1	1	1	1	1	A	1	0	1	1	1	1	1	1	1	1	0.7	1.3
23-Oct	1	1	1	1	0	0	1	1	1	0	1	0	A	1	0	1	1	1	1	0	1	1	1	1	0.6	1.0
24-Oct	1	1	1	1	0	0	0	0	0	0	0	A	0	0	0	0	0	0	1	1	0	0	0	1	0.5	0.9
25-Oct	0	0	0	0	0	0	0	0	0	1	A	1	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.7
26-Oct	0	0	0	0	0	0	0	0	0	A	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0.4	0.7
27-Oct	0	0	0	0	0	0	1	1	A	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0.4	0.7
28-Oct	0	0	0	0	0	0	0	A	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0.4	0.7
29-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0.4	0.6
30-Oct	0	0	0	0	0	A	0	0	0	0	0	0	1	1	1	1	0	0	0	0	0	0	0	0	0.4	1.0
31-Oct	0	0	0	0	A	0	1	1	1	1	1	0	0	0	0	0	0	0	1	1	1	1	1	1	0.5	0.6
0.4 0.4 0.4 0.4 0.3 0.4 0.4 0.4 0.4 0.5 0.5 0.5 0.4 0.5 0.5 0.4 0.5 0.5 0.5 0.5 0.4 0.4 0.4 0.4 0.4																								Diurnal Average		
0.9 0.8 0.8 0.7 0.6 0.6 0.7 0.7 1.1 1.1 0.8 0.7 0.7 0.7 1.0 0.7 1.9 1.3 1.3 1.0 1.1 1.3 1.2 1.0																								Diurnal Maximum		
C - Calibration																								A - Automated Daily Zero Span		

Hourly Averages

Total Reduced Sulphur (TRS) - ppb
Evergreen Park - October 2018



Hourly Maximums


Total Reduced Sulphur (TRS) - ppb

Evergreen Park - October 2018

Maximum Value: 5.1 ppb on Oct 10 18:00																				Maximum Daily Average: 1.3 ppb on Oct 22										Hours in Service: 744									
Minimum Value: 0 ppb on Oct 19 10:00																				Minimum Daily Average: 0.8 ppb on Oct 2										Hours of Data: 709									
Maximum Diurnal Average: 1.4 ppb at hour 18																				Minimum Diurnal Average: 0.9 ppb at hour 5										Hours of Missing Data: 35									
Monthly Average: 1.01 ppb																				Percentiles: P ₁ = 0.6 P ₁₀ = 0.8 Q ₁ = 0.9 Median = 1.0 Q ₃ = 1.1 P ₉₀ = 1.3 P ₉₉ = 2.0										Hours of Calibration: 35									
																				Percent Operational Time: 100.0																			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum													
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24															
1-Oct	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	4	1	1	1	1	1	1	1.0	4.0													
2-Oct	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	2	1	1	1	1	1	1	1	1	0.8	1.8													
3-Oct	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1.0	1.3													
4-Oct	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1.1	1.5													
5-Oct	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.2													
6-Oct	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.1													
7-Oct	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.2													
8-Oct	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.2													
9-Oct	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.2													
10-Oct	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	5	2	1	1	1	1	1	1.2	5.1													
11-Oct	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	4	3	1	1	1	1	1	1	1.2	4.3													
12-Oct	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	0.9	1.2													
13-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1.0	1.4													
14-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	3	2	2	A	1	1	1	1.1	2.8													
15-Oct	1	1	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1.0	1.5													
16-Oct	1	1	1	1	1	1	1	1	1	2	2	1	1	1	1	1	1	3	1	A	1	1	1	1	1.1	2.6													
17-Oct	1	1	1	1	1	1	1	1	2	2	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1.1	1.9													
18-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	0.9	1.1													
19-Oct	1	1	1	1	0	1	1	1	1	0	0	C	C	C	2	1	A	1	1	1	1	1	1	1	0.8	1.6													
20-Oct	1	1	2	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1.1	1.5													
21-Oct	2	1	1	1	1	1	1	1	1	1	2	1	1	2	A	1	1	1	1	1	1	1	1	1	1.1	1.7													
22-Oct	1	1	1	1	1	1	1	2	1	1	1	1	1	A	1	1	2	2	2	2	2	2	2	2	1.3	2.0													
23-Oct	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	2	1	1	1	1	1	1	2	1	1.2	1.6													
24-Oct	2	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1.0	1.6													
25-Oct	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1.0	1.3													
26-Oct	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.3													
27-Oct	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1.0	1.3													
28-Oct	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.3													
29-Oct	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1.0	1.3													
30-Oct	1	1	1	1	1	A	1	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	1.0	1.7													
31-Oct	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1	1	1.1	1.5													
1.0 0.9 0.9 0.9 0.9 0.9 0.9 1.0 1.0 1.1 1.1 1.0 1.0 1.0 1.0 1.1 1.1 1.4 1.1 1.0 1.0 1.0 1.0 1.0																									Diurnal Average														
1.6 1.3 1.5 1.3 1.1 1.2 1.3 1.5 1.9 1.8 1.7 1.3 1.3 1.6 1.7 1.8 4.3 5.1 2.8 1.7 1.9 2.0 1.7 1.7																									Diurnal Maximum														
C - Calibration																									A - Automated Daily Zero Span														


Hourly Maximums

Total Reduced Sulphur (TRS) - ppb
Evergreen Park - October 2018




Pollutant Rose

Total Reduced Sulphur (TRS) - ppb
Evergreen Park - October 2018



Span Responses

Total Reduced Sulphur (TRS)
Evergreen Park - October 2018



Hourly Averages


Particulate Matter 2.5 (PM_{2.5}) - µg/m³

Evergreen Park - October 2018

Number of Exceedences: 1-hr: 0 24-hr: 0 Maximum Value: 56.5 µg/m³ on Oct 20 10:00 Maximum Daily Average: 15.4 µg/m³ on Oct 20																								Hours in Service: 744 Hours of Data: 741 Hours of Missing Data: 3 Hours of Calibration: 3 Percent Operational Time: 100.0		
Minimum Value: 0 µg/m³ on Oct 1 02:00 Maximum Diurnal Average: 8.8 µg/m³ at hour 10 Monthly Average: 4.97 µg/m³ Minimum Daily Average: 0.5 µg/m³ on Oct 2 Minimum Diurnal Average: 2.8 µg/m³ at hour 4 Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.8 Median = 2.9 Q ₃ = 6.8 P ₉₀ = 11.6 P ₉₉ = 30.4																										
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	1	1	1	1	2	4	5	5	2	1	2	1	1	2	3	3	2	1	1	0	0	0	1.7	5.5
2-Oct	0	0	0	0	0	0	0	0	0	0	0	0	2	0	0	0	2	1	0	1	0	1	1	3	0.5	2.5
3-Oct	4	10	4	0	0	0	1	3	5	2	8	2	1	0	0	0	0	0	0	0	1	0	3	2	1.9	9.8
4-Oct	2	4	3	4	4	3	5	7	7	4	16	14	10	8	9	2	2	3	4	5	9	9	8	9	6.3	15.5
5-Oct	6	6	5	3	1	1	0	1	1	1	0	1	1	1	1	1	1	1	2	1	2	2	3	4	1.9	6.4
6-Oct	3	3	4	6	7	5	6	7	8	6	8	6	4	2	3	4	4	3	4	3	3	3	2	3	4.4	7.8
7-Oct	4	3	3	3	3	3	3	4	4	4	3	3	1	1	3	4	3	3	2	1	3	2	2	2	2.7	4.4
8-Oct	2	2	0	0	1	0	0	0	0	1	0	0	0	1	1	0	1	1	1	1	1	2	4	5	1.0	5.1
9-Oct	7	7	7	7	7	6	5	4	4	3	3	3	2	2	3	2	2	2	2	2	3	4	4	4	4.0	7.3
10-Oct	4	4	4	4	4	5	5	5	6	5	11	12	7	6	5	7	8	7	5	8	7	5	3	3	5.8	11.8
11-Oct	3	3	3	3	3	5	4	5	5	12	4	5	3	7	13	4	11	0	0	0	0	1	2	1	4.0	13.3
12-Oct	1	1	1	1	2	5	5	5	4	3	2	2	2	2	1	2	1	2	1	1	2	2	2	2	2.2	4.9
13-Oct	1	1	1	0	1	0	0	1	1	1	1	0	0	0	0	0	0	1	3	4	3	2	1	0	0.9	4.0
14-Oct	0	0	0	0	0	0	1	2	3	4	3	0	6	30	11	0	15	0	19	11	7	6	0	1	4.9	30.0
15-Oct	1	0	0	0	0	0	0	6	25	28	14	11	0	0	0	0	0	1	1	1	3	5	10	4	4.6	28.0
16-Oct	7	6	7	7	6	7	5	4	10	11	6	7	25	24	11	43	33	2	0	0	2	2	4	2	9.6	43.1
17-Oct	2	3	4	5	5	4	7	11	27	29	14	10	9	19	20	22	18	13	4	11	9	5	5	10	11.0	29.2
18-Oct	6	5	5	7	9	7	9	10	15	9	3	7	0	0	0	0	1	4	8	13	16	8	9	9	6.7	16.3
19-Oct	6	5	4	3	3	5	7	7	7	5	4	C	C	C	9	7	4	3	2	2	3	4	6	7	4.8	8.6
20-Oct	9	8	9	7	7	3	9	20	27	56	43	37	39	30	24	25	12	0	0	0	0	0	1	1	15.4	56.5
21-Oct	1	0	0	0	0	1	4	5	4	5	6	3	0	0	0	0	1	1	0	1	3	7	4	5	2.2	7.0
22-Oct	5	3	4	3	3	4	6	9	8	10	20	21	37	12	4	2	8	15	31	25	30	23	16	15	13.0	37.0
23-Oct	11	11	11	8	8	8	10	13	20	16	13	6	28	21	3	0	9	8	6	7	9	12	13	15	11.1	28.4
24-Oct	10	12	10	8	7	6	7	7	7	8	8	8	12	16	16	17	13	12	9	10	10	10	10	9	10.0	17.0
25-Oct	7	8	7	6	6	6	8	9	6	9	8	11	6	1	1	6	0	0	0	0	0	0	0	0	4.3	10.5
26-Oct	0	0	0	0	2	2	1	2	2	3	2	2	1	8	7	3	2	3	2	1	3	0	0	0	1.9	7.6
27-Oct	0	0	0	0	1	8	17	12	27	19	12	0	0	0	0	0	0	5	3	2	1	4	3	3	4.9	26.8
28-Oct	3	1	0	0	0	0	0	0	0	0	2	3	4	3	5	3	3	3	3	3	2	3	3	2	1.8	4.6
29-Oct	2	2	2	1	1	2	3	2	3	5	12	5	4	12	7	14	18	0	0	3	1	0	1	1	4.2	18.1
30-Oct	1	0	1	0	0	0	0	0	0	0	1	13	20	12	20	19	0	0	0	0	0	0	0	0	3.6	19.9
31-Oct	0	0	0	0	1	2	3	5	7	7	1	1	2	2	2	2	5	4	3	5	4	2	3	4	2.7	7.4
3.5 3.4 3.2 2.8 2.9 3.2 4.4 5.4 8.0 8.8 7.3 6.5 7.6 7.4 5.8 6.2 5.8 3.3 3.8 4.0 4.4 4.0 3.9 4.0																								Diurnal Average		
11.4 11.8 10.6 8.4 8.7 8.3 16.9 20.1 27.5 56.5 42.8 36.9 39.4 30.0 23.7 43.1 33.2 15.2 30.5 25.5 30.1 22.9 15.5 14.9																								Diurnal Maximum		
C - Calibration Alberta Ambient Air Quality Guideline (AAAQG): 1-hr 80 µg/m³ Alberta Ambient Air Quality Objective (AAAQO): 24-hr 30 µg/m³																										

Hourly Averages

Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Evergreen Park - October 2018



Hourly Maximums


Particulate Matter 2.5 (PM_{2.5}) - µg/m³

Evergreen Park - October 2018

Maximum Value: 90.0 μg/m ³ on Oct 23 13:00																								Maximum Daily Average: 29.1 μg/m ³ on Oct 20										Hours in Service: 744	
Minimum Value: 0 μg/m ³ on Oct 3 19:00																								Minimum Daily Average: 3.5 μg/m ³ on Oct 13										Hours of Data: 741	
Maximum Diurnal Average: 16.7 μg/m ³ at hour 13																								Minimum Diurnal Average: 5.5 μg/m ³ at hour 4										Hours of Missing Data: 3	
Monthly Average: 9.56 μg/m ³																								Percentiles: P ₁ = 0.0 P ₁₀ = 1.8 Q ₁ = 3.7 Median = 6.1 Q ₃ = 10.4 P ₉₀ = 21.3 P ₉₉ = 61.6										Hours of Calibration: 3	
																								Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum									
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24											
1-Oct	5	6	5	3	6	4	6	10	8	9	6	5	4	4	5	5	7	8	6	5	5	5	4	4	5.6	9.8									
2-Oct	3	5	5	4	2	5	3	4	2	2	1	2	8	1	2	2	4	6	2	4	3	4	5	8	3.7	8.1									
3-Oct	11	21	12	2	4	5	5	6	14	4	21	5	3	1	3	2	1	1	0	1	4	2	6	6	5.9	21.4									
4-Oct	6	7	5	8	7	8	9	10	12	9	24	32	30	11	23	6	4	4	6	7	14	12	11	11	11.5	31.8									
5-Oct	12	8	9	5	3	4	4	4	2	3	3	4	3	3	3	3	3	3	3	3	4	6	6	6	4.6	11.8									
6-Oct	6	6	7	10	12	9	10	10	12	8	11	11	6	5	6	6	5	5	5	4	5	4	4	6	7.2	11.8									
7-Oct	7	5	6	7	5	6	7	6	8	5	5	6	4	4	5	8	6	7	4	4	6	4	4	5	5.7	7.6									
8-Oct	5	4	3	3	3	3	3	3	3	4	3	3	2	3	3	3	6	3	3	4	5	5	6	8	3.9	7.9									
9-Oct	11	10	10	9	9	9	7	7	6	6	6	6	5	5	6	4	3	4	4	5	6	6	7	6	6.6	10.8									
10-Oct	6	6	6	7	7	7	10	9	9	8	17	19	16	9	10	12	11	10	7	11	11	7	6	6	9.4	19.4									
11-Oct	7	6	5	5	6	8	7	8	8	19	8	7	8	23	22	9	19	2	0	1	1	3	4	4	7.9	23.2									
12-Oct	3	3	3	3	5	7	8	8	6	5	4	4	5	5	3	5	4	5	3	5	4	4	3	5	4.7	8.1									
13-Oct	4	4	4	2	3	2	3	6	3	3	3	2	2	2	2	2	2	5	4	7	4	6	4	2	3.5	7.2									
14-Oct	1	3	2	2	2	4	4	3	5	7	7	2	33	43	35	3	34	0	48	42	34	37	1	6	14.9	47.9									
15-Oct	6	0	0	3	3	2	1	27	38	53	23	30	0	0	1	1	2	2	2	5	5	11	17	8	10.0	52.6									
16-Oct	8	10	9	9	10	9	8	11	13	16	12	20	38	34	33	61	77	25	0	4	4	4	5	4	17.7	76.8									
17-Oct	5	4	5	7	6	7	10	16	43	34	33	17	17	41	35	32	28	20	10	21	13	7	9	16	18.1	43.3									
18-Oct	12	9	7	18	20	11	20	21	26	35	7	14	4	1	1	1	4	8	10	16	28	10	14	11	12.9	34.7									
19-Oct	8	7	7	6	6	7	14	9	11	9	7	C	C	C	12	11	5	5	4	4	5	9	10	9	7.8	13.6									
20-Oct	11	11	10	10	9	8	22	55	75	85	62	65	70	46	38	47	32	21	0	4	4	0	8	5	29.1	85.2									
21-Oct	7	3	5	2	2	3	8	12	9	11	18	9	0	3	1	1	4	3	2	3	7	10	8	8	5.8	17.8									
22-Oct	8	6	8	5	6	7	13	14	11	14	24	35	63	17	6	4	13	28	40	33	33	28	18	16	18.8	63.1									
23-Oct	14	13	14	10	10	10	12	19	31	23	21	10	90	35	9	9	13	14	7	10	11	15	16	20	18.2	90.0									
24-Oct	13	19	13	11	9	9	9	9	10	10	10	9	17	23	20	27	16	17	11	12	11	12	11	11	13.3	26.9									
25-Oct	9	13	9	7	9	8	10	11	9	12	10	17	7	3	2	17	5	3	2	0	0	1	0	0	6.9	17.2									
26-Oct	0	1	1	0	8	4	3	5	5	5	5	6	5	14	11	7	4	6	4	2	5	4	1	1	4.4	14.3									
27-Oct	2	1	1	2	2	41	38	19	40	34	21	9	5	8	12	5	2	11	4	5	3	7	5	5	11.8	40.9									
28-Oct	5	3	1	1	0	0	1	3	1	3	4	9	7	5	8	4	4	4	4	5	4	5	6	4	3.8	8.8									
29-Oct	4	4	3	2	2	7	6	4	5	10	18	9	10	21	16	26	34	22	0	19	13	1	2	3	10.1	34.2									
30-Oct	3	3	2	2	1	0	2	3	1	1	4	32	34	25	28	26	20	0	0	0	0	0	0	0	7.8	34.0									
31-Oct	0	0	1	1	4	4	7	7	10	11	4	3	4	5	4	4	6	7	5	7	5	5	4	7	4.8	10.6									
6.6 6.5 5.8 5.5 5.9 7.0 8.7 10.9 14.1 14.8 13.0 13.4 16.7 13.4 11.7 11.4 12.2 8.4 6.5 8.1 8.4 7.6 6.7 6.8																								Diurnal Average											
14.2 21.4 13.5 18.4 19.8 40.9 38.1 54.5 75.2 85.2 62.1 65.5 90.0 45.8 37.7 60.9 76.8 27.6 47.9 42.3 34.0 36.6 18.4 19.6																								Diurnal Maximum											
C - Calibration																																			


Hourly Maximums

Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Evergreen Park - October 2018



Pollutant Rose

Particulate Matter 2.5 (PM_{2.5}) - µg/m³
Evergreen Park - October 2018



Hourly Averages


External Temperature (ET) - °C

Evergreen Park - October 2018

Maximum Value: 21.4 °C on Oct 16 15:00																				Maximum Daily Average: 12.0 °C on Oct 15										Hours in Service: 744	
Minimum Value: -6 °C on Oct 23 08:00																				Minimum Daily Average: -2.1 °C on Oct 2										Hours of Data: 744	
Maximum Diurnal Average: 8.7 °C at hour 15																				Minimum Diurnal Average: -0.5 °C at hour 7										Hours of Missing Data: 0	
Monthly Average: 3.09 °C																				Percentiles: P ₁ = -5.4 P ₁₀ = -3.0 Q ₁ = -1.5 Median = 1.1 Q ₃ = 7.4 P ₉₀ = 12.2 P ₉₉ = 18.7										Hours of Calibration: 0	
																				Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
1-Oct	0	-1	-1	-1	-1	-1	-1	-1	-1	0	1	1	2	1	1	1	0	0	-1	-2	-2	-2	-3	-3	-0.6	1.7					
2-Oct	-3	-4	-4	-4	-3	-4	-4	-4	-3	-2	-2	-1	0	0	0	1	1	0	-1	-2	-3	-3	-3	-4	-2.1	1.0					
3-Oct	-4	-4	-4	-5	-5	-4	-5	-5	-4	-3	-1	2	4	6	7	7	8	6	2	-1	-2	-2	-2	-2	-0.5	7.9					
4-Oct	-2	-2	-2	-3	-4	-4	-5	-4	-3	-1	3	6	8	9	10	10	10	9	6	2	0	-1	-1	-2	1.5	10.0					
5-Oct	-3	-3	-3	-4	-4	-5	-6	-5	-4	-2	-1	0	2	4	5	6	5	4	1	-1	-2	-3	-3	-3	-1.0	5.7					
6-Oct	-4	-4	-3	-3	-2	-1	-1	-1	0	2	3	7	8	9	9	7	6	5	4	3	3	2	2	1	2.2	9.1					
7-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	-1	-1	-1	-1	-1	-1	-2	-0.3	0.4						
8-Oct	-2	-2	-2	-2	-2	-2	-2	-2	-2	-1	0	1	1	2	1	1	1	0	0	0	-1	-1	-1	-1	-0.6	1.9					
9-Oct	-1	-1	-1	-2	-2	-2	-2	-2	-2	-1	-1	0	0	1	1	1	1	0	0	-1	-1	-1	-1	-1	-0.7	1.4					
10-Oct	-1	-1	-1	-1	-2	-2	-2	-2	-2	-1	-1	0	0	0	1	1	0	0	0	0	0	0	-1	-1	-0.7	0.6					
11-Oct	-2	-3	-4	-4	-5	-5	-4	-3	-2	-1	1	4	6	7	8	9	9	9	9	8	7	5	5	4	2.4	9.4					
12-Oct	4	5	5	5	5	5	5	3	1	0	1	1	0	0	0	1	1	1	1	0	0	0	0	0	1.8	5.0					
13-Oct	0	1	1	2	1	1	1	2	1	2	5	6	7	8	9	10	10	9	4	1	0	1	1	1	3.5	9.7					
14-Oct	1	-2	-3	-3	-3	-3	-3	-3	-2	1	8	12	12	13	14	14	13	13	13	12	12	12	11	10	6.2	13.9					
15-Oct	11	10	9	9	8	8	8	10	12	13	14	15	16	16	17	17	17	15	14	12	11	10	9	8	12.0	17.1					
16-Oct	5	4	2	1	0	-1	-1	-2	-1	4	10	15	19	20	21	21	20	19	17	15	15	15	15	14	10.3	21.4					
17-Oct	13	12	10	5	3	2	3	3	3	5	13	16	18	20	20	18	17	19	13	10	10	13	10	7	10.9	19.8					
18-Oct	9	10	11	14	14	13	10	9	9	12	13	13	14	14	14	14	13	11	8	4	3	2	0	-1	9.7	14.4					
19-Oct	-2	-2	-2	-3	-3	-3	-3	-3	-2	0	3	7	10	12	13	14	15	13	10	9	8	6	3	2	4.2	14.5					
20-Oct	1	2	3	2	6	11	11	10	11	11	12	13	13	14	14	13	13	12	10	9	7	5	3	2	8.7	13.6					
21-Oct	3	3	2	-1	-1	-1	-2	-3	-2	1	5	7	10	11	12	12	11	7	3	1	-1	-2	-2	-3	3.0	11.7					
22-Oct	-3	-3	-4	-4	-5	-5	-6	-6	-6	-2	3	5	7	10	10	11	12	7	3	0	-2	-2	-2	-3	0.7	12.2					
23-Oct	-3	-4	-4	-5	-5	-5	-6	-6	-6	-2	3	6	8	11	13	13	14	9	3	0	-1	-2	-2	-2	1.2	13.7					
24-Oct	-2	-2	-3	-3	-3	-3	-4	-4	-3	0	4	8	12	13	13	13	13	11	7	4	2	1	0	-1	3.1	13.4					
25-Oct	-2	-2	-2	-3	-3	-2	-2	-1	0	1	4	10	13	16	15	12	8	8	8	8	7	6	5	5	4.5	15.5					
26-Oct	5	4	4	0	-2	-2	-3	-3	-3	-2	-1	0	2	3	4	4	6	4	0	-1	-1	4	5	5	1.3	6.0					
27-Oct	6	6	6	6	6	6	6	6	6	6	7	9	9	10	10	10	10	7	4	2	1	-2	-2	-3	5.5	10.2					
28-Oct	-3	-3	-3	-3	-2	-1	-1	0	0	1	2	2	1	0	0	1	1	1	1	1	1	0	0	-1	-0.1	1.7					
29-Oct	-1	-2	-2	-2	-3	-3	-3	-4	-4	-3	2	5	8	10	11	10	10	8	5	1	-1	-2	0	1	1.6	10.6					
30-Oct	2	2	1	2	2	4	4	5	5	5	7	8	10	10	10	10	9	8	6	6	4	4	4	3	5.5	10.3					
31-Oct	3	2	1	1	0	0	0	1	1	2	4	5	6	6	6	5	4	3	3	2	2	2	2	2	2.6	6.2					
0.8 0.5 0.2 -0.3 -0.4 -0.4 -0.5 -0.5 -0.1 1.6 3.9 5.8 7.3 8.3 8.7 8.6 8.3 7.0 4.9 3.3 2.4 2.1 1.6 1.1																								Diurnal Average	Diurnal Maximum						
12.5 11.9 11.4 13.6 14.4 12.6 10.6 10.2 11.7 13.2 13.8 16.1 19.1 20.4 21.4 21.1 19.9 18.8 16.6 14.9 14.9 15.0 15.0 13.7																															

Hourly Averages

External Temperature (ET) - °C
Evergreen Park - October 2018




Hourly Averages

Relative Humidity (RH) - %
Evergreen Park - October 2018

Maximum Value: 98.8 % on Oct 12 07:00																					Maximum Daily Average: 95.3 % on Oct 7										Hours in Service: 744	
Minimum Value: 14 % on Oct 16 15:00																					Minimum Daily Average: 37.4 % on Oct 15										Hours of Data: 744	
Maximum Diurnal Average: 81.7 % at hour 7																					Minimum Diurnal Average: 48.1 % at hour 15										Hours of Missing Data: 0	
Monthly Average: 68.37 %																					Percentiles: P ₁ = 20.0 P ₁₀ = 33.4 Q ₁ = 49.7 Median = 74.6 Q ₃ = 87.1 P ₉₀ = 93.9 P ₉₉ = 98.0										Hours of Calibration: 0	
																					Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum						
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24								
1-Oct	65	67	70	72	75	79	81	82	74	62	55	53	56	57	64	77	83	86	95	94	93	90	85	83	74.9	95.3						
2-Oct	79	79	79	78	78	77	77	76	69	62	56	52	50	48	47	47	47	53	61	67	70	71	71	70	65.1	79.0						
3-Oct	72	73	73	76	74	71	71	73	70	64	58	50	42	36	34	31	30	37	55	69	77	80	77	77	61.1	79.6						
4-Oct	78	83	84	89	91	91	92	92	91	79	63	46	38	33	32	28	28	30	38	54	66	70	73	79	64.6	92.4						
5-Oct	83	85	86	83	81	83	86	86	82	75	72	66	58	52	48	46	48	53	63	72	79	82	84	86	72.5	86.4						
6-Oct	88	89	88	89	89	92	94	93	89	85	78	65	59	55	58	68	76	78	79	81	82	83	83	88	80.4	94.2						
7-Oct	92	93	93	94	94	94	96	96	96	96	96	96	97	96	95	95	96	96	96	96	96	96	96	96	95.3	96.5						
8-Oct	95	95	95	95	95	95	95	96	96	95	93	90	88	86	85	88	85	83	82	82	82	82	83	83	89.4	95.6						
9-Oct	83	84	83	84	84	84	84	84	83	81	78	75	72	69	66	64	65	67	69	71	74	76	78	79	76.6	84.0						
10-Oct	79	80	81	82	83	83	84	84	84	82	81	79	77	75	70	72	73	76	79	82	82	83	84	85	80.1	85.3						
11-Oct	86	91	93	93	93	94	94	94	93	84	72	63	59	60	60	56	53	54	57	61	67	75	90	96	76.6	95.5						
12-Oct	97	98	98	98	98	99	99	98	98	98	98	98	98	98	92	87	86	86	86	89	94	95	96	96	95.0	98.8						
13-Oct	96	93	92	91	91	92	94	90	94	91	80	70	63	54	50	45	42	46	70	85	88	85	78	73	77.3	95.6						
14-Oct	75	90	94	95	95	95	95	95	95	87	61	40	35	31	29	28	30	31	33	35	38	40	44	48	60.0	95.3						
15-Oct	48	48	51	51	53	50	47	40	34	26	25	24	24	24	23	24	25	30	32	36	40	44	46	52	37.4	52.5						
16-Oct	66	70	78	81	79	85	85	87	82	65	41	26	16	14	14	16	18	20	23	26	25	24	22	24	45.2	86.8						
17-Oct	26	27	32	48	58	61	62	65	66	64	34	24	21	20	18	23	23	18	33	43	37	29	36	48	38.2	66.3						
18-Oct	43	46	40	35	31	35	44	48	50	48	42	39	35	33	30	29	31	36	46	63	69	76	81	87	46.6	87.3						
19-Oct	91	92	93	93	93	93	93	94	93	87	69	50	40	35	32	29	27	31	36	41	45	55	66	73	64.6	93.7						
20-Oct	78	77	69	73	67	57	52	48	39	39	37	33	31	29	29	30	29	31	35	40	48	54	63	68	48.0	77.9						
21-Oct	59	59	62	76	80	78	84	84	81	65	53	44	36	30	27	26	26	42	54	61	69	74	78	79	59.5	84.3						
22-Oct	81	82	86	88	90	91	92	92	92	84	61	51	44	39	36	34	33	48	65	75	81	84	87	89	71.1	92.0						
23-Oct	91	92	92	93	93	93	93	93	93	88	62	48	40	32	28	27	30	44	62	70	76	82	85	86	70.5	93.0						
24-Oct	89	91	91	89	87	89	91	91	91	76	61	50	41	39	42	42	43	50	63	72	80	85	88	91	72.1	91.1						
25-Oct	91	92	93	94	95	94	93	93	94	92	78	54	43	36	37	57	77	78	73	62	60	64	67	65	74.2	94.6						
26-Oct	62	60	58	77	86	88	91	92	92	90	87	83	76	68	67	75	58	69	83	89	89	66	61	60	76.1	92.2						
27-Oct	56	52	50	47	46	45	46	48	48	48	45	39	36	34	33	33	34	42	54	60	66	76	80	81	50.0	81.4						
28-Oct	82	84	83	82	80	75	72	69	68	63	62	64	67	80	83	82	80	81	82	81	77	83	87	89	77.4	89.4						
29-Oct	87	91	94	94	94	95	94	93	94	94	79	69	58	48	44	43	44	51	62	75	81	84	82	81	76.2	95.1						
30-Oct	83	91	94	94	88	73	71	69	66	62	57	51	44	43	42	40	43	48	52	55	60	61	61	66	63.1	94.1						
31-Oct	67	69	72	75	79	80	80	76	79	75	62	62	62	73	77	86	94	95	95	96	96	96	95	95	80.7	96.1						
																									Diurnal Average							
																									Diurnal Maximum							

Hourly Averages

Relative Humidity (RH) - %
Evergreen Park - October 2018



Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Evergreen Park - October 2018

Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1 Spd	6	3	2	1	0	0	1	0	2	3	4	7	7	4	7	6	2	2	8	12	12	12	12	12	4.6	12.2
Dir	36	35	22	0	322	209	12	220	311	321	357	28	2	5	330	339	223	313	37	38	28	26	32	31	17.5	31.5
2 Spd	13	12	5	9	7	5	6	4	5	4	6	4	8	11	8	15	18	18	16	17	14	14	17	16	4.7	18.1
Dir	31	33	16	20	18	18	22	21	19	30	6	278	263	248	270	234	232	225	222	223	225	228	234	235	255.3	232.0
3 Spd	19	17	10	9	12	13	9	6	10	8	7	6	6	7	8	7	4	2	2	1	0	1	2	1	5.0	18.8
Dir	236	239	260	261	259	249	251	271	246	241	238	198	161	148	173	160	156	134	55	33	218	53	33	28	231.0	236.1
4 Spd	1	0	1	0	2	1	6	4	5	7	7	11	11	11	10	15	16	16	5	3	2	7	4	3	4.6	16.2
Dir	219	56	206	254	189	196	207	205	206	235	252	249	252	277	291	303	323	323	338	327	259	291	227	248	276.2	322.7
5 Spd	2	1	1	3	3	2	3	3	5	7	6	4	5	6	7	7	8	7	4	2	3	1	1	1	3.6	7.9
Dir	350	315	57	54	68	40	30	40	53	60	51	50	26	23	36	47	48	48	56	58	63	31	40	103	45.1	47.7
6 Spd	1	0	0	1	1	1	1	1	4	2	1	1	2	6	4	8	4	7	7	6	4	8	9	9	2.6	9.0
Dir	17	125	33	196	227	116	166	297	207	179	204	357	324	9	55	82	86	63	64	78	57	53	56	53	62.7	52.5
7 Spd	8	5	8	8	7	6	7	4	5	8	8	7	9	9	11	9	10	9	8	7	7	7	7	7	7.4	10.7
Dir	50	54	49	49	50	52	51	41	57	45	39	27	23	26	27	25	27	21	27	33	28	36	42	42	36.9	26.9
8 Spd	5	7	6	6	6	7	6	7	6	5	8	9	8	13	12	14	11	13	13	13	12	12	14	12	9.0	14.0
Dir	47	44	50	54	53	55	49	58	64	66	81	77	77	89	99	95	65	75	80	81	81	80	80	77	74.4	95.0
9 Spd	11	10	9	9	7	7	8	8	7	9	11	12	13	14	13	15	13	14	13	10	9	10	10	9	10.1	14.5
Dir	74	71	71	70	67	66	70	73	76	79	86	81	88	90	90	101	108	103	102	94	101	99	107	103	88.2	100.7
10 Spd	7	5	4	2	3	4	2	2	2	2	4	5	8	7	7	6	3	2	0	3	3	1	1	2	1.0	8.2
Dir	93	79	76	64	52	54	54	11	32	24	322	329	301	281	296	301	282	300	360	201	197	178	144	165	342.7	300.6
11 Spd	2	1	1	1	1	0	3	3	1	5	6	7	15	20	19	19	23	23	20	12	2	2	1	1	6.5	23.3
Dir	153	199	149	50	51	171	173	166	140	186	181	191	222	234	248	257	255	251	262	275	302	357	171	140	241.0	254.8
12 Spd	2	3	3	3	0	2	5	21	15	14	15	18	11	14	19	17	15	15	7	1	3	2	3	3	6.4	20.9
Dir	169	171	158	186	352	31	14	18	7	316	305	317	331	322	315	306	305	302	296	263	191	185	191	215	319.4	18.2
13 Spd	5	7	10	11	9	6	5	8	2	9	12	22	20	17	19	14	12	5	1	4	3	8	12	13	7.7	21.6
Dir	226	248	255	260	265	268	262	275	256	252	293	315	319	316	314	294	293	292	196	219	215	220	208	211	277.7	314.9
14 Spd	8	2	2	1	3	1	1	0	1	0	7	20	24	32	40	50	47	33	30	27	24	22	19	18	16.0	50.3
Dir	195	104	74	98	41	65	54	91	66	203	238	242	245	252	247	246	247	247	250	258	256	260	265	261	249.4	246.0
15 Spd	16	14	8	9	7	7	8	15	34	38	33	35	29	28	30	30	22	17	8	7	5	5	5	2	15.9	37.6
Dir	260	279	304	273	293	277	302	249	242	246	253	257	260	273	284	302	309	316	302	286	265	278	268	306	271.6	245.7
16 Spd	1	1	1	1	1	1	1	2	1	1	1	5	28	31	29	22	17	22	18	16	22	22	23	20	11.4	30.9
Dir	44	296	70	250	20	118	191	197	207	134	202	242	255	254	253	248	252	251	258	246	243	246	238	229	247.5	254.2
17 Spd	15	11	3	1	0	3	1	2	1	2	9	8	14	8	12	4	11	11	4	2	9	12	2	2	5.6	14.7
Dir	227	238	277	150	148	189	183	200	207	198	215	239	232	229	240	235	237	251	339	268	235	248	259	280	236.9	227.4
18 Spd	5	4	9	24	33	14	8	9	16	17	28	29	23	25	24	23	18	9	3	4	4	1	1	1	12.1	33.4
Dir	235	249	245	259	247	258	261	229	236	258	264	263	285	302	304	312	316	310	321	204	200	40	200	190	271.1	247.3
19 Spd	0	1	0	2	3	2	1	2	3	3	4	5	6	10	11	11	11	7	7	6	5	3	1	1	3.4	11.1
Dir	128	185	174	191	194	201	215	194	191	163	165	112	123	109	108	106	108	88	80	95	85	40	348	255	113.6	105.7
20 Spd	1	4	7	5	16	29	32	35	45	41	52	51	54	51	47	42	39	25	19	11	7	6	4	5	26.0	54.3
Dir	248	246	272	233	242	242	238	238	240	247	244	251	245	249	250	249	252	253	239	253	253	221	225	230	245.7	245.0
21 Spd	10	9	7	4	3	3	3	2	4	8	7	8	3	3	3	3	0	1	0	0	1	2	2	0	2.9	10.5
Dir	236	235	233	199	220	230	236	283	277	245	244	244	286	225	307	201	71	42	42	35	46	37	40	119	242.5	236.0
22 Spd	0	1	0	0	1	0	1	0	1	1	1	2	1	4	5	3	2	0	0	1	0	2	1	2	0.4	5.3
Dir	151	207	89	59	167	200	190	176	57	117	294	292	311	313	337	30	195	209	13	214	102	276	218	216	287.2	337.5


Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Evergreen Park - October 2018

Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum								
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24										
23 Spd	1	1	2	1	1	1	0	0	0	0	3	7	3	2	4	3	1	0	0	1	0	2	0	4	0.7	6.8								
Dir	202	192	193	189	198	198	223	204	59	270	245	273	263	323	37	36	336	118	26	16	228	212	227	251	257.9	272.8								
24 Spd	1	4	3	4	4	1	1	3	3	7	7	9	12	15	10	8	7	1	3	3	1	0	1	2	4.4	14.9								
Dir	260	220	223	244	210	205	180	188	190	215	204	206	239	245	227	227	221	218	190	203	211	209	203	191	220.8	244.9								
25 Spd	0	2	1	1	1	1	1	1	2	1	3	4	5	3	5	15	16	17	19	25	18	17	21	20	7.2	25.2								
Dir	257	205	283	184	199	1	231	170	108	229	19	145	166	264	313	256	244	208	217	218	217	209	210	221	220.3	217.6								
26 Spd	19	18	12	3	4	2	2	1	1	2	1	2	3	3	4	3	3	6	2	3	11	18	21	22	5.5	22.3								
Dir	225	220	230	55	176	165	50	127	44	51	80	56	333	277	179	219	245	187	179	159	206	214	226	225	217.1	225.5								
27 Spd	22	25	29	34	37	38	36	37	36	35	34	35	32	27	25	21	14	8	2	4	2	1	1	1	21.5	38.0								
Dir	226	227	232	235	240	240	243	242	246	247	250	261	264	266	267	282	263	239	233	248	214	27	157	64	247.4	240.4								
28 Spd	0	4	4	4	7	9	8	8	11	7	2	7	13	16	12	12	9	6	6	8	5	4	3	2	1.6	15.6								
Dir	68	40	57	51	50	59	61	66	61	66	186	270	236	242	241	241	245	253	253	240	262	206	213	222	246.2	242.1								
29 Spd	3	1	6	6	2	4	5	0	2	2	5	11	10	24	27	26	25	14	2	0	1	1	2	3	7.4	27.4								
Dir	243	193	206	217	241	236	232	138	181	186	226	236	253	247	239	242	250	243	211	214	201	218	206	214	237.6	239.4								
30 Spd	3	1	2	5	12	20	18	17	19	24	30	38	41	42	38	37	27	23	14	10	10	15	12	4	18.4	42.3								
Dir	292	169	195	182	196	228	232	230	228	231	236	244	247	245	245	249	255	254	236	208	221	211	203	201	236.8	245.1								
31 Spd	4	3	6	2	0	2	4	3	1	1	6	8	5	2	1	3	10	8	7	3	6	11	11	17	2.0	17.1								
Dir	189	175	191	57	27	204	189	187	305	162	193	184	200	166	206	14	20	321	319	289	268	303	310	320	277.9	319.6								
Spd	2.5	2.3	2.0	2.0	3.0	3.1	2.8	2.9	4.2	5.3	6.8	8.4	9.5	10.0	9.6	9.0	8.1	6.0	3.2	3.0	3.1	3.1	3.0	3.0	Diurnal Average									
Dir	229.4	233.8	240.9	248.7	244.1	242.9	244.3	245.6	242.5	245.9	250.1	257.6	258.3	261.7	265.1	264.1	263.8	261.4	253.4	236.7	230.6	238.4	226.2	233.9	Diurnal Maximum									
Spd	22.2	24.9	29.3	33.6	36.5	38.0	35.7	36.9	45.5	40.9	51.7	51.1	54.3	50.6	47.3	50.3	47.0	32.8	30.5	26.5	23.6	22.4	22.6	22.3	Diurnal Maximum									
Dir	225.7	226.9	231.8	234.7	240.4	240.4	242.5	242.1	240.3	247.0	243.6	251.4	245.0	249.2	250.1	246.0	247.2	247.3	249.5	257.8	256.0	245.7	238.4	225.5										
Maximum Speed Value: 54 km/h on Oct 20 13:00																				Minimum Speed Value: 0 km/h on Oct 22 21:00										Hours in Service:				744
Maximum Daily Speed Average: 26.0 km/h on Oct 20																				Minimum Daily Speed Average: 0.4 km/h on Oct 22										Hours of Data:				744
Maximum Diurnal Speed Average: 10.0 km/h at hour 14																				Minimum Diurnal Speed Average: 2.0 km/h at hour 4										Hours of Missing Data:				0
Monthly Average Velocity: 4.74 km/h 252.09 deg										Speed Percentiles: P ₁ = 0.2 P ₁₀ = 0.8 Q ₁ = 2.2 Median = 5.7 Q ₃ = 11.9 P ₉₀ = 22.2 P ₉₉ = 44.1										Percent Operational Time:				100.0										
All monthly, daily, and diurnal averages have been calculated using vector methods																																		
Frequency Distribution																																		
Speed Range (km/h)																																		
Direction	0 to 5		5 to 11		11 to 19		19 to 28		28 to 38		> 38		Total																					
North	23		12		1		1		0		0		37																					
NorthEast	66		56		7		0		0		0		129																					
East	16		27		20		0		0		0		63																					
SouthEast	22		2		0		0		0		0		24																					
South	83		13		1		0		0		0		97																					
SouthWest	70		39		43		22		17		10		201																					
West	36		34		19		19		14		5		127																					
NorthWest	23		16		19		7		1		0		66																					
Total	339		199		110		49		32		15		744																					

Wind Rose

Wind Speed (WS) (km/h)
Evergreen Park - October 2018



Hourly Averages - Wind Speed (Scalar)

Wind Speed (km/h)

Evergreen Park - October 2018

Maximum Speed: 55 km/h on Oct 20 13:00																				Maximum Daily Speed Average: 26.7 km/h on Oct 20										Hours in Service: 744	
Minimum Speed: 0 km/h on Oct 22 21:00																				Minimum Daily Speed Average: 1.9 km/h on Oct 22										Hours of Data: 744	
Maximum Diurnal Speed Average: 16.2 km/h at hour 15																				Minimum Diurnal Speed Average: 5.8 km/h at hour 3										Hours of Missing Data: 0	
Monthly Average Speed: 9.57 km/h																				Percentiles: P ₁ = 0.4 P ₁₀ = 1.4 Q ₁ = 2.8 Median = 6.4 Q ₃ = 12.4 P ₉₀ = 22.5 P ₉₉ = 44.5										Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	6	3	2	1	0	0	2	0	2	3	5	8	9	6	8	6	4	6	8	12	12	12	12	13	6.0	12.5					
2-Oct	14	12	6	9	7	5	6	4	5	4	8	7	11	12	10	16	19	18	16	17	14	14	17	17	11.2	18.7					
3-Oct	19	17	10	9	12	13	10	6	11	9	8	7	8	10	11	9	6	3	2	1	0	1	2	1	7.7	18.9					
4-Oct	1	1	2	1	3	1	6	4	6	7	8	12	12	13	11	16	17	16	5	3	3	7	5	4	6.8	16.8					
5-Oct	3	1	1	3	3	2	3	3	6	8	7	6	7	8	9	8	8	7	4	2	3	2	2	3	4.5	8.8					
6-Oct	1	0	1	3	6	4	2	2	4	2	2	2	4	7	6	8	5	7	7	6	5	8	9	10	4.6	9.5					
7-Oct	8	6	9	8	8	6	7	5	5	9	9	8	10	9	11	10	10	9	9	7	7	7	7	8	7.9	11.0					
8-Oct	6	7	7	6	6	7	6	7	7	6	8	10	9	14	13	15	11	13	13	13	12	12	14	13	9.8	14.9					
9-Oct	11	11	10	9	8	8	8	8	8	9	11	12	14	15	14	16	14	14	14	11	10	10	11	9	11.0	15.7					
10-Oct	8	6	5	3	3	5	3	2	2	2	5	6	9	8	8	6	4	4	3	4	4	2	2	3	4.3	8.8					
11-Oct	2	1	1	1	1	0	4	4	2	6	7	7	16	21	19	19	24	23	21	12	3	2	1	2	8.2	23.7					
12-Oct	3	3	3	3	4	2	7	21	16	15	15	19	11	15	19	17	16	15	7	2	3	2	3	4	9.4	21.4					
13-Oct	5	7	11	11	9	6	5	9	6	9	13	22	21	18	20	16	12	5	1	4	3	8	12	13	10.3	22.1					
14-Oct	9	3	2	1	3	1	2	2	2	1	8	21	25	32	40	51	47	33	31	27	24	22	19	18	17.6	50.7					
15-Oct	17	15	8	10	8	8	8	16	34	38	34	36	30	29	32	31	23	17	8	8	5	5	5	3	17.8	38.0					
16-Oct	1	2	1	2	2	2	1	3	1	1	2	6	29	31	29	22	17	22	19	16	22	23	23	20	12.4	31.4					
17-Oct	15	11	4	4	1	4	1	2	2	2	9	9	14	8	12	4	12	13	4	3	10	12	4	4	6.8	14.8					
18-Oct	6	5	10	25	34	15	8	9	17	18	28	30	24	26	25	24	18	9	4	4	4	1	2	1	14.5	33.6					
19-Oct	1	2	2	2	3	2	1	3	3	4	4	5	7	11	12	12	11	7	7	6	5	4	3	2	4.9	11.9					
20-Oct	2	4	8	5	17	30	32	35	46	41	52	52	55	51	48	43	39	25	19	11	8	6	5	5	26.7	55.0					
21-Oct	11	9	7	4	4	4	4	4	5	9	8	9	6	4	5	6	2	1	1	0	1	2	2	1	4.4	10.7					
22-Oct	1	1	1	1	1	1	2	2	1	1	2	3	3	5	6	4	3	0	0	1	0	2	1	3	1.9	6.1					
23-Oct	1	2	2	2	2	1	0	0	1	0	3	7	4	3	5	5	2	1	1	3	1	2	2	4	2.3	7.3					
24-Oct	2	4	3	4	4	2	1	3	4	7	8	10	12	15	11	9	7	2	3	3	1	1	1	2	4.9	15.3					
25-Oct	1	2	2	1	1	1	1	4	3	4	3	5	6	5	6	18	17	17	19	25	18	17	21	21	9.1	25.4					
26-Oct	19	18	12	4	5	5	2	2	2	2	3	5	5	5	5	5	4	6	2	4	11	18	22	22	7.8	22.4					
27-Oct	22	25	30	34	37	38	36	37	36	35	34	36	33	29	25	22	14	9	3	6	3	2	2	1	22.9	38.2					
28-Oct	0	4	4	5	7	10	9	8	11	7	6	8	13	16	12	12	9	6	6	8	5	4	4	3	7.4	15.9					
29-Oct	3	2	6	6	2	4	5	1	3	3	5	11	11	24	28	26	25	14	2	1	1	2	3	5	8.1	27.6					
30-Oct	5	4	4	5	13	20	18	17	20	24	30	39	41	43	38	38	27	24	14	10	11	15	12	4	19.7	42.7					
31-Oct	4	3	7	3	1	2	4	3	1	2	6	8	5	3	2	3	10	9	7	5	7	11	12	17	5.7	17.4					
6.6 6.2 5.8 6.0 6.9 6.7 6.6 7.3 8.6 9.3 11.3 13.6 14.9 16.0 16.2 16.1 14.1 11.5 8.4 7.6 7.0 7.7 7.7 7.6																									Diurnal Average						
22.3 25.1 29.5 33.8 36.7 38.2 36.0 37.1 45.8 41.3 52.2 51.9 55.0 51.3 48.1 50.7 47.4 33.2 30.7 27.0 23.8 22.6 22.7 22.4																									Diurnal Maximum						
All monthly, daily, and diurnal averages have been calculated using scalar methods																															

Hourly Standard Deviations

Wind Direction (WD) - deg
Evergreen Park - October 2018

Maximum Value: 99.0 deg on Oct 10 19:00																			Hours in Service: 744						
Minimum Value: 4.1 deg on Oct 13 23:00																			Hours of Data: 744						
																			Hours of Missing Data: 0						
Percentiles: P ₁ = 5.4 P ₁₀ = 9.0 Q ₁ = 15.2 Median = 24.6 Q ₃ = 54.3 P ₉₀ = 79.1 P ₉₉ = 94.2																			Hours of Calibration: 0						
																			Percent Operational Time: 100.0						
Day	Hourly Period Ending At (MST)																								Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
1-Oct	16	22	18	34	73	75	39	95	38	33	39	27	48	53	37	33	69	71	21	15	14	15	15	14	95.3
2-Oct	16	14	24	12	15	27	16	20	21	42	38	71	42	38	38	24	15	11	9	8	10	8	8	8	71.1
3-Oct	7	8	20	17	18	11	16	28	19	18	31	45	56	47	46	37	54	49	25	40	79	39	31	42	78.6
4-Oct	75	87	89	92	46	43	14	14	31	15	24	24	26	35	36	30	17	12	17	24	55	22	39	53	92.4
5-Oct	35	59	84	19	24	36	20	17	24	38	36	55	60	48	55	33	23	19	14	25	23	59	78	90	90.2
6-Oct	56	95	83	75	88	78	76	54	25	36	89	83	92	41	47	20	24	18	21	28	28	19	19	22	95.2
7-Oct	30	27	20	20	24	29	35	29	24	18	19	20	15	16	21	16	16	16	17	19	18	21	31	24	34.6
8-Oct	21	21	24	27	21	20	25	17	28	33	23	26	26	20	21	22	17	15	18	19	17	17	13	16	33.4
9-Oct	17	22	20	20	19	24	16	16	19	19	20	22	20	24	25	24	20	16	19	22	24	19	20	19	25.3
10-Oct	23	22	23	45	29	28	45	32	35	46	53	29	23	31	37	28	37	56	99	23	28	67	74	43	99.0
11-Oct	36	69	72	50	70	68	21	47	61	27	23	30	17	11	13	13	11	11	12	21	78	37	50	52	78.2
12-Oct	69	42	29	36	92	45	58	14	19	10	15	19	19	14	10	16	14	11	17	52	34	53	48	31	92.0
13-Oct	38	27	19	13	15	26	16	29	87	11	23	12	14	20	18	25	18	32	58	26	19	12	4	5	87.0
14-Oct	25	31	41	52	21	81	72	82	76	95	62	9	8	8	7	6	7	9	8	10	9	10	10	11	94.8
15-Oct	12	16	13	26	19	23	16	21	6	7	9	11	13	14	22	15	20	10	13	23	18	23	27	66	65.6
16-Oct	45	86	64	92	90	81	77	41	81	90	92	33	12	11	10	9	10	7	12	13	7	7	6	7	92.5
17-Oct	7	22	81	82	95	54	91	78	65	54	16	32	11	23	17	38	35	41	27	54	14	8	47	50	95.4
18-Oct	39	25	15	10	6	21	36	14	7	23	12	12	20	16	18	17	10	15	77	26	27	78	81	67	81.4
19-Oct	84	69	84	60	35	45	73	52	32	36	44	44	46	28	25	23	16	14	10	13	21	39	78	78	84.5
20-Oct	65	22	23	19	14	8	6	7	7	8	8	10	9	9	10	9	7	11	7	17	28	62	76	26	75.7
21-Oct	10	11	13	23	59	69	64	56	41	16	17	21	81	53	64	91	97	22	90	98	64	27	63	83	97.7
22-Oct	78	77	86	80	92	92	78	85	70	59	90	51	91	52	39	50	52	82	91	53	92	51	65	53	92.4
23-Oct	71	92	29	65	80	77	85	61	69	65	67	24	48	69	27	57	83	91	89	74	96	39	93	33	96.1
24-Oct	78	26	19	23	19	62	81	49	32	13	20	22	17	13	10	10	8	46	71	30	68	93	87	79	93.4
25-Oct	89	74	71	66	59	75	92	81	81	91	34	54	41	62	46	58	20	5	7	7	7	6	5	7	91.8
26-Oct	7	8	53	60	78	70	22	60	90	59	80	48	64	69	47	55	38	31	90	61	13	9	9	6	90.0
27-Oct	5	7	6	7	5	6	7	6	6	8	9	10	13	16	15	19	13	16	63	43	76	61	86	71	86.1
28-Oct	83	6	11	13	11	13	14	15	13	24	67	25	16	9	13	12	13	15	17	12	22	28	23	25	82.9
29-Oct	41	54	21	18	47	43	17	89	53	73	31	15	20	11	8	10	11	10	70	88	74	60	71	64	88.7
30-Oct	57	82	80	12	11	8	5	7	6	7	7	10	8	8	8	9	10	12	10	7	22	5	9	51	81.5
31-Oct	15	52	16	68	79	68	18	34	47	46	20	15	16	36	76	24	9	24	16	55	22	16	25	10	79.3
88.6 95.2 88.6 92.4 95.4 92.4 91.8 95.3 90.0 94.8 92.5 83.1 91.7 69.3 76.5 91.0 96.9 91.1 99.0 97.7 96.1 93.4 92.8 90.2																									

PAZA

Henry Pirker Station

Monthly Summary Tables, Graphs and
Roses

Hourly Averages


Sulphur Dioxide (SO₂) - ppb

Henry Pirker - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 2.4 ppb on Oct 22 15:00 Maximum Daily Average: 0.7 ppb on Oct 22																			Hours in Service: 744 Hours of Data: 708 Hours of Missing Data: 36 Hours of Calibration: 36 Percent Operational Time: 100.0							
Minimum Value: 0 ppb on Oct 3 00:00 Minimum Daily Average: 0.1 ppb on Oct 20 Maximum Diurnal Average: 0.4 ppb at hour 13 Minimum Diurnal Average: 0.1 ppb at hour 3 Monthly Average: 0.26 ppb Percentiles: P ₁ = 0.0 P ₁₀ = 0.1 Q ₁ = 0.1 Median = 0.2 Q ₃ = 0.3 P ₉₀ = 0.6 P ₉₉ = 1.5																										
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	0	0	0	1	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
3-Oct	0	0	0	0	0	0	0	0	0	A	1	1	1	1	2	0	0	0	0	0	0	0	0	0	0.4	1.6
4-Oct	0	0	0	0	0	0	0	1	A	0	0	0	0	0	0	2	0	0	0	0	0	0	1	1	0.4	1.7
5-Oct	0	0	0	0	0	0	0	A	C	C	C	C	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5
6-Oct	0	0	0	0	0	0	A	0	0	0	0	1	2	1	0	0	0	0	0	0	0	0	0	0	0.3	2.0
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4
10-Oct	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	1	1	1	1	1	1	0.4	0.9
11-Oct	0	A	0	0	0	0	0	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.7
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	A	0.2	1.1
13-Oct	0	0	0	1	1	1	1	1	0	1	1	1	0	0	0	0	0	0	0	0	0	0	A	0	0.4	1.0
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0.2	0.5
15-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	1	0	0.2	0.6
16-Oct	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0.2	0.8
17-Oct	0	0	0	0	0	0	0	1	1	1	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0.3	1.2
18-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.2	0.4
19-Oct	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	A	0	0	0	0	0	0	0	0.2	0.6
20-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0.1	0.1
21-Oct	0	0	0	0	0	0	0	0	0	0	0	0	1	1	A	0	0	0	0	0	0	0	0	0	0.3	1.3
22-Oct	0	0	0	0	0	0	0	0	1	1	1	1	1	A	2	2	2	2	1	1	0	0	0	0	0.7	2.4
23-Oct	0	0	0	0	0	0	0	1	1	1	0	1	A	0	0	1	1	1	1	0	0	0	0	0	0.4	0.8
24-Oct	0	0	0	0	0	0	0	0	1	1	1	A	0	0	0	0	0	0	0	0	0	1	1	0	0.3	0.7
25-Oct	0	0	0	0	0	0	0	1	1	0	A	1	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.6
26-Oct	0	0	0	0	0	0	0	0	0	A	0	0	1	1	1	1	1	0	0	0	0	0	0	0	0.4	1.4
27-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
28-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5
29-Oct	0	0	0	0	0	0	A	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0.2	1.4
30-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
31-Oct	0	0	0	0	A	0	1	1	1	1	1	2	1	1	0	0	0	0	0	0	0	0	0	0	0.4	1.8
0.2 0.1 0.1 0.2 0.2 0.2 0.3 0.3 0.4 0.3 0.3 0.4 0.4 0.3 0.3 0.3 0.3 0.3 0.3 0.2 0.2 0.2 0.2 0.2																									Diurnal Average	
0.5 0.4 0.3 0.7 0.7 0.7 1.0 1.0 1.2 1.0 1.1 1.8 2.0 1.5 2.4 1.7 2.3 1.7 1.2 1.0 0.7 1.1 0.7 0.6																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 172 ppb 24-hr 48 ppb 30-day 11 ppb																										

Hourly Averages

Sulphur Dioxide (SO₂) - ppb
Henry Pirker - October 2018



Hourly Maximums


Sulphur Dioxide (SO₂) - ppb

Henry Pirker - October 2018

Maximum Value: 4.0 ppb on Oct 4 16:00																		Maximum Daily Average: 1.0 ppb on Oct 22										Hours in Service: 744	
Minimum Value: 0 ppb on Oct 3 02:00																		Minimum Daily Average: 0.2 ppb on Oct 20										Hours of Data: 708	
Maximum Diurnal Average: 0.8 ppb at hour 13																		Minimum Diurnal Average: 0.2 ppb at hour 3										Hours of Missing Data: 36	
Monthly Average: 0.43 ppb																		Percentiles: P ₁ = 0.1 P ₁₀ = 0.1 Q ₁ = 0.2 Median = 0.3 Q ₃ = 0.5 P ₉₀ = 0.9 P ₉₉ = 2.4										Hours of Calibration: 36	
																		Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum			
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24					
1-Oct	0	0	0	1	0	1	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.9			
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	1	0	0	0	0	0	0	0	0	0	0	0	0.2	0.6			
3-Oct	0	0	0	0	0	0	0	0	0	A	2	2	2	2	3	0	0	0	0	0	0	0	0	0	0.6	2.8			
4-Oct	0	0	0	0	0	0	0	1	A	1	0	0	1	1	1	4	1	1	1	1	1	1	1	1	0.7	4.0			
5-Oct	1	0	0	0	0	0	0	A	C	C	C	C	0	1	0	0	0	0	0	0	0	0	0	0	0.3	0.7			
6-Oct	0	0	0	0	0	0	A	0	0	0	0	1	4	1	1	0	0	0	0	0	0	0	0	0	0.5	3.7			
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3			
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4			
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.5			
10-Oct	0	0	A	0	0	0	0	0	0	0	0	0	0	1	1	1	1	1	1	1	1	1	1	1	0.6	1.3			
11-Oct	0	A	0	0	0	0	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0.4	1.1			
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	2	1	A	0.4	1.7			
13-Oct	0	0	1	1	1	1	2	1	1	1	1	1	1	0	0	0	1	0	0	0	0	0	A	0	0.7	1.5			
14-Oct	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0.3	0.7			
15-Oct	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	1	0	0	A	1	1	1	0.3	0.9			
16-Oct	0	0	0	0	0	0	1	1	1	1	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0.4	1.1			
17-Oct	0	0	0	0	0	0	1	1	2	1	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0.4	1.9			
18-Oct	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	A	1	0	1	0	0	0	0.3	0.9			
19-Oct	0	0	0	0	0	0	0	0	1	0	1	1	1	1	1	0	A	0	0	0	0	0	0	0	0.4	0.9			
20-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0.2	0.3			
21-Oct	0	0	0	0	0	0	1	1	1	0	0	1	2	1	A	0	1	0	1	0	0	0	0	0	0.5	2.0			
22-Oct	0	0	0	0	0	0	0	1	1	1	1	1	2	A	3	2	3	2	1	2	1	0	0	0	1.0	3.4			
23-Oct	0	0	0	0	0	1	1	1	1	1	1	1	A	0	1	1	1	1	1	1	1	1	0	0	0.6	1.0			
24-Oct	0	0	0	0	0	1	1	1	1	1	1	A	0	0	0	0	0	0	0	1	1	1	1	1	0.5	0.9			
25-Oct	0	0	0	0	0	1	1	1	1	1	A	1	1	1	0	1	0	0	1	0	0	0	0	0	0.5	1.0			
26-Oct	0	0	0	0	0	0	0	1	0	A	1	1	1	1	3	2	1	1	1	1	0	0	0	0	0.6	2.5			
27-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0.2	0.5				
28-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0.3	0.7			
29-Oct	0	0	0	0	0	0	A	0	0	0	1	1	3	0	0	0	0	0	0	0	0	0	0	0	0.5	3.5			
30-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	2	0	0	0	0	0	0.3	1.8			
31-Oct	0	0	0	0	A	1	1	2	1	1	2	2	2	1	1	0	0	0	0	0	0	0	1	0	0.7	2.3			
0.3 0.3 0.2 0.3 0.3 0.4 0.4 0.5 0.5 0.5 0.5 0.6 0.8 0.5 0.6 0.6 0.5 0.4 0.5 0.4 0.4 0.4 0.3 0.3																								Diurnal Average					
0.7 0.9 0.6 1.1 0.9 1.3 1.5 1.9 1.9 1.3 1.6 2.4 3.7 2.4 3.4 4.0 2.7 2.2 1.8 1.6 1.4 1.7 1.0 0.9																								Diurnal Maximum					
C - Calibration A - Automated Daily Zero Span																													

Hourly Maximums


Sulphur Dioxide (SO₂) - ppb
Henry Pirker - October 2018




Pollutant Rose

Sulphur Dioxide (SO₂) - ppb

Henry Pirker - October 2018




Pollutant Classes (ppb)



Span Responses

Sulphur Dioxide (SO₂)
Henry Pirker - October 2018



Hourly Averages


Total Reduced Sulphur (TRS) - ppb

Henry Pirker - October 2018

Maximum Value: 1.6 ppb on Oct 17 10:00																				Maximum Daily Average: 0.6 ppb on Oct 22										Hours in Service: 744									
Minimum Value: 0 ppb on Oct 1 14:00																				Minimum Daily Average: 0.0 ppb on Oct 2										Hours of Data: 708									
Maximum Diurnal Average: 0.5 ppb at hour 9																				Minimum Diurnal Average: 0.1 ppb at hour 17										Hours of Missing Data: 36									
Monthly Average: 0.23 ppb																				Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.1 Median = 0.2 Q ₃ = 0.3 P ₉₀ = 0.5 P ₉₉ = 1.0										Hours of Calibration: 36									
																				Percent Operational Time: 100.0																			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum													
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24															
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.4													
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.2													
3-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2													
4-Oct	0	0	0	0	0	0	0	1	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.9													
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2													
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.4													
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2													
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1													
9-Oct	0	0	0	A	0	0	0	0	0	0	0	C	C	C	C	0	0	0	0	0	0	0	0	0	0.1	0.3													
10-Oct	0	0	A	0	0	0	0	0	0	0	0	0	0	C	0	0	0	0	0	0	0	0	0	0	0.2	0.3													
11-Oct	0	A	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.6													
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0.1	0.3													
13-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	A	0	0.3	0.6													
14-Oct	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0.2	0.6													
15-Oct	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0.2	0.6													
16-Oct	0	0	0	0	0	0	0	1	1	1	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0.3	1.0													
17-Oct	0	0	0	0	0	0	0	1	1	2	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0.3	1.6													
18-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.1	0.4													
19-Oct	0	0	0	0	0	1	0	0	1	1	1	0	0	0	0	0	A	0	0	0	0	0	0	0	0.4	0.7													
20-Oct	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0.2	0.6													
21-Oct	0	0	0	0	0	0	1	0	1	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0.2	0.5													
22-Oct	0	0	0	0	0	0	1	1	1	1	1	1	1	A	1	0	0	0	1	1	1	1	0	0	0.6	1.3													
23-Oct	1	1	0	0	0	1	1	1	1	1	0	0	A	0	0	0	0	0	0	0	0	1	0	0	0.4	1.1													
24-Oct	0	0	1	0	1	1	1	1	1	1	1	A	0	0	0	0	0	0	0	1	0	1	1	1	0.5	1.0													
25-Oct	1	1	1	1	1	1	1	1	1	1	A	1	0	0	0	0	0	0	0	0	0	0	0	0	0.5	1.0													
26-Oct	0	0	0	0	0	0	0	1	1	A	1	0	0	1	1	1	0	0	0	0	0	0	0	0	0.3	0.7													
27-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.4													
28-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5													
29-Oct	0	0	0	0	0	0	A	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.7													
30-Oct	0	0	0	0	0	A	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.7													
31-Oct	0	0	0	0	A	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.8													
0.2 0.2 0.2 0.2 0.2 0.3 0.3 0.4 0.5 0.4 0.3 0.2 0.2 0.1 0.2 0.1 0.1 0.2 0.2 0.2 0.2 0.2 0.2																									Diurnal Average														
0.7 0.8 0.9 0.7 0.7 0.8 0.9 1.0 1.5 1.6 1.1 0.8 0.8 0.6 0.5 0.6 0.4 0.4 0.7 0.9 0.8 0.8 0.9 0.6																									Diurnal Maximum														
C - Calibration																									A - Automated Daily Zero Span														

Hourly Averages

Total Reduced Sulphur (TRS) - ppb
Henry Pirker - October 2018



Hourly Maximums

Total Reduced Sulphur (TRS) - ppb


Henry Pirker - October 2018

Maximum Value: 2.8 ppb on Oct 17 10:00																				Maximum Daily Average: 1.2 ppb on Oct 22										Hours in Service: 744									
Minimum Value: 0 ppb on Oct 2 20:00																				Minimum Daily Average: 0.5 ppb on Oct 2										Hours of Data: 708									
Maximum Diurnal Average: 1.0 ppb at hour 9																				Minimum Diurnal Average: 0.6 ppb at hour 17										Hours of Missing Data: 36									
Monthly Average: 0.72 ppb																				Percentiles: P ₁ = 0.3 P ₁₀ = 0.5 Q ₁ = 0.5 Median = 0.6 Q ₃ = 0.8 P ₉₀ = 1.1 P ₉₉ = 1.7										Hours of Calibration: 36									
																				Percent Operational Time: 100.0																			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum													
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24															
1-Oct	1	1	1	1	1	1	1	1	1	1	1	A	1	0	0	0	0	0	0	1	0	1	1	0	0.6	1.1													
2-Oct	0	1	1	0	1	0	1	1	1	1	A	0	1	0	0	0	0	1	0	0	1	0	0	1	0.5	0.9													
3-Oct	0	0	0	1	0	1	0	1	1	A	1	1	1	1	1	0	1	1	1	1	0	0	1	0	0.5	0.8													
4-Oct	0	1	1	1	1	1	1	1	A	1	0	1	1	1	1	0	1	1	1	1	1	1	1	1	0.7	1.4													
5-Oct	1	1	1	1	1	1	1	A	1	0	1	0	0	1	1	1	0	1	0	1	1	1	1	0	0.6	0.8													
6-Oct	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.7	1.2													
7-Oct	0	1	1	1	1	A	1	1	1	1	1	1	1	1	0	1	1	1	1	1	1	1	1	1	0.6	0.8													
8-Oct	1	1	1	1	A	1	1	1	1	0	1	0	0	0	1	1	1	0	1	1	0	0	1	0	0.5	0.6													
9-Oct	0	1	0	A	0	0	0	0	0	0	0	C	C	C	C	1	1	1	1	1	1	1	0	0	0.5	0.7													
10-Oct	0	0	A	1	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.6	1.0													
11-Oct	1	A	1	1	1	1	1	1	1	1	1	1	1	0	1	1	1	1	1	1	0	0	1	1	0.7	1.0													
12-Oct	A	1	1	1	1	1	1	1	0	0	0	1	0	0	1	0	1	0	1	1	1	1	1	A	0.6	0.9													
13-Oct	1	1	1	0	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	0.8	1.0													
14-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	0	1	0	A	0	1	0.7	1.1													
15-Oct	1	1	1	1	1	1	1	1	0	0	1	0	1	0	0	1	1	1	1	1	A	1	1	1	0.6	1.1													
16-Oct	1	1	1	1	1	1	1	2	2	1	1	1	1	0	1	0	1	1	1	A	1	1	1	0	0.7	1.6													
17-Oct	0	1	1	1	1	1	1	1	3	3	1	1	1	1	1	1	1	1	A	1	1	1	1	1	0.9	2.8													
18-Oct	1	0	1	0	1	1	1	1	1	0	0	1	1	1	1	0	0	A	1	1	1	1	1	1	0.6	0.8													
19-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	0.9	1.3													
20-Oct	1	1	1	1	1	1	1	1	1	1	1	1	0	1	0	A	1	0	1	1	1	1	1	1	0.7	1.1													
21-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	0	1	1	1	1	1	1	1	0.7	1.1													
22-Oct	1	1	1	1	1	1	1	2	2	2	2	1	1	A	1	1	1	1	1	1	1	1	1	1	1.2	2.0													
23-Oct	1	1	1	1	1	1	1	2	2	1	1	1	A	1	1	0	0	1	1	1	1	1	1	1	1.0	1.7													
24-Oct	1	1	1	1	1	1	1	1	1	1	2	A	1	1	1	1	1	1	1	1	1	1	2	1	1.0	1.7													
25-Oct	1	1	1	1	1	1	2	1	1	1	A	1	1	1	1	1	0	1	1	0	1	1	0	0	0.9	1.6													
26-Oct	1	0	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.3													
27-Oct	1	1	0	0	1	1	1	0	A	0	1	1	1	1	1	0	1	1	1	1	1	1	1	1	0.6	1.0													
28-Oct	1	1	1	1	1	0	1	A	1	0	1	1	1	1	1	1	1	1	1	1	1	1	0	1	0.7	1.1													
29-Oct	1	1	1	1	1	1	A	1	1	1	2	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.6													
30-Oct	1	1	1	1	1	A	1	1	1	2	1	1	1	1	0	1	1	1	1	1	0	0	0	1	1	0.7	2.5												
31-Oct	0	0	1	1	A	1	1	1	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	2.4													
0.7 0.7 0.7 0.7 0.7 0.8 0.8 0.9 1.0 1.0 0.8 0.7 0.7 0.6 0.6 0.6 0.6 0.6 0.7 0.6 0.7 0.7 0.7 0.6																									Diurnal Average														
1.3 1.2 1.4 1.3 1.2 1.3 1.6 1.7 2.5 2.8 1.7 1.4 1.3 1.2 1.4 1.1 1.0 0.9 1.2 1.3 1.5 1.4 1.5 1.0																									Diurnal Maximum														
C - Calibration																									A - Automated Daily Zero Span														

Hourly Maximums

Total Reduced Sulphur (TRS) - ppb


Henry Pirker - October 2018



Pollutant Rose


Total Reduced Sulphur (TRS) - ppb

Henry Pirker - October 2018



Span Responses

Total Reduced Sulphur (TRS)
Henry Pirker - October 2018



Hourly Averages

Nitrogen Dioxide (NO₂) - ppb


Henry Pirker - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 60.1 ppb on Oct 17 09:00 Maximum Daily Average: 22.6 ppb on Oct 22																				Hours in Service: 744 Hours of Data: 706 Hours of Missing Data: 38 Hours of Calibration: 38 Percent Operational Time: 100.0							
Minimum Value: 1 ppb on Oct 13 15:00 Maximum Diurnal Average: 17.8 ppb at hour 8 Monthly Average: 10.93 ppb										Minimum Daily Average: 2.6 ppb on Oct 8 Minimum Diurnal Average: 5.6 ppb at hour 14 Percentiles: P ₁ = 1.2 P ₁₀ = 2.5 Q ₁ = 3.8 Median = 8.1 Q ₃ = 16.1 P ₉₀ = 23.7 P ₉₉ = 34.8																	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	4	3	3	6	7	14	17	19	13	8	11	A	3	3	5	4	8	8	5	3	4	3	2	2	6.8	19.2	
2-Oct	2	3	3	2	3	7	10	15	12	4	A	2	1	2	2	2	2	4	4	6	7	6	5	4	4.6	14.6	
3-Oct	3	3	3	4	6	8	9	13	14	A	8	6	4	3	3	2	2	4	17	15	10	8	7	8	7.0	17.2	
4-Oct	8	15	18	14	17	18	20	25	A	18	9	5	3	2	2	2	3	4	8	14	17	15	13	11	11.4	25.3	
5-Oct	14	10	8	5	6	9	8	A	C	C	C	C	C	C	3	3	3	4	7	8	9	9	11	10	--	13.6	
6-Oct	13	16	15	15	15	20	A	21	19	19	17	10	3	4	5	7	7	6	5	5	4	3	3	3	10.2	20.6	
7-Oct	3	3	3	2	2	A	3	3	3	3	3	3	3	3	3	4	4	5	4	4	4	4	3	3	3.3	4.6	
8-Oct	3	3	3	3	A	3	3	3	2	3	2	2	2	2	2	2	3	2	3	3	2	2	2	3	2.6	3.3	
9-Oct	3	3	3	A	4	5	5	6	5	5	4	3	3	3	3	3	4	4	5	4	4	5	5	5	4.0	5.7	
10-Oct	5	5	A	6	7	8	10	12	10	9	12	9	5	10	4	5	6	7	8	10	11	9	8	12	8.2	12.0	
11-Oct	11	A	17	17	16	15	14	15	15	13	11	10	5	5	5	5	5	8	7	6	8	8	15	9	10.5	17.3	
12-Oct	A	8	8	7	10	18	15	6	7	4	3	3	4	3	2	2	3	2	4	7	8	10	15	A	6.7	18.4	
13-Oct	8	7	4	2	3	7	8	9	10	7	2	1	1	1	1	1	1	7	20	19	9	9	A	7	6.2	19.6	
14-Oct	4	5	10	15	17	19	19	17	18	20	15	6	5	5	4	3	3	3	2	3	2	A	3	3	8.7	20.5	
15-Oct	3	5	3	3	8	21	28	16	5	4	4	3	2	1	1	2	2	3	8	15	A	15	15	6	7.5	28.3	
16-Oct	4	9	8	7	13	27	36	40	41	33	16	14	4	4	5	5	6	9	13	A	11	5	5	6	14.0	41.2	
17-Oct	6	5	7	15	26	35	42	53	60	54	27	16	8	8	13	19	22	18	A	16	18	12	5	5	21.3	60.1	
18-Oct	5	3	2	1	3	6	5	6	8	5	3	2	2	1	1	2	5	A	24	25	29	26	25	26	9.3	28.8	
19-Oct	22	21	17	15	16	19	23	24	24	21	23	15	11	5	6	4	A	8	9	9	8	9	17	16	14.9	24.3	
20-Oct	14	15	13	15	7	3	3	4	3	3	2	2	2	2	2	A	4	5	5	6	14	14	9	10	6.8	15.5	
21-Oct	6	8	4	4	7	18	27	28	25	16	9	8	2	3	A	5	3	13	22	27	23	16	13	15	13.2	28.2	
22-Oct	19	19	19	20	15	17	17	24	28	30	29	26	26	A	27	23	16	24	31	30	22	20	19	18	22.6	30.9	
23-Oct	17	17	15	14	15	19	18	20	27	25	21	15	A	13	18	12	11	25	30	29	27	22	19	17	19.5	29.5	
24-Oct	16	16	17	16	17	20	22	24	26	20	22	A	11	11	18	19	25	32	33	33	31	31	29	21	22.2	33.4	
25-Oct	17	15	17	13	13	18	19	23	25	23	A	20	14	10	14	13	8	9	6	6	5	5	4	3	13.1	25.2	
26-Oct	3	2	5	9	16	28	33	31	29	A	24	23	24	24	29	25	22	23	25	27	25	7	5	3	19.2	32.9	
27-Oct	2	1	1	2	2	3	3	3	A	5	4	2	2	1	1	2	5	18	12	15	14	20	26	27	7.4	26.5	
28-Oct	26	23	14	13	8	6	5	A	5	4	5	12	10	9	10	11	13	15	17	12	11	14	9	13	11.5	25.7	
29-Oct	16	17	15	17	16	15	A	16	15	15	15	9	6	4	3	4	6	6	17	24	29	30	23	11	14.3	29.8	
30-Oct	17	15	5	8	7	A	10	9	12	12	6	3	3	3	3	3	4	5	9	10	13	9	10	8	7.9	16.7	
31-Oct	11	8	7	10	A	23	32	31	27	28	23	23	20	16	15	17	18	9	8	9	7	5	5	4	15.4	31.6	
9.4 9.5 8.9 9.4 10.5 14.8 16.0 17.8 17.4 14.7 11.9 9.2 6.5 5.6 7.0 7.0 7.4 9.6 12.3 13.3 12.9 11.6 10.9 9.6 25.7 23.4 19.2 20.2 26.3 34.9 41.5 52.7 60.1 53.7 29.0 25.6 26.4 23.8 29.2 25.0 25.4 31.8 33.3 33.4 31.3 30.9 28.9 26.5																										Diurnal Average	Diurnal Maximum
C - Calibration A - Automated Daily Zero Span Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 159 ppb 24-hr 106 ppb																											

Hourly Averages

Nitrogen Dioxide (NO₂) - ppb

Henry Pirker - October 2018



Hourly Maximums

Nitrogen Dioxide (NO₂) - ppb


Henry Pirker - October 2018

Maximum Value: 79.0 ppb on Oct 17 09:00																				Maximum Daily Average: 28.9 ppb on Oct 17					Hours in Service: 744	
Minimum Value: 2 ppb on Oct 13 15:00																				Minimum Daily Average: 4.3 ppb on Oct 8					Hours of Data: 706	
Maximum Diurnal Average: 25.0 ppb at hour 8																				Minimum Diurnal Average: 9.1 ppb at hour 13					Hours of Missing Data: 38	
Monthly Average: 14.80 ppb																				Percentiles: P ₁ = 2.1 P ₁₀ = 4.0 Q ₁ = 6.1 Median = 12.0 Q ₃ = 20.5 P ₉₀ = 29.2 P ₉₉ = 50.0					Hours of Calibration: 38	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	5	4	5	8	10	20	21	57	20	12	13	A	6	5	17	6	13	13	11	6	9	5	4	3	11.8	56.9
2-Oct	3	4	5	4	7	11	25	19	17	9	A	2	2	4	3	3	4	7	6	8	10	8	7	5	7.5	24.9
3-Oct	4	5	5	6	11	14	11	18	21	A	12	12	5	5	5	3	4	6	23	23	13	11	9	13	10.3	23.3
4-Oct	11	19	20	17	20	20	25	32	A	27	11	7	6	3	4	3	5	6	12	22	21	22	17	16	15.0	31.6
5-Oct	16	13	14	7	8	15	11	A	C	C	C	C	C	C	4	7	5	6	11	11	11	17	13	12	--	17.0
6-Oct	18	19	19	19	19	29	A	23	23	20	19	15	5	11	8	10	9	8	8	7	7	4	5	4	13.4	29.2
7-Oct	4	4	4	3	3	A	5	7	5	5	5	5	5	4	5	5	5	7	7	5	5	5	5	6	5.0	6.8
8-Oct	5	4	4	4	A	5	7	6	4	10	3	4	3	3	3	4	5	3	4	4	3	3	3	3	4.3	9.7
9-Oct	5	4	5	A	5	7	7	13	7	7	5	4	4	4	4	4	7	14	12	5	6	6	6	8	6.4	13.7
10-Oct	7	9	A	8	10	12	13	14	13	13	14	12	6	68	6	7	7	12	11	13	14	12	10	18	13.3	68.1
11-Oct	13	A	18	18	18	16	16	16	17	16	13	13	9	7	8	8	8	11	9	11	11	13	18	12	13.0	18.3
12-Oct	A	11	9	9	16	21	20	12	10	7	6	5	13	5	5	3	6	3	7	11	11	17	19	A	10.1	20.7
13-Oct	11	9	9	3	5	14	13	12	15	12	3	2	2	3	2	2	3	18	25	28	11	11	A	8	9.6	28.5
14-Oct	7	9	16	18	19	21	20	20	20	24	19	9	6	11	7	4	4	5	5	5	5	A	5	7	11.5	24.3
15-Oct	6	7	7	7	15	33	38	35	8	5	6	5	4	2	3	3	4	7	23	27	A	21	24	12	13.1	37.6
16-Oct	6	16	11	9	33	34	42	56	46	38	21	21	8	6	8	8	9	14	22	A	17	8	10	16	19.9	55.7
17-Oct	8	11	8	22	37	43	50	63	79	66	38	25	19	13	23	26	29	23	A	21	24	22	9	7	28.9	79.0
18-Oct	8	5	6	2	5	10	11	11	13	8	4	4	4	2	3	3	8	A	28	30	31	32	28	28	12.4	31.6
19-Oct	28	24	20	18	17	23	25	28	38	23	28	18	15	8	8	9	A	11	14	12	14	11	21	18	18.8	38.1
20-Oct	18	19	19	18	18	5	4	8	4	4	3	3	3	4	4	A	5	7	8	9	21	20	16	15	10.1	20.5
21-Oct	12	13	7	8	12	28	32	30	32	59	15	9	4	5	A	9	6	23	26	35	31	27	18	20	19.9	58.8
22-Oct	21	22	25	25	20	20	20	28	34	34	33	28	28	A	30	31	20	31	35	36	27	20	20	19	26.5	36.3
23-Oct	19	19	18	15	18	25	25	24	33	31	24	19	A	15	21	18	17	32	32	31	30	26	21	20	23.2	32.8
24-Oct	17	17	19	19	20	24	28	29	31	27	24	A	14	14	23	26	31	35	36	37	33	34	35	28	26.2	37.3
25-Oct	22	19	20	16	16	23	26	42	30	28	A	27	16	13	19	25	13	15	9	9	9	10	5	5	18.2	42.4
26-Oct	6	3	9	16	26	32	38	34	33	A	27	26	27	27	32	29	24	26	28	30	28	11	7	3	22.7	38.2
27-Oct	4	2	4	2	3	4	4	5	A	8	5	4	3	3	3	4	8	29	23	20	24	25	30	30	10.8	30.1
28-Oct	28	29	20	16	11	8	6	A	7	6	10	15	16	11	12	13	16	18	20	14	15	17	10	20	14.6	28.7
29-Oct	19	20	19	21	19	16	A	21	17	16	19	18	8	6	4	7	8	12	21	30	34	33	30	14	17.9	33.7
30-Oct	19	20	17	12	9	A	18	25	16	16	9	6	5	4	6	6	7	8	13	15	18	11	17	14	12.6	24.8
31-Oct	17	9	10	13	A	32	38	38	33	31	28	25	21	25	24	25	27	14	11	12	8	7	6	5	19.9	38.3
12.2 12.3 12.4 12.1 14.8 19.4 20.6 25.0 22.3 20.1 14.9 12.3 9.1 10.0 10.1 10.4 10.6 14.0 16.6 17.6 16.7 15.7 14.2 13.0																								Diurnal Average		
28.5 28.7 24.8 25.3 36.7 42.7 50.2 62.5 79.0 65.7 38.2 28.2 27.8 68.1 32.1 31.2 31.4 34.8 36.2 37.3 33.7 34.3 35.3 30.1																								Diurnal Maximum		
C - Calibration A - Automated Daily Zero Span																										

Hourly Maximums

Nitrogen Dioxide (NO₂) - ppb


Henry Pirker - October 2018



Pollutant Rose


Nitrogen Dioxide (NO₂) - ppb

Henry Pirker - October 2018



Span Responses

Nitrogen Dioxide (NO₂)
Henry Pirker - October 2018




Hourly Averages

Nitrogen Oxide (NO) - ppb
Henry Pirker - October 2018

Maximum Value: 152.7 ppb on Oct 17 09:00																					Maximum Daily Average: 34.5 ppb on Oct 22					Hours in Service: 744	
Minimum Value: 0 ppb on Oct 15 03:00																					Minimum Daily Average: 0.4 ppb on Oct 2					Hours of Data: 706	
Maximum Diurnal Average: 33.9 ppb at hour 9																					Minimum Diurnal Average: 1.3 ppb at hour 18					Hours of Missing Data: 38	
Monthly Average: 8.09 ppb																					Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.2 Median = 0.7 Q ₃ = 5.4 P ₉₀ = 26.3 P ₉₉ = 101.5					Hours of Calibration: 38	
																										Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	0	0	0	0	0	1	4	6	4	2	3	A	0	0	1	0	1	1	1	1	0	0	0	0	1.1	6.4	
2-Oct	0	0	0	0	0	0	1	2	2	1	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	2.4	
3-Oct	0	0	0	0	0	0	1	2	6	A	4	4	2	1	1	0	0	0	1	1	0	0	0	0	1.0	5.8	
4-Oct	0	2	3	4	11	18	39	99	A	44	10	2	1	0	0	0	0	0	0	0	1	0	0	0	10.2	98.6	
5-Oct	0	0	0	0	0	1	1	A	C	C	C	C	C	C	1	1	1	1	1	0	1	1	1	0	--	1.3	
6-Oct	1	1	3	9	5	10	A	13	24	19	15	5	1	1	1	1	0	1	1	0	0	0	0	0	4.8	24.3	
7-Oct	0	0	0	0	0	A	1	1	1	2	1	2	2	2	2	1	1	1	1	0	0	0	0	0	1.0	2.3	
8-Oct	1	0	0	0	A	1	1	1	1	2	1	2	1	1	0	1	1	0	0	0	0	0	0	0	0.7	2.0	
9-Oct	0	0	0	A	0	1	1	2	1	1	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0.8	1.8	
10-Oct	0	0	A	0	1	1	2	2	2	2	6	3	1	13	1	1	0	1	0	0	1	0	0	0	1.7	13.4	
11-Oct	0	A	3	8	18	22	23	25	37	16	10	8	2	1	1	1	0	0	0	0	0	0	0	0	7.6	36.8	
12-Oct	A	0	0	0	1	5	8	2	1	1	1	1	1	1	1	0	0	0	0	0	0	0	1	A	1.0	7.8	
13-Oct	0	0	0	0	0	0	0	1	2	4	1	1	0	0	0	0	0	0	2	8	0	0	A	0	0.9	7.6	
14-Oct	0	0	0	1	2	7	15	22	35	43	14	1	0	0	0	0	0	0	0	0	0	A	0	0	6.1	43.5	
15-Oct	0	0	0	0	0	6	38	11	0	0	0	1	0	0	0	0	0	0	0	1	A	0	0	0	2.5	37.8	
16-Oct	0	0	0	0	1	4	38	102	107	43	6	7	1	1	1	0	1	0	0	A	0	0	0	0	13.5	107.3	
17-Oct	0	0	0	0	1	3	21	80	153	120	10	4	1	1	1	2	1	0	A	0	0	0	0	0	17.3	152.7	
18-Oct	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	A	2	3	10	4	8	13	1.8	12.9	
19-Oct	21	18	14	10	11	27	19	16	48	43	46	17	6	2	2	1	A	0	0	0	0	0	1	1	13.2	48.2	
20-Oct	0	2	1	2	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0.4	2.2	
21-Oct	0	0	0	0	0	1	17	18	36	9	5	4	0	1	A	1	0	0	0	3	1	1	0	2	4.3	35.9	
22-Oct	4	3	6	13	14	35	36	81	114	102	73	39	34	A	21	13	3	6	42	73	51	12	9	10	34.5	113.7	
23-Oct	9	18	10	10	25	54	65	86	114	80	40	18	A	7	9	3	2	6	10	12	28	35	17	6	28.8	113.7	
24-Oct	5	6	16	19	26	37	55	54	80	59	52	A	5	4	7	6	8	10	21	38	40	88	103	46	34.2	103.3	
25-Oct	32	22	28	18	25	47	62	86	76	58	A	22	11	4	3	2	0	0	0	0	0	0	0	0	21.6	85.8	
26-Oct	0	0	0	0	0	9	29	38	37	A	19	18	21	31	30	27	15	6	3	9	5	0	0	0	13.0	38.1	
27-Oct	0	0	0	0	0	0	0	0	A	1	1	1	0	0	0	0	0	1	0	0	1	1	13	29	2.1	28.7	
28-Oct	41	13	1	1	1	0	0	A	0	1	0	3	2	2	2	3	3	2	5	1	0	1	0	0	3.6	40.8	
29-Oct	1	3	6	19	24	31	A	36	42	39	28	11	4	1	1	1	0	0	0	1	10	17	4	0	12.2	42.2	
30-Oct	1	0	0	0	0	A	1	1	1	3	2	1	1	1	1	1	0	0	0	0	0	0	0	0	0.6	3.1	
31-Oct	0	0	0	0	A	1	20	36	22	27	14	13	10	6	5	4	3	0	0	1	0	0	0	0	7.1	35.6	
3.9 3.0 3.1 3.8 5.7 11.1 17.2 28.4 33.9 25.8 13.0 6.7 3.8 2.8 3.1 2.3 1.5 1.3 3.1 5.1 5.1 5.4 5.3 3.6																									Diurnal Average		
40.8 21.7 28.4 18.8 25.5 53.6 64.8 102.1 152.7 120.2 72.8 39.1 34.0 30.6 30.0 27.4 15.0 9.9 41.9 73.2 50.9 88.1 103.3 45.5																									Diurnal Maximum		
C - Calibration																									A - Automated Daily Zero Span		

Hourly Averages

Nitrogen Oxide (NO) - ppb
Henry Pirker - October 2018



Hourly Maximums


Nitrogen Oxide (NO) - ppb

Henry Pirker - October 2018

Maximum Value: 283.5 ppb on Oct 17 09:00																								Hours in Service: 744		
Maximum Daily Average: 50.6 ppb on Oct 24																								Hours of Data: 706		
Minimum Value: 0 ppb on Oct 14 20:00																								Hours of Missing Data: 38		
Maximum Diurnal Average: 51.3 ppb at hour 9																								Hours of Calibration: 38		
Monthly Average: 14.41 ppb																								Percent Operational Time: 100.0		
Minimum Daily Average: 1.6 ppb on Oct 2																										
Minimum Diurnal Average: 3.3 ppb at hour 17																										
Percentiles: P ₁ = 0.0 P ₁₀ = 0.3 Q ₁ = 0.9 Median = 2.5 Q ₃ = 13.6 P ₉₀ = 44.7 P ₉₉ = 131.3																										
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	2	2	1	1	1	7	9	18	11	4	5	A	7	1	11	1	3	3	5	3	2	1	0	2	4.4	18.3
2-Oct	1	0	0	0	1	1	14	4	4	2	A	0	0	2	1	1	1	1	0	0	1	0	0	0	1.6	14.4
3-Oct	0	0	0	0	0	2	2	5	35	A	8	9	3	2	2	1	1	0	2	3	1	1	0	1	3.3	35.2
4-Oct	0	14	4	7	23	29	77	143	A	88	14	4	5	1	5	1	1	1	0	7	3	3	2	1	18.8	142.6
5-Oct	3	0	1	1	1	7	5	A	C	C	C	C	C	C	2	5	4	2	5	1	2	18	2	2	--	17.7
6-Oct	4	3	9	22	19	30	A	38	35	27	22	9	2	6	3	2	1	1	3	2	4	2	2	1	10.8	38.1
7-Oct	1	2	3	3	3	A	5	7	3	7	7	4	5	4	4	3	3	4	7	2	3	2	4	3	3.9	7.2
8-Oct	5	3	2	3	A	3	5	4	2	19	3	5	2	3	1	1	2	1	1	2	1	1	1	1	3.1	19.0
9-Oct	1	1	1	A	1	2	2	36	2	2	2	2	4	2	3	2	3	6	5	1	1	1	0	1	3.6	36.4
10-Oct	1	2	A	2	4	5	10	7	4	4	9	7	2	88	2	2	1	11	2	1	2	1	0	3	7.3	87.5
11-Oct	0	A	10	21	26	25	29	34	55	32	15	11	4	2	1	2	1	1	0	0	1	1	1	1	11.9	54.6
12-Oct	A	0	0	0	10	14	23	9	7	3	2	2	7	2	3	1	1	0	1	1	1	2	3	A	4.1	22.7
13-Oct	1	1	1	0	0	1	3	7	6	8	2	1	1	2	0	1	0	2	10	31	1	1	A	0	3.4	31.0
14-Oct	0	1	1	4	6	16	26	29	47	68	26	3	1	2	1	0	0	0	0	0	0	A	0	0	10.0	68.0
15-Oct	0	0	0	0	0	53	64	81	3	1	1	1	1	0	1	0	0	0	3	6	A	2	1	0	9.5	80.5
16-Oct	0	0	0	0	7	15	95	202	154	82	9	14	2	1	2	1	1	1	2	A	1	0	1	19	26.5	201.6
17-Oct	0	0	0	1	13	16	48	127	284	178	23	11	6	1	2	3	2	1	A	0	1	0	0	0	31.2	283.5
18-Oct	1	0	0	0	0	0	0	6	1	2	1	1	1	1	1	1	1	A	3	13	22	15	20	21	4.8	22.4
19-Oct	26	27	25	13	25	46	43	25	85	50	67	23	10	4	3	3	A	1	1	1	1	0	2	3	21.1	85.4
20-Oct	2	9	5	5	2	0	0	0	4	1	1	1	1	1	1	A	1	1	0	0	1	0	0	0	1.6	9.0
21-Oct	0	0	0	1	0	6	35	34	54	28	11	6	1	1	A	2	0	2	2	22	8	6	1	12	10.2	54.3
22-Oct	12	7	42	27	28	42	47	121	133	131	106	46	41	A	25	22	5	24	77	96	69	21	21	14	50.3	132.7
23-Oct	16	25	18	13	36	85	83	112	142	108	59	23	A	9	13	7	2	13	20	29	66	64	27	12	42.7	141.8
24-Oct	6	12	27	34	40	52	77	75	92	102	63	A	8	6	12	9	17	16	42	54	90	120	128	83	50.6	128.3
25-Oct	44	29	34	26	34	62	102	118	99	69	A	27	14	8	7	12	1	1	0	0	0	0	0	0	29.8	118.4
26-Oct	0	0	1	1	3	47	52	61	56	A	27	22	56	52	40	35	26	13	7	17	11	0	0	0	22.9	60.8
27-Oct	0	0	1	0	0	0	0	0	A	2	1	2	1	1	1	1	1	5	2	1	11	3	32	52	5.1	51.8
28-Oct	55	41	3	2	2	1	1	A	1	1	2	4	5	4	3	3	4	4	10	4	2	5	1	3	7.0	54.6
29-Oct	4	7	18	33	42	58	A	61	59	56	40	34	5	3	2	1	1	1	1	4	23	31	14	1	21.7	61.4
30-Oct	4	1	1	0	1	A	3	24	3	5	3	2	2	2	2	1	1	1	1	0	1	0	1	0	2.6	24.4
31-Oct	0	0	0	1	A	9	37	66	57	38	62	17	16	17	14	11	14	2	1	2	1	1	0	1	16.0	66.5
6.3 6.2 6.9 7.3 11.4 21.8 31.0 50.2 51.3 40.0 21.1 10.4 7.4 7.8 5.6 4.5 3.3 3.9 7.2 10.1 11.0 10.1 8.8 7.9																								Diurnal Average		
54.6 40.9 41.6 34.5 41.9 84.6 101.9 201.6 283.5 178.4 105.6 45.7 55.6 87.5 40.3 34.7 26.4 24.1 77.3 96.0 89.9 120.3 128.3 82.8																								Diurnal Maximum		
C - Calibration A - Automated Daily Zero Span																										

Hourly Maximums

Nitrogen Oxide (NO) - ppb
Henry Pirker - October 2018



Hourly Averages

Oxides of Nitrogen (NO_x) - ppb


Henry Pirker - October 2018

Maximum Value: 213.5 ppb on Oct 17 09:00																				Maximum Daily Average: 57.5 ppb on Oct 22					Hours in Service: 744	
Minimum Value: 1 ppb on Oct 13 15:00																				Minimum Daily Average: 3.4 ppb on Oct 8					Hours of Data: 706	
Maximum Diurnal Average: 51.5 ppb at hour 9																				Minimum Diurnal Average: 8.5 ppb at hour 14					Hours of Missing Data: 38	
Monthly Average: 19.18 ppb																				Percentiles: P ₁ = 1.3 P ₁₀ = 2.8 Q ₁ = 4.4 Median = 8.6 Q ₃ = 23.5 P ₉₀ = 49.0 P ₉₉ = 132.7					Hours of Calibration: 38	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	4	4	3	5	7	16	21	26	17	10	14	A	3	3	6	4	10	9	7	4	5	4	2	2	8.1	25.9
2-Oct	2	3	3	2	3	7	11	16	14	5	A	2	1	2	2	2	3	4	4	6	8	6	5	3	5.1	16.5
3-Oct	3	3	3	4	7	8	10	16	20	A	12	10	5	5	4	2	2	4	18	16	10	8	7	9	8.1	20.1
4-Oct	8	17	21	19	28	36	59	124	A	63	20	7	4	2	2	2	3	4	8	15	18	16	13	11	21.8	124.4
5-Oct	14	11	8	5	7	11	10	A	C	C	C	C	C	C	3	4	4	5	8	9	10	11	11	10	--	14.2
6-Oct	13	18	18	24	21	30	A	34	44	38	32	15	4	4	7	8	7	7	6	5	5	4	3	3	15.2	44.0
7-Oct	3	3	3	3	3	A	4	5	4	5	4	5	6	5	5	5	5	6	5	4	4	4	4	4	4.3	5.6
8-Oct	4	4	4	3	A	4	4	4	3	5	4	4	3	3	3	3	3	3	3	3	3	2	2	3	3.4	5.0
9-Oct	4	4	3	A	5	6	6	8	5	6	5	4	4	4	4	4	6	6	6	4	5	5	5	5	4.9	7.6
10-Oct	5	6	A	7	7	9	12	14	12	11	18	13	6	24	5	6	6	8	9	10	12	10	8	12	10.0	23.8
11-Oct	11	A	20	26	34	37	37	40	52	30	22	19	7	6	6	6	6	8	7	6	8	8	15	9	18.3	52.4
12-Oct	A	8	8	7	10	24	23	8	8	5	4	4	6	4	3	2	3	2	4	7	8	10	16	A	7.9	24.0
13-Oct	8	7	4	2	3	8	8	10	12	11	3	2	2	1	1	1	2	7	22	27	9	9	A	7	7.2	26.6
14-Oct	4	5	10	16	19	26	34	39	53	64	29	7	5	6	4	3	3	3	2	3	2	A	3	3	14.9	64.3
15-Oct	3	5	3	3	8	28	66	27	5	4	4	4	2	1	1	2	2	3	8	16	A	15	15	6	10.1	66.4
16-Oct	4	9	8	7	14	31	74	142	149	76	22	21	5	5	5	5	7	9	14	A	11	5	5	6	27.6	149.0
17-Oct	6	5	7	15	28	38	63	133	214	175	37	21	9	8	15	21	23	19	A	16	18	13	5	5	38.8	213.5
18-Oct	5	3	2	1	3	6	5	6	8	5	3	2	2	1	2	2	5	A	26	28	40	30	33	39	11.2	39.5
19-Oct	43	39	31	26	27	47	43	41	72	65	69	32	17	8	7	5	A	8	9	9	9	10	18	18	28.2	72.4
20-Oct	14	17	15	18	8	3	3	4	3	3	2	3	2	2	2	A	4	5	5	6	14	14	10	10	7.3	17.6
21-Oct	6	8	4	4	7	19	45	46	62	25	15	12	2	4	A	5	3	14	23	30	25	16	14	17	17.6	61.6
22-Oct	24	22	25	33	29	53	54	105	142	132	102	65	61	A	48	37	19	31	73	104	74	32	28	28	57.5	142.1
23-Oct	27	35	26	24	40	73	84	107	141	106	62	33	A	20	27	15	13	31	40	41	55	57	36	24	48.6	141.3
24-Oct	21	22	33	35	43	58	78	78	106	80	74	A	17	15	25	25	34	42	55	72	71	120	133	67	56.7	132.7
25-Oct	50	37	46	32	38	65	81	109	102	82	A	42	25	15	17	15	8	9	7	6	5	5	4	3	34.9	109.5
26-Oct	3	2	5	9	17	37	63	69	67	A	43	41	46	55	60	53	37	29	29	37	31	7	5	3	32.4	69.4
27-Oct	2	1	1	2	2	3	3	3	A	5	4	3	2	2	2	3	5	20	13	15	15	21	39	56	9.6	55.5
28-Oct	67	37	15	14	9	7	5	A	5	5	6	15	12	11	13	13	15	18	23	13	11	15	9	13	15.3	66.9
29-Oct	17	21	22	36	40	46	A	53	57	54	44	20	10	5	3	5	6	6	17	25	40	47	28	11	26.7	57.2
30-Oct	18	15	6	8	7	A	11	10	14	15	8	5	4	4	4	4	5	6	9	10	13	9	10	8	8.7	17.5
31-Oct	11	8	7	10	A	25	52	67	50	56	37	36	30	22	20	21	21	9	9	10	7	5	5	4	22.7	67.2
13.4 12.6 12.2 13.3 16.3 26.1 33.4 46.4 51.5 40.7 25.1 16.0 10.4 8.5 10.3 9.5 9.0 11.1 15.6 18.6 18.2 17.2 16.3 13.3																									Diurnal Average	
66.9 39.1 46.0 36.2 43.2 73.4 83.7 142.3 213.5 174.5 102.2 65.1 60.8 54.7 59.5 52.7 36.8 42.0 73.3 103.5 73.6 119.5 132.7 67.3																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										

Hourly Averages

Oxides of Nitrogen (NO_x) - ppb

Henry Pirker - October 2018



Hourly Maximums

Oxides of Nitrogen (NO_x) - ppb


Henry Pirker - October 2018

Maximum Value: 362.0 ppb on Oct 17 09:00																					Maximum Daily Average: 75.9 ppb on Oct 24										Hours in Service: 744	
Minimum Value: 2 ppb on Oct 27 02:00																					Minimum Daily Average: 7.2 ppb on Oct 8										Hours of Data: 706	
Maximum Diurnal Average: 74.2 ppb at hour 8																					Minimum Diurnal Average: 13.7 ppb at hour 17										Hours of Missing Data: 38	
Monthly Average: 28.61 ppb																					Percentiles: P ₁ = 2.5 P ₁₀ = 5.0 Q ₁ = 7.5 Median = 14.0 Q ₃ = 35.4 P ₉₀ = 72.2 P ₉₉ = 164.9										Hours of Calibration: 38	
																					Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum						
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24								
1-Oct	6	5	6	9	10	27	30	76	30	16	18	A	8	6	28	7	15	14	14	8	10	5	4	6	15.5	75.7						
2-Oct	4	5	5	4	8	11	40	22	21	11	A	3	2	5	4	4	5	8	6	8	11	9	7	5	9.1	39.5						
3-Oct	4	5	5	6	11	16	13	23	51	A	20	21	8	7	6	3	4	5	24	24	13	12	9	14	13.3	51.1						
4-Oct	11	33	25	22	43	47	102	175	A	116	25	12	11	4	9	4	6	7	12	28	22	26	19	17	33.7	174.8						
5-Oct	19	14	15	8	9	22	14	A	C	C	C	C	C	C	5	13	8	7	12	12	13	35	14	14	--	35.0						
6-Oct	22	22	24	38	34	59	A	57	58	47	41	24	7	17	11	12	10	9	10	8	11	7	6	5	23.4	59.2						
7-Oct	5	6	7	7	6	A	10	14	7	11	9	8	10	8	8	7	8	9	13	7	7	7	9	9	8.3	13.7						
8-Oct	10	7	6	6	A	7	11	10	6	29	6	8	5	6	4	5	7	4	5	6	4	4	4	4	7.2	28.9						
9-Oct	6	5	5	A	7	9	9	50	8	9	8	6	6	6	7	6	10	20	17	6	6	7	6	9	9.9	49.5						
10-Oct	8	11	A	10	14	17	22	20	16	17	23	19	8	100	7	9	8	21	12	14	16	13	10	21	18.0	100.1						
11-Oct	13	A	28	38	45	41	45	49	72	49	26	23	14	9	9	10	8	11	9	11	12	13	19	12	24.7	72.4						
12-Oct	A	11	9	9	23	34	42	20	15	10	7	7	15	6	7	4	7	3	7	11	11	20	21	A	13.6	42.2						
13-Oct	12	10	9	3	5	14	16	19	21	20	6	3	2	4	2	3	3	20	34	60	12	12	A	8	13.0	59.8						
14-Oct	7	10	17	20	23	37	46	46	67	92	45	12	7	12	7	4	4	5	5	5	4	A	6	7	21.3	92.2						
15-Oct	6	7	7	7	15	87	91	116	9	6	7	7	5	2	3	3	4	7	27	31	A	22	25	12	22.0	116.0						
16-Oct	6	16	11	9	41	50	131	258	199	118	30	35	9	8	10	9	10	14	25	A	17	8	11	35	46.1	257.9						
17-Oct	8	11	8	23	43	55	98	180	362	236	62	37	26	14	25	29	31	24	A	21	25	22	9	7	59.0	362.0						
18-Oct	8	4	5	2	5	11	11	12	14	10	5	5	5	3	5	4	9	A	31	42	52	46	47	50	16.8	52.3						
19-Oct	50	49	44	28	41	68	64	49	112	74	95	41	25	13	11	12	A	11	15	13	15	11	23	21	38.5	111.9						
20-Oct	20	28	24	22	20	5	4	8	6	5	4	5	4	5	5	A	6	7	8	10	22	20	17	15	11.6	27.5						
21-Oct	12	13	7	8	12	34	67	64	83	86	25	15	5	6	A	11	7	25	27	56	38	33	19	32	29.8	86.3						
22-Oct	31	26	63	52	44	59	67	150	165	163	139	73	69	A	55	54	24	55	108	126	94	40	40	32	75.2	165.0						
23-Oct	34	45	35	27	53	105	107	136	176	139	81	41	A	23	34	25	20	46	53	59	92	89	48	31	65.1	175.5						
24-Oct	23	28	46	53	60	76	104	104	121	128	87	A	22	19	35	36	49	51	76	88	121	152	161	109	75.9	160.5						
25-Oct	64	46	52	42	49	81	127	161	127	96	A	52	30	21	25	37	14	17	10	9	10	10	5	5	47.3	161.3						
26-Oct	6	3	9	17	28	79	89	95	87	A	53	48	83	79	73	61	49	39	34	46	37	12	7	4	45.1	95.3						
27-Oct	4	2	5	2	3	4	4	5	A	10	6	5	5	4	4	4	9	34	26	21	35	28	60	82	15.8	81.8						
28-Oct	81	69	22	17	12	9	7	A	7	7	12	19	19	15	15	16	20	22	30	19	17	22	10	23	21.4	80.9						
29-Oct	23	26	38	51	58	73	A	83	76	72	58	53	14	9	6	8	9	13	22	34	55	65	42	15	39.2	82.9						
30-Oct	21	20	19	12	10	A	22	50	20	21	12	8	7	6	8	7	7	8	13	15	19	11	19	14	15.2	49.5						
31-Oct	17	9	10	14	A	40	75	101	82	69	61	41	37	40	38	34	37	16	12	14	10	8	6	5	33.8	101.0						
18.0 18.2 18.9 18.9 25.3 40.6 50.6 74.2 72.1 59.5 34.6 22.5 16.0 15.7 15.6 14.7 13.7 17.7 23.2 27.1 27.1 25.6 22.8 20.8																									Diurnal Average							
80.9 68.7 62.6 52.8 59.8 104.9 131.2 257.9 362.0 235.5 138.5 72.9 82.8 100.1 72.6 61.0 49.1 55.2 108.4 125.5 120.5 151.9 160.5 109.4																									Diurnal Maximum							
C - Calibration																									A - Automated Daily Zero Span							

Hourly Maximums


Oxides of Nitrogen (NO_x) - ppb

Henry Pirker - October 2018



Span Responses

Oxides of Nitrogen (NO_x)
Henry Pirker - October 2018



Hourly Averages


Ozone (O₃) - ppb

Henry Pirker - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 49.1 ppb on Oct 16 14:00 Maximum Daily Average: 28.2 ppb on Oct 18																				Hours in Service: 744 Hours of Data: 708 Hours of Missing Data: 36 Hours of Calibration: 36 Percent Operational Time: 100.0						
Minimum Value: 0 ppb on Oct 19 06:00 Maximum Diurnal Average: 28.2 ppb at hour 14 Monthly Average: 17.53 ppb										Minimum Daily Average: 4.3 ppb on Oct 22 Minimum Diurnal Average: 9.0 ppb at hour 8 Percentiles: P ₁ = 0.4 P ₁₀ = 1.1 Q ₁ = 7.2 Median = 17.2 Q ₃ = 26.8 P ₉₀ = 34.2 P ₉₉ = 41.5																
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	25	25	25	23	22	11	4	5	10	18	17	A	31	31	28	29	25	24	24	26	24	25	27	26	22.0	31.0
2-Oct	26	26	26	27	28	25	22	17	19	27	A	33	33	32	33	33	33	31	28	25	23	23	22	23	26.8	33.2
3-Oct	23	22	22	22	19	17	15	12	12	A	20	23	29	32	33	35	37	35	21	22	26	26	28	25	24.2	36.8
4-Oct	24	5	2	2	2	1	1	2	A	6	13	25	31	35	35	35	34	32	26	17	9	9	11	13	16.1	35.3
5-Oct	9	11	13	18	17	15	16	A	19	22	22	C	C	C	C	29	28	26	22	20	16	15	12	13	18.1	28.9
6-Oct	9	5	4	1	3	1	A	1	2	4	8	19	29	31	28	19	19	17	17	16	19	17	18	17	13.3	30.7
7-Oct	15	13	13	12	12	A	11	10	10	10	12	12	12	13	13	13	13	13	14	15	16	16	16	15	13.0	16.1
8-Oct	15	15	15	15	A	15	15	14	14	13	13	15	17	21	24	21	21	22	21	20	20	21	21	19	17.7	24.1
9-Oct	18	18	17	A	17	16	16	16	17	17	18	19	20	20	20	21	19	18	17	19	19	18	18	17	18.0	20.6
10-Oct	17	16	A	14	14	12	10	8	10	11	7	9	14	15	15	14	13	12	10	9	7	9	13	8	11.6	16.8
11-Oct	8	A	2	1	1	0	0	1	2	5	9	13	26	28	28	29	29	25	24	24	20	19	12	15	14.0	29.3
12-Oct	A	13	12	13	9	2	5	15	14	15	20	20	19	20	22	23	21	22	18	13	12	10	4	A	14.7	22.6
13-Oct	9	7	14	20	18	12	11	9	10	12	19	21	24	27	29	30	30	22	7	7	14	13	A	17	16.6	30.2
14-Oct	21	18	13	9	4	1	0	0	2	4	11	32	36	37	40	40	40	40	40	37	35	A	27	26	22.3	40.3
15-Oct	27	25	27	25	19	6	1	17	35	39	40	40	41	41	41	40	38	35	27	18	A	17	17	25	28.0	41.5
16-Oct	27	23	26	27	23	10	1	1	2	8	27	29	47	49	47	40	34	30	24	A	28	36	35	35	26.5	49.1
17-Oct	37	41	38	28	12	7	1	1	2	3	20	30	44	43	36	30	26	29	A	30	24	29	34	30	25.0	44.5
18-Oct	29	36	39	45	43	36	35	35	34	34	36	38	38	38	38	37	32	A	10	5	2	5	3	1	28.2	45.4
19-Oct	0	0	0	1	1	0	1	1	1	2	5	12	23	31	32	35	A	31	27	28	28	26	11	4	13.1	35.2
20-Oct	5	3	4	2	23	36	36	34	37	38	38	38	39	38	37	A	36	34	34	31	22	21	25	24	27.7	38.9
21-Oct	27	25	28	27	22	10	1	0	2	13	19	23	36	38	A	37	38	27	19	11	13	20	21	13	20.4	38.1
22-Oct	1	3	2	0	0	0	0	1	1	3	5	9	11	A	14	16	19	8	1	2	1	0	1	1	4.3	18.9
23-Oct	1	1	0	0	0	1	1	1	1	2	7	11	A	22	20	26	24	9	3	2	1	1	0	1	5.9	26.3
24-Oct	0	0	0	0	0	1	1	1	1	2	4	A	19	22	15	16	10	4	1	1	1	1	1	1	4.5	22.1
25-Oct	1	1	1	0	0	1	1	1	1	2	A	10	16	25	23	22	30	26	24	26	32	31	32	32	14.6	31.9
26-Oct	31	31	27	22	11	5	1	1	1	A	4	6	6	5	3	4	6	5	2	1	4	24	28	31	11.3	31.4
27-Oct	33	34	34	34	33	33	32	32	A	31	32	34	34	35	34	33	29	15	21	16	15	9	1	1	26.4	34.7
28-Oct	1	5	14	11	17	22	22	A	23	24	23	14	14	14	13	11	9	6	3	8	11	6	10	9	12.6	23.5
29-Oct	4	2	2	1	0	1	A	1	2	3	6	13	21	28	33	32	30	30	17	8	1	1	6	16	11.1	32.9
30-Oct	9	10	24	20	19	A	19	21	18	18	26	30	33	32	32	34	33	31	25	24	20	24	22	24	23.8	34.2
31-Oct	20	22	22	18	A	5	1	1	1	2	6	6	9	13	14	12	12	19	16	14	13	12	13	14	11.6	22.1
15.7 15.2 15.5 14.7 13.4 10.4 9.7 9.0 10.5 13.5 16.8 20.9 25.9 28.2 26.9 26.5 25.7 22.6 18.2 16.5 15.8 16.2 16.3 16.6																								Diurnal Average		
37.0 41.4 38.7 45.4 43.1 36.3 35.9 35.0 37.3 39.5 39.8 40.2 47.4 49.1 47.2 40.2 40.3 39.6 39.8 37.5 35.1 36.3 34.6 35.2																								Diurnal Maximum		
C - Calibration A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 82 ppb 24-hr na																										

Hourly Averages

Ozone (O₃) - ppb
Henry Pirker - October 2018



Hourly Maximums


Ozone (O₃) - ppb

Henry Pirker - October 2018

Maximum Value: 52.2 ppb on Oct 16 13:00																				Maximum Daily Average: 32.0 ppb on Oct 15					Hours in Service: 744	
Minimum Value: 1 ppb on Oct 19 06:00																				Minimum Daily Average: 6.0 ppb on Oct 24					Hours of Data: 708	
Maximum Diurnal Average: 30.8 ppb at hour 14																				Minimum Diurnal Average: 11.1 ppb at hour 8					Hours of Missing Data: 36	
Monthly Average: 20.46 ppb																				Percentiles: P ₁ = 0.7 P ₁₀ = 2.3 Q ₁ = 11.7 Median = 20.3 Q ₃ = 29.9 P ₉₀ = 36.4 P ₉₉ = 44.4					Hours of Calibration: 36	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	26	27	27	25	23	18	9	8	15	20	20	A	33	33	31	31	28	27	28	28	27	27	29	28	24.7	33.1
2-Oct	28	27	29	29	29	28	27	25	24	32	A	34	34	34	34	34	35	33	30	27	26	25	24	25	29.2	35.1
3-Oct	24	24	24	23	22	20	17	15	16	A	23	27	32	34	36	37	38	37	32	27	28	30	31	27	27.0	37.8
4-Oct	27	13	3	3	4	2	2	3	A	11	22	29	36	36	36	36	36	35	30	23	12	13	18	19	19.6	36.4
5-Oct	12	14	21	20	18	17	18	A	21	23	23	C	C	C	C	30	30	27	25	23	18	17	15	15	20.3	30.0
6-Oct	13	11	9	3	11	3	A	2	3	5	17	27	31	33	30	26	21	19	18	19	21	19	20	19	16.5	33.4
7-Oct	16	14	13	13	13	A	12	11	11	11	13	13	13	14	14	14	14	15	16	17	17	17	16	16	14.0	16.8
8-Oct	15	15	16	16	A	16	16	15	15	14	15	16	19	25	25	24	23	23	23	21	21	22	22	20	19.1	25.3
9-Oct	19	19	18	A	17	18	18	18	18	19	20	20	20	21	21	22	21	20	20	20	20	19	19	18	19.3	21.9
10-Oct	18	17	A	16	15	14	13	11	13	13	11	14	16	17	16	16	15	15	12	12	9	15	18	11	14.1	17.8
11-Oct	9	A	3	2	1	1	1	1	3	7	10	18	29	31	30	32	32	29	27	27	24	22	17	20	16.4	32.3
12-Oct	A	18	13	14	12	5	13	18	17	21	23	22	22	22	23	24	22	23	21	16	15	15	6	A	17.5	23.9
13-Oct	12	8	20	21	20	15	15	13	13	17	22	23	26	29	30	31	32	30	13	16	16	15	A	19	19.8	31.7
14-Oct	22	22	20	13	8	3	1	1	3	7	24	36	37	42	42	41	41	40	41	39	37	A	29	27	25.2	42.1
15-Oct	30	30	29	27	24	13	2	31	38	41	42	43	42	42	42	42	39	38	32	29	A	22	27	29	32.0	42.9
16-Oct	29	28	29	30	31	19	4	2	3	23	31	40	52	51	50	44	36	34	32	A	39	38	39	37	31.4	52.2
17-Oct	42	44	42	37	20	18	3	2	3	9	23	36	47	50	40	36	31	36	A	35	28	40	38	32	30.2	49.5
18-Oct	32	38	42	47	45	40	39	38	36	36	39	39	39	39	39	39	36	A	15	10	7	9	12	1	31.2	47.0
19-Oct	1	1	3	2	3	1	3	2	2	3	9	19	27	34	34	37	A	34	30	31	29	28	26	7	15.9	37.0
20-Oct	8	9	7	4	36	37	37	36	38	39	39	39	40	40	39	A	37	36	36	34	28	26	32	29	30.7	40.0
21-Oct	30	29	30	31	28	19	7	1	7	16	22	32	38	39	A	39	41	36	24	18	18	28	26	24	25.4	40.7
22-Oct	3	6	6	1	1	1	1	1	2	4	7	9	12	A	16	24	24	15	2	20	1	1	1	2	7.0	24.5
23-Oct	1	1	1	1	1	3	2	2	2	4	13	13	A	24	24	30	29	18	5	4	3	3	1	2	8.1	29.9
24-Oct	1	1	1	1	1	1	1	1	2	4	6	A	23	25	19	19	14	8	4	1	2	2	1	1	6.0	24.9
25-Oct	1	1	1	1	1	1	1	2	2	5	A	13	20	28	28	33	36	30	27	35	35	34	33	33	17.4	36.2
26-Oct	32	33	31	28	17	12	2	1	2	A	7	8	8	9	8	6	8	8	5	2	19	27	31	32	14.6	32.7
27-Oct	34	35	35	35	34	34	33	33	A	32	34	36	35	36	35	35	33	27	26	26	21	15	6	2	29.2	35.6
28-Oct	2	12	17	14	21	25	24	A	23	26	25	17	18	16	15	12	12	9	5	12	13	9	11	12	15.2	25.8
29-Oct	8	5	6	1	1	1	A	4	3	5	8	19	28	32	35	35	33	33	29	14	3	4	16	19	14.8	35.2
30-Oct	14	17	29	27	21	A	22	23	22	23	28	32	34	34	35	36	36	34	29	28	24	27	27	26	27.2	35.8
31-Oct	24	24	25	21	A	10	4	2	3	3	10	8	14	26	21	20	19	21	19	17	14	14	15	17	15.3	25.7
17.8 18.0 18.3 16.8 16.5 13.5 12.0 11.1 12.4 16.3 20.2 24.4 28.5 30.8 29.3 29.6 28.5 26.3 21.9 21.0 19.3 19.4 20.2 18.9																									Diurnal Average	
42.5 43.7 42.4 47.0 45.4 39.9 38.8 37.5 38.5 41.1 41.8 42.9 52.2 51.5 49.9 44.4 41.4 40.5 41.1 39.3 38.6 40.2 38.7 37.5																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										


Hourly Maximums

Ozone (O₃) - ppb
Henry Pirker - October 2018



Pollutant Rose

Ozone (O_3) - ppb
Henry Pirker - October 2018



Eight Hour Running Averages


Ozone (O₃) - ppb

Henry Pirker - October 2018

Maximum Value: 40.1 ppb on Oct 15 17:00																								Hours in Service: 744	
Minimum Value: 0.5 ppb on Oct 24 06:00																								Hours of Data: 736	
																								Hours of Missing Data: 8	
Percentiles: P ₁ = 0.5 P ₁₀ = 3.5 Q ₁ = 10.7 Median = 17.0 Q ₃ = 25.2 P ₉₀ = 31.5 P ₉₉ = 38.1																								Hours of Calibration: 8	
																								Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																							Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
1-Oct	26	25	25	25	25	23	20	17	16	15	14	12	14	17	20	23	26	26	27	27	27	26	26	25	27.4
2-Oct	26	26	26	26	27	26	26	25	24	24	24	24	25	26	28	30	32	33	32	31	30	29	27	26	32.6
3-Oct	25	24	23	23	22	21	21	19	18	17	17	17	18	20	23	26	30	31	31	31	30	29	29	27	30.8
4-Oct	26	22	20	17	14	11	8	5	2	2	4	7	11	16	21	26	27	30	32	31	28	24	21	19	31.6
5-Oct	16	13	12	12	13	13	14	14	16	17	18	18	N	N	N	N	N	N	N	N	24	22	21	19	23.5
6-Oct	17	14	12	10	8	6	5	4	3	2	3	6	9	13	15	17	20	21	22	22	21	19	18	18	22.4
7-Oct	17	17	16	16	15	14	13	12	12	11	11	11	11	11	12	12	12	13	13	13	14	14	15	15	17.1
8-Oct	15	15	15	15	15	15	15	15	15	15	14	14	15	15	16	17	18	19	20	21	21	21	21	21	21.3
9-Oct	20	20	19	19	19	18	17	17	17	17	17	17	18	18	18	19	19	19	19	19	19	19	18	18	20.3
10-Oct	18	18	18	17	16	15	14	13	12	11	11	10	10	11	11	12	12	12	13	13	12	11	11	10	17.8
11-Oct	9	9	8	7	6	5	3	2	1	2	2	4	7	10	14	17	21	23	25	27	26	25	23	21	26.7
12-Oct	20	18	16	15	13	11	10	10	11	11	12	13	14	16	18	19	20	21	21	20	19	18	15	14	20.9
13-Oct	13	11	10	11	12	12	13	12	12	13	14	14	15	17	19	22	24	25	24	22	21	19	18	16	25.4
14-Oct	14	14	14	15	13	12	10	8	6	4	4	7	11	15	20	25	30	35	38	39	39	39	37	35	38.9
15-Oct	33	31	29	27	25	23	19	18	19	21	23	25	27	32	37	40	40	40	38	35	34	31	28	25	40.1
16-Oct	24	22	22	23	23	22	20	17	14	12	12	13	16	21	26	31	35	38	37	39	36	34	32	32	38.7
17-Oct	32	34	36	35	33	29	25	21	16	11	9	10	14	18	22	26	29	32	34	34	31	29	29	29	35.7
18-Oct	29	30	31	33	36	37	37	37	38	38	37	36	36	36	37	36	37	37	33	28	23	18	13	8	37.9
19-Oct	4	3	2	1	1	1	1	1	1	1	1	3	6	9	13	18	20	24	27	30	30	30	27	22	30.3
20-Oct	20	17	14	10	10	11	14	18	22	26	31	35	37	37	38	38	38	37	37	36	33	31	29	28	38.1
21-Oct	27	26	25	25	25	23	20	17	14	13	12	11	13	17	19	24	29	31	31	29	26	24	23	20	31.2
22-Oct	16	13	11	9	8	5	3	1	1	1	1	2	4	4	6	8	11	12	11	10	9	8	6	4	15.7
23-Oct	2	1	1	1	0	1	1	1	1	1	2	3	3	6	9	13	16	17	17	15	14	11	8	5	17.1
24-Oct	2	1	1	1	1	0	1	0	1	1	1	1	4	7	9	11	13	13	12	11	9	6	4	2	12.8
25-Oct	1	1	1	1	1	1	1	1	1	1	1	2	4	8	11	14	18	22	22	24	26	27	28	29	29.1
26-Oct	29	30	30	30	27	24	20	16	13	10	7	4	3	3	4	4	5	5	5	4	4	6	9	13	30.2
27-Oct	16	20	24	28	31	32	33	33	33	33	33	33	33	33	33	33	33	31	29	27	25	22	17	13	33.4
28-Oct	10	9	8	7	7	9	12	13	16	19	20	21	20	19	18	17	15	13	10	10	9	8	8	8	20.6
29-Oct	7	7	6	6	4	4	3	1	1	1	2	4	7	11	13	17	21	24	25	25	22	19	16	14	25.4
30-Oct	11	8	9	11	13	15	17	17	19	20	20	22	23	25	26	28	30	31	31	31	29	28	27	25	31.4
31-Oct	24	23	22	21	22	19	16	13	10	7	5	3	4	5	7	8	9	11	13	14	14	14	14	14	23.7
Diurnal Maximums																									
N - Not Valid																									

Span Responses

Ozone (O₃)
Henry Pirker - October 2018



Hourly Averages


Carbon Monoxide (CO) - ppm

Henry Pirker - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 1.43 ppm on Oct 17 09:00 Maximum Daily Average: 0.39 ppm on Oct 24																								Hours in Service: 744 Hours of Data: 708 Hours of Missing Data: 36 Hours of Calibration: 36 Percent Operational Time: 100.0		
Minimum Value: 0.1 ppm on Oct 9 18:00 Minimum Daily Average: 0.14 ppm on Oct 2 Maximum Diurnal Average: 0.36 ppm at hour 9 Minimum Diurnal Average: 0.16 ppm at hour 14 Monthly Average: 0.214 ppm Percentiles: P ₁ = 0.11 P ₁₀ = 0.13 Q ₁ = 0.14 Median = 0.17 Q ₃ = 0.23 P ₉₀ = 0.35 P ₉₉ = 0.81																										
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.1	0.2	A	0.1	0.1	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.14	0.22
2-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.1	A	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.1	0.2	0.2	0.1	0.1	0.14	0.19
3-Oct	0.1	0.1	0.1	0.1	0.2	0.1	0.1	0.2	0.2	A	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.3	0.2	0.2	0.2	0.2	0.1	0.15	0.26
4-Oct	0.1	0.1	0.2	0.1	0.2	0.2	0.3	0.4	A	0.3	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.19	0.41
5-Oct	0.2	0.2	0.1	0.1	0.2	0.2	0.2	A	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.16	0.19
6-Oct	0.2	0.2	0.2	0.2	0.2	0.2	A	0.3	0.3	0.3	0.3	0.2	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.19	0.29
7-Oct	0.1	0.1	0.1	0.1	0.1	A	0.1	0.1	0.1	0.2	0.1	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.14	0.17
8-Oct	0.1	0.1	0.1	0.1	A	0.1	0.1	0.1	0.2	0.2	0.1	0.2	0.2	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.14	0.16
9-Oct	0.1	0.1	0.1	A	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	C	C	C	C	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.16	0.25
10-Oct	0.1	0.1	A	0.1	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.2	0.1	0.2	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.17	0.26
11-Oct	0.2	A	0.3	0.3	0.4	0.3	0.3	0.3	0.4	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.2	0.1	0.2	0.2	0.22	0.35
12-Oct	A	0.2	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	A	0.17	0.30
13-Oct	0.1	0.2	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.3	0.3	0.2	0.2	A	0.1	0.16	0.32
14-Oct	0.1	0.1	0.2	0.2	0.2	0.3	0.4	0.3	0.4	0.4	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	A	0.1	0.1	0.21	0.42
15-Oct	0.1	0.1	0.1	0.1	0.2	0.3	0.6	0.3	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	A	0.2	0.2	0.2	0.18	0.59
16-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.6	1.0	1.2	0.6	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.2	A	0.2	0.1	0.1	0.2	0.29	1.18
17-Oct	0.2	0.2	0.2	0.2	0.4	0.4	0.4	1.0	1.4	1.1	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	A	0.2	0.2	0.2	0.1	0.1	0.35	1.43
18-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	A	0.2	0.2	0.3	0.3	0.3	0.3	0.18	0.34
19-Oct	0.3	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.4	0.3	0.4	0.2	0.2	0.1	0.2	0.2	A	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.23	0.39
20-Oct	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	A	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.15	0.21
21-Oct	0.2	0.1	0.1	0.1	0.2	0.2	0.3	0.3	0.3	0.2	0.2	0.2	0.1	0.1	A	0.1	0.1	0.2	0.5	0.4	0.3	0.2	0.2	0.2	0.22	0.46
22-Oct	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.5	0.6	0.7	0.5	0.4	0.4	A	0.4	0.3	0.2	0.3	0.8	0.8	0.3	0.2	0.2	0.2	0.36	0.81
23-Oct	0.2	0.3	0.2	0.2	0.3	0.6	0.4	0.5	0.6	0.5	0.3	0.2	A	0.2	0.2	0.2	0.2	0.3	0.4	0.4	0.4	0.3	0.2	0.2	0.31	0.59
24-Oct	0.2	0.2	0.3	0.2	0.3	0.4	0.5	0.4	0.5	0.4	0.4	A	0.2	0.2	0.2	0.2	0.3	0.3	0.4	0.5	0.6	1.0	1.0	0.4	0.39	0.98
25-Oct	0.3	0.4	0.4	0.3	0.3	0.5	0.6	0.5	0.6	0.5	A	0.3	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.29	0.63
26-Oct	0.1	0.1	0.1	0.2	0.2	0.3	0.7	0.6	0.5	A	0.3	0.4	0.4	0.3	0.4	0.3	0.3	0.3	0.2	0.3	0.3	0.1	0.1	0.1	0.30	0.70
27-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	A	0.2	0.2	0.1	0.1	0.1	0.2	0.1	0.2	0.2	0.2	0.2	0.3	0.5	0.6	0.19	0.60
28-Oct	0.6	0.4	0.2	0.2	0.2	0.2	0.1	A	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.21	0.65
29-Oct	0.2	0.3	0.3	0.3	0.3	0.3	A	0.3	0.4	0.4	0.4	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.4	0.4	0.3	0.2	0.27	0.40
30-Oct	0.2	0.2	0.2	0.1	0.2	A	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.16	0.24
31-Oct	0.2	0.2	0.2	0.2	A	0.3	0.3	0.5	0.4	0.4	0.2	0.2	0.2	0.1	0.2	0.2	0.2	0.1	0.1	0.1	0.2	0.1	0.2	0.2	0.21	0.51
0.18 0.18 0.17 0.17 0.20 0.25 0.29 0.34 0.36 0.30 0.23 0.19 0.18 0.16 0.17 0.17 0.16 0.18 0.22 0.23 0.21 0.21 0.21 0.19																								Diurnal Average		
0.65 0.35 0.37 0.31 0.39 0.56 0.70 1.05 1.43 1.06 0.53 0.41 0.42 0.32 0.37 0.33 0.28 0.34 0.81 0.78 0.60 0.98 0.96 0.60																								Diurnal Maximum		
C - Calibration A - Automated Daily Zero Span Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 13 ppm 24-hr na																										

Hourly Averages

Carbon Monoxide (CO) - ppm
Henry Pirker - October 2018



Hourly Maximums

Carbon Monoxide (CO) - ppm


Henry Pirker - October 2018

Maximum Value: 2.80 ppm on Oct 23 06:00																					Maximum Daily Average: 0.61 ppm on Oct 24				Hours in Service: 744	
Minimum Value: 0.1 ppm on Oct 9 18:00																					Minimum Daily Average: 0.17 ppm on Oct 2				Hours of Data: 708	
Maximum Diurnal Average: 0.60 ppm at hour 8																					Minimum Diurnal Average: 0.19 ppm at hour 14				Hours of Missing Data: 36	
Monthly Average: 0.303 ppm																					Percentiles: P ₁ = 0.12 P ₁₀ = 0.15 Q ₁ = 0.17 Median = 0.21 Q ₃ = 0.32 P ₉₀ = 0.52 P ₉₉ = 1.69				Hours of Calibration: 36	
																					Percent Operational Time: 100.0					
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.3	0.3	0.2	0.2	A	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.2	0.1	0.1	0.18	0.28
2-Oct	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.4	0.2	0.2	A	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.17	0.44
3-Oct	0.1	0.2	0.2	0.1	0.2	0.1	0.2	0.2	0.2	A	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.4	0.4	0.2	0.2	0.2	0.2	0.19	0.40
4-Oct	0.2	0.2	0.2	0.3	0.3	0.4	0.6	0.8	A	0.4	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.2	0.2	0.3	0.2	0.3	0.3	0.2	0.27	0.77
5-Oct	0.3	0.2	0.2	0.2	0.2	0.2	0.3	A	0.2	0.2	0.1	0.2	0.4	0.1	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.21	0.37
6-Oct	0.2	0.2	0.4	0.2	0.2	0.4	A	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.25	0.42
7-Oct	0.2	0.2	0.2	0.2	0.2	A	0.2	0.2	0.2	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.18	0.30
8-Oct	0.1	0.1	0.1	0.1	A	0.1	0.2	0.3	0.4	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.19	0.36
9-Oct	0.2	0.2	0.2	A	0.3	0.3	0.4	0.3	0.3	0.3	0.3	0.3	0.3	C	C	C	C	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.21	0.38
10-Oct	0.1	0.1	A	0.2	0.3	0.3	0.3	0.5	0.3	0.2	0.2	0.2	0.2	0.4	0.2	0.2	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.23	0.54
11-Oct	0.2	A	0.3	0.5	0.5	0.3	0.4	0.4	0.4	0.4	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.27	0.50
12-Oct	A	0.2	0.2	0.2	0.6	0.4	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	A	0.22	0.57
13-Oct	0.2	0.2	0.2	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.2	0.2	0.1	0.1	0.2	0.2	0.7	0.6	0.2	0.2	A	0.2	0.22	0.73
14-Oct	0.2	0.2	0.2	0.3	0.5	0.4	0.8	0.3	0.7	0.5	0.4	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	A	0.1	0.1	0.29	0.80
15-Oct	0.2	0.2	0.3	0.1	0.3	0.5	1.1	1.1	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.3	A	0.2	0.3	0.3	0.29	1.09
16-Oct	0.2	0.2	0.2	0.3	0.3	0.4	1.2	1.9	1.8	0.8	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	A	0.3	0.2	0.2	0.2	0.43	1.88
17-Oct	0.2	0.2	0.2	0.3	0.8	0.8	1.2	1.7	1.9	1.4	0.4	0.3	0.2	0.2	0.2	0.2	0.3	0.2	A	0.2	0.2	0.2	0.2	0.2	0.51	1.88
18-Oct	0.2	0.2	0.2	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.2	0.2	0.1	A	0.3	0.3	0.5	0.4	0.5	0.5	0.23	0.52
19-Oct	0.3	0.4	0.3	0.2	0.3	0.4	0.3	0.4	1.0	0.5	0.5	0.3	0.2	0.2	0.2	0.2	A	0.2	0.3	0.2	0.2	0.2	0.3	0.2	0.32	1.01
20-Oct	0.2	0.3	0.2	0.3	0.3	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	A	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.19	0.29
21-Oct	0.2	0.3	0.2	0.2	0.4	0.3	0.4	0.5	0.5	0.2	0.2	0.2	0.1	0.1	A	0.2	0.1	0.4	0.6	0.5	0.5	0.4	0.3	0.3	0.31	0.64
22-Oct	0.2	0.2	0.3	0.3	0.3	0.4	0.5	0.8	0.8	0.9	0.8	0.5	0.5	A	0.4	0.4	0.2	0.5	2.4	1.6	0.4	0.3	0.3	0.3	0.57	2.40
23-Oct	0.4	0.5	0.2	0.2	0.6	2.8	1.0	2.0	0.8	0.6	0.4	0.3	A	0.2	0.3	0.3	0.2	0.5	0.5	0.5	0.5	0.5	0.3	0.2	0.60	2.80
24-Oct	0.2	0.3	0.3	0.3	1.5	0.5	0.7	0.7	0.7	0.5	0.4	A	0.2	0.2	0.3	0.3	0.3	0.3	0.5	0.7	1.1	1.8	1.3	0.7	0.61	1.84
25-Oct	0.4	0.4	0.5	0.4	0.8	0.8	0.7	0.8	0.8	0.7	A	0.4	0.3	0.2	0.4	0.3	0.2	0.2	0.2	0.2	0.1	0.2	0.1	0.1	0.39	0.79
26-Oct	0.1	0.2	0.2	0.2	0.2	0.7	1.5	0.8	0.6	A	0.4	0.5	0.5	0.4	0.4	0.4	0.3	0.8	0.3	0.4	0.3	0.2	0.2	0.1	0.42	1.46
27-Oct	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.1	A	0.2	0.2	0.2	0.2	0.1	0.2	0.2	0.2	0.3	0.3	0.3	0.4	0.4	0.8	0.8	0.25	0.80
28-Oct	0.8	0.6	0.3	0.2	0.2	0.2	0.2	A	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.26	0.83
29-Oct	0.3	0.4	0.4	0.4	0.6	0.4	A	0.4	0.6	0.5	0.9	0.4	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.4	0.5	0.5	0.3	0.2	0.38	0.90
30-Oct	0.3	0.3	0.2	0.2	0.2	A	0.2	0.2	0.2	0.3	0.2	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.2	0.2	0.2	0.2	0.21	0.33
31-Oct	0.2	0.2	0.2	0.2	A	0.3	0.4	1.0	1.0	0.5	0.3	0.3	0.3	0.2	0.3	0.3	0.2	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.31	1.03
0.23 0.24 0.23 0.23 0.37 0.43 0.49 0.60 0.53 0.39 0.30 0.24 0.22 0.19 0.22 0.20 0.20 0.25 0.35 0.31 0.28 0.28 0.28 0.24																									Diurnal Average	
0.83 0.55 0.49 0.47 1.52 2.80 1.46 2.01 1.88 1.36 0.90 0.49 0.48 0.45 0.43 0.38 0.33 0.81 2.40 1.56 1.05 1.84 1.32 0.80																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										

Hourly Maximums

Carbon Monoxide (CO) - ppm


Henry Pirker - October 2018



Pollutant Rose

Carbon Monoxide (CO) - ppm

Henry Pirker - October 2018



Eight Hour Running Averages


Carbon Monoxide (CO) - ppm

Henry Pirker - October 2018

Number of Exceedences (AAAQO): 8-hr: 0 Maximum Value: 0.66 ppm on Oct 17 11:00																							Hours in Service: 744			
Minimum Value: 0.08 ppm on Oct 9 23:00																							Hours of Data: 737			
Percentiles: P ₁ = 0.12 P ₁₀ = 0.14 Q ₁ = 0.15 Median = 0.17 Q ₃ = 0.25 P ₉₀ = 0.36 P ₉₉ = 0.56																							Hours of Missing Data: 7			
																							Hours of Calibration: 7			
																							Percent Operational Time: 100.0			
Day	Hourly Period Ending At (MST)																								Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.16
2-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.15
3-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.17
4-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.2	0.2	0.2	0.26
5-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.18
6-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.25
7-Oct	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.16
8-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.15
9-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	N	N	N	N	N	N	N	N	0.1	0.1	0.23
10-Oct	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.19
11-Oct	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.32
12-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.20
13-Oct	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.19
14-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.32
15-Oct	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.24
16-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.5	0.5	0.5	0.5	0.5	0.5	0.5	0.4	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.53
17-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.4	0.5	0.6	0.7	0.7	0.6	0.6	0.6	0.5	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.66
18-Oct	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.3	0.26
19-Oct	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.30
20-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.19
21-Oct	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.28
22-Oct	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.4	0.4	0.4	0.5	0.5	0.5	0.5	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.51
23-Oct	0.4	0.4	0.3	0.2	0.2	0.3	0.3	0.3	0.4	0.4	0.4	0.4	0.4	0.4	0.4	0.3	0.3	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.44
24-Oct	0.3	0.3	0.3	0.2	0.2	0.2	0.3	0.3	0.4	0.4	0.4	0.4	0.4	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.4	0.5	0.6	0.55
25-Oct	0.6	0.6	0.6	0.5	0.5	0.4	0.4	0.4	0.4	0.5	0.5	0.5	0.5	0.4	0.4	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.57	
26-Oct	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.3	0.3	0.4	0.4	0.4	0.5	0.5	0.4	0.4	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.46
27-Oct	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.30
28-Oct	0.4	0.4	0.4	0.4	0.4	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.38
29-Oct	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.35
30-Oct	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.27
31-Oct	0.1	0.1	0.1	0.2	0.2	0.2	0.2	0.2	0.3	0.3	0.3	0.3	0.3	0.3	0.3	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.1	0.2	0.2	0.32
0.56 0.57 0.56 0.53 0.50 0.43 0.38 0.40 0.54 0.65 0.66 0.66 0.63 0.60 0.57 0.48 0.41 0.36 0.40 0.45 0.44 0.44 0.53 0.55																										
Diurnal Maximums																										
N - Not Valid																										
Alberta Ambient Air Quality Objectives (AAAQO): 8-hr 5 ppm																										

Span Responses

Carbon Monoxide (CO)
Henry Pirker - October 2018



Hourly Averages


Total Hydrocarbons (THC) - ppm

Henry Pirker - October 2018

Maximum Value: 3.05 ppm on Oct 23 05:00																					Maximum Daily Average: 2.44 ppm on Oct 23				Hours in Service: 744	
Minimum Value: 1.8 ppm on Oct 15 15:00																					Minimum Daily Average: 1.89 ppm on Oct 15				Hours of Data: 708	
Maximum Diurnal Average: 2.17 ppm at hour 7																					Minimum Diurnal Average: 1.92 ppm at hour 17				Hours of Missing Data: 36	
Monthly Average: 2.038 ppm																					Percentiles: P ₁ = 1.85 P ₁₀ = 1.88 Q ₁ = 1.92 Median = 1.96 Q ₃ = 2.08 P ₉₀ = 2.31 P ₉₉ = 2.78				Hours of Calibration: 36	
																					Percent Operational Time: 100.0					
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.0	2.0	2.0	2.0	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.96	2.12
2-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.94	1.96
3-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.9	2.0	1.9	1.9	2.0	2.0	1.95	2.01
4-Oct	2.0	2.1	2.1	2.1	2.4	2.5	2.5	2.3	A	2.1	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.1	2.2	2.1	2.08	2.49
5-Oct	2.1	2.1	2.0	2.0	2.0	2.0	2.0	A	1.9	1.9	1.9	1.9	1.9	C	C	C	C	1.9	1.9	1.9	2.0	2.0	2.0	2.0	1.97	2.09
6-Oct	2.0	2.0	2.2	2.2	2.1	2.1	A	2.5	2.5	2.1	2.1	2.1	1.9	1.9	1.9	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.0	2.05	2.48
7-Oct	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.98	1.99
8-Oct	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.97	1.99
9-Oct	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.99	2.00
10-Oct	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.1	2.1	2.1	2.1	2.03	2.13
11-Oct	2.3	A	2.2	2.2	2.3	2.3	2.7	2.4	2.3	2.3	2.2	2.2	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.12	2.74
12-Oct	A	2.3	2.3	2.1	2.4	2.3	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	A	2.01	2.39
13-Oct	2.0	2.0	2.0	1.9	1.9	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.9	1.9	A	1.9	1.95	2.02
14-Oct	1.9	1.9	2.0	2.1	2.1	2.1	2.2	2.2	2.2	2.2	2.1	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.97	2.20
15-Oct	1.9	1.9	1.9	1.9	1.9	2.0	2.2	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.9	1.8	1.9	1.9	A	1.9	1.9	1.9	1.89	2.20
16-Oct	1.9	1.9	1.9	1.9	2.0	2.0	2.1	2.2	2.3	2.6	2.1	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.98	2.60
17-Oct	1.9	1.9	2.0	2.0	2.1	2.1	2.2	2.2	2.4	2.4	2.0	2.0	2.0	1.9	1.9	1.9	1.9	1.9	A	1.9	2.0	1.9	1.9	1.9	2.02	2.43
18-Oct	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	2.1	2.2	2.1	2.2	2.2	1.95	2.18
19-Oct	2.2	2.2	2.3	2.6	2.7	2.7	2.3	2.1	2.2	2.2	2.3	2.1	2.0	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	2.0	2.3	2.2	2.16	2.75
20-Oct	2.2	2.4	2.3	2.4	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.97	2.39
21-Oct	1.9	1.9	1.9	1.9	2.0	2.1	2.1	2.1	2.3	2.1	2.0	2.0	1.9	1.9	A	1.9	1.9	1.9	2.0	2.0	2.0	2.1	2.1	2.4	2.02	2.37
22-Oct	2.4	2.3	2.3	2.4	2.9	2.3	2.5	2.6	2.5	2.5	2.5	2.3	2.3	A	2.1	2.1	2.0	2.0	2.1	2.1	2.2	2.2	2.4	2.7	2.34	2.87
23-Oct	2.7	2.7	2.8	3.0	3.0	3.0	2.8	2.7	2.7	2.6	2.4	2.1	A	2.0	2.0	1.9	1.9	2.0	2.1	2.1	2.1	2.4	2.4	2.6	2.44	3.05
24-Oct	2.6	2.8	2.6	2.4	2.4	2.5	2.5	2.4	2.4	2.6	2.4	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.1	2.2	2.0	2.22	2.83
25-Oct	2.3	2.6	2.7	2.7	2.8	2.7	2.7	2.5	2.4	2.4	A	2.2	2.1	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.22	2.80
26-Oct	1.9	1.9	1.9	1.9	2.1	2.2	2.1	2.1	2.1	A	2.1	2.1	2.1	2.1	2.2	2.3	2.1	2.1	2.1	2.3	2.3	1.9	1.9	1.9	2.08	2.33
27-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.1	2.2	1.92	2.19
28-Oct	2.2	2.1	2.0	2.0	2.0	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	1.99	2.20
29-Oct	2.1	2.1	2.2	2.2	2.2	2.2	A	2.3	2.6	2.4	2.1	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.1	2.0	2.0	2.07	2.59
30-Oct	2.0	2.0	1.9	2.0	2.0	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.92	2.04
31-Oct	2.0	2.0	2.0	2.0	A	2.2	2.5	2.3	2.1	2.1	2.1	2.1	2.0	2.0	1.9	2.0	1.9	1.9	1.9	1.9	2.1	2.0	1.9	1.9	2.04	2.47
2.07 2.09 2.10 2.11 2.17 2.16 2.17 2.14 2.14 2.13 2.05 2.00 1.96 1.93 1.94 1.94 1.92 1.93 1.94 1.97 1.99 2.00 2.02 2.05																									Diurnal Average	
2.72 2.83 2.78 3.03 3.05 3.02 2.82 2.73 2.69 2.60 2.49 2.32 2.30 2.15 2.19 2.33 2.06 2.08 2.10 2.33 2.28 2.36 2.44 2.66																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										

Hourly Averages

Total Hydrocarbons (THC) - ppm
Henry Pirker - October 2018



Hourly Maximums


Total Hydrocarbons (THC) - ppm

Henry Pirker - October 2018

Maximum Value: 3.84 ppm on Oct 19 23:00																					Maximum Daily Average: 2.70 ppm on Oct 23					Hours in Service: 744	
Minimum Value: 1.9 ppm on Oct 15 16:00																					Minimum Daily Average: 2.00 ppm on Oct 15					Hours of Data: 708	
Maximum Diurnal Average: 2.42 ppm at hour 5																					Minimum Diurnal Average: 2.01 ppm at hour 17					Hours of Missing Data: 36	
Monthly Average: 2.203 ppm																					Percentiles: P ₁ = 1.88 P ₁₀ = 1.93 Q ₁ = 2.00 Median = 2.08 Q ₃ = 2.30 P ₉₀ = 2.65 P ₉₉ = 3.32					Hours of Calibration: 36	
																					Percent Operational Time: 100.0						
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	2.0	2.0	2.0	2.0	2.0	2.2	2.2	2.2	2.3	2.1	2.2	A	2.1	2.0	1.9	1.9	2.0	2.0	2.0	2.2	2.0	2.0	2.0	2.0	2.06	2.34	
2-Oct	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	A	1.9	2.1	2.0	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.2	2.0	2.03	2.19	
3-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.4	A	2.0	2.0	2.0	2.0	2.0	2.0	1.9	2.1	2.1	2.1	2.1	2.0	2.0	2.1	2.04	2.37	
4-Oct	2.3	2.2	2.2	2.5	2.7	3.0	2.8	2.7	A	2.5	2.1	2.0	2.1	2.0	2.0	1.9	1.9	1.9	1.9	2.1	2.5	2.6	3.0	2.5	2.32	3.04	
5-Oct	2.3	2.4	2.3	2.2	2.0	2.0	2.2	A	2.0	2.1	2.0	2.0	2.0	C	C	C	C	2.0	2.0	2.0	2.0	2.0	2.3	2.0	2.09	2.42	
6-Oct	2.0	2.2	2.4	2.4	2.5	2.2	A	2.8	3.3	2.2	2.2	2.3	2.0	2.0	2.0	2.2	2.0	2.0	2.0	2.0	2.1	2.0	2.2	2.1	2.22	3.25	
7-Oct	2.1	2.1	2.0	2.0	2.0	A	2.2	2.1	2.1	2.1	2.1	2.0	2.0	2.1	2.1	2.2	2.0	2.1	2.1	2.2	2.1	2.1	2.1	2.2	2.09	2.22	
8-Oct	2.0	2.1	2.0	2.1	A	2.1	2.1	2.0	2.1	2.1	2.0	2.1	2.0	2.0	2.0	2.1	2.0	2.1	2.0	2.0	2.1	2.1	2.0	2.0	2.05	2.14	
9-Oct	2.1	2.1	2.1	A	2.2	2.0	2.0	2.2	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.2	2.0	2.1	2.0	2.1	2.1	2.1	2.0	2.0	2.06	2.20	
10-Oct	2.0	2.1	A	2.2	2.1	2.1	2.2	2.1	2.2	2.1	2.1	2.1	2.0	2.2	2.1	2.1	2.1	2.1	2.3	2.2	2.1	2.2	2.3	2.5	2.15	2.46	
11-Oct	2.6	A	2.5	2.4	2.5	2.6	3.7	2.8	2.5	2.5	2.5	2.4	2.1	2.0	2.1	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.1	2.4	2.32	3.66	
12-Oct	A	2.8	2.6	2.3	2.8	2.6	2.1	2.1	2.1	2.1	2.1	2.1	2.0	2.0	2.0	2.0	2.1	2.2	2.0	2.1	2.1	2.1	2.3	A	2.20	2.83	
13-Oct	2.2	2.0	2.0	2.2	2.1	2.0	2.0	2.2	2.0	2.0	2.1	2.1	2.0	2.0	2.0	2.1	1.9	1.9	2.0	2.0	2.0	2.1	A	2.1	2.05	2.18	
14-Oct	1.9	2.0	2.4	2.3	2.3	2.3	2.4	2.4	2.4	2.4	2.2	2.0	1.9	1.9	2.1	2.0	2.0	2.0	2.0	2.0	1.9	A	1.9	1.9	2.11	2.43	
15-Oct	1.9	1.9	2.1	2.0	2.1	2.2	2.5	2.1	2.0	1.9	2.0	1.9	1.9	1.9	1.9	1.9	2.1	1.9	1.9	2.2	A	1.9	1.9	1.9	2.00	2.54	
16-Oct	1.9	2.0	2.0	1.9	2.2	2.5	2.3	2.5	2.7	3.2	2.2	2.3	1.9	1.9	2.0	2.0	1.9	1.9	1.9	A	2.0	2.0	2.0	1.9	2.13	3.19	
17-Oct	1.9	2.1	2.2	2.1	2.2	2.3	2.5	2.3	2.6	2.9	2.2	2.2	2.3	2.0	2.0	2.2	2.1	1.9	A	2.0	2.2	2.0	1.9	2.0	2.19	2.87	
18-Oct	2.0	2.1	2.0	1.9	1.9	2.0	2.0	2.0	1.9	2.0	2.0	1.9	2.0	1.9	2.1	1.9	1.9	A	2.0	2.4	2.5	2.8	3.7	2.5	2.15	3.66	
19-Oct	2.9	2.5	2.6	2.8	3.6	3.5	2.7	2.3	2.4	2.4	2.5	2.3	2.1	2.1	2.1	1.9	A	2.0	2.0	2.1	2.1	2.1	3.8	2.7	2.51	3.84	
20-Oct	2.5	2.6	2.5	2.6	3.1	2.0	2.0	1.9	1.9	1.9	2.0	2.1	1.9	1.9	1.9	A	2.1	1.9	2.0	1.9	2.0	2.0	2.0	2.0	2.12	3.15	
21-Oct	2.1	2.0	2.0	2.0	2.3	2.3	2.4	2.3	2.6	2.2	2.1	2.0	2.0	2.0	A	2.1	2.2	2.0	2.0	2.0	2.1	2.8	2.3	3.1	2.21	3.09	
22-Oct	3.0	2.7	3.1	2.9	3.2	2.6	3.1	3.1	2.8	2.8	2.9	2.5	2.5	A	2.3	2.3	2.2	2.2	2.3	2.4	2.5	2.5	2.8	3.1	2.69	3.23	
23-Oct	3.0	3.0	3.0	3.3	3.3	3.5	3.2	3.3	3.0	2.8	2.6	2.4	A	2.1	2.1	2.1	2.0	2.2	2.3	2.3	2.3	2.6	2.6	3.1	2.70	3.45	
24-Oct	2.9	3.0	2.9	2.7	2.7	2.8	2.9	2.7	2.7	2.9	2.7	A	2.0	2.1	2.1	2.1	1.9	2.1	2.0	2.0	2.2	2.3	2.4	2.2	2.44	3.05	
25-Oct	2.6	2.9	3.0	3.0	3.1	3.1	3.2	2.8	2.8	2.7	A	2.6	2.2	2.1	2.1	2.1	2.1	1.9	1.9	1.9	1.9	1.9	2.0	1.9	2.42	3.22	
26-Oct	2.0	2.1	2.0	2.0	2.5	2.3	2.3	2.4	2.3	A	2.5	2.3	2.3	2.3	2.3	2.6	2.2	2.3	2.5	2.5	2.5	2.1	1.9	1.9	2.27	2.65	
27-Oct	2.0	1.9	2.2	1.9	2.0	1.9	1.9	1.9	A	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.3	2.1	2.3	2.2	2.7	2.02	2.72	
28-Oct	2.4	2.3	2.1	2.1	2.2	2.1	2.0	A	2.0	1.9	2.0	2.1	2.0	2.3	2.2	2.1	2.1	2.2	2.2	2.1	2.2	2.1	2.1	2.2	2.13	2.38	
29-Oct	2.4	2.4	2.5	2.4	2.5	2.4	A	2.5	2.8	2.7	2.5	2.1	2.0	1.9	1.9	2.1	1.9	1.9	1.9	2.0	2.1	2.2	2.1	2.3	2.24	2.82	
30-Oct	2.2	2.4	2.1	2.1	2.0	A	2.0	1.9	2.1	2.0	2.3	2.0	1.9	1.9	2.4	1.9	1.9	2.0	1.9	2.0	1.9	2.1	2.0	2.0	2.04	2.38	
31-Oct	2.2	2.2	2.2	2.2	A	2.4	2.7	2.6	2.2	2.4	2.2	2.3	2.2	2.2	2.1	2.1	2.0	1.9	2.0	2.2	2.3	2.1	2.0	2.2	2.21	2.73	
2.25 2.27 2.30 2.29 2.42 2.37 2.40 2.35 2.36 2.31 2.21 2.14 2.05 2.02 2.05 2.08 2.01 2.04 2.04 2.10 2.13 2.18 2.27 2.24																									Diurnal Average		
3.02 3.05 3.05 3.27 3.59 3.47 3.66 3.32 3.25 3.19 2.90 2.63 2.55 2.31 2.38 2.65 2.19 2.29 2.48 2.54 2.50 2.84 3.84 3.12																									Diurnal Maximum		
C - Calibration A - Automated Daily Zero Span																											

Hourly Maximums


Total Hydrocarbons (THC) - ppm
Henry Pirker - October 2018



Pollutant Rose


Total Hydrocarbons (THC) - ppm

Henry Pirker - October 2018



Span Responses

Total Hydrocarbons (THC)
Henry Pirker - October 2018



Hourly Averages


Methane (CH₄) - ppm

Henry Pirker - October 2018

Maximum Value: 2.87 ppm on Oct 22 05:00																				Maximum Daily Average: 2.31 ppm on Oct 23					Hours in Service: 744	
Minimum Value: 1.8 ppm on Oct 15 15:00																				Minimum Daily Average: 1.88 ppm on Oct 15					Hours of Data: 708	
Maximum Diurnal Average: 2.14 ppm at hour 7																				Minimum Diurnal Average: 1.92 ppm at hour 17					Hours of Missing Data: 36	
Monthly Average: 2.025 ppm																				Percentiles: P ₁ = 1.85 P ₁₀ = 1.88 Q ₁ = 1.92 Median = 1.96 Q ₃ = 2.06 P ₉₀ = 2.29 P ₉₉ = 2.60					Hours of Calibration: 36	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.0	2.0	2.0	2.0	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.96	2.11
2-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.94	1.96
3-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.9	2.0	1.9	2.0	1.9	1.9	1.95	1.98
4-Oct	2.0	2.1	2.1	2.1	2.3	2.5	2.4	2.3	A	2.1	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.1	2.2	2.1	2.07	2.46	
5-Oct	2.1	2.1	2.0	2.0	2.0	2.0	2.0	A	1.9	1.9	1.9	1.9	1.9	C	C	C	C	1.9	1.9	1.9	2.0	2.0	2.0	2.0	1.97	2.08
6-Oct	2.0	2.0	2.2	2.2	2.1	2.1	A	2.5	2.5	2.1	2.1	2.1	1.9	1.9	1.9	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.0	2.05	2.48
7-Oct	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.97	1.99
8-Oct	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.97	1.99
9-Oct	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.98	2.00
10-Oct	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.0	2.1	2.1	2.1	2.02	2.12
11-Oct	2.3	A	2.2	2.2	2.3	2.3	2.7	2.4	2.2	2.2	2.2	2.2	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.11	2.73	
12-Oct	A	2.2	2.3	2.0	2.4	2.2	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	A	2.01	2.38
13-Oct	2.0	2.0	2.0	1.9	1.9	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.9	1.9	A	1.9	1.95	2.02
14-Oct	1.9	1.9	2.0	2.1	2.1	2.1	2.2	2.2	2.1	2.2	2.1	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.97	2.19
15-Oct	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.9	1.9	A	1.9	1.9	1.9	1.88	2.04
16-Oct	1.9	1.9	1.9	1.9	2.0	2.0	2.1	2.2	2.2	2.6	2.1	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.97	2.60
17-Oct	1.9	1.9	2.0	2.0	2.1	2.1	2.2	2.2	2.3	2.4	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	2.0	1.9	1.9	1.9	2.01	2.42
18-Oct	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	2.1	2.1	2.1	2.2	2.2	1.95	2.17
19-Oct	2.2	2.2	2.3	2.5	2.5	2.6	2.2	2.1	2.1	2.2	2.3	2.1	2.0	1.9	1.9	1.9	A	1.9	1.9	1.9	2.0	2.2	2.2	2.2	2.13	2.57
20-Oct	2.2	2.4	2.2	2.3	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.96	2.36	
21-Oct	1.9	1.9	1.9	1.9	2.0	2.1	2.1	2.1	2.3	2.1	2.0	2.0	1.9	1.9	A	1.9	1.9	1.9	2.0	2.0	2.0	2.1	2.1	2.4	2.02	2.37
22-Oct	2.4	2.3	2.3	2.4	2.9	2.3	2.5	2.5	2.4	2.4	2.5	2.3	2.3	A	2.1	2.0	2.0	2.0	2.0	2.1	2.2	2.2	2.3	2.4	2.30	2.87
23-Oct	2.5	2.5	2.4	2.6	2.7	2.6	2.6	2.6	2.6	2.5	2.3	2.1	A	2.0	2.0	1.9	2.0	2.0	2.1	2.1	2.4	2.4	2.4	2.4	2.31	2.66
24-Oct	2.4	2.5	2.5	2.4	2.4	2.4	2.4	2.3	2.4	2.4	2.3	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.1	2.0	2.16	2.49	
25-Oct	2.3	2.6	2.6	2.6	2.6	2.6	2.6	2.4	2.4	2.4	A	2.2	2.1	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.18	2.63
26-Oct	1.9	1.9	1.9	1.9	2.1	2.2	2.1	2.1	2.1	A	2.1	2.1	2.1	2.1	2.2	2.3	2.1	2.1	2.1	2.3	2.3	1.9	1.9	1.9	2.07	2.32
27-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.1	2.2	1.92	2.16
28-Oct	2.2	2.1	2.0	2.0	2.0	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.99	2.18
29-Oct	2.1	2.1	2.2	2.2	2.2	2.2	A	2.2	2.4	2.3	2.1	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.1	2.0	2.0	2.0	2.05	2.42
30-Oct	2.0	2.0	1.9	2.0	2.0	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.92	2.03	
31-Oct	2.0	2.0	2.0	2.0	A	2.2	2.5	2.2	2.0	2.1	2.0	2.1	2.0	1.9	1.9	2.0	1.9	1.9	1.9	2.1	2.0	1.9	1.9	2.03	2.46	
2.05 2.07 2.08 2.09 2.13 2.13 2.14 2.12 2.12 2.11 2.04 2.00 1.95 1.93 1.94 1.94 1.92 1.93 1.94 1.96 1.99 1.99 2.01 2.03																								Diurnal Average		
2.45 2.58 2.63 2.60 2.87 2.63 2.73 2.58 2.60 2.60 2.45 2.31 2.28 2.14 2.17 2.31 2.06 2.07 2.09 2.32 2.27 2.35 2.39 2.40																								Diurnal Maximum		
C - Calibration A - Automated Daily Zero Span																										

Hourly Averages


Methane (CH₄) - ppm
Henry Pirker - October 2018



Maximum Value: 3.85 ppm on Oct 19 23:00																				Maximum Daily Average: 2.55 ppm on Oct 22					Hours in Service: 744	
Minimum Value: 1.9 ppm on Oct 15 16:00																				Minimum Daily Average: 1.92 ppm on Oct 15					Hours of Data: 708	
Maximum Diurnal Average: 2.31 ppm at hour 5																				Minimum Diurnal Average: 1.96 ppm at hour 17					Hours of Missing Data: 36	
Monthly Average: 2.127 ppm																				Percentiles: P ₁ = 1.88 P ₁₀ = 1.90 Q ₁ = 1.95 Median = 2.01 Q ₃ = 2.20 P ₉₀ = 2.53 P ₉₉ = 3.18					Hours of Calibration: 36	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	2.0	2.0	2.0	2.0	2.0	2.1	2.2	2.1	2.2	2.1	2.0	A	2.0	1.9	1.9	1.9	2.0	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.01	2.18
2-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	1.9	1.9	1.9	2.0	1.9	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.96	2.00
3-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	1.9	2.1	2.0	2.0	2.0	2.0	2.0	2.0	1.99	2.12
4-Oct	2.3	2.2	2.2	2.5	2.5	3.0	2.5	2.4	A	2.3	2.0	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.5	2.6	3.0	2.4	2.27	3.04
5-Oct	2.2	2.2	2.1	2.0	2.0	2.0	2.0	A	2.0	2.1	2.0	2.0	2.0	C	C	C	C	1.9	2.0	2.0	2.0	2.0	2.3	2.0	2.03	2.26
6-Oct	2.0	2.2	2.4	2.4	2.4	2.2	A	2.8	3.2	2.1	2.1	2.2	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.0	2.2	2.0	2.19	3.25
7-Oct	2.1	2.1	2.0	2.0	2.0	A	2.1	2.1	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.0	2.0	2.0	2.1	2.2	2.03	2.19
8-Oct	2.0	2.1	2.0	2.0	A	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.99	2.08
9-Oct	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.00	2.02
10-Oct	2.0	2.0	A	2.0	2.1	2.0	2.1	2.0	2.0	2.0	2.0	2.1	2.0	2.0	2.0	2.1	2.0	2.0	2.2	2.1	2.1	2.2	2.3	2.5	2.09	2.45
11-Oct	2.6	A	2.5	2.3	2.5	2.5	3.7	2.8	2.4	2.3	2.3	2.4	2.1	2.0	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.9	2.4	2.26	3.66
12-Oct	A	2.7	2.6	2.2	2.7	2.6	2.1	1.9	1.9	2.0	2.0	1.9	2.0	1.9	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.2	A	2.11	2.73
13-Oct	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.2	2.0	2.0	2.0	2.0	2.0	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	2.0	A	2.0	1.99	2.16
14-Oct	1.9	2.0	2.4	2.2	2.3	2.2	2.4	2.4	2.2	2.4	2.2	2.0	1.9	1.9	1.9	1.9	1.9	2.0	1.9	1.9	A	1.9	1.9	1.9	2.06	2.43
15-Oct	1.9	1.9	1.9	1.9	1.9	2.0	2.1	2.1	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.2	A	1.9	1.9	1.9	1.92	2.15
16-Oct	1.9	2.0	1.9	1.9	2.2	2.1	2.1	2.3	2.7	3.2	2.2	2.2	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	2.0	1.9	2.07	3.19
17-Oct	1.9	2.1	2.2	2.0	2.1	2.3	2.4	2.3	2.6	2.9	2.3	2.2	2.1	1.9	2.0	2.0	2.0	2.0	1.9	A	2.0	2.2	2.0	1.9	2.14	2.86
18-Oct	2.0	2.0	1.9	1.9	1.9	1.9	2.0	1.9	1.9	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	2.0	2.3	2.5	2.8	3.7	2.5	2.11	3.65
19-Oct	2.9	2.5	2.5	2.6	3.3	3.2	2.5	2.3	2.4	2.3	2.4	2.2	2.0	2.0	1.9	1.9	A	1.9	1.9	1.9	2.0	2.1	3.8	2.7	2.41	3.85
20-Oct	2.4	2.6	2.5	2.5	3.2	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	2.0	2.0	2.0	2.0	2.07	3.15
21-Oct	2.1	2.0	2.0	2.0	2.3	2.3	2.4	2.3	2.6	2.1	2.0	2.0	1.9	1.9	A	2.1	2.0	2.0	2.0	2.0	2.1	2.8	2.3	3.1	2.19	3.09
22-Oct	3.0	2.7	3.1	2.9	3.2	2.6	3.1	2.9	2.6	2.6	2.7	2.4	2.3	A	2.2	2.1	2.1	2.1	2.2	2.3	2.5	2.2	2.4	2.6	2.55	3.23
23-Oct	2.6	2.6	2.5	2.7	2.8	2.8	2.9	3.1	2.7	2.6	2.5	2.2	A	2.0	2.0	2.1	2.0	2.1	2.1	2.1	2.1	2.5	2.5	2.6	2.44	3.06
24-Oct	2.5	2.5	2.5	2.5	2.6	2.5	2.8	2.7	2.4	2.5	2.5	A	2.0	2.0	2.0	1.9	1.9	1.9	2.0	2.0	2.0	2.2	2.3	2.2	2.27	2.78
25-Oct	2.6	2.9	2.8	2.8	2.6	2.9	2.8	2.6	2.5	2.6	A	2.3	2.1	2.1	2.0	2.0	1.9	1.9	1.9	1.9	1.9	2.0	1.9	1.9	2.30	2.86
26-Oct	1.9	1.9	1.9	2.0	2.5	2.3	2.3	2.4	2.2	A	2.2	2.1	2.2	2.2	2.3	2.6	2.1	2.2	2.5	2.5	2.4	1.9	1.9	1.9	2.20	2.63
27-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.3	2.1	2.3	2.2	2.4	1.98	2.43
28-Oct	2.4	2.2	2.1	2.1	2.0	1.9	1.9	A	1.9	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.2	2.0	2.1	2.03	2.38
29-Oct	2.1	2.3	2.4	2.3	2.3	2.4	A	2.3	2.5	2.4	2.2	2.1	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.1	2.1	2.1	2.3	2.14	2.51
30-Oct	2.2	2.4	2.0	2.1	2.0	A	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.4	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.99	2.38	
31-Oct	2.2	2.1	2.2	2.1	A	2.3	2.7	2.4	2.1	2.3	2.2	2.2	2.2	2.0	1.9	2.0	2.0	1.9	2.0	2.0	2.1	2.1	2.0	2.0	2.13	2.66
																									Diurnal Average	
																									Diurnal Maximum	
C - Calibration																									A - Automated Daily Zero Span	


Hourly Maximums

Methane (CH₄) - ppm
Henry Pirker - October 2018




Pollutant Rose

Methane (CH₄) - ppm
Henry Pirker - October 2018



Span Responses

Methane (CH₄)
Henry Pirker - October 2018




Hourly Averages

Non Methane Hydrocarbon (NMHC) - ppm

Henry Pirker - October 2018

Maximum Value: 0.43 ppm on Oct 23 04:00																					Maximum Daily Average: 0.13 ppm on Oct 23				Hours in Service: 744	
Minimum Value: 0.0 ppm on Oct 1 03:00																					Minimum Daily Average: 0.00 ppm on Oct 6				Hours of Data: 708	
Maximum Diurnal Average: 0.03 ppm at hour 5																					Minimum Diurnal Average: 0.00 ppm at hour 17				Hours of Missing Data: 36	
Monthly Average: 0.013 ppm																					Percentiles: P ₁ = 0.00 P ₁₀ = 0.00 Q ₁ = 0.00 Median = 0.00 Q ₃ = 0.01 P ₉₀ = 0.02 P ₉₉ = 0.24				Hours of Calibration: 36	
																					Percent Operational Time: 100.0					
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.01
2-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.00
3-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.04
4-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.01	0.04
5-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	C	C	C	C	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.02
6-Oct	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.00
7-Oct	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.01
8-Oct	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.01
9-Oct	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.01
10-Oct	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.01
11-Oct	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.01	0.03
12-Oct	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.00	0.01
13-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.00	0.01
14-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.00	0.02
15-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.01	0.16
16-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.01	0.04
17-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.01	0.06
18-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.01
19-Oct	0.0	0.0	0.0	0.1	0.2	0.2	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.03	0.17
20-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.01	0.03
21-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.01
22-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.1	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.1	0.2	0.3	0.04	0.26
23-Oct	0.3	0.2	0.3	0.4	0.4	0.4	0.2	0.1	0.1	0.0	0.1	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.2	0.13	0.43
24-Oct	0.2	0.3	0.2	0.1	0.1	0.1	0.1	0.0	0.0	0.1	0.1	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.06	0.34
25-Oct	0.0	0.0	0.1	0.1	0.2	0.1	0.1	0.1	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.04	0.22
26-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.01	0.02
27-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.03
28-Oct	0.0	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.02
29-Oct	0.0	0.0	0.0	0.0	0.0	0.0	A	0.0	0.2	0.1	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.02	0.17
30-Oct	0.0	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.00	0.03
31-Oct	0.0	0.0	0.0	0.0	A	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.0	0.01	0.02
0.02 0.02 0.02 0.03 0.03 0.03 0.03 0.02 0.02 0.02 0.01 0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.01 0.01 0.02																									Diurnal Average	
0.27 0.34 0.32 0.43 0.39 0.38 0.23 0.15 0.17 0.11 0.11 0.02 0.05 0.01 0.01 0.02 0.01 0.01 0.02 0.02 0.02 0.02 0.05 0.16 0.26																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										




Hourly Maximums

Non Methane Hydrocarbon (NMHC) - ppm

Henry Pirker - October 2018


Maximum Value: 0.64 ppm on Oct 23 04:00																					Maximum Daily Average: 0.33 ppm on Oct 23					Hours in Service: 744	
Minimum Value: 0.0 ppm on Oct 2 12:00																					Minimum Daily Average: 0.07 ppm on Oct 27					Hours of Data: 708	
Maximum Diurnal Average: 0.18 ppm at hour 9																					Minimum Diurnal Average: 0.08 ppm at hour 17					Hours of Missing Data: 36	
Monthly Average: 0.124 ppm																					Percentiles: P ₁ = 0.00 P ₁₀ = 0.01 Q ₁ = 0.04 Median = 0.10 Q ₃ = 0.18 P ₉₀ = 0.26 P ₉₉ = 0.51					Hours of Calibration: 36	
																					Percent Operational Time: 100.0						
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	0.1	0.1	0.0	0.1	0.1	0.0	0.1	0.1	0.2	0.1	0.2	A	0.1	0.1	0.0	0.0	0.0	0.1	0.0	0.2	0.0	0.1	0.0	0.1	0.08	0.22	
2-Oct	0.1	0.2	0.0	0.1	0.1	0.0	0.0	0.1	0.1	0.1	A	0.0	0.2	0.1	0.0	0.2	0.1	0.1	0.0	0.1	0.1	0.1	0.2	0.0	0.08	0.24	
3-Oct	0.1	0.1	0.0	0.1	0.0	0.0	0.1	0.1	0.4	A	0.0	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.2	0.1	0.1	0.1	0.0	0.1	0.07	0.38	
4-Oct	0.1	0.1	0.2	0.0	0.2	0.3	0.3	0.3	A	0.2	0.1	0.1	0.1	0.1	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.1	0.0	0.0	0.11	0.33	
5-Oct	0.2	0.3	0.3	0.2	0.0	0.1	0.2	A	0.0	0.2	0.0	0.0	0.1	C	C	C	C	0.1	0.1	0.1	0.0	0.0	0.1	0.0	0.10	0.34	
6-Oct	0.0	0.2	0.1	0.2	0.1	0.1	A	0.1	0.2	0.1	0.2	0.2	0.0	0.0	0.0	0.3	0.0	0.1	0.0	0.0	0.1	0.1	0.0	0.1	0.09	0.27	
7-Oct	0.1	0.0	0.0	0.0	0.1	A	0.1	0.1	0.2	0.1	0.2	0.0	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.2	0.2	0.1	0.1	0.0	0.10	0.20	
8-Oct	0.1	0.1	0.0	0.1	A	0.1	0.1	0.1	0.1	0.1	0.1	0.2	0.0	0.1	0.1	0.1	0.0	0.1	0.1	0.0	0.1	0.2	0.0	0.0	0.08	0.17	
9-Oct	0.1	0.1	0.1	A	0.2	0.0	0.0	0.2	0.0	0.1	0.0	0.0	0.0	0.0	0.0	0.2	0.0	0.1	0.0	0.1	0.1	0.1	0.0	0.0	0.07	0.20	
10-Oct	0.0	0.1	A	0.2	0.1	0.0	0.2	0.1	0.2	0.1	0.0	0.1	0.0	0.2	0.0	0.1	0.1	0.1	0.2	0.1	0.1	0.1	0.2	0.2	0.11	0.24	
11-Oct	0.2	A	0.2	0.2	0.2	0.2	0.1	0.2	0.2	0.2	0.3	0.2	0.1	0.0	0.2	0.1	0.1	0.1	0.0	0.1	0.1	0.0	0.1	0.0	0.13	0.27	
12-Oct	A	0.2	0.3	0.2	0.1	0.1	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.1	0.0	0.0	0.1	0.2	0.0	0.1	0.2	0.1	0.2	A	0.13	0.27	
13-Oct	0.2	0.0	0.1	0.3	0.1	0.1	0.0	0.1	0.1	0.0	0.1	0.1	0.1	0.1	0.1	0.2	0.0	0.0	0.0	0.0	0.1	0.2	A	0.2	0.09	0.25	
14-Oct	0.0	0.0	0.0	0.3	0.2	0.2	0.2	0.1	0.3	0.1	0.1	0.1	0.0	0.0	0.2	0.1	0.1	0.1	0.1	0.1	0.0	A	0.0	0.0	0.10	0.32	
15-Oct	0.0	0.0	0.3	0.1	0.2	0.3	0.5	0.1	0.1	0.1	0.1	0.0	0.1	0.0	0.0	0.0	0.2	0.0	0.1	0.1	A	0.0	0.0	0.0	0.10	0.49	
16-Oct	0.0	0.0	0.1	0.0	0.1	0.4	0.2	0.3	0.2	0.2	0.1	0.1	0.0	0.0	0.1	0.2	0.0	0.0	0.0	A	0.0	0.1	0.0	0.0	0.11	0.44	
17-Oct	0.0	0.1	0.2	0.1	0.1	0.2	0.1	0.2	0.2	0.1	0.1	0.0	0.4	0.1	0.0	0.2	0.1	0.0	A	0.1	0.0	0.1	0.0	0.0	0.11	0.37	
18-Oct	0.1	0.2	0.0	0.0	0.0	0.0	0.1	0.1	0.0	0.1	0.2	0.0	0.2	0.0	0.2	0.0	0.0	A	0.1	0.2	0.1	0.2	0.1	0.1	0.09	0.23	
19-Oct	0.2	0.1	0.2	0.3	0.4	0.5	0.3	0.1	0.1	0.2	0.2	0.2	0.1	0.2	0.1	0.0	A	0.1	0.1	0.2	0.2	0.1	0.1	0.1	0.19	0.45	
20-Oct	0.2	0.2	0.2	0.3	0.1	0.1	0.1	0.0	0.0	0.0	0.2	0.2	0.0	0.0	0.0	A	0.2	0.0	0.1	0.0	0.0	0.0	0.0	0.1	0.10	0.29	
21-Oct	0.1	0.1	0.0	0.0	0.1	0.1	0.0	0.0	0.2	0.1	0.1	0.1	0.1	0.1	A	0.0	0.3	0.1	0.0	0.0	0.1	0.0	0.1	0.1	0.08	0.27	
22-Oct	0.1	0.1	0.2	0.2	0.1	0.2	0.1	0.3	0.3	0.3	0.3	0.1	0.2	A	0.2	0.2	0.1	0.2	0.2	0.2	0.2	0.3	0.5	0.5	0.23	0.54	
23-Oct	0.4	0.5	0.5	0.6	0.6	0.6	0.4	0.4	0.3	0.2	0.4	0.3	A	0.1	0.2	0.0	0.0	0.2	0.2	0.2	0.2	0.2	0.3	0.5	0.33	0.64	
24-Oct	0.4	0.5	0.4	0.3	0.3	0.3	0.3	0.2	0.3	0.4	0.3	A	0.1	0.2	0.1	0.3	0.1	0.2	0.0	0.0	0.2	0.3	0.2	0.2	0.24	0.54	
25-Oct	0.2	0.2	0.3	0.3	0.5	0.3	0.5	0.3	0.3	0.2	A	0.3	0.1	0.1	0.1	0.0	0.2	0.1	0.1	0.0	0.0	0.0	0.0	0.0	0.19	0.51	
26-Oct	0.1	0.2	0.1	0.0	0.0	0.2	0.2	0.2	0.2	A	0.3	0.2	0.2	0.2	0.1	0.2	0.1	0.2	0.1	0.1	0.1	0.2	0.0	0.1	0.14	0.29	
27-Oct	0.1	0.0	0.3	0.0	0.1	0.0	0.0	0.0	A	0.1	0.0	0.0	0.0	0.0	0.1	0.0	0.0	0.1	0.1	0.1	0.1	0.1	0.1	0.3	0.07	0.29	
28-Oct	0.2	0.2	0.1	0.1	0.2	0.1	0.1	A	0.1	0.0	0.1	0.2	0.1	0.3	0.2	0.1	0.1	0.2	0.2	0.1	0.2	0.1	0.1	0.1	0.14	0.31	
29-Oct	0.3	0.2	0.2	0.2	0.3	0.1	A	0.2	0.4	0.3	0.3	0.0	0.1	0.0	0.1	0.2	0.0	0.0	0.0	0.1	0.2	0.1	0.1	0.2	0.16	0.36	
30-Oct	0.0	0.2	0.2	0.1	0.0	A	0.0	0.0	0.1	0.0	0.4	0.1	0.0	0.1	0.2	0.1	0.0	0.1	0.0	0.1	0.1	0.2	0.0	0.0	0.08	0.38	
31-Oct	0.0	0.2	0.1	0.1	A	0.2	0.2	0.2	0.2	0.2	0.1	0.1	0.1	0.2	0.1	0.1	0.0	0.0	0.1	0.2	0.2	0.1	0.1	0.2	0.14	0.24	
0.12 0.15 0.16 0.16 0.16 0.17 0.16 0.15 0.18 0.14 0.16 0.11 0.09 0.09 0.09 0.11 0.08 0.09 0.08 0.10 0.10 0.12 0.09 0.11																									Diurnal Average		
0.44 0.54 0.53 0.64 0.57 0.62 0.49 0.36 0.38 0.36 0.44 0.33 0.37 0.31 0.23 0.27 0.27 0.23 0.24 0.25 0.23 0.31 0.49 0.55																									Diurnal Maximum		
C - Calibration A - Automated Daily Zero Span																											




Pollutant Rose

Non Methane Hydrocarbon (NMHC) - ppm

Henry Pirker - October 2018





Hourly Averages

Particulate Matter 2.5 (PM_{2.5}) - µg/m³


Henry Pirker - October 2018

Number of Exceedences: 1-hr: 1 24-hr: 0 Maximum Value: 82.3 µg/m³ on Oct 22 20:00 Maximum Daily Average: 16.0 µg/m³ on Oct 22																								Hours in Service: 744 Hours of Data: 744 Hours of Missing Data: 0 Hours of Calibration: 0 Percent Operational Time: 100.0			
Minimum Value: 1 µg/m³ on Oct 29 15:00 Maximum Diurnal Average: 9.8 µg/m³ at hour 9 Monthly Average: 6.57 µg/m³												Minimum Daily Average: 1.8 µg/m³ on Oct 30 Minimum Diurnal Average: 3.4 µg/m³ at hour 17 Percentiles: P ₁ = 1.0 P ₁₀ = 1.6 Q ₁ = 2.4 Median = 4.9 Q ₃ = 8.9 P ₉₀ = 13.5 P ₉₉ = 24.7															
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	2	2	2	2	2	2	3	3	3	3	3	2	2	2	2	3	3	3	3	2	3	2	2	2	2	2.3	3.5
2-Oct	2	2	2	2	1	2	2	2	2	2	1	1	1	1	2	2	2	2	3	3	8	9	5	3	3	2.6	8.6
3-Oct	2	2	2	2	3	3	3	3	4	3	2	2	2	2	2	1	1	2	9	8	3	2	3	4	2.9	8.6	
4-Oct	6	5	5	5	5	5	6	11	10	10	10	11	11	7	4	2	3	3	4	7	10	9	7	6	6.7	11.4	
5-Oct	9	7	6	4	3	3	3	3	3	2	2	2	2	2	2	2	2	3	4	5	5	9	6	8	4.1	9.2	
6-Oct	9	12	9	10	9	11	10	10	9	8	8	5	3	2	3	6	5	6	6	6	6	5	6	6	7.1	11.9	
7-Oct	7	7	7	7	7	7	7	7	6	5	5	4	2	3	5	5	4	3	4	4	4	4	4	4	5.1	7.4	
8-Oct	4	5	4	4	4	3	2	2	2	2	2	2	2	2	2	2	2	2	2	2	2	3	3	8	2.9	7.6	
9-Oct	13	13	13	13	12	10	8	8	7	6	5	4	4	4	4	3	3	3	4	4	5	5	6	7	6.8	12.8	
10-Oct	8	8	9	10	10	10	10	10	9	8	10	9	8	8	9	11	10	9	8	7	10	9	6	8	8.9	10.6	
11-Oct	9	11	15	20	14	10	9	7	18	22	6	5	3	2	1	1	1	1	1	1	2	2	2	2	7.0	22.3	
12-Oct	2	3	5	7	8	9	5	3	2	1	1	1	2	2	2	2	2	2	2	2	2	3	4	5	3.3	9.4	
13-Oct	5	5	3	2	2	2	2	2	3	3	2	2	2	2	2	2	2	2	5	10	3	3	3	3	3.0	9.9	
14-Oct	2	6	5	5	10	8	8	7	7	7	5	2	2	2	2	1	1	2	2	2	2	3	3	3	4.0	9.8	
15-Oct	3	3	3	3	4	6	8	6	3	3	2	2	1	1	1	1	1	2	3	8	7	7	10	9	4.0	9.8	
16-Oct	9	9	8	8	7	8	11	19	21	14	8	6	4	6	8	4	3	3	6	6	9	9	11	11	8.7	21.5	
17-Oct	13	14	13	14	17	17	18	20	31	27	11	8	4	5	5	6	7	7	7	11	9	8	8	10	11.9	30.6	
18-Oct	11	12	13	10	7	8	8	8	5	4	2	2	1	1	1	1	2	8	12	14	15	11	12	11	7.5	15.0	
19-Oct	12	12	11	10	9	9	7	6	9	9	10	17	8	3	2	2	1	2	3	7	5	5	7	24	7.9	23.9	
20-Oct	18	14	10	10	7	3	2	2	2	2	3	2	2	2	1	1	1	2	2	3	5	5	4	4	4.6	17.8	
21-Oct	4	4	4	4	6	5	7	7	8	5	3	3	2	1	1	1	2	5	21	13	8	5	6	8	5.5	21.0	
22-Oct	15	13	9	9	9	9	8	11	14	18	16	12	13	11	10	10	7	19	14	82	24	20	14	15	16.0	82.3	
23-Oct	13	13	12	11	11	11	12	13	15	13	10	8	6	6	6	4	6	12	20	17	25	37	28	21	13.8	37.0	
24-Oct	15	14	14	14	12	13	13	11	18	15	17	11	9	9	10	10	12	12	15	17	18	22	23	21	14.4	22.5	
25-Oct	25	25	24	21	20	20	19	19	16	15	19	20	11	5	4	4	2	3	1	1	1	1	2	2	11.7	25.3	
26-Oct	2	2	3	3	6	5	9	11	19	13	14	13	17	12	10	10	4	4	5	7	6	2	1	1	7.4	18.9	
27-Oct	1	1	2	2	2	2	2	3	3	4	3	2	2	2	1	2	2	4	5	7	7	8	12	19	4.1	19.0	
28-Oct	22	14	5	5	4	3	3	3	3	3	3	5	4	4	5	5	5	5	5	5	5	6	5	6	5.5	21.9	
29-Oct	7	8	9	8	6	6	9	18	42	17	7	2	1	1	1	1	1	1	2	4	6	5	3	2	6.9	42.2	
30-Oct	2	2	1	2	2	1	1	2	2	5	2	1	1	1	1	1	1	2	2	2	2	2	2	2	1.8	4.7	
31-Oct	4	3	5	4	6	6	7	8	8	8	6	5	5	4	4	5	6	4	5	5	6	4	4	6	5.4	8.4	
8.3 8.1 7.5 7.4 7.2 7.1 7.1 7.9 9.8 8.2 6.4 5.5 4.5 3.6 3.6 3.5 3.4 4.5 6.0 8.9 7.3 7.3 6.8 7.8																								Diurnal Average			
24.5 25.3 23.9 21.0 19.7 20.2 18.9 20.3 42.2 26.6 18.8 20.3 16.9 11.8 10.5 10.6 11.6 19.4 21.0 82.3 24.9 37.0 27.6 23.9																								Diurnal Maximum			
Alberta Ambient Air Quality Guideline (AAAQG): 1-hr 80 µg/m³																								Alberta Ambient Air Quality Objective (AAAQO): 24-hr 30 µg/m³			

Hourly Averages

Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$

Henry Pirker - October 2018



Hourly Maximums

Particulate Matter 2.5 (PM_{2.5}) - µg/m³


Henry Pirker - October 2018

Maximum Value: 450.0 $\mu\text{g}/\text{m}^3$ on Oct 22 20:00																					Maximum Daily Average: 39.1 $\mu\text{g}/\text{m}^3$ on Oct 22										Hours in Service: 744	
Minimum Value: 1 $\mu\text{g}/\text{m}^3$ on Oct 29 15:00																					Minimum Daily Average: 2.9 $\mu\text{g}/\text{m}^3$ on Oct 30										Hours of Data: 744	
Maximum Diurnal Average: 25.6 $\mu\text{g}/\text{m}^3$ at hour 20																					Minimum Diurnal Average: 4.3 $\mu\text{g}/\text{m}^3$ at hour 17										Hours of Missing Data: 0	
Monthly Average: 9.59 $\mu\text{g}/\text{m}^3$																					Percentiles: P ₁ = 1.2 P ₁₀ = 2.0 Q ₁ = 2.9 Median = 6.5 Q ₃ = 11.5 P ₉₀ = 17.8 P ₉₉ = 43.9										Hours of Calibration: 0	
																					Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum						
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24								
1-Oct	2	2	2	2	2	3	4	4	4	4	4	3	2	3	2	3	5	5	3	3	3	3	2	2	3.1	5.4						
2-Oct	3	3	3	2	2	2	2	3	3	2	1	1	1	2	2	2	2	5	4	5	11	11	8	3	3.4	10.6						
3-Oct	3	3	3	3	3	3	3	4	7	3	3	4	3	3	2	2	1	2	32	29	7	3	5	8	5.7	32.5						
4-Oct	8	6	5	6	6	6	8	12	14	13	12	12	12	11	7	2	4	4	6	13	11	10	9	7	8.5	13.8						
5-Oct	16	10	7	4	4	4	3	3	4	4	3	2	2	2	2	2	3	4	5	7	7	27	8	10	5.9	26.6						
6-Oct	11	15	11	12	13	13	12	11	10	9	11	6	3	3	4	12	6	7	6	7	7	6	8	7	8.8	14.8						
7-Oct	8	8	7	7	7	7	7	8	8	6	6	4	3	4	6	5	5	4	7	5	5	4	4	5	5.9	8.4						
8-Oct	5	5	5	4	4	3	3	3	3	3	3	2	2	2	2	3	3	4	2	2	2	3	5	12	3.5	11.6						
9-Oct	14	14	14	13	13	11	9	8	8	7	5	5	5	4	5	3	4	4	5	5	6	6	6	8	7.5	14.0						
10-Oct	8	9	9	10	10	11	10	15	10	11	14	10	9	9	10	12	11	10	9	8	11	11	7	10	10.1	14.5						
11-Oct	11	15	22	23	20	11	11	8	30	44	8	7	5	2	2	2	1	2	1	1	3	2	2	3	9.9	43.7						
12-Oct	3	4	13	13	12	12	9	6	3	1	2	2	2	2	2	3	3	3	3	3	3	5	5	6	5.0	13.2						
13-Oct	6	5	4	2	2	2	2	3	4	7	2	2	2	2	2	2	2	5	9	32	5	4	3	4	4.7	32.4						
14-Oct	3	10	9	9	11	17	13	7	8	8	6	2	2	2	3	1	2	2	2	2	3	3	14	6	6.1	16.5						
15-Oct	3	3	3	4	4	9	10	10	6	3	3	2	2	1	1	1	1	2	5	15	8	9	12	10	5.3	15.4						
16-Oct	9	11	10	8	9	10	19	25	26	18	9	9	5	7	8	6	3	4	10	10	12	10	13	12	10.9	26.0						
17-Oct	13	14	14	16	18	20	21	25	44	32	12	11	7	9	6	8	8	22	11	25	11	10	10	11	15.7	44.1						
18-Oct	12	13	14	13	8	9	9	9	7	5	3	2	2	2	2	4	16	77	17	20	14	15	15	15	12.0	77.2						
19-Oct	12	13	11	12	11	11	11	7	12	12	13	29	19	5	9	2	2	3	4	11	6	7	13	29	10.9	29.0						
20-Oct	25	18	13	11	11	5	3	3	2	3	7	6	4	2	2	1	2	2	3	5	7	7	5	5	6.3	24.8						
21-Oct	5	5	6	5	6	6	8	7	10	8	6	4	3	2	1	1	2	15	33	30	12	7	7	14	8.5	32.6						
22-Oct	18	16	10	9	10	10	9	14	18	24	22	14	16	13	11	16	9	119	21	450	36	31	19	21	39.1	450.0						
23-Oct	14	16	13	12	13	13	13	14	17	14	13	10	8	7	7	5	9	16	34	31	58	61	35	29	19.2	61.2						
24-Oct	17	15	15	15	15	15	14	14	31	23	22	15	12	10	14	11	13	16	18	19	27	25	25	23	17.5	30.9						
25-Oct	27	26	26	22	20	22	20	21	21	19	23	34	16	8	6	7	2	6	3	2	2	2	2	2	14.1	33.6						
26-Oct	2	2	3	5	13	7	11	16	35	18	17	22	21	17	12	18	6	6	6	10	11	2	2	1	11.0	35.3						
27-Oct	1	1	2	2	2	2	3	3	3	4	4	3	2	2	2	2	2	6	7	23	17	11	16	43	6.8	43.4						
28-Oct	26	23	8	6	8	3	3	3	5	4	4	8	6	5	5	5	6	6	6	7	8	7	5	13	7.6	26.3						
29-Oct	9	11	10	9	8	7	11	81	74	30	12	7	3	1	1	1	1	2	4	4	8	7	4	3	12.9	80.9						
30-Oct	3	3	2	6	4	2	2	2	2	11	5	1	1	1	1	1	2	2	3	2	3	2	2	6	2.9	11.2						
31-Oct	6	5	8	5	6	6	8	12	17	13	7	6	8	5	6	6	9	6	8	9	10	7	13	8	8.1	16.8						
																									Diurnal Average							
																									Diurnal Maximum							

Hourly Maximums


Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$

Henry Pirker - October 2018



Pollutant Rose

Particulate Matter 2.5 (PM_{2.5}) - µg/m³
Henry Pirker - October 2018



Hourly Averages


External Temperature (ET) - °C

Henry Pirker - October 2018

Maximum Value: 21.3 °C on Oct 16 15:00																				Maximum Daily Average: 11.9 °C on Oct 17					Hours in Service: 744	
Minimum Value: -5 °C on Oct 3 07:00																				Minimum Daily Average: -1.4 °C on Oct 2					Hours of Data: 744	
Maximum Diurnal Average: 9.4 °C at hour 15																				Minimum Diurnal Average: 0.0 °C at hour 7					Hours of Missing Data: 0	
Monthly Average: 3.81 °C																				Percentiles: P ₁ = -4.8 P ₁₀ = -2.3 Q ₁ = -0.5 Median = 2.2 Q ₃ = 7.9 P ₉₀ = 12.2 P ₉₉ = 18.5					Hours of Calibration: 0	
																				Percent Operational Time: 100.0						
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	0	0	0	0	-1	-1	-1	-1	0	1	2	2	2	2	1	1	0	-1	-1	-2	-2	-2	-2	-0.1	2.0	
2-Oct	-3	-3	-3	-3	-3	-3	-3	-3	-3	-2	0	2	2	3	3	2	2	1	-1	-2	-3	-4	-4	-4	-1.4	2.8
3-Oct	-5	-5	-5	-5	-5	-5	-5	-5	-4	-2	0	4	6	7	8	8	8	6	5	3	2	2	1	1	0.6	8.0
4-Oct	0	-1	-2	-3	-4	-4	-5	-5	0	3	5	8	10	11	11	10	10	10	6	3	0	-1	-1	-1	2.6	11.0
5-Oct	-2	-2	-2	-3	-4	-4	-4	-4	-3	-1	0	1	5	6	6	6	5	5	3	2	0	0	-1	-1	0.2	6.3
6-Oct	-2	-2	-2	-2	-1	0	0	0	1	3	5	8	9	10	10	8	6	6	4	3	3	3	2	1	3.1	10.2
7-Oct	1	0	0	0	0	0	0	0	1	0	0	0	1	1	0	0	0	-1	-1	-1	-1	-1	-2	-0.1	1.0	
8-Oct	-2	-2	-2	-2	-2	-2	-2	-2	-1	-1	0	1	2	3	2	2	1	1	1	0	0	0	-1	-1	-0.2	2.8
9-Oct	-1	-1	-1	-1	-1	-1	-2	-2	-2	-1	-1	0	0	1	2	2	2	1	0	0	-1	-1	-1	-1	-0.5	1.8
10-Oct	-1	-1	-1	-1	-1	-2	-2	-2	-1	-1	0	0	0	1	1	1	1	0	0	0	0	0	-1	-1	-0.5	1.1
11-Oct	-2	-3	-3	-4	-4	-4	-4	-3	-2	0	2	4	7	8	8	9	9	9	8	7	6	6	6	5	2.7	9.2
12-Oct	5	5	5	5	5	5	5	3	1	0	1	1	1	0	1	1	1	1	1	1	1	1	1	1	2.2	5.2
13-Oct	1	1	2	2	1	0	1	1	1	3	6	7	8	10	10	10	10	10	5	4	3	2	2	1	4.3	10.5
14-Oct	2	0	0	0	-1	-2	-2	-2	1	5	8	12	12	13	13	14	13	12	12	11	11	11	8	8	6.6	13.6
15-Oct	9	9	8	7	5	4	3	7	11	13	13	15	16	16	18	17	16	14	12	9	9	8	8	8	10.6	17.8
16-Oct	8	7	7	5	5	4	2	1	4	8	12	16	20	20	21	21	20	18	15	13	12	13	11	10	11.3	21.3
17-Oct	10	10	8	7	5	5	6	6	7	9	14	16	20	20	19	18	17	17	15	13	11	12	11	8	11.9	19.9
18-Oct	9	12	12	14	13	10	9	10	11	12	13	13	14	14	14	14	12	9	7	5	4	4	2	1	9.8	14.1
19-Oct	0	-1	-1	-2	-2	-2	-1	-1	0	1	4	7	10	12	14	14	14	14	11	10	9	8	5	2	5.3	14.5
20-Oct	2	2	2	1	8	12	11	10	10	11	12	13	13	14	14	14	13	12	10	9	7	6	5	5	9.0	13.8
21-Oct	5	4	3	3	1	0	-1	-2	0	4	6	10	12	13	13	12	11	9	7	5	4	4	2	1	5.2	13.5
22-Oct	-1	-2	-2	-3	-3	-4	-4	-4	-3	1	4	8	10	12	13	13	11	9	5	2	0	-1	-2	-2	2.4	13.4
23-Oct	-3	-3	-3	-4	-4	-5	-5	-5	-3	1	5	8	11	13	15	14	13	11	8	6	3	0	-1	-1	2.9	14.8
24-Oct	-2	-2	-2	-3	-3	-2	-3	-2	-1	2	7	11	14	15	14	14	13	12	9	6	5	3	2	0	4.5	15.1
25-Oct	-1	-1	-1	-2	-2	-1	-1	0	2	3	7	10	13	16	15	11	8	8	8	7	7	6	5	5	5.2	16.5
26-Oct	4	4	3	1	0	-1	-1	-2	-1	0	1	2	3	4	4	4	6	4	3	2	2	4	4	5	2.3	5.7
27-Oct	5	5	6	6	6	6	5	5	6	6	8	9	10	10	11	10	9	7	6	4	2	0	0	-1	5.8	10.6
28-Oct	-2	0	0	0	0	1	1	1	1	2	3	2	2	1	1	1	2	2	1	1	1	0	0	0	0.9	2.8
29-Oct	-1	-2	-3	-3	-4	-4	-5	-5	-4	0	5	7	9	11	11	11	10	8	6	3	1	1	2	3	2.5	11.3
30-Oct	3	3	5	4	4	4	4	4	4	5	7	9	10	11	11	10	9	7	6	6	5	4	4	4	5.9	10.7
31-Oct	4	3	3	2	2	1	1	2	2	3	4	5	7	6	6	5	4	3	2	2	2	2	2	2	3.0	6.6
1.3 1.1 0.9 0.5 0.3 0.2 0.0 0.1 1.1 2.8 5.0 6.8 8.3 9.1 9.4 8.9 8.2 7.3 5.5 4.3 3.3 2.8 2.2 1.7																									Diurnal Average	
10.1 11.6 12.2 13.8 13.1 11.5 10.6 9.6 11.5 12.7 13.5 16.5 19.9 20.4 21.3 20.6 19.7 18.3 15.1 13.5 12.2 13.3 11.0 9.8																									Diurnal Maximum	

Hourly Averages

External Temperature (ET) - °C
Henry Pirker - October 2018



Hourly Averages


Relative Humidity (RH) - %

Henry Pirker - October 2018

Maximum Value: 89.3 % on Oct 12 06:00																								Maximum Daily Average: 84.2 % on Oct 12						Hours in Service: 744	
Minimum Value: 15 % on Oct 16 14:00																								Minimum Daily Average: 34.8 % on Oct 17						Hours of Data: 744	
Maximum Diurnal Average: 73.8 % at hour 7																								Minimum Diurnal Average: 46.5 % at hour 15						Hours of Missing Data: 0	
Monthly Average: 62.26 %																								Percentiles: P ₁ = 22.4 P ₁₀ = 35.6 Q ₁ = 49.4 Median = 66.6 Q ₃ = 75.6 P ₉₀ = 83.0 P ₉₉ = 87.2						Hours of Calibration: 0	
																								Percent Operational Time: 100.0							
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	59	58	58	59	59	63	70	68	65	60	55	52	53	52	61	73	73	79	84	82	81	74	68	67	65.5	84.2					
2-Oct	67	68	67	67	66	65	64	65	63	57	48	44	45	44	43	43	46	49	59	65	67	69	71	72	58.9	71.7					
3-Oct	73	72	71	70	70	70	71	70	67	64	57	51	43	38	36	34	33	35	40	45	49	52	52	54	54.9	72.6					
4-Oct	58	71	74	79	81	81	83	83	71	60	56	45	37	32	30	30	30	31	39	49	61	70	73	68	58.0	83.2					
5-Oct	72	70	69	68	69	71	73	74	72	67	66	61	52	49	48	48	48	50	55	59	65	67	70	71	63.0	74.1					
6-Oct	74	75	77	77	78	80	82	81	80	75	69	62	55	52	55	65	68	71	73	76	74	77	78	82	72.3	81.9					
7-Oct	84	84	84	85	85	85	85	85	84	84	85	85	83	79	80	82	83	83	83	83	83	83	84	84	83.5	85.4					
8-Oct	83	82	82	83	83	84	84	83	83	82	80	77	73	70	69	71	71	70	71	72	72	72	73	73	76.8	83.7					
9-Oct	74	74	74	74	75	75	75	75	75	74	71	69	67	65	62	60	59	62	65	66	69	71	72	73	69.8	75.1					
10-Oct	73	74	74	75	75	76	77	76	76	75	74	74	71	67	67	68	70	72	73	73	73	72	71	73	72.8	76.5					
11-Oct	77	80	82	84	84	85	85	84	80	75	67	60	58	58	59	55	54	56	60	64	67	70	73	84	70.9	84.6					
12-Oct	87	88	89	89	89	89	89	87	87	88	87	87	86	85	79	75	77	77	79	80	80	81	82	82	84.2	89.3					
13-Oct	82	83	84	81	82	85	85	85	85	81	69	65	61	53	49	46	44	48	62	68	68	72	71	70	69.9	85.4					
14-Oct	67	70	73	71	77	79	81	83	75	65	59	44	38	35	33	33	34	36	38	42	43	47	56	58	55.7	82.8					
15-Oct	56	56	55	59	65	69	73	55	37	31	29	27	27	26	24	27	31	34	39	48	49	52	51	52	44.6	72.7					
16-Oct	52	54	51	54	50	52	63	70	56	43	31	26	17	15	15	18	20	23	30	32	33	29	35	36	37.7	69.6					
17-Oct	35	34	39	42	49	49	48	47	48	44	34	29	20	20	22	23	25	24	27	29	34	34	37	45	34.8	49.4					
18-Oct	47	41	41	37	38	46	51	49	50	49	45	40	38	35	34	33	38	47	56	63	64	63	69	75	47.8	74.8					
19-Oct	78	80	82	84	83	84	81	77	75	75	67	56	46	38	33	31	29	32	38	40	43	46	59	71	59.6	84.2					
20-Oct	75	76	76	79	63	54	53	52	43	40	39	35	34	32	31	32	32	34	38	43	49	52	52	52	48.6	79.2					
21-Oct	54	57	58	59	66	71	73	74	72	58	51	45	33	27	28	28	30	33	38	43	48	48	52	61	50.4	74.2					
22-Oct	69	71	73	76	80	80	80	82	77	67	59	47	41	37	33	35	36	41	54	63	70	73	76	76	62.4	81.7					
23-Oct	76	78	80	80	82	82	83	83	80	67	56	46	42	34	30	29	32	38	47	53	62	71	77	76	61.9	83.2					
24-Oct	80	78	81	82	83	81	80	78	75	66	55	49	43	41	45	45	46	50	58	67	69	76	79	81	66.2	82.9					
25-Oct	82	83	84	84	85	84	84	82	80	76	62	57	49	39	41	58	70	72	73	67	61	63	64	64	69.4	84.8					
26-Oct	64	64	65	70	76	76	77	79	77	74	71	66	65	67	62	67	61	64	68	73	74	63	64	61	68.6	79.2					
27-Oct	58	54	51	50	51	50	52	52	51	50	46	42	38	37	35	36	40	45	49	54	60	65	70	73	50.4	73.0					
28-Oct	76	68	65	68	66	60	60	60	61	59	58	60	63	75	76	76	74	74	75	73	70	74	78	77	68.4	78.1					
29-Oct	77	80	84	84	85	85	85	85	85	81	63	62	56	50	44	44	46	50	58	65	71	73	70	69	68.9	85.5					
30-Oct	72	78	76	75	71	72	72	70	70	67	61	53	46	45	44	43	45	50	55	56	59	61	62	61	61.0	77.9					
31-Oct	62	62	63	67	69	69	69	68	70	67	64	62	60	68	73	78	83	85	86	87	87	87	86	86	73.3	87.0					
69.1 69.9 70.4 71.3 72.1 72.7 73.8 73.0 70.0 65.1 59.2 54.1 49.7 47.3 46.5 47.9 49.3 52.1 57.1 60.5 63.0 64.8 67.0 68.5																									Diurnal Average						
86.9 88.0 88.7 89.0 89.2 89.3 89.2 87.3 87.1 87.5 87.2 86.7 86.4 85.0 80.1 82.1 83.0 85.4 86.2 86.6 87.0 87.0 86.5 85.7																									Diurnal Maximum						

Hourly Averages

Relative Humidity (RH) - %
Henry Pirker - October 2018



Hourly Averages


Solar Radiation (SR) - W/m²

Henry Pirker - October 2018

Maximum Value: 549.0 W/m ² on Oct 2 13:00																				Maximum Daily Average: 132.9 W/m ² on Oct 2					Hours in Service: 744	
Minimum Value: 0 W/m ² on Oct 1 01:00																				Minimum Daily Average: 15.7 W/m ² on Oct 28					Hours of Data: 744	
Maximum Diurnal Average: 288.0 W/m ² at hour 13																				Minimum Diurnal Average: 0.0 W/m ² at hour 21					Hours of Missing Data: 0	
Monthly Average: 71.04 W/m ²																				Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.1 Q ₃ = 92.0 P ₉₀ = 290.0 P ₉₉ = 469.0					Hours of Calibration: 0	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	42.1	130.6
2-Oct	0	0	0	0	0	0	0	13	57	124	323	485	549	459	471	368	208	120	13	0	0	0	0	0	132.9	549.0
3-Oct	0	0	0	0	0	0	0	16	59	109	238	444	500	493	433	343	173	54	6	0	0	0	0	0	119.5	499.6
4-Oct	0	0	0	0	0	0	0	10	148	257	365	449	463	472	389	302	174	124	10	0	0	0	0	0	131.8	471.5
5-Oct	0	0	0	0	0	0	0	9	80	204	154	400	494	479	467	337	187	106	8	0	0	0	0	0	121.9	494.1
6-Oct	0	0	0	0	0	0	0	3	24	65	165	201	274	217	177	44	21	19	1	0	0	0	0	0	50.5	273.8
7-Oct	0	0	0	0	0	0	0	5	20	35	53	32	48	81	69	36	33	10	1	0	0	0	0	0	17.7	81.4
8-Oct	0	0	0	0	0	0	0	3	25	59	111	222	258	260	151	64	58	17	1	0	0	0	0	0	51.2	260.4
9-Oct	0	0	0	0	0	0	0	2	15	40	91	127	160	176	284	194	123	37	2	0	0	0	0	0	52.1	283.7
10-Oct	0	0	0	0	0	0	0	3	18	36	54	61	107	122	78	42	19	7	1	0	0	0	0	0	22.9	121.6
11-Oct	0	0	0	0	0	0	0	5	82	128	188	239	228	132	107	156	86	42	1	0	0	0	0	0	58.1	238.8
12-Oct	0	0	0	0	0	0	0	1	12	30	52	69	60	54	113	56	28	12	1	0	0	0	0	0	20.4	112.9
13-Oct	0	0	0	0	0	0	0	4	45	150	331	406	394	445	396	286	139	64	1	0	0	0	0	0	110.8	444.6
14-Oct	0	0	0	0	0	0	0	6	81	191	253	180	119	148	165	168	80	17	0	0	0	0	0	0	58.7	253.0
15-Oct	0	0	0	0	0	0	0	3	24	111	101	373	215	239	408	171	132	22	1	0	0	0	0	0	75.0	407.9
16-Oct	0	0	0	0	0	0	0	5	89	194	303	388	428	362	335	148	91	30	1	0	0	0	0	0	99.0	427.9
17-Oct	0	0	0	0	0	0	0	5	27	157	202	269	327	165	98	78	46	30	0	0	0	0	0	0	58.5	327.1
18-Oct	0	0	0	0	0	0	0	3	54	187	294	280	291	330	272	198	59	18	0	0	0	0	0	0	82.8	329.9
19-Oct	0	0	0	0	0	0	0	9	46	92	208	333	398	389	342	225	118	41	0	0	0	0	0	0	91.7	397.9
20-Oct	0	0	0	0	0	0	0	3	60	196	296	368	393	380	326	201	92	21	0	0	0	0	0	0	97.4	393.5
21-Oct	0	0	0	0	0	0	0	3	56	190	292	362	391	378	314	180	78	19	0	0	0	0	0	0	94.3	391.2
22-Oct	0	0	0	0	0	0	0	3	55	181	278	350	379	371	301	205	74	14	0	0	0	0	0	0	92.2	379.2
23-Oct	0	0	0	0	0	0	0	2	47	181	277	347	374	361	304	185	75	15	0	0	0	0	0	0	90.3	373.8
24-Oct	0	0	0	0	0	0	0	1	35	141	243	309	322	271	136	98	66	20	0	0	0	0	0	0	68.4	322.0
25-Oct	0	0	0	0	0	0	0	1	20	110	257	324	352	332	126	13	8	2	0	0	0	0	0	0	64.4	351.6
26-Oct	0	0	0	0	0	0	0	2	17	49	63	100	84	133	48	91	96	10	0	0	0	0	0	0	28.9	132.9
27-Oct	0	0	0	0	0	0	0	2	20	115	254	302	352	345	288	168	62	6	0	0	0	0	0	0	79.7	352.2
28-Oct	0	0	0	0	0	0	0	0	9	28	39	59	43	51	59	49	33	7	0	0	0	0	0	0	15.7	59.5
29-Oct	0	0	0	0	0	0	0	0	17	132	255	310	342	314	311	177	62	4	0	0	0	0	0	0	80.2	342.3
30-Oct	0	0	0	0	0	0	0	0	26	112	167	313	347	342	262	142	48	4	0	0	0	0	0	0	73.5	346.9
31-Oct	0	0	0	0	0	0	0	0	5	41	53	86	105	72	67	39	8	1	0	0	0	0	0	0	19.9	104.6
0.0 0.0 0.0 0.0 0.0 0.0 0.0 4.7 43.1 120.8 196.2 268.1 288.0 274.0 238.7 156.6 83.6 29.4 1.6 0.0 0.0 0.0 0.0 0.0																									Diurnal Average	
0.1 0.2 0.1 0.2 0.1 0.1 0.2 20.2 148.3 256.9 365.3 485.1 549.0 493.2 470.9 367.8 208.4 124.3 12.9 0.0 0.1 0.1 0.1 0.3																									Diurnal Maximum	

Hourly Averages

Solar Radiation (SR) - W/m²
Henry Pirker - October 2018




Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Henry Pirker - October 2018


Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1 Spd	4	4	2	2	1	1	3	2	3	3	2	6	7	4	3	3	2	2	4	5	5	6	7	5	2.8	7.2
Dir	35	34	39	35	23	329	323	318	306	316	316	319	331	337	358	307	212	3	37	40	25	28	29	34	3.0	28.7
2 Spd	5	4	4	6	6	4	2	3	3	4	4	3	5	6	6	8	9	10	10	9	8	10	8	9	3.3	10.1
Dir	29	28	16	5	2	7	18	341	310	320	334	278	274	261	240	235	238	227	212	218	219	230	234	235	258.0	230.1
3 Spd	9	8	9	10	8	9	8	6	6	5	4	4	5	5	6	5	4	3	3	5	4	5	2	2	3.1	10.2
Dir	239	255	254	249	246	246	244	249	238	219	239	189	148	133	161	155	159	156	93	88	103	101	75	67	213.1	249.1
4 Spd	1	2	1	1	2	0	1	2	0	2	3	5	4	8	11	12	11	11	6	4	2	2	3	3	3.6	12.0
Dir	48	292	285	280	289	255	270	296	332	231	232	228	245	288	288	309	309	315	326	325	308	300	277	286	293.5	308.9
5 Spd	2	3	2	2	2	3	3	4	4	4	3	3	2	3	3	5	4	6	5	5	4	2	3	3	3.0	5.7
Dir	343	325	57	75	89	73	64	64	69	70	74	85	45	40	78	67	71	80	81	79	82	70	97	107	70.6	80.1
6 Spd	1	1	0	1	1	1	2	0	1	0	1	3	6	3	3	7	6	7	7	5	4	6	6	6	2.0	7.1
Dir	82	81	277	249	249	125	145	295	231	215	195	305	314	335	51	96	88	85	85	85	65	68	69	65	70.7	84.7
7 Spd	5	4	5	5	5	6	5	4	3	4	4	5	6	5	7	7	7	5	5	5	6	5	5	5	4.9	7.3
Dir	64	65	60	60	61	71	63	61	59	47	44	31	25	22	30	21	28	22	36	32	32	41	47	56	43.2	28.4
8 Spd	5	5	5	6	5	5	5	6	9	8	9	9	9	11	11	12	12	14	11	11	12	11	11	11	8.7	13.5
Dir	57	53	55	59	61	58	57	67	76	79	87	86	79	98	108	92	85	87	87	88	84	82	82	82	81.3	87.1
9 Spd	11	9	8	8	7	8	8	8	8	8	8	10	10	10	11	12	12	13	11	11	10	9	11	8	9.3	12.6
Dir	84	83	78	79	77	76	78	81	82	88	92	95	93	94	95	111	116	111	107	105	107	109	110	111	95.9	111.2
10 Spd	6	6	5	2	3	5	3	2	2	1	3	4	4	5	6	5	2	2	3	2	3	2	1	2	0.6	6.2
Dir	100	100	98	80	67	75	75	69	67	57	323	322	300	293	301	303	284	299	313	254	206	201	153	146	20.5	99.8
11 Spd	3	2	2	1	2	2	4	3	3	5	5	5	10	14	14	16	16	12	9	7	4	4	3	2	4.2	16.3
Dir	153	145	149	144	120	143	149	151	141	155	148	171	230	239	246	255	255	254	266	284	301	286	279	310	236.1	254.8
12 Spd	1	2	4	3	1	1	5	12	11	10	10	11	7	12	14	14	12	11	5	3	3	2	1	3	5.1	14.1
Dir	285	159	151	150	128	9	8	9	340	305	304	315	328	315	310	312	302	297	295	273	278	268	201	220	312.7	311.6
13 Spd	3	3	7	13	8	2	4	2	2	5	10	11	10	10	12	10	8	3	1	5	7	6	6	5	5.7	12.8
Dir	232	246	283	287	285	294	282	288	300	279	293	295	294	300	297	299	291	284	256	233	230	225	217	206	278.2	286.7
14 Spd	4	1	2	3	2	1	2	3	1	2	1	11	14	16	23	25	24	19	20	13	16	15	10	10	8.7	24.8
Dir	222	188	177	118	107	91	119	143	107	140	201	248	252	248	247	244	250	256	264	270	265	272	274	276	252.6	244.3
15 Spd	8	5	5	4	1	0	1	6	21	24	22	24	22	23	28	20	16	12	7	3	4	3	3	5	10.5	27.6
Dir	269	248	275	288	290	347	351	276	251	247	258	267	279	287	291	292	301	300	302	309	282	269	270	259	276.2	291.2
16 Spd	3	5	3	2	4	3	1	1	1	2	3	5	18	18	19	18	19	16	10	10	8	12	8	3	7.4	19.0
Dir	268	240	265	263	290	297	339	144	137	141	174	239	259	258	262	265	260	250	266	281	241	251	254	281	258.2	260.2
17 Spd	8	5	3	1	2	1	3	0	1	3	3	7	12	8	7	6	4	9	7	5	6	6	6	2	4.0	12.2
Dir	250	250	220	122	118	235	174	143	147	207	192	187	242	240	264	270	263	267	275	240	229	262	300	310	245.9	242.2
18 Spd	6	10	11	20	14	8	11	11	17	10	17	21	19	17	16	13	9	6	4	2	2	1	1	1	9.9	20.6
Dir	309	288	281	276	260	285	285	275	262	289	272	277	286	286	285	302	305	316	319	237	235	227	266	288	282.1	277.3
19 Spd	1	2	1	2	2	1	3	1	1	0	3	4	4	8	9	11	9	7	7	7	6	3	3	2	2.6	11.2
Dir	300	312	320	283	272	204	218	236	294	140	127	125	113	115	116	109	109	100	91	101	96	96	301	311	109.4	108.7
20 Spd	2	2	3	0	16	24	21	17	21	23	29	32	34	32	33	28	23	19	16	11	7	6	6	8	17.0	33.8
Dir	300	271	303	231	252	251	243	241	245	249	253	259	247	250	260	257	256	256	249	251	234	232	244	251	251.8	247.0
21 Spd	6	5	8	6	2	1	2	1	1	2	3	3	3	1	1	2	2	2	2	2	3	3	1	0	1.2	7.7
Dir	255	256	272	293	283	246	272	249	250	222	226	241	271	238	164	177	160	83	76	97	99	99	69	50	247.4	271.6
22 Spd	1	2	1	1	0	2	1	1	0	2	2	1	1	2	1	3	2	1	1	2	3	3	2	2	0.5	3.1
Dir	280	287	343	299	328	300	285	160	342	132	155	255	224	235	2	95	123	103	318	319	305	320	313	308	296.2	320.5

Wind Rose

Wind Speed (WS) (km/h)
Henry Pirker - October 2018



Wind Speed Classes (km/h)



Hourly Averages - Wind Speed (Scalar)

Wind Speed (km/h)

Henry Pirker - October 2018

Maximum Speed: 34 km/h on Oct 20 13:00																			Maximum Daily Speed Average: 17.5 km/h on Oct 20										Hours in Service: 744	
Minimum Speed: 0 km/h on Oct 24 22:00																			Minimum Daily Speed Average: 1.7 km/h on Oct 22										Hours of Data: 744	
Maximum Diurnal Speed Average: 10.7 km/h at hour 16																			Minimum Diurnal Speed Average: 4.2 km/h at hour 2										Hours of Missing Data: 0	
Monthly Average Speed: 6.44 km/h																			Percentiles: P ₁ = 0.3 P ₁₀ = 1.2 Q ₁ = 2.2 Median = 4.6 Q ₃ = 8.7 P ₉₀ = 14.1 P ₉₉ = 26.1										Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum				
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24						
1-Oct	4	4	2	2	1	2	3	2	3	3	2	6	8	4	3	3	2	4	5	5	5	6	7	6	3.8	7.5				
2-Oct	5	4	4	6	6	4	2	3	4	4	4	4	6	6	6	8	10	10	10	9	8	10	8	9	6.2	10.2				
3-Oct	9	9	9	10	8	9	9	6	6	5	4	4	5	5	7	6	5	3	3	5	4	5	2	2	5.8	10.2				
4-Oct	1	2	1	2	2	1	2	2	0	2	3	6	5	9	12	12	12	11	6	4	2	3	3	3	4.3	12.2				
5-Oct	2	3	3	2	2	3	3	4	4	4	4	3	3	4	4	6	4	6	5	5	4	2	3	3	3.6	5.9				
6-Oct	1	1	0	2	4	2	2	1	1	0	1	3	6	4	4	7	6	7	7	5	4	6	6	6	3.5	7.2				
7-Oct	5	4	5	5	5	6	5	4	3	4	5	5	6	6	7	7	7	5	5	6	6	5	5	5	5.2	7.5				
8-Oct	5	5	5	6	5	5	5	6	9	8	9	10	12	11	12	12	14	11	11	12	11	11	11	11	9.0	13.6				
9-Oct	11	9	8	8	7	8	8	8	8	8	8	10	11	10	11	12	12	13	11	11	10	9	11	9	9.7	12.7				
10-Oct	6	6	6	2	3	5	3	2	2	2	4	4	4	5	6	5	2	2	3	2	3	2	2	2	3.5	6.3				
11-Oct	3	2	2	1	2	2	4	3	3	5	6	5	10	14	14	16	16	12	9	7	4	4	3	2	6.3	16.5				
12-Oct	3	3	4	3	1	2	5	12	11	10	11	11	7	13	14	14	12	11	5	3	4	2	1	3	6.8	14.3				
13-Oct	3	3	7	13	8	2	5	3	3	5	10	11	10	10	12	10	9	3	1	5	7	6	6	5	6.6	12.8				
14-Oct	4	1	3	3	2	1	2	3	1	2	2	11	15	16	23	25	24	19	20	13	16	15	10	10	10.0	24.9				
15-Oct	8	5	6	5	1	1	1	6	21	24	22	24	22	23	28	20	16	12	7	3	4	3	3	5	11.3	27.8				
16-Oct	3	5	3	3	4	3	1	1	1	2	3	6	18	18	19	18	19	16	10	10	8	12	8	4	8.2	19.1				
17-Oct	8	6	3	2	2	2	3	0	1	3	3	8	12	9	8	6	5	9	7	6	6	6	6	3	5.1	12.4				
18-Oct	6	10	11	21	14	9	11	11	17	11	18	21	19	17	13	9	6	4	2	2	1	1	1	1	10.5	20.8				
19-Oct	1	2	1	2	2	1	3	1	2	1	3	4	5	8	9	11	9	7	7	7	6	3	3	2	4.1	11.3				
20-Oct	2	2	4	2	16	24	21	17	21	24	30	32	34	33	34	28	23	19	16	11	7	6	6	8	17.5	34.0				
21-Oct	6	6	8	6	2	1	3	1	1	2	3	3	3	2	3	2	2	3	3	2	3	3	1	1	2.9	7.7				
22-Oct	1	3	1	1	0	2	1	2	0	2	2	1	1	2	1	3	2	1	1	2	3	3	2	2	1.7	3.2				
23-Oct	3	2	2	1	0	1	1	0	0	1	2	2	3	2	2	3	4	2	3	2	2	4	3	4	2.0	4.0				
24-Oct	2	3	0	1	1	1	1	2	1	1	1	3	4	5	4	5	4	3	1	1	1	0	0	1	1.9	5.4				
25-Oct	1	2	2	0	0	1	1	1	3	2	3	3	4	2	4	12	11	11	10	9	7	7	9	11	4.8	11.8				
26-Oct	10	7	4	3	4	1	3	2	1	1	1	1	3	5	3	3	2	3	4	4	4	10	11	13	4.3	12.8				
27-Oct	13	14	18	21	20	21	21	21	21	20	19	27	23	22	20	15	9	6	8	5	3	1	1	1	14.6	26.9				
28-Oct	1	2	2	2	6	10	13	11	13	10	7	5	6	8	7	6	5	3	5	6	6	2	4	1	5.8	12.8				
29-Oct	0	0	1	1	1	1	2	3	1	0	1	4	9	13	17	16	16	13	3	2	3	1	1	2	4.7	17.3				
30-Oct	1	1	8	7	9	9	12	13	10	10	14	21	23	24	21	23	21	19	11	8	6	6	5	5	12.0	24.2				
31-Oct	4	4	4	1	2	2	2	1	2	2	1	1	2	2	3	4	6	5	6	6	6	6	9	13	3.9	12.6				
4.3 4.2 4.4 4.6 4.5 4.6 5.1 5.0 5.6 5.8 6.6 8.4 9.6 10.1 10.7 10.7 9.6 8.3 6.7 5.7 5.4 5.2 4.9 4.7																									Diurnal Average					
13.5 14.4 18.0 20.8 20.3 23.8 20.9 21.3 21.1 23.9 29.8 32.3 34.0 32.8 33.6 28.0 24.4 19.2 19.8 13.1 15.8 14.7 11.3 12.8																									Diurnal Maximum					
All monthly, daily, and diurnal averages have been calculated using scalar methods																														

Hourly Standard Deviations

Wind Direction (WD) - deg

Henry Pirker - October 2018

Maximum Value: 94.3 deg on Oct 30 02:00																		Hours in Service: 744							
Minimum Value: 3.7 deg on Oct 15 00:00																		Hours of Data: 744							
																		Hours of Missing Data: 0							
Percentiles: P ₁ = 4.3 P ₁₀ = 6.7 Q ₁ = 9.1 Median = 15.3 Q ₃ = 33.5 P ₉₀ = 56.5 P ₉₉ = 84.9																		Hours of Calibration: 0							
																		Percent Operational Time: 100.0							
Day	Hourly Period Ending At (MST)																								Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
1-Oct	14	14	17	20	35	37	6	7	11	22	66	16	21	38	30	26	34	69	19	14	14	14	12	13	69.3
2-Oct	18	18	21	14	14	17	30	23	9	19	28	38	35	30	30	18	16	11	7	8	8	6	11	10	37.9
3-Oct	9	11	10	7	8	8	9	12	14	12	18	28	32	29	27	34	37	19	7	13	7	10	20	15	36.8
4-Oct	77	41	51	47	17	81	50	12	78	29	19	28	33	20	21	11	12	8	6	10	34	63	27	31	81.1
5-Oct	22	9	53	16	18	11	16	15	14	21	33	47	65	47	45	22	25	17	9	7	7	29	20	33	65.5
6-Oct	47	43	74	70	94	61	48	55	40	50	51	38	20	31	28	11	10	8	10	20	17	10	11	11	93.8
7-Oct	11	14	16	13	11	15	18	16	15	14	16	15	13	15	12	16	15	21	14	14	11	13	14	11	20.6
8-Oct	12	10	11	11	10	11	11	11	8	7	9	10	10	12	12	11	12	7	7	9	8	8	7	7	12.4
9-Oct	6	9	9	7	10	9	9	7	9	11	13	12	12	14	14	12	12	7	8	8	9	9	8	9	13.9
10-Oct	10	10	13	28	18	13	23	16	23	46	19	21	21	13	14	21	19	29	59	47	14	21	43	18	58.6
11-Oct	11	8	10	26	12	15	15	9	13	16	11	16	12	11	8	8	6	8	8	10	23	18	12	55	55.3
12-Oct	67	36	14	35	78	67	39	14	13	4	8	12	10	12	9	9	9	6	10	14	8	30	27	16	77.7
13-Oct	14	25	9	4	7	22	22	59	35	8	10	9	11	17	11	10	13	12	43	12	6	6	15	11	59.0
14-Oct	18	56	61	29	38	54	29	13	45	17	53	8	7	9	5	5	6	6	5	7	7	4	4	4	61.0
15-Oct	13	19	19	22	68	80	67	20	10	5	7	7	8	6	6	6	5	6	12	24	22	9	12	7	79.9
16-Oct	14	12	39	27	15	25	54	64	23	17	18	24	9	7	6	10	6	5	6	12	18	22	23	35	63.5
17-Oct	13	43	22	71	42	55	22	37	83	39	20	15	10	16	12	14	21	13	8	17	8	18	23	85	84.7
18-Oct	13	6	6	12	8	23	17	11	5	20	8	8	8	9	14	7	6	7	7	33	60	80	53	52	80.1
19-Oct	60	34	44	35	34	20	15	58	56	84	19	21	21	17	12	8	8	7	10	7	8	35	41	38	84.3
20-Oct	23	31	15	83	46	6	6	9	5	8	9	8	8	8	8	8	6	6	7	7	14	9	14	10	82.5
21-Oct	15	19	5	18	72	52	41	39	40	13	13	28	44	56	74	57	24	22	13	8	7	18	18	74	74.5
22-Oct	81	60	37	48	78	26	93	65	62	34	35	38	63	21	75	42	17	19	38	12	10	10	13	28	93.4
23-Oct	25	31	21	66	54	35	40	72	56	21	43	23	25	45	60	32	16	11	7	13	26	15	19	19	72.3
24-Oct	19	24	72	24	37	68	44	43	62	41	78	44	17	16	12	9	10	39	34	51	43	85	51	66	85.1
25-Oct	55	75	31	74	64	63	40	64	52	76	31	18	24	85	36	33	23	8	11	7	7	8	9	6	85.1
26-Oct	6	14	34	33	43	20	18	21	33	51	37	70	81	55	38	45	15	15	11	17	33	7	9	5	81.1
27-Oct	6	6	4	4	4	4	4	4	4	6	7	7	8	7	7	12	5	17	11	30	64	58	40	32	63.8
28-Oct	45	18	15	31	12	10	5	6	6	8	46	20	22	12	12	11	14	12	8	8	13	64	11	77	77.4
29-Oct	94	86	78	46	34	57	25	22	69	73	33	20	12	9	10	11	7	7	32	19	19	42	24	48	94.2
30-Oct	82	94	14	29	12	6	6	5	7	6	6	8	9	7	7	6	6	6	7	10	11	12	10	14	94.3
31-Oct	7	22	53	89	30	48	20	40	54	45	38	25	61	62	19	13	6	12	8	10	7	8	9	5	89.3
94.2 94.3 78.0 89.3 93.8 81.1 93.4 72.3 82.9 84.3 78.2 70.3 81.1 85.1 75.5 56.7 36.8 69.3 58.6 51.1 63.8 85.1 53.2 84.7																									

PAZA
Smoky Heights Station
Monthly Summary Tables, Graphs and
Roses


Hourly Averages

Sulphur Dioxide (SO₂) - ppb Smoky Heights - October 2018

Number of Exceedences (AAAQO): Maximum Value: 3.4 ppb on Oct 3 00:00																			1-hr: 0 24-hr: 0 Maximum Daily Average: 0.9 ppb on Oct 16					Hours in Service: 744 Hours of Data: 709 Hours of Missing Data: 35 Hours of Calibration: 35 Percent Operational Time: 100.0				
Minimum Value: 0 ppb on Oct 1 18:00 Maximum Diurnal Average: 0.3 ppb at hour 23 Monthly Average: 0.25 ppb																			Minimum Daily Average: 0.1 ppb on Oct 8 Minimum Diurnal Average: 0.2 ppb at hour 20 Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.1 Median = 0.1 Q ₃ = 0.3 P ₉₀ = 0.6 P ₉₉ = 2.1									
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum		
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24				
1-Oct	0	0	0	0	1	0	0	1	1	1	1	A	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.8		
2-Oct	0	0	0	0	0	0	0	1	1	0	A	0	0	0	0	0	0	0	0	0	0	0	0	3	0.2	3.4		
3-Oct	3	1	1	1	1	1	1	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.5	3.1		
4-Oct	0	0	0	0	0	0	0	0	A	0	0	1	1	0	1	1	0	0	0	0	0	0	1	0	0.4	1.1		
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0.1	0.7		
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1		
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1		
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0.2	0.7		
10-Oct	1	1	A	1	1	1	0	0	0	0	0	0	0	0	0	1	0	1	0	0	0	0	0	0	0.4	0.8		
11-Oct	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	0	0	0	1	0.3	0.8		
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0.1	0.4		
13-Oct	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	A	0	0.1	0.7		
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	1	0	1	0	1	A	0	0	0.3	1.0		
15-Oct	0	0	1	1	1	1	1	0	0	1	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0.3	1.3		
16-Oct	0	0	0	0	0	0	0	0	1	1	2	1	1	2	0	0	3	0	0	A	1	1	3	1	0.9	2.9		
17-Oct	2	1	2	0	0	0	0	0	0	0	0	0	0	0	2	0	0	0	A	1	1	0	0	0	0.6	2.3		
18-Oct	0	1	0	0	0	1	0	2	0	1	0	0	0	0	0	0	0	A	0	0	0	0	1	1	0.4	1.5		
19-Oct	1	1	1	1	1	1	1	1	1	C	C	C	0	0	0	0	A	0	0	0	0	0	0	0	0.5	1.1		
20-Oct	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	A	0	0	0	0	0	2	2	0	0.3	2.5		
21-Oct	0	2	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0.2	1.6		
22-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0.2	0.5		
23-Oct	0	0	0	0	0	0	0	0	0	0	1	0	A	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5		
24-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3		
25-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3		
26-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.4		
27-Oct	0	0	0	0	0	0	0	1	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.9		
28-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0.1	0.6		
29-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0.2	0.7		
30-Oct	0	0	0	0	1	A	0	1	1	1	1	0	0	0	0	1	0	0	0	0	0	0	0	1	0.3	1.0		
31-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3		
0.3 0.3 0.3 0.2 0.3 0.2 0.2 0.3 0.3 0.3 0.3 0.3 0.3 0.3 0.3 0.2 0.3 0.2 0.2 0.2 0.2 0.2 0.3 0.3 0.3																									Diurnal Average			
3.1 1.6 1.6 0.9 1.3 0.9 0.9 1.5 1.1 0.8 1.7 1.3 1.2 2.1 2.3 0.7 2.7 0.8 0.8 0.8 0.8 1.2 2.5 2.9 3.4																									Diurnal Maximum			
C - Calibration A - Automated Daily Zero Span																												
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 172 ppb 24-hr 48 ppb 30-day 11 ppb																												

Hourly Averages

Sulphur Dioxide (SO₂) - ppb
Smoky Heights - October 2018




Hourly Maximums

Sulphur Dioxide (SO₂) - ppb Smoky Heights - October 2018

Maximum Value: 7.5 ppb on Oct 3 01:00																			Maximum Daily Average: 2.0 ppb on Oct 16										Hours in Service: 744	
Minimum Value: 0 ppb on Oct 27 02:00																			Minimum Daily Average: 0.5 ppb on Oct 25										Hours of Data: 709	
Maximum Diurnal Average: 1.1 ppb at hour 24																			Minimum Diurnal Average: 0.6 ppb at hour 20										Hours of Missing Data: 35	
Monthly Average: 0.82 ppb																			Percentiles: P ₁ = 0.2 P ₁₀ = 0.4 Q ₁ = 0.4 Median = 0.6 Q ₃ = 0.8 P ₉₀ = 1.4 P ₉₉ = 4.9										Hours of Calibration: 35	
																													Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum				
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24						
1-Oct	0	0	1	1	1	1	1	1	1	2	1	A	1	0	1	0	1	0	1	0	0	1	0	0	0.7	1.5				
2-Oct	0	0	0	0	0	0	1	4	4	0	A	1	0	0	0	1	1	0	0	0	0	0	0	7	1.1	7.2				
3-Oct	8	2	2	3	2	2	1	1	1	A	1	1	1	1	1	1	1	1	1	1	0	1	1	0	1.4	7.5				
4-Oct	0	0	1	1	1	1	1	0	A	1	1	2	1	1	1	2	1	1	1	1	1	1	3	1	1.0	2.8				
5-Oct	0	0	0	0	1	0	0	A	0	1	1	1	0	0	1	1	1	1	1	1	1	1	1	0	0.6	1.2				
6-Oct	0	1	1	0	0	0	A	1	1	1	1	1	1	1	1	0	1	0	1	1	1	0	0	0	0.6	1.0				
7-Oct	1	0	1	0	0	A	0	0	0	1	0	0	1	1	0	1	1	0	0	1	1	1	1	0	0.5	0.7				
8-Oct	1	0	1	1	A	1	1	0	1	1	0	0	1	1	1	1	0	0	0	0	1	0	0	1	0.5	0.7				
9-Oct	0	0	0	A	1	0	0	1	1	1	0	0	1	1	1	0	1	1	1	1	1	1	1	1	0.6	1.1				
10-Oct	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.4				
11-Oct	1	A	1	1	1	1	1	0	1	0	1	1	1	1	1	1	1	2	1	1	1	1	1	3	0.9	2.7				
12-Oct	A	1	1	1	1	1	1	1	1	0	1	1	0	0	1	0	0	1	1	1	1	1	1	A	0.6	1.0				
13-Oct	0	0	0	0	1	1	1	0	1	1	1	2	1	1	0	0	1	0	1	0	0	0	A	1	0.6	1.7				
14-Oct	1	0	0	0	1	1	1	1	1	1	1	1	1	2	2	1	1	1	2	1	2	A	1	1	1.0	2.3				
15-Oct	1	1	1	3	3	1	1	1	1	1	1	1	1	1	0	0	1	0	0	0	A	1	1	1	1.0	3.3				
16-Oct	0	1	1	1	1	1	1	1	1	1	4	3	3	6	1	0	5	1	1	A	2	2	5	5	2.0	5.6				
17-Oct	4	3	5	1	1	1	1	1	1	1	1	1	1	2	5	1	0	2	A	2	2	1	1	1	1.5	5.0				
18-Oct	1	2	1	1	1	1	1	6	1	1	1	1	1	1	0	1	1	A	1	1	1	1	1	1	1.1	6.0				
19-Oct	1	1	1	1	1	1	1	2	2	C	C	C	0	1	1	1	A	0	0	1	1	1	1	1	0.9	1.5				
20-Oct	0	0	1	1	1	1	0	0	1	3	1	1	1	0	0	A	0	0	2	0	0	4	7	1	1.2	7.5				
21-Oct	1	4	1	0	0	0	1	1	1	1	1	1	1	1	A	0	0	0	0	0	0	0	0	0	0.7	3.5				
22-Oct	0	0	0	0	0	0	0	0	0	0	1	1	1	A	1	1	1	1	0	0	0	1	1	1	0.5	1.0				
23-Oct	1	1	1	1	0	0	0	0	0	1	2	1	A	1	1	1	0	1	0	1	1	1	1	1	0.7	1.5				
24-Oct	0	1	0	1	0	0	0	0	0	0	0	A	1	1	1	1	1	1	0	0	0	0	0	0	0.5	1.4				
25-Oct	0	0	0	0	0	0	0	0	0	1	A	1	1	0	0	0	1	0	0	0	0	0	0	0	0.5	0.8				
26-Oct	0	0	0	0	0	0	1	0	0	A	0	0	1	4	1	1	0	1	0	0	0	0	0	0	0.6	3.6				
27-Oct	0	0	0	0	0	0	0	2	A	1	0	0	0	0	0	1	0	0	0	0	1	0	0	0	0.5	2.2				
28-Oct	1	0	1	0	0	0	0	A	0	0	0	0	1	1	1	1	2	2	1	1	1	0	0	1	0.6	1.6				
29-Oct	0	0	0	0	0	0	A	0	0	1	1	1	1	1	1	2	0	0	1	1	1	0	0	0	0.6	1.7				
30-Oct	1	0	1	0	3	A	1	2	1	2	4	1	1	1	1	2	1	0	0	0	0	0	0	2	1.1	3.8				
31-Oct	0	0	0	0	A	0	0	0	1	1	1	1	1	1	1	1	0	0	0	0	0	0	1	0	0.5	0.8				
0.9		0.8	0.9	0.7	0.8	0.7	0.7	1.0	0.8	0.8	0.9	0.8	0.7	1.0	0.8	0.8	0.8	0.7	0.6	0.6	0.7	0.7	1.1	1.1	Diurnal Average					
7.5		3.5	5.0	3.0	3.3	2.0	1.5	6.0	4.0	3.2	4.3	2.7	2.5	5.6	4.8	2.2	4.7	1.7	1.7	2.1	2.1	3.8	7.5	7.2	Diurnal Maximum					
C - Calibration		A - Automated Daily Zero Span																												


Hourly Maximums

Sulphur Dioxide (SO₂) - ppb
Smoky Heights - October 2018




Pollutant Rose

Sulphur Dioxide (SO₂) - ppb
Smoky Heights - October 2018




Pollutant Classes (ppb)



Span Responses

Sulphur Dioxide (SO₂)
Smoky Heights - October 2018



Hourly Averages


Total Reduced Sulphur (TRS) - ppb

Smoky Heights - October 2018

Maximum Value: 2.3 ppb on Oct 13 19:00																			Maximum Daily Average: 0.9 ppb on Oct 29										Hours in Service: 744	
Minimum Value: 0 ppb on Oct 1 13:00																			Minimum Daily Average: 0.3 ppb on Oct 1										Hours of Data: 709	
Maximum Diurnal Average: 0.7 ppb at hour 21																			Minimum Diurnal Average: 0.4 ppb at hour 13										Hours of Missing Data: 35	
Monthly Average: 0.53 ppb																			Percentiles: P ₁ = 0.2 P ₁₀ = 0.3 Q ₁ = 0.3 Median = 0.4 Q ₃ = 0.6 P ₉₀ = 0.9 P ₉₉ = 1.6										Hours of Calibration: 35	
																													Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum				
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24						
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.4				
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.4				
3-Oct	0	0	0	0	0	1	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.5				
4-Oct	1	1	1	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.6				
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.4				
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0.4	0.5				
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.4				
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0.3	0.6				
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.4				
10-Oct	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0.4	0.6				
11-Oct	0	A	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.7				
12-Oct	A	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	1	A	0.5	1.3				
13-Oct	1	0	0	0	1	1	1	1	1	1	1	0	0	0	0	0	1	1	2	2	1	1	A	1	0.8	2.3				
14-Oct	1	1	1	1	1	1	1	1	1	1	0	1	0	1	1	0	0	0	0	1	1	A	0	0	0.6	1.3				
15-Oct	1	1	1	0	1	1	1	1	1	1	0	0	0	1	1	1	1	0	1	0	A	1	1	1	0.6	0.9				
16-Oct	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0	0	1	1	1	A	1	1	1	1	0.6	1.3				
17-Oct	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	1	1	1	A	1	1	1	0	0	0.5	0.6				
18-Oct	0	0	1	0	0	0	1	1	1	1	0	0	0	0	0	0	0	A	1	0	0	0	1	1	0.5	0.8				
19-Oct	1	1	1	1	1	1	1	1	1	1	1	C	C	C	0	0	A	0	0	0	0	0	0	0	0.5	0.9				
20-Oct	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	1	0	0.4	1.0				
21-Oct	0	0	1	1	1	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0.4	0.6				
22-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	1	1	1	2	1	1	1	0.5	1.6				
23-Oct	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	0.7	1.0				
24-Oct	1	1	1	1	1	1	1	1	1	1	1	A	1	1	0	0	1	1	1	1	2	1	1	1	0.8	1.6				
25-Oct	1	1	1	1	1	1	1	1	1	1	A	1	1	1	0	0	1	1	1	1	0	0	1	1	0.8	1.4				
26-Oct	1	1	1	1	1	1	1	1	1	A	1	0	1	0	0	0	1	1	0	1	1	1	0	0	0.6	1.2				
27-Oct	0	0	0	0	1	0	1	0	A	0	0	0	0	0	0	0	0	1	1	1	1	1	1	1	0.6	1.2				
28-Oct	2	2	1	1	0	0	0	A	0	0	0	0	0	0	0	0	1	1	1	1	2	2	2	1	0.8	2.0				
29-Oct	1	2	1	1	2	1	A	2	1	1	0	1	1	0	0	0	0	1	1	1	1	1	1	1	0.9	1.5				
30-Oct	1	1	0	0	1	A	1	1	1	1	0	1	0	0	0	0	0	0	0	0	0	0	0	1	0.5	1.0				
31-Oct	1	1	1	1	A	1	1	1	1	1	1	1	0	0	1	1	1	1	1	1	1	1	1	1	0.8	1.5				
																								Diurnal Average						
																								Diurnal Maximum						
C - Calibration																								A - Automated Daily Zero Span						

Hourly Averages

Total Reduced Sulphur (TRS) - ppb
Smoky Heights - October 2018



Hourly Maximums


Total Reduced Sulphur (TRS) - ppb

Smoky Heights - October 2018

Maximum Value: 3.9 ppb on Oct 13 19:00																				Maximum Daily Average: 1.7 ppb on Oct 28										Hours in Service: 744									
Minimum Value: 1 ppb on Oct 1 13:00																				Minimum Daily Average: 0.7 ppb on Oct 1										Hours of Data: 709									
Maximum Diurnal Average: 1.3 ppb at hour 21																				Minimum Diurnal Average: 0.8 ppb at hour 15										Hours of Missing Data: 35									
Monthly Average: 1.07 ppb																				Percentiles: P ₁ = 0.6 P ₁₀ = 0.7 Q ₁ = 0.8 Median = 0.9 Q ₃ = 1.2 P ₉₀ = 1.7 P ₉₉ = 3.0										Hours of Calibration: 35									
																				Percent Operational Time: 100.0																			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum													
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24															
1-Oct	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	0.7	1.2													
2-Oct	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.0													
3-Oct	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.0													
4-Oct	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	0.9	2.2													
5-Oct	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.0													
6-Oct	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.2													
7-Oct	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.2													
8-Oct	1	1	1	1	A	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	2	1	0.9	2.1													
9-Oct	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.1													
10-Oct	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.1													
11-Oct	1	A	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.5													
12-Oct	A	2	1	4	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1	1	A	1.2	3.6													
13-Oct	2	1	1	1	1	1	1	1	2	2	1	1	1	1	1	1	1	2	4	3	2	2	A	2	1.5	3.9													
14-Oct	2	1	2	1	2	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1.1	2.0													
15-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	2	1	1	1.1	1.5													
16-Oct	2	1	1	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1.1	1.9													
17-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	0.9	1.2													
18-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1.0	1.3													
19-Oct	1	1	1	1	1	1	1	1	1	1	1	C	C	C	1	1	A	1	1	1	1	1	1	1	1.0	1.4													
20-Oct	1	1	1	2	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	0.9	1.7													
21-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	0.9	1.2													
22-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	2	2	2	2	1	1	1.1	2.3													
23-Oct	1	1	1	1	1	2	1	1	1	2	1	1	A	1	1	1	1	3	1	1	1	1	1	1	1.3	2.7													
24-Oct	1	2	2	2	1	2	2	1	1	1	1	A	1	1	1	1	1	1	1	2	3	2	1	1	1.5	2.6													
25-Oct	2	1	2	2	1	2	2	2	2	2	A	1	1	1	1	1	1	1	1	1	1	1	2	1	1.3	1.9													
26-Oct	2	1	1	1	1	1	1	2	2	A	1	1	1	1	1	1	1	1	1	2	2	1	1	1	1.3	2.5													
27-Oct	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	2	2	2	2	2	2	1.2	2.1													
28-Oct	4	3	2	1	1	1	1	A	1	1	1	1	1	1	1	1	1	2	1	2	3	3	4	3	1.7	3.5													
29-Oct	2	2	2	2	2	2	A	2	3	1	1	1	1	1	1	1	1	1	1	1	2	3	2	1	1.6	2.7													
30-Oct	2	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1.0	2.1													
31-Oct	1	1	1	2	A	1	1	2	2	1	1	1	1	1	1	1	2	1	2	3	3	2	1	1	1.5	3.1													
		1.2	1.2	1.1	1.2	1.1	1.1	1.1	1.1	1.0	0.9	0.9	0.8	0.8	0.8	0.9	0.9	1.1	1.1	1.2	1.3	1.2	1.2	1.1	Diurnal Average														
		3.5	3.2	2.4	3.6	2.4	2.0	1.9	2.2	2.7	1.8	1.5	1.4	1.5	1.1	1.2	1.2	1.7	2.7	3.9	3.1	2.9	3.3	3.0	Diurnal Maximum														
C - Calibration		A - Automated Daily Zero Span																																					


Hourly Maximums

Total Reduced Sulphur (TRS) - ppb
Smoky Heights - October 2018




Pollutant Rose

Total Reduced Sulphur (TRS) - ppb
Smoky Heights - October 2018



Span Responses

Total Reduced Sulphur (TRS)
Smoky Heights - October 2018



Hourly Averages


Particulate Matter 2.5 (PM_{2.5}) - µg/m³

Smoky Heights - October 2018

Number of Exceedences: 1-hr: 0 24-hr: 0 Maximum Value: 44.5 µg/m³ on Oct 22 20:00 Maximum Daily Average: 11.3 µg/m³ on Oct 24																								Hours in Service: 744 Hours of Data: 744 Hours of Missing Data: 0 Hours of Calibration: 0 Percent Operational Time: 100.0			
Minimum Value: 0 µg/m³ on Oct 2 07:00 Maximum Diurnal Average: 6.9 µg/m³ at hour 20 Monthly Average: 4.79 µg/m³												Minimum Daily Average: 0.2 µg/m³ on Oct 2 Minimum Diurnal Average: 3.3 µg/m³ at hour 14 Percentiles: P ₁ = 0.1 P ₁₀ = 0.9 Q ₁ = 1.9 Median = 3.9 Q ₃ = 6.2 P ₉₀ = 9.1 P ₉₉ = 25.4															
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	2	1	1	2	2	3	3	4	3	1	1	1	1	3	1	1	1	0	0	0	0	0	0	0	0	1.3	4.1
2-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	1	1	1	0.2	1.0
3-Oct	1	0	0	0	0	0	1	1	1	0	1	1	1	0	0	1	0	0	3	1	1	5	2	1	1	0.9	4.6
4-Oct	1	1	1	1	1	1	2	3	2	3	10	16	10	4	6	4	4	5	6	4	4	3	3	5	4.1	15.7	
5-Oct	5	3	2	2	2	1	1	1	2	2	2	3	3	3	3	4	4	5	8	8	7	6	6	6	3.7	8.3	
6-Oct	4	7	4	4	5	5	5	5	5	5	5	5	5	5	5	5	5	4	5	5	4	5	5	5	4.9	7.4	
7-Oct	6	7	7	7	7	6	6	4	4	3	3	2	3	4	4	3	2	3	4	3	4	3	3	3	4.1	7.2	
8-Oct	3	2	2	2	1	1	1	1	1	1	1	1	1	3	2	7	2	1	1	14	4	1	7	24	3.5	24.4	
9-Oct	32	9	12	13	14	2	3	5	5	5	11	4	6	5	4	4	4	3	3	3	3	4	5	6	6.9	32.2	
10-Oct	6	7	7	7	7	7	7	7	7	6	6	6	6	4	5	5	6	7	9	8	8	7	7	7	6.6	8.9	
11-Oct	7	8	9	9	12	12	12	7	7	6	5	5	5	4	3	1	1	1	1	1	1	1	2	1	5.0	12.0	
12-Oct	1	1	2	3	3	3	4	1	0	1	2	1	1	3	3	2	1	2	2	1	1	1	1	2	1.7	4.0	
13-Oct	1	1	1	2	5	3	1	1	1	1	1	1	1	1	1	1	1	2	1	3	1	1	2	2	1.5	4.8	
14-Oct	1	1	2	2	3	3	2	2	2	2	1	1	1	1	1	1	1	1	1	1	2	2	2	2	1.6	2.9	
15-Oct	2	3	3	3	4	4	4	4	5	5	5	3	3	5	3	4	5	8	6	6	5	5	6	8	4.6	8.1	
16-Oct	8	9	9	8	7	6	5	5	6	5	4	3	4	4	7	11	17	11	7	7	7	7	7	6	7.1	17.2	
17-Oct	5	5	4	6	12	10	6	6	7	7	5	5	6	5	5	11	7	9	7	8	13	9	8	12	7.4	12.6	
18-Oct	12	8	10	13	12	13	10	9	8	31	2	2	2	1	2	4	6	3	3	4	2	2	5	6	7.2	31.1	
19-Oct	4	4	5	4	3	5	6	4	5	7	6	4	5	4	4	3	2	3	4	4	6	5	8	6	4.7	8.1	
20-Oct	5	5	5	7	6	4	4	2	2	3	3	3	3	3	4	2	3	4	5	7	7	7	10	6	4.7	10.2	
21-Oct	4	4	4	4	4	3	3	4	4	3	2	2	2	1	1	1	8	23	5	5	5	9	5	3	4.4	23.3	
22-Oct	4	2	2	3	3	3	3	5	3	3	3	3	4	2	2	3	3	29	42	45	9	6	7	5	8.0	44.5	
23-Oct	6	5	5	5	6	6	7	9	11	10	10	10	8	6	7	6	10	14	9	7	6	7	7	6	7.5	13.8	
24-Oct	6	6	6	5	6	7	9	10	11	11	10	10	9	8	10	7	7	17	29	27	18	15	14	14	11.3	28.7	
25-Oct	12	13	13	13	13	14	14	13	12	15	13	11	7	4	3	2	2	2	2	2	2	2	2	2	7.8	14.9	
26-Oct	2	3	3	4	4	4	4	4	5	7	6	4	5	4	3	3	5	9	7	7	5	7	3	2	4.6	9.2	
27-Oct	2	3	3	3	3	3	4	4	3	3	4	4	3	2	3	2	4	7	6	8	7	7	7	7	4.3	7.8	
28-Oct	8	7	6	5	5	4	4	3	3	4	4	4	3	3	4	4	5	5	5	7	7	6	5	5	4.8	7.7	
29-Oct	4	3	4	4	4	5	5	6	7	7	6	5	4	2	2	1	2	3	3	8	5	5	5	5	4.4	8.1	
30-Oct	4	3	2	1	1	2	2	2	2	3	3	3	3	2	1	1	4	4	5	4	3	3	3	3	2.7	4.7	
31-Oct	3	2	2	2	3	3	3	5	4	7	6	9	7	4	4	5	7	10	7	6	16	18	26	8	7.0	26.1	
5.2 4.3 4.4 4.6 5.0 4.6 4.5 4.4 4.4 5.4 4.5 4.2 3.9 3.3 3.3 3.5 4.2 6.4 6.3 6.9 5.3 5.2 5.6 5.4																								Diurnal Average			
32.2 12.5 12.8 13.4 14.4 13.6 14.2 13.2 12.4 31.1 13.3 15.7 10.1 8.0 9.7 11.2 17.2 29.2 42.1 44.5 18.0 18.2 26.1 24.4																								Diurnal Maximum			
Alberta Ambient Air Quality Guideline (AAAQG): 1-hr 80 µg/m³ Alberta Ambient Air Quality Objective (AAAQO): 24-hr 30 µg/m³																											

Hourly Averages

Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Smoky Heights - October 2018



Hourly Maximums

Particulate Matter 2.5 (PM_{2.5}) - µg/m³


Smoky Heights - October 2018

Maximum Value: 346.3 $\mu\text{g}/\text{m}^3$ on Oct 22 19:00																								Maximum Daily Average: 44.9 $\mu\text{g}/\text{m}^3$ on Oct 22										Hours in Service: 744	
Minimum Value: 0 $\mu\text{g}/\text{m}^3$ on Oct 2 10:00																								Minimum Daily Average: 1.2 $\mu\text{g}/\text{m}^3$ on Oct 2										Hours of Data: 744	
Maximum Diurnal Average: 29.2 $\mu\text{g}/\text{m}^3$ at hour 18																								Minimum Diurnal Average: 5.6 $\mu\text{g}/\text{m}^3$ at hour 4										Hours of Missing Data: 0	
Monthly Average: 10.59 $\mu\text{g}/\text{m}^3$																								Percentiles: $P_1 = 0.5$ $P_{10} = 1.4$ $Q_1 = 2.9$ Median = 5.8 $Q_3 = 10.3$ $P_{90} = 20.2$ $P_{99} = 91.8$										Hours of Calibration: 0	
																								Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum									
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24											
1-Oct	3	2	2	2	4	4	17	11	11	6	1	2	3	25	10	1	3	0	0	0	0	0	0	0	4.6	24.9									
2-Oct	0	0	0	0	0	0	0	0	0	0	0	2	1	2	5	1	7	0	2	2	1	0	1	1	1.2	6.6									
3-Oct	1	1	0	0	1	1	1	4	1	1	1	1	1	1	1	3	1	1	6	3	1	27	4	5	2.7	27.1									
4-Oct	1	1	1	1	5	1	5	10	3	6	17	21	15	7	13	21	20	10	12	9	13	5	4	8	8.8	21.4									
5-Oct	6	4	2	2	2	2	1	3	2	5	3	3	3	4	6	10	10	13	23	42	10	29	15	17	9.0	42.5									
6-Oct	5	45	5	11	15	5	6	5	8	6	5	5	5	7	5	5	6	5	6	5	5	6	5	6	7.8	45.2									
7-Oct	7	9	7	7	8	7	6	5	5	3	3	3	4	4	4	3	3	4	4	4	4	3	3	3	4.7	9.5									
8-Oct	3	3	2	2	1	1	1	2	1	1	1	1	2	31	10	39	20	4	1	49	19	1	43	54	12.3	54.4									
9-Oct	46	38	25	20	24	6	8	7	5	10	20	8	10	8	8	11	9	4	4	3	4	5	6	7	12.3	45.9									
10-Oct	7	7	8	8	8	7	12	9	9	7	6	6	6	6	6	7	8	11	17	9	9	8	9	7	8.1	17.1									
11-Oct	9	9	10	10	13	13	15	8	16	7	6	6	7	6	3	2	1	2	1	1	3	1	3	2	6.4	16.1									
12-Oct	1	1	3	3	3	3	7	3	0	2	2	1	2	3	3	2	1	2	2	1	1	1	2	2	2.3	6.8									
13-Oct	2	1	2	3	6	4	2	1	1	1	1	1	2	2	1	1	1	4	2	8	1	2	2	2	2.2	7.6									
14-Oct	2	2	2	2	4	3	3	2	2	2	2	1	1	1	1	1	1	1	2	1	2	2	2	2	1.9	3.6									
15-Oct	3	4	3	4	4	4	4	4	6	5	5	4	5	9	5	8	9	18	8	14	9	6	8	9	6.6	18.0									
16-Oct	10	10	9	8	7	7	6	7	6	6	7	6	4	9	11	17	30	18	14	20	7	8	8	8	10.2	30.5									
17-Oct	5	5	5	7	18	21	16	23	11	9	11	6	20	7	9	57	13	30	8	15	18	17	9	13	14.7	57.4									
18-Oct	19	9	13	13	12	42	11	15	11	99	34	8	16	8	12	30	43	22	14	29	20	6	10	27	21.8	99.4									
19-Oct	9	5	7	9	4	5	8	6	7	8	7	5	7	5	4	4	3	4	5	6	6	6	24	19	7.2	24.2									
20-Oct	14	6	6	8	7	5	6	3	4	8	9	12	24	19	13	9	19	37	20	25	18	22	34	29	14.8	36.9									
21-Oct	4	5	4	4	4	4	9	9	10	14	3	2	10	1	1	2	82	226	21	11	11	105	35	3	24.3	225.7									
22-Oct	42	2	3	3	8	4	20	19	4	6	4	6	27	3	2	20	18	262	346	222	29	9	11	7	44.9	346.3									
23-Oct	10	5	5	7	9	20	9	12	14	10	11	10	9	7	11	6	44	38	15	9	7	14	24	6	13.0	44.0									
24-Oct	6	6	8	6	8	9	9	19	16	12	10	11	10	9	29	13	20	74	109	115	30	16	15	20	24.3	115.3									
25-Oct	14	13	16	14	13	14	16	17	14	16	15	13	10	5	4	4	3	3	2	2	2	2	2	2	9.0	16.6									
26-Oct	4	3	4	4	4	4	4	4	7	7	7	4	6	6	4	10	32	42	11	9	8	8	4	2	8.4	42.0									
27-Oct	3	3	3	4	4	4	4	5	4	4	5	9	9	12	33	9	35	23	16	23	16	7	7	8	10.3	34.6									
28-Oct	11	8	9	5	5	5	4	3	4	4	6	6	4	5	8	9	9	6	7	14	15	11	7	8	7.2	14.6									
29-Oct	6	4	5	4	5	8	5	14	10	9	8	6	6	2	14	4	6	8	4	26	8	9	6	7	7.6	26.1									
30-Oct	6	4	3	1	2	2	2	2	3	4	5	4	3	2	2	2	10	5	6	9	5	4	3	4	3.9	10.2									
31-Oct	3	3	3	3	4	5	6	31	7	10	8	13	12	4	6	6	10	31	19	7	61	66	50	17	16.0	65.9									
8.4 7.1 5.6 5.6 6.8 7.1 7.3 8.5 6.5 9.3 7.2 6.1 7.9 7.1 7.9 10.3 15.3 29.2 22.9 22.4 11.1 13.2 11.5 9.8																									Diurnal Average										
45.9 45.2 25.1 19.5 23.9 42.0 20.0 30.5 16.3 99.4 34.2 21.0 27.3 31.2 33.2 57.4 81.8 262.3 346.3 221.8 60.8 105.0 50.1 54.4																									Diurnal Maximum										

Hourly Maximums


Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$

Smoky Heights - October 2018



Pollutant Rose

Particulate Matter 2.5 (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Smoky Heights - October 2018



Hourly Averages

External Temperature (ET) - °C


Smoky Heights - October 2018

Maximum Value: 20.5 °C on Oct 16 16:00																				Maximum Daily Average: 11.7 °C on Oct 17					Hours in Service: 744				
Minimum Value: -6 °C on Oct 23 06:00																				Minimum Daily Average: -2.0 °C on Oct 2					Hours of Data: 744				
Maximum Diurnal Average: 8.3 °C at hour 15																				Minimum Diurnal Average: -1.0 °C at hour 8					Hours of Missing Data: 0				
Monthly Average: 2.85 °C																				Percentiles: P ₁ = -5.5 P ₁₀ = -3.0 Q ₁ = -1.2 Median = 1.3 Q ₃ = 6.2 P ₉₀ = 11.1 P ₉₉ = 16.6					Hours of Calibration: 0				
																				Percent Operational Time: 100.0									
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum			
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24					
1-Oct	0	-1	-1	-1	-1	-1	-1	-1	-1	0	0	1	2	2	2	2	0	-1	-1	-1	-2	-2	-2	-3	-0.5	1.9			
2-Oct	-3	-3	-3	-4	-4	-4	-4	-5	-3	-2	0	0	1	1	2	2	2	1	-1	-3	-5	-4	-5	-5	-2.0	2.3			
3-Oct	-5	-4	-4	-5	-6	-5	-5	-4	-3	-1	0	3	4	5	6	7	7	5	2	1	1	-1	-1	-1	-0.2	7.1			
4-Oct	-3	-3	-2	-3	-4	-5	-5	-6	-4	0	3	6	8	9	9	10	9	8	5	3	3	3	1	-1	1.8	9.7			
5-Oct	-2	-3	-4	-4	-5	-5	-6	-5	-4	-3	-2	0	1	2	3	4	3	2	1	0	-2	-2	-3	-3	-1.5	3.6			
6-Oct	-5	-3	-3	-2	-1	-1	1	0	0	1	2	2	3	4	4	5	4	4	3	2	2	1	0	0	1.0	4.7			
7-Oct	0	0	0	0	0	0	-1	-1	-1	0	0	-1	-1	-1	-1	-1	-1	-1	-1	-1	-2	-2	-2	-2	-0.8	0.0			
8-Oct	-2	-2	-2	-2	-2	-2	-2	-2	-2	-2	-1	0	0	1	1	1	1	0	0	-1	-1	-1	-1	-2	-1.0	1.3			
9-Oct	-2	-2	-2	-3	-3	-3	-3	-3	-3	-3	-2	-1	0	0	1	2	2	1	0	-1	-1	-1	-1	-2	-1.2	1.8			
10-Oct	-2	-2	-2	-2	-2	-2	-2	-2	-1	-1	0	0	1	1	1	0	0	0	-1	-1	-1	-1	-1	-2	-0.9	0.9			
11-Oct	-2	-3	-3	-4	-6	-5	-4	-3	-3	-1	0	3	6	7	7	7	8	9	8	6	6	5	5	5	2.0	8.6			
12-Oct	4	4	4	4	4	4	3	0	0	0	0	0	0	0	1	1	1	0	0	0	0	-1	0	0	1.2	4.3			
13-Oct	0	0	0	0	0	1	1	1	1	2	4	4	5	6	7	8	9	7	2	2	0	0	-1	-2	2.4	8.6			
14-Oct	-2	-2	-3	-3	-3	-3	-3	-2	-1	2	6	9	11	11	12	13	12	11	11	10	9	10	10	10	5.1	13.1			
15-Oct	10	10	11	10	8	7	6	7	9	12	14	15	16	16	16	16	15	14	12	12	13	11	7	5	11.4	16.1			
16-Oct	4	4	4	3	2	3	3	2	4	9	12	14	16	18	20	20	18	14	12	12	12	11	10	9	9.9	20.5			
17-Oct	11	9	10	6	5	6	7	5	6	9	13	16	17	19	19	17	15	15	15	13	11	10	12	15	11.7	19.2			
18-Oct	13	10	11	12	10	6	3	5	4	8	13	13	13	14	14	13	12	9	8	7	4	5	2	1	8.7	14.1			
19-Oct	0	-1	-1	-2	-3	-2	-2	-3	-1	0	3	6	7	9	11	13	12	11	9	9	9	9	6	5	4.2	12.5			
20-Oct	5	4	3	3	5	8	7	7	7	9	12	13	13	14	14	13	13	11	8	5	3	3	1	2	7.6	13.7			
21-Oct	3	2	0	0	-1	-2	-1	-3	-1	2	6	8	10	11	11	11	10	6	4	3	2	2	2	2	3.6	11.0			
22-Oct	1	1	1	0	-1	-1	-3	-4	-3	1	3	6	8	10	11	12	11	8	3	1	0	1	4	3	3.1	12.2			
23-Oct	1	0	1	-2	-4	-6	-6	-6	-4	-1	2	6	8	10	12	12	13	9	6	5	5	4	3	1	2.9	13.1			
24-Oct	-1	-2	-2	-4	-4	-4	-4	-5	-4	0	2	4	8	12	16	15	13	10	7	6	4	2	1	1	3.0	16.4			
25-Oct	1	1	0	0	-1	-1	-2	-1	0	2	4	7	10	13	15	13	9	8	8	7	7	6	3	1	4.6	14.7			
26-Oct	0	0	1	-1	-2	-2	-3	-3	-3	-1	0	1	3	3	3	4	4	3	1	-1	-2	-1	5	5	0.7	4.8			
27-Oct	5	5	5	4	4	3	3	4	5	5	6	8	9	10	10	11	9	5	3	1	-1	-2	-2	-2	4.6	10.5			
28-Oct	-3	-2	-1	-1	-1	-1	0	0	0	0	1	1	2	3	3	3	2	2	1	1	-1	-1	-2	-3	0.2	3.3			
29-Oct	-3	-4	-4	-5	-5	-5	-5	-5	-5	-3	-1	2	6	9	10	10	8	6	3	1	0	-2	-2	0	0.3	9.9			
30-Oct	1	2	4	5	5	4	3	3	3	4	5	6	9	10	11	10	9	7	5	4	4	5	3	1	5.1	11.1			
31-Oct	1	2	1	1	0	0	0	0	0	1	2	4	5	5	5	4	2	1	0	0	1	1	2	1	1.7	5.4			
0.8 0.6 0.5 0.0 -0.6 -0.6 -0.8 -1.0 -0.2 1.7 3.5 5.1 6.5 7.6 8.3 8.3 7.5 6.0 4.2 3.3 2.6 2.2 1.7 1.2																								Diurnal Average					
12.6 10.4 10.8 11.7 9.6 8.5 6.9 7.0 9.0 11.7 14.0 15.7 17.0 19.2 20.0 20.5 18.2 15.0 14.7 12.6 13.1 11.3 12.4 14.6																								Diurnal Maximum					

Hourly Averages

External Temperature (ET) - °C

Smoky Heights - October 2018






Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Smoky Heights - October 2018

Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
23 Spd	9	7	7	7	8	9	9	7	8	7	6	5	6	7	3	4	0	3	4	6	7	7	5	6	2.6	9.3
Dir	242	231	215	200	207	182	198	206	219	216	195	174	164	162	162	115	286	17	358	355	359	2	353	16	210.2	182.2
24 Spd	1	5	1	7	4	4	6	8	9	8	7	6	6	6	23	19	16	14	14	11	9	10	8	8	7.1	22.7
Dir	135	238	271	237	181	175	203	194	195	184	168	140	152	150	254	242	250	257	258	262	252	233	214	202	225.7	254.0
25 Spd	5	7	9	7	9	9	9	10	7	6	9	10	7	7	8	13	32	18	20	16	26	21	16	14	11.5	32.0
Dir	199	218	231	227	217	234	236	212	221	207	196	193	169	168	197	240	246	211	201	219	229	222	235	252	221.1	246.2
26 Spd	13	16	15	12	15	12	8	8	9	5	6	5	3	5	5	6	6	8	7	6	4	11	25	21	9.2	24.9
Dir	242	230	237	232	216	236	212	212	247	241	191	207	221	172	176	215	218	201	189	191	180	200	228	218	218.4	227.7
27 Spd	23	24	25	24	20	20	15	24	28	29	30	33	26	23	22	19	15	11	13	12	13	9	8	5	18.8	33.1
Dir	223	227	232	234	237	236	250	243	252	256	257	260	266	286	281	271	259	253	251	256	260	254	227	247	251.2	259.9
28 Spd	2	4	1	4	8	10	11	14	11	10	10	10	8	6	4	6	11	12	4	3	2	2	1	9	3.2	13.9
Dir	266	254	292	14	14	16	23	41	26	17	20	16	16	3	16	240	245	238	241	308	224	247	226	249	350.4	41.3
29 Spd	11	6	6	8	6	7	6	4	3	6	6	7	13	21	23	19	14	13	13	11	12	10	12	9	9.4	23.1
Dir	257	240	204	233	228	222	239	240	200	178	182	174	203	235	250	246	246	256	246	255	258	254	254	254	238.4	250.4
30 Spd	8	13	20	20	16	14	14	13	15	16	18	17	21	27	35	33	23	22	20	17	16	19	12	11	18.2	34.5
Dir	233	234	233	237	240	239	247	242	236	239	239	233	239	257	245	245	241	243	233	235	240	234	245	231	240.2	245.1
31 Spd	11	12	9	9	8	5	12	7	7	7	6	7	4	2	3	1	6	5	2	3	1	3	10	14	4.0	13.6
Dir	215	244	242	239	253	238	203	218	233	229	189	208	182	152	26	335	290	339	320	264	157	276	327	351	247.2	351.3
Spd	6.3	6.4	6.4	6.2	5.4	5.5	4.9	4.3	4.3	4.8	5.5	5.8	6.1	6.2	7.9	7.9	7.6	6.2	5.2	5.4	5.4	5.8	6.3	6.6	Diurnal Average	
Dir	250.0	248.4	243.9	244.8	240.1	243.3	240.4	238.8	248.7	241.3	240.0	244.9	243.2	246.6	256.7	254.2	255.2	251.7	246.7	255.2	259.7	252.6	255.1	258.2	Diurnal Maximum	
Spd	23.5	25.0	25.3	23.9	19.8	22.7	21.8	24.0	28.5	29.3	44.4	50.6	48.8	46.1	46.7	45.4	40.2	27.8	26.5	26.5	29.4	27.0	28.4	26.4	Diurnal Maximum	
Dir	251.5	274.7	270.4	266.6	237.3	243.0	225.6	243.2	252.0	256.1	252.0	252.6	248.4	248.5	257.8	257.0	256.4	250.4	248.0	248.9	254.1	251.5	263.8	255.2		
Maximum Speed Value: 51 km/h on Oct 20 12:00							Minimum Speed Value: 0 km/h on Oct 2 09:00													Hours in Service:			744			
Maximum Daily Speed Average: 24.9 km/h on Oct 20							Minimum Daily Speed Average: 2.6 km/h on Oct 10													Hours of Data:			744			
Maximum Diurnal Speed Average: 7.9 km/h at hour 16							Minimum Diurnal Speed Average: 4.3 km/h at hour 8													Hours of Missing Data:			0			
Monthly Average Velocity: 5.90 km/h 248.85 deg							Speed Percentiles: P ₁ = 1.0 P ₁₀ = 4.1 Q ₁ = 6.3 Median = 9.8 Q ₃ = 14.7 P ₉₀ = 20.8 P ₉₉ = 37.8													Percent Operational Time:			100.0			
All monthly, daily, and diurnal averages have been calculated using vector methods																										
Frequency Distribution																										
Speed Range (km/h)																										
Direction	0 to 5		5 to 11		11 to 19		19 to 28		28 to 38		> 38		Total													
North	13		43		16		0		0		0		72													
NorthEast	6		61		29		0		0		0		96													
East	4		6		6		0		0		0		16													
SouthEast	10		17		17		0		0		0		44													
South	15		61		8		1		0		0		85													
SouthWest	23		93		83		36		4		0		239													
West	16		29		62		33		9		8		157													
NorthWest	14		11		7		3		0		0		35													
Total	101		321		228		73		13		8		744													

Wind Rose

Wind Speed (WS) (km/h)
Smoky Heights - October 2018



Hourly Averages - Wind Speed (Scalar)

Wind Speed (km/h)

Smoky Heights - October 2018

Maximum Speed: 51 km/h on Oct 20 12:00																				Maximum Daily Speed Average: 25.4 km/h on Oct 20										Hours in Service: 744	
Minimum Speed: 1 km/h on Oct 23 17:00																				Minimum Daily Speed Average: 5.7 km/h on Oct 10										Hours of Data: 744	
Maximum Diurnal Speed Average: 15.0 km/h at hour 16																				Minimum Diurnal Speed Average: 9.8 km/h at hour 9										Hours of Missing Data: 0	
Monthly Average Speed: 11.75 km/h																				Percentiles: P ₁ = 2.1 P ₁₀ = 4.4 Q ₁ = 6.6 Median = 10.0 Q ₃ = 14.9 P ₉₀ = 20.9 P ₉₉ = 37.9										Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	6	6	6	5	2	3	4	3	5	7	3	8	9	8	7	6	9	11	12	14	13	13	12	12	7.8	14.0					
2-Oct	11	11	11	10	8	7	7	5	4	6	5	7	11	9	8	14	22	20	18	15	14	18	18	18	11.4	21.8					
3-Oct	19	24	20	14	14	15	7	5	7	7	8	6	7	8	8	5	6	6	5	6	7	4	2	3	8.9	24.0					
4-Oct	3	2	6	6	6	7	7	6	5	6	7	10	11	19	15	11	11	9	8	14	15	10	8	6	8.7	19.0					
5-Oct	8	10	8	9	10	9	5	7	9	10	9	7	9	8	6	9	8	7	7	6	6	5	4	4	7.5	10.0					
6-Oct	5	6	5	7	7	8	8	4	3	3	7	6	7	8	7	9	10	6	5	7	13	12	10	10	7.2	13.2					
7-Oct	9	9	10	11	12	14	11	12	13	13	12	14	15	13	14	14	15	16	15	14	12	10	10	11	12.5	16.2					
8-Oct	10	10	11	11	13	13	10	9	10	9	11	10	8	11	13	14	14	13	15	11	10	9	9	9	10.9	14.6					
9-Oct	9	10	10	10	11	11	9	10	11	10	10	11	9	14	16	19	19	19	19	18	17	18	16	16	13.4	19.1					
10-Oct	14	12	12	10	7	6	5	3	4	6	3	3	2	4	7	7	6	3	5	4	3	3	4	4	5.7	14.3					
11-Oct	3	2	4	6	6	7	6	7	6	9	11	14	19	20	19	17	14	22	17	10	6	7	10	9	10.5	21.9					
12-Oct	6	4	4	3	4	7	14	19	16	17	12	18	19	13	10	14	13	6	3	4	4	3	4	6	9.3	18.7					
13-Oct	7	5	3	4	8	8	7	7	6	5	6	7	5	5	11	14	10	7	8	10	12	13	16	14	8.3	16.1					
14-Oct	12	10	11	12	12	13	14	13	15	16	19	19	23	16	30	41	34	26	27	24	21	24	28	23	20.1	41.3					
15-Oct	22	25	25	24	18	18	18	18	19	25	29	29	23	26	22	18	15	15	15	16	16	7	6	15	19.3	29.4					
16-Oct	15	14	14	14	14	15	14	14	13	15	20	21	25	18	27	26	15	15	19	27	29	27	25	26	19.3	29.5					
17-Oct	24	19	18	15	12	14	10	10	6	8	14	17	18	19	15	9	12	18	21	21	18	17	15	17	15.2	23.6					
18-Oct	11	12	15	15	12	13	15	15	12	13	26	26	23	22	20	16	13	11	11	14	11	15	16	15	15.6	26.0					
19-Oct	12	10	11	11	8	6	4	9	7	8	8	7	8	8	9	14	14	11	13	14	13	10	5	8	9.5	13.9					
20-Oct	11	10	12	13	18	23	22	19	17	24	45	51	49	46	47	46	40	28	17	15	13	14	12	19	25.4	51.0					
21-Oct	20	13	16	17	15	17	16	10	7	8	7	7	6	9	11	12	9	3	1	2	4	6	4	5	9.4	19.6					
22-Oct	5	6	6	9	9	11	10	9	8	9	7	9	9	9	7	6	7	9	5	2	2	6	7	8	7.2	10.6					
23-Oct	9	7	7	7	8	9	9	7	8	7	6	5	7	7	3	4	1	3	4	6	7	7	5	6	6.3	9.3					
24-Oct	4	5	2	7	5	4	6	8	9	8	7	6	6	6	24	19	16	14	14	11	10	11	8	8	9.1	23.8					
25-Oct	6	8	10	7	9	9	9	10	8	6	9	10	7	7	10	13	32	18	20	16	26	21	16	14	12.5	32.5					
26-Oct	13	16	15	12	15	12	9	8	9	5	6	5	3	7	7	7	6	8	8	7	6	12	25	22	10.1	25.0					
27-Oct	23	24	25	24	20	20	15	24	29	29	30	33	27	23	22	20	15	11	13	12	13	10	9	6	19.9	33.3					
28-Oct	3	4	2	4	8	10	12	14	11	10	10	10	8	6	4	7	11	12	4	4	2	3	2	10	7.1	14.0					
29-Oct	11	6	6	8	6	7	6	4	3	7	6	7	14	21	24	19	14	13	13	12	12	11	12	10	10.4	23.5					
30-Oct	8	13	20	20	16	14	14	13	15	16	18	17	22	27	35	33	23	23	20	17	16	19	12	11	18.5	34.7					
31-Oct	11	12	9	9	8	6	12	7	7	7	7	7	5	3	4	2	7	5	3	3	4	5	11	15	6.9	14.7					
																									Diurnal Average						
																									Diurnal Maximum						

All monthly, daily, and diurnal averages have been calculated using scalar methods

Hourly Standard Deviations

Wind Direction (WD) - deg
Smoky Heights - October 2018

Maximum Value: 96.4 deg on Oct 23 17:00																			Hours in Service: 744						
Minimum Value: 2.0 deg on Oct 16 00:00																			Hours of Data: 744						
																			Hours of Missing Data: 0						
																			Hours of Calibration: 0						
																			Percent Operational Time: 100.0						
Percentiles: P ₁ = 2.6 P ₁₀ = 4.5 Q ₁ = 6.2 Median = 9.5 Q ₃ = 16.8 P ₉₀ = 30.2 P ₉₉ = 83.3																									
Day	Hourly Period Ending At (MST)																							Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	
1-Oct	10	10	10	13	77	32	54	56	16	27	70	41	15	19	19	28	26	8	5	6	8	6	7	7	77.4
2-Oct	9	8	9	11	13	9	9	9	94	29	55	34	30	33	48	21	7	7	3	4	5	5	5	3	94.4
3-Oct	3	4	4	37	22	17	14	18	15	24	17	28	30	23	33	38	42	21	6	12	26	39	48	37	47.9
4-Oct	42	62	12	14	16	10	23	11	8	16	13	14	24	12	56	40	29	10	8	4	4	18	25	36	62.1
5-Oct	10	8	6	5	5	5	7	7	7	11	13	28	18	19	37	17	14	11	11	6	8	12	8	12	36.8
6-Oct	26	11	9	10	19	21	42	47	29	56	17	17	22	15	17	11	9	13	18	8	7	8	6	5	56.4
7-Oct	8	17	6	6	6	6	7	6	6	6	6	5	6	7	7	7	6	8	5	5	5	6	8	6	17.5
8-Oct	6	6	8	6	5	5	7	6	5	7	6	7	13	22	11	9	15	6	6	9	14	7	5	15	22.3
9-Oct	7	10	5	6	6	5	5	5	5	7	8	7	17	9	11	14	12	9	10	5	6	5	6	5	17.3
10-Oct	6	8	6	8	8	9	11	20	24	17	85	68	78	36	23	17	20	24	22	38	15	52	25	32	85.0
11-Oct	42	56	17	18	9	10	14	8	10	7	13	16	11	8	6	9	18	5	4	7	12	21	5	4	55.9
12-Oct	7	19	57	89	23	14	8	7	8	8	11	9	7	6	11	8	17	20	16	19	12	22	16	7	89.2
13-Oct	9	12	14	14	11	9	7	6	8	8	25	10	26	75	14	11	8	13	6	7	6	6	3	7	75.2
14-Oct	10	8	6	5	5	4	7	5	7	4	5	7	6	10	4	5	4	3	4	6	5	3	2	2	10.1
15-Oct	4	3	3	3	5	3	7	7	3	4	5	5	8	6	8	8	7	8	4	4	5	94	46	2	94.3
16-Oct	8	5	3	6	5	4	7	5	4	6	5	8	6	11	6	4	5	6	6	2	3	3	3	3	11.1
17-Oct	6	9	12	7	10	5	19	11	21	20	6	8	10	20	13	22	8	7	8	10	6	4	10	15	21.6
18-Oct	20	5	8	8	8	5	4	13	28	10	12	8	10	11	9	9	6	6	4	5	8	13	7	5	27.7
19-Oct	12	24	12	18	14	14	11	17	23	10	15	16	14	14	13	9	8	4	4	9	7	13	25	11	25.0
20-Oct	17	11	9	8	10	7	6	8	8	8	6	6	5	6	9	6	4	6	6	7	11	6	8	7	16.6
21-Oct	3	13	5	4	6	3	3	37	12	8	18	14	22	22	22	13	17	35	53	36	11	10	14	7	53.4
22-Oct	6	7	7	6	16	8	8	16	21	11	16	11	15	17	14	18	12	9	13	43	63	21	15	10	62.7
23-Oct	5	7	8	11	12	7	17	16	12	11	16	19	14	17	35	18	96	9	11	11	9	12	13	9	96.4
24-Oct	79	17	86	8	28	27	18	12	9	8	14	13	13	13	28	10	7	7	2	9	25	13	9	13	86.1
25-Oct	15	10	24	13	11	8	6	10	24	19	6	10	12	16	34	11	9	13	8	6	3	6	13	4	33.5
26-Oct	11	4	4	6	5	8	13	20	9	15	17	19	38	36	39	25	20	7	13	29	63	10	5	5	62.8
27-Oct	7	3	3	2	3	3	13	4	4	4	4	5	15	11	11	15	11	7	7	11	10	22	28	28	28.4
28-Oct	62	24	58	9	6	6	10	7	7	11	12	7	10	10	30	64	6	8	50	55	65	56	81	22	81.1
29-Oct	4	28	18	9	15	19	17	28	35	13	18	14	17	7	11	8	5	3	7	12	12	12	6	14	35.3
30-Oct	14	4	5	4	5	7	3	6	4	4	5	8	8	6	6	5	5	4	7	4	5	3	11	17	17.3
31-Oct	8	7	12	9	11	33	14	17	6	11	17	18	39	72	53	85	17	11	50	26	93	48	8	22	93.0
	79.4	62.1	86.1	89.2	77.4	32.6	54.0	56.4	94.4	56.4	85.0	68.2	78.4	75.2	55.6	85.2	96.4	34.7	53.4	54.9	93.0	94.3	81.1	36.5	

PAZA
Valleyview Station
Monthly Summary Tables, Graphs and
Roses

Hourly Averages


Sulphur Dioxide (SO₂) - ppb

Valleyview - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0																				Hours in Service: 744						
Maximum Value: 8.6 ppb on Oct 30 11:00																				Maximum Daily Average: 1.4 ppb on Oct 30		Hours of Data: 710				
Minimum Value: 0 ppb on Oct 1 19:00																				Minimum Daily Average: 0.0 ppb on Oct 2		Hours of Missing Data: 34				
Maximum Diurnal Average: 0.7 ppb at hour 11																				Minimum Diurnal Average: 0.0 ppb at hour 7		Hours of Calibration: 34				
Monthly Average: 0.17 ppb																				Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.1 Q ₃ = 0.1 P ₉₀ = 0.3 P ₉₉ = 2.3		Percent Operational Time: 100.0				
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	A	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
3-Oct	0	0	0	0	0	0	0	0	0	A	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0.2	1.2
4-Oct	0	0	0	0	0	0	0	0	A	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0.2	0.6
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.2
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0.1	0.5
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
10-Oct	0	0	A	0	0	0	0	0	0	0	1	2	0	0	0	0	0	0	0	0	0	0	0	0	0.3	1.7
11-Oct	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0.0	0.0
13-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0.1	0.3
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0.1	0.4
15-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0.1	0.3
16-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0.1	0.4
17-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0.1	0.3
18-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.1	0.5
19-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	A	0	0	0	0	0	0	0	0.1	0.5
20-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	2	2	0	0	0	0.3	2.3
21-Oct	0	0	0	0	0	0	0	0	0	0	1	0	1	0	A	0	0	0	0	0	0	0	0	0	0.2	0.7
22-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	A	7	2	1	0	0	0	0	0	1	1	0.6	7.0
23-Oct	1	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0.2	0.6
24-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	1	0	0	0	0	0	0	0	0	0	0	0.1	0.7
25-Oct	0	0	0	0	0	0	0	0	0	0	A	R	C	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
26-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.3
27-Oct	0	0	0	0	0	0	0	0	A	4	3	1	0	0	0	0	0	1	0	0	0	0	0	0	0.4	3.6
28-Oct	0	0	0	0	0	0	0	A	0	0	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.9
29-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
30-Oct	0	0	0	0	0	A	0	0	2	4	9	7	1	0	0	0	1	2	3	2	0	0	0	0	1.4	8.6
31-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
0.1 0.1 0.1 0.1 0.0 0.0 0.0 0.1 0.1 0.3 0.7 0.5 0.2 0.2 0.4 0.2 0.1 0.2 0.2 0.2 0.1 0.1 0.1 0.1																									Diurnal Average	
0.6 0.4 0.5 0.3 0.3 0.2 0.2 0.4 1.9 3.6 8.6 7.1 1.1 0.7 7.0 2.1 0.9 1.6 3.0 2.0 2.3 0.4 0.7 0.8																									Diurnal Maximum	
C - Calibration R - Recovery A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 172 ppb 24-hr 48 ppb 30-day 11 ppb																										


Hourly Averages

Sulphur Dioxide (SO₂) - ppb
Valleyview - October 2018




Hourly Maximums

Sulphur Dioxide (SO₂) - ppb
Valleyview - October 2018




Pollutant Rose

Sulphur Dioxide (SO₂) - ppb
Valleyview - October 2018




Pollutant Classes (ppb)



Span Responses


Sulphur Dioxide (SO₂)
Valleyview - October 2018



Number of Exceedences (AAAQO): Maximum Value: 1.3 ppb on Oct 4 02:00																			1-hr: 0 24-hr: 0 Maximum Daily Average: 0.3 ppb on Oct 4					Hours in Service: 744 Hours of Data: 710 Hours of Missing Data: 34 Hours of Calibration: 34 Percent Operational Time: 100.0				
Minimum Value: 0 ppb on Oct 1 14:00 Maximum Diurnal Average: 0.2 ppb at hour 2 Monthly Average: 0.11 ppb																			Minimum Daily Average: 0.0 ppb on Oct 7 Minimum Diurnal Average: 0.1 ppb at hour 15 Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.1 Q ₃ = 0.2 P ₉₀ = 0.3 P ₉₉ = 0.6									
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum		
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24				
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3		
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.2		
3-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0.2	0.9		
4-Oct	1	1	1	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0.3	1.3		
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3		
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.4		
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1		
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1		
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.2		
10-Oct	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.4		
11-Oct	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	A	0.1	0.6		
13-Oct	1	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	A	0	0.2	0.8		
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0.2	0.4		
15-Oct	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0.2	0.5		
16-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0.0	0.2		
17-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0.0	0.2		
18-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.0	0.1		
19-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0.0	0.2		
20-Oct	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0.0	0.6		
21-Oct	0	0	0	0	0	0	0	0	0	0	1	0	0	0	A	0	0	0	0	0	0	0	0	0	0.1	0.9		
22-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
23-Oct	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5		
24-Oct	1	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5		
25-Oct	0	0	0	0	0	0	0	0	0	C	C	C	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5		
26-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.3		
27-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.5		
28-Oct	0	0	0	0	1	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	1.0		
29-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
30-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5		
31-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2		
																									Diurnal Average			
																									Diurnal Maximum			
C - Calibration A - Automated Daily Zero Span																												
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 10 ppb 24-hr 3 ppb																												

Hourly Averages

Hydrogen Sulphide (H₂S) - ppb
Valleyview - October 2018



Hourly Maximums


Hydrogen Sulphide (H₂S) - ppb

Valleyview - October 2018

Maximum Value: 28.7 ppb on Oct 21 11:00		Maximum Daily Average: 1.9 ppb on Oct 21		Hours in Service: 744																							
Minimum Value: 0 ppb on Oct 17 23:00		Minimum Daily Average: 0.3 ppb on Oct 17		Hours of Data: 710																							
Maximum Diurnal Average: 1.5 ppb at hour 11		Minimum Diurnal Average: 0.5 ppb at hour 16		Hours of Missing Data: 34																							
Monthly Average: 0.60 ppb		Percentiles: P ₁ = 0.2 P ₁₀ = 0.3 Q ₁ = 0.4 Median = 0.5 Q ₃ = 0.6 P ₉₀ = 0.8 P ₉₉ = 1.9		Hours of Calibration: 34																							
				Percent Operational Time: 100.0																							
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
1-Oct	1	1	1	1	1	1	1	1	1	1	1	1	A	1	0	0	1	0	0	0	0	0	0	1	0	0.6	1.1
2-Oct	1	1	0	0	0	0	0	0	0	0	A	1	1	1	1	0	1	1	1	1	0	0	0	0	0	0.5	0.7
3-Oct	0	1	1	1	0	0	0	0	0	A	1	1	1	1	1	1	0	1	1	2	0	1	5	1	0.8	4.8	
4-Oct	2	2	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	2	1	1	1	0	0.8	2.4	
5-Oct	1	1	1	1	1	1	1	A	1	0	1	1	1	1	0	1	1	1	0	1	1	0	0	1	0.7	1.3	
6-Oct	1	1	1	1	1	0	A	1	0	0	1	1	1	1	1	1	1	1	1	0	0	0	1	0	0.6	1.3	
7-Oct	0	0	1	0	0	A	1	0	0	0	0	0	0	1	0	0	1	0	0	0	0	1	0	0	0.4	0.7	
8-Oct	0	0	0	0	A	1	0	0	0	0	0	1	0	0	0	0	1	0	0	0	0	1	1	0	0.4	0.6	
9-Oct	0	1	0	A	1	0	0	0	0	1	1	0	0	1	0	1	0	1	0	0	0	0	1	0	0.5	0.6	
10-Oct	0	1	A	1	1	0	0	1	0	0	0	0	1	1	1	1	0	1	1	1	1	1	1	1	0.6	1.3	
11-Oct	0	A	1	1	1	1	1	1	1	0	1	1	0	0	0	0	1	1	0	0	1	1	0	1	0.5	0.8	
12-Oct	A	1	1	1	1	1	1	1	0	0	1	1	0	1	1	1	0	0	1	1	0	1	1	A	0.6	1.3	
13-Oct	2	1	1	1	1	1	1	0	1	1	1	1	1	0	1	0	0	1	1	0	0	1	A	0	0.7	1.9	
14-Oct	0	0	1	0	0	0	1	0	1	0	1	1	1	1	1	1	1	1	1	1	1	A	1	1	0.6	0.8	
15-Oct	1	1	1	1	1	1	0	0	1	1	1	1	1	0	0	0	0	1	1	1	A	1	1	1	0.6	1.0	
16-Oct	1	0	0	0	1	0	1	0	0	0	0	0	0	0	0	0	1	0	1	A	1	0	0	0	0.4	0.7	
17-Oct	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	A	0	0	0	0	0	0.3	1.2	
18-Oct	0	0	0	0	0	1	0	0	0	0	0	0	0	1	0	0	0	A	1	0	1	1	0	1	0.4	0.6	
19-Oct	0	0	0	0	0	1	0	1	0	1	1	0	0	1	0	0	A	1	1	0	0	0	0	0	0.5	0.8	
20-Oct	0	0	2	0	0	0	0	1	1	0	0	0	0	0	A	0	0	0	0	1	0	0	0	0	0.5	2.3	
21-Oct	0	0	0	0	0	0	0	0	0	0	29	6	0	1	A	0	0	1	0	1	1	0	0	0	1.9	28.7	
22-Oct	0	0	0	0	0	0	0	0	1	1	0	0	0	A	1	1	0	0	0	0	0	0	1	1	0.5	0.8	
23-Oct	1	0	1	1	1	0	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	0.7	1.1	
24-Oct	1	1	1	1	1	1	1	0	1	1	1	A	1	1	0	0	1	1	1	1	1	1	1	1	0.7	1.1	
25-Oct	1	1	1	1	1	1	1	1	1	C	C	C	1	0	1	1	0	1	0	1	0	0	1	0	0.6	1.0	
26-Oct	0	0	0	0	0	0	0	0	1	A	0	0	0	0	1	0	0	1	1	1	1	0	0	0	0.4	0.7	
27-Oct	1	1	0	1	0	1	0	1	A	1	1	1	1	0	1	0	1	2	1	0	1	1	1	1	0.7	1.6	
28-Oct	1	1	1	1	2	1	1	A	1	0	1	1	1	1	0	0	0	0	1	1	0	0	0	0	0.6	1.8	
29-Oct	0	0	1	0	0	0	A	0	1	0	0	0	0	0	0	0	1	1	1	0	1	1	1	1	0.4	0.7	
30-Oct	1	1	1	0	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	0	0	0	1	1	0.7	1.3	
31-Oct	1	0	1	1	A	1	1	0	0	1	1	1	1	0	0	0	1	1	0	0	0	0	1	1	0.5	1.3	
0.6		0.7	0.6	0.5	0.6	0.5	0.5	0.5	0.5	0.5	1.5	0.7	0.5	0.5	0.5	0.5	0.6	0.6	0.5	0.6	0.5	0.5	0.7	0.5	Diurnal Average		
1.9		2.2	2.3	1.0	1.8	1.0	0.9	0.9	1.1	1.0	28.7	5.8	1.0	0.8	0.7	1.2	1.3	1.6	1.3	2.4	0.9	1.3	4.8	1.0	Diurnal Maximum		
C - Calibration		A - Automated Daily Zero Span																									


Hourly Maximums

Hydrogen Sulphide (H₂S) - ppb
Valleyview - October 2018




Pollutant Rose

Hydrogen Sulphide (H₂S) - ppb
Valleyview - October 2018



Span Responses


Hydrogen Sulphide (H₂S)
Valleyview - October 2018



Hourly Averages


External Temperature (ET) - °C

Valleyview - October 2018



Hourly Averages

Relative Humidity (RH) - %
Valleyview - October 2018



Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Valleyview - October 2018

Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1 Spd	0	0	1	0	2	1	1	1	1	2	2	3	3	5	7	6	8	10	3	8	11	13	12	10	4.6	12.5
Dir	152	3	346	10	339	337	339	334	9	0	25	4	2	334	340	355	4	6	347	3	7	10	8	3	1	10
2 Spd	11	12	11	10	7	6	5	6	5	5	7	6	3	4	6	6	6	7	6	6	6	3	1	1	4.2	12.0
Dir	1	4	6	3	10	3	1	359	2	4	9	356	347	259	316	256	289	284	262	255	255	238	165	220	335	4
3 Spd	2	3	3	3	1	2	1	1	3	4	3	1	1	2	1	2	0	1	1	1	0	0	1	0	1.0	3.9
Dir	227	235	233	224	176	219	162	243	252	272	281	252	271	26	246	240	287	53	113	78	312	311	349	171	249	272
4 Spd	0	0	0	1	3	2	2	1	2	1	2	2	1	3	5	6	7	7	1	1	0	1	0	1	1.0	6.6
Dir	216	252	195	176	170	172	172	166	163	194	177	160	149	261	292	298	306	324	298	326	298	166	280	186	266	306
5 Spd	0	0	1	4	1	1	2	2	2	4	3	4	3	4	5	6	3	2	2	2	2	1	0	0	2.1	5.7
Dir	327	68	55	29	20	3	340	346	18	23	24	7	358	4	21	33	54	68	23	29	9	334	173	193	19	33
6 Spd	0	0	0	1	1	1	1	1	1	0	0	0	1	1	2	4	5	3	6	3	3	3	5	5	1.2	5.9
Dir	292	300	180	189	179	241	172	158	172	172	146	224	158	270	359	41	19	32	17	22	21	22	7	9	21	17
7 Spd	3	2	3	4	3	3	3	4	3	2	4	4	7	7	9	9	8	9	8	7	7	5	6	5	5.1	9.1
Dir	8	11	2	15	18	8	15	3	33	45	24	25	15	18	28	19	14	15	22	17	14	18	22	28	18	15
8 Spd	5	2	3	1	2	1	1	1	3	3	4	5	7	7	7	7	7	7	7	6	5	7	5	5	4.0	7.4
Dir	37	38	34	60	38	81	126	128	134	136	155	147	135	133	131	133	139	135	136	138	140	141	145	138	130	133
9 Spd	6	7	7	5	5	4	4	4	6	6	7	5	7	8	8	8	7	7	5	6	5	6	7	5	5.9	8.2
Dir	134	130	131	137	135	138	148	141	147	147	157	149	141	148	144	152	137	148	159	165	156	156	150	155	146	144
10 Spd	4	5	4	3	4	5	5	4	6	6	5	5	4	3	2	1	0	1	3	3	2	1	1	1	2.7	6.3
Dir	159	145	154	162	157	160	158	164	165	165	170	158	161	163	164	179	203	268	302	287	271	68	76	141	165	165
11 Spd	0	0	0	0	1	1	1	1	4	4	5	5	5	4	3	2	6	2	1	2	1	1	1	1	1.9	5.9
Dir	155	164	166	164	157	173	177	176	165	169	167	169	167	168	173	175	255	234	179	170	163	186	191	188	180	255
12 Spd	1	1	2	1	1	1	4	8	16	14	18	16	14	14	15	14	11	8	8	7	5	4	3	4	6.9	18.4
Dir	174	183	166	172	174	8	5	14	11	4	352	346	350	353	348	339	328	331	336	322	305	274	298	275	344	352
13 Spd	4	4	3	1	1	0	0	1	0	1	2	4	5	7	7	5	5	1	1	1	1	2	2	3	1.7	7.3
Dir	273	257	244	215	228	281	89	159	169	282	277	293	301	312	309	314	323	5	165	183	184	186	178	172	283	312
14 Spd	3	3	2	2	1	2	3	3	2	4	3	3	11	12	8	11	16	14	7	11	12	11	16	16	6.0	16.4
Dir	168	166	177	173	187	176	175	184	167	171	163	191	261	273	267	267	275	269	273	268	279	285	284	283	262	284
15 Spd	16	13	11	9	6	3	2	1	2	2	1	9	15	14	14	12	11	5	3	4	6	4	3	2	6.2	15.9
Dir	280	280	278	273	271	253	216	178	155	132	60	311	307	308	307	307	317	327	306	305	306	290	264	282	294	280
16 Spd	1	6	2	1	1	1	1	2	2	3	2	8	8	10	10	8	6	1	4	3	4	10	10	11	4.0	10.6
Dir	206	250	220	191	175	177	176	165	183	178	186	267	279	283	296	290	302	307	247	241	233	258	256	254	259	254
17 Spd	6	3	1	2	3	2	1	0	2	3	3	4	3	9	6	4	2	1	1	2	1	1	1	0	1.9	9.3
Dir	230	216	155	136	144	160	177	182	198	181	175	180	202	245	268	302	276	215	223	232	207	200	156	137	215	245
18 Spd	2	1	2	2	4	1	7	6	5	1	9	16	18	17	16	14	9	5	1	1	1	1	1	1	5.1	18.2
Dir	278	219	225	240	249	217	264	273	315	44	302	304	316	320	314	292	297	277	185	193	186	209	183	196	297	316
19 Spd	1	0	0	0	1	0	1	0	0	1	2	2	2	3	4	5	5	3	3	3	6	11	9	8	3.0	11.0
Dir	167	175	196	202	200	45	194	218	188	154	171	172	156	168	168	162	155	150	143	161	168	168	167	168	165	168
20 Spd	9	6	2	3	3	2	2	8	6	13	17	24	25	20	22	21	17	11	10	8	9	10	11	8	9.8	24.5
Dir	173	174	245	202	185	183	245	270	268	268	270	274	282	292	275	292	283	281	273	261	262	252	253	250	268	282
21 Spd	6	11	8	6	3	2	1	2	2	2	1	2	2	2	4	2	1	1	0	0	0	0	0	0	1.9	10.6
Dir	249	246	259	256	218	184	164	167	149	135	288	294	194	217	252	254	274	300	141	269	186	184	174	244	239	246
22 Spd	0	0	0	0	0	0	0	0	0	1	5	4	5	6	4	6	5	4	6	3	4	6	4	4	2.8	6.0
Dir	295	207	217	314	349	158	248	230	215	169	172	177	165	169	168	167	168	172	168	170	174	170	176	184	172	170




Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Valleyview - October 2018


Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
23 Spd	2	0	0	0	0	0	2	0	0	1	2	2	2	0	2	3	4	4	2	0	1	0	0	0	0	0.3	4.2
Dir	192	260	190	163	207	201	200	206	187	181	179	185	192	260	18	332	337	351	329	69	176	145	260	349	284	337	7.5
24 Spd	0	0	1	1	0	1	2	1	1	2	2	3	3	2	8	3	5	2	1	1	1	1	2	1	0.7	7.5	
Dir	181	94	167	176	189	205	195	194	168	173	173	172	143	328	317	347	296	262	201	196	165	190	169	201	233	317	
25 Spd	0	1	1	1	0	0	0	0	0	0	0	1	2	3	3	2	1	13	9	2	2	3	2	2	0.9	13.2	
Dir	215	192	199	170	204	287	156	210	142	165	161	154	170	168	159	138	104	304	303	262	163	185	189	182	236	304	
26 Spd	3	3	3	2	3	2	3	2	2	1	0	0	1	1	2	1	0	2	0	1	2	1	1	2	1.1	3.1	
Dir	180	181	186	158	198	175	188	187	177	183	210	107	310	169	4	212	55	331	180	194	199	186	185	180	186	180	
27 Spd	2	2	2	3	3	3	3	5	8	15	15	16	17	17	14	12	8	5	5	2	1	1	2	2	5.3	17.5	
Dir	164	183	176	174	183	188	207	239	247	270	272	279	288	306	303	308	300	267	245	212	186	170	181	170	271	288	
28 Spd	1	0	1	1	1	4	5	3	1	3	7	2	2	6	4	2	2	1	1	0	2	1	0	0	0.7	7.3	
Dir	196	13	342	306	2	339	344	326	337	126	133	174	175	176	181	176	170	167	246	207	244	239	170	173	180	133	
29 Spd	0	0	1	1	1	1	0	0	1	2	2	2	2	3	1	2	1	1	1	2	2	2	1	2	1.2	2.5	
Dir	212	169	168	157	163	203	208	200	196	183	171	180	170	168	164	150	164	173	185	216	205	166	168	173	177	168	
30 Spd	1	1	1	1	1	1	2	2	5	6	6	7	13	11	10	13	11	6	9	6	2	2	2	2	4.3	13.4	
Dir	177	212	170	167	189	193	190	214	257	251	268	272	288	287	301	297	284	261	261	260	230	199	187	188	269	288	
31 Spd	2	2	2	1	1	1	1	2	2	1	2	0	2	2	2	2	3	2	8	6	3	2	4	6	0.6	7.9	
Dir	144	153	166	151	177	196	281	197	168	310	6	120	154	146	173	223	261	3	345	357	345	319	291	315	307	345	
Spd	1.0	1.1	0.7	0.5	0.7	0.5	0.6	0.5	0.4	0.8	1.1	2.4	3.0	3.3	3.6	3.3	3.2	2.2	1.6	1.3	1.1	1.2	1.2	1.2	Diurnal Average		

Wind Rose

Wind Speed (WS) (km/h)
Valleyview - October 2018



Wind Speed Classes (km/h)



Hourly Averages - Wind Speed (Scalar)

Wind Speed (km/h)

Valleyview - October 2018

Maximum Speed: 25 km/h on Oct 20 13:00																				Maximum Daily Speed Average: 11.4 km/h on Oct 20										Hours in Service: 744	
Minimum Speed: 0 km/h on Oct 24 00:00																				Minimum Daily Speed Average: 1.3 km/h on Oct 29										Hours of Data: 744	
Maximum Diurnal Speed Average: 7.2 km/h at hour 15																				Minimum Diurnal Speed Average: 1.9 km/h at hour 6										Hours of Missing Data: 0	
Monthly Average Speed: 4.14 km/h																				Percentiles: P ₁ = 0.2 P ₁₀ = 0.6 Q ₁ = 1.3 Median = 2.7 Q ₃ = 5.7 P ₉₀ = 9.6 P ₉₉ = 17.7										Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	0	1	1	0	2	1	1	1	1	2	2	4	4	6	8	6	8	10	3	8	12	13	12	10	4.9	12.7					
2-Oct	11	12	11	10	7	6	5	6	6	5	7	6	4	5	7	7	6	7	6	6	6	3	1	1	6.4	12.1					
3-Oct	3	3	3	3	1	2	1	1	3	4	3	2	3	3	3	2	1	1	1	1	1	1	1	0	2.0	4.5					
4-Oct	0	1	0	1	3	2	2	1	2	1	2	2	3	4	6	7	7	7	2	2	1	1	1	1	2.4	7.1					
5-Oct	1	1	1	4	1	1	2	2	2	4	3	4	3	5	6	6	3	2	2	2	3	1	0	0	2.5	6.2					
6-Oct	0	1	0	2	2	2	2	1	2	0	0	0	1	1	2	4	5	3	6	3	3	3	6	5	2.2	6.0					
7-Oct	3	2	3	4	3	3	3	4	3	3	4	4	7	7	9	9	8	9	8	7	7	5	6	5	5.3	9.2					
8-Oct	5	2	3	1	2	1	2	2	3	4	4	5	7	8	7	7	8	7	7	7	6	8	5	5	4.7	7.6					
9-Oct	6	7	7	5	5	4	4	5	6	6	7	6	7	8	9	8	7	7	5	6	5	7	7	5	6.2	8.7					
10-Oct	5	5	4	4	5	5	5	5	6	6	5	5	4	3	2	1	0	1	3	4	2	1	1	1	3.4	6.5					
11-Oct	0	1	1	0	1	1	1	1	4	4	5	5	5	4	3	2	6	2	1	2	1	1	1	1	2.2	6.2					
12-Oct	1	1	2	1	1	2	4	8	16	14	19	16	14	14	15	14	11	8	9	7	5	4	4	4	8.1	18.6					
13-Oct	4	4	4	1	1	0	1	1	1	1	2	4	6	8	7	6	6	1	1	1	1	2	3	3	2.7	7.7					
14-Oct	3	3	2	2	1	2	3	3	2	4	4	3	11	12	8	12	16	15	8	12	12	12	17	16	7.6	16.6					
15-Oct	16	13	11	9	6	3	2	1	2	2	2	10	15	15	14	12	11	5	3	4	6	4	3	2	7.2	16.0					
16-Oct	1	6	2	1	1	1	2	2	2	3	3	8	8	11	11	8	6	2	4	3	5	10	10	11	5.1	10.9					
17-Oct	6	3	1	2	3	3	1	1	2	4	3	4	3	9	6	4	2	1	1	2	1	1	2	1	2.8	9.4					
18-Oct	2	1	2	3	4	2	7	6	5	2	10	17	19	17	16	14	9	5	1	1	1	1	1	1	6.1	18.6					
19-Oct	1	0	0	0	1	1	1	0	0	1	2	2	3	4	4	5	5	3	3	3	6	11	9	8	3.2	11.1					
20-Oct	9	6	2	3	3	2	4	8	6	13	17	24	25	20	23	21	17	12	10	8	9	10	12	9	11.4	25.0					
21-Oct	6	11	8	6	3	2	2	2	2	2	3	3	3	3	4	2	2	1	0	0	0	0	0	0	2.8	10.7					
22-Oct	1	0	0	0	0	0	0	1	1	1	5	4	5	6	4	6	5	4	6	3	4	6	4	4	3.0	6.0					
23-Oct	2	1	0	0	0	0	2	0	0	1	2	2	2	1	2	3	4	4	2	0	1	0	0	0	1.4	4.2					
24-Oct	0	0	1	1	0	1	2	1	2	2	2	3	3	3	8	3	6	2	1	1	1	1	2	1	2.0	7.8					
25-Oct	0	1	1	1	1	0	0	0	0	1	0	1	2	3	3	2	2	15	10	4	2	3	2	2	2.4	14.7					
26-Oct	3	3	3	2	3	2	3	2	2	1	2	1	1	1	2	2	1	2	1	1	2	1	1	2	1.8	3.1					
27-Oct	2	2	2	3	3	3	3	6	8	16	16	16	18	17	14	12	8	5	5	2	1	1	2	2	6.9	18.0					
28-Oct	1	0	2	1	1	4	5	3	1	3	8	4	2	6	4	2	3	1	1	0	2	1	1	0	2.3	7.5					
29-Oct	1	1	1	1	1	1	0	0	1	2	2	2	2	3	1	2	1	1	1	2	2	2	1	2	1.3	2.6					
30-Oct	1	1	1	2	1	1	2	2	5	6	6	8	14	12	11	13	11	7	9	6	2	2	2	2	5.2	13.8					
31-Oct	2	2	2	1	1	1	1	2	2	2	2	1	2	2	3	2	3	2	8	6	3	2	4	6	2.6	8.2					
3.2 3.0 2.6 2.4 2.1 1.9 2.4 2.5 3.1 3.8 5.0 5.7 6.7 7.2 7.2 6.6 6.1 4.9 4.1 3.7 3.6 3.8 3.9 3.6																									Diurnal Average						
16.0 13.3 11.2 10.3 6.8 6.4 7.0 7.8 15.9 15.7 18.6 23.9 25.0 20.3 22.6 20.9 17.1 14.7 9.7 11.5 12.5 12.7 16.6 16.1																									Diurnal Maximum						

All monthly, daily, and diurnal averages have been calculated using scalar methods

Hourly Standard Deviations

Wind Direction (WD) - deg

Valleyview - October 2018

Maximum Value: 96.4 deg on Oct 3 17:00																		Hours in Service: 744							
Minimum Value: 4.8 deg on Oct 27 04:00																		Hours of Data: 744							
																		Hours of Missing Data: 0							
Percentiles: P ₁ = 6.0 P ₁₀ = 8.6 Q ₁ = 11.8 Median = 17.5 Q ₃ = 35.3 P ₉₀ = 62.7 P ₉₉ = 90.1																		Hours of Calibration: 0							
																		Percent Operational Time: 100.0							
Day	Hourly Period Ending At (MST)																								Daily
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	Maximum
1-Oct	76	80	46	51	8	30	22	19	35	23	36	32	54	36	20	23	11	10	26	16	8	8	8	8	80.0
2-Oct	8	8	8	8	12	9	10	9	16	17	18	35	54	52	27	28	16	15	11	8	14	24	23	21	53.7
3-Oct	42	37	32	32	32	35	31	65	46	32	40	80	77	75	86	61	96	44	37	34	63	66	52	85	96.4
4-Oct	55	54	78	17	10	9	8	55	12	32	36	42	76	51	39	27	22	10	39	52	67	12	89	49	88.9
5-Oct	82	83	67	11	86	51	36	15	23	25	46	43	53	35	39	26	33	42	21	9	17	15	77	39	86.0
6-Oct	85	78	28	14	16	58	49	15	16	54	59	87	26	68	49	21	21	15	9	15	34	19	8	11	86.7
7-Oct	51	21	28	13	19	16	11	16	19	30	20	22	13	12	11	11	9	8	11	9	8	10	15	12	50.9
8-Oct	17	19	12	29	8	26	14	16	15	12	14	14	12	11	10	13	11	9	9	12	12	12	15	12	29.4
9-Oct	13	9	8	10	11	12	14	12	11	11	12	19	18	17	19	17	12	12	12	11	14	13	15	13	19.4
10-Oct	12	14	21	21	15	12	11	11	9	12	13	16	17	18	26	31	39	25	25	14	35	65	43	46	65.0
11-Oct	30	56	46	30	17	10	8	14	9	10	11	10	13	13	18	15	26	30	24	13	11	20	7	15	56.4
12-Oct	12	13	14	8	18	68	17	13	8	7	9	8	9	8	8	6	7	11	13	9	15	12	16	9	68.0
13-Oct	12	14	18	23	28	52	67	24	31	36	19	19	15	17	19	24	21	66	17	25	22	8	14	10	67.1
14-Oct	7	5	6	9	15	10	8	11	28	12	15	31	13	14	23	16	10	9	16	12	12	11	8	8	31.3
15-Oct	7	8	7	7	10	21	16	19	17	22	76	17	11	11	11	11	13	18	23	10	9	15	11	24	75.8
16-Oct	41	13	15	9	12	9	13	10	7	9	34	15	19	20	15	19	25	91	12	23	21	10	6	7	91.2
17-Oct	14	13	32	22	13	22	39	81	33	15	12	11	34	13	20	13	29	18	24	12	18	23	51	90	90.4
18-Oct	26	14	15	56	26	34	9	15	13	72	22	12	12	11	12	14	12	14	48	31	47	25	37	45	72.0
19-Oct	25	26	17	58	19	92	12	90	36	38	23	20	29	19	15	17	13	12	12	12	8	6	7	6	92.4
20-Oct	6	8	48	15	14	14	69	18	12	12	13	11	11	12	13	10	11	13	8	14	13	8	6	6	68.7
21-Oct	9	8	10	10	26	25	42	28	32	15	82	52	42	76	42	52	56	44	43	84	36	31	68	93	93.0
22-Oct	76	28	69	84	72	57	50	51	75	17	10	13	15	11	19	10	9	8	6	8	7	6	8	5	83.9
23-Oct	32	50	40	22	61	22	16	71	76	17	18	40	41	86	41	22	8	10	17	92	38	84	74	93	93.0
24-Oct	57	89	32	41	55	54	18	20	20	15	15	12	24	71	16	18	14	25	23	46	42	28	10	16	89.0
25-Oct	53	16	52	48	57	88	35	47	77	55	68	24	15	21	24	38	59	33	17	78	27	10	10	8	88.4
26-Oct	7	19	18	25	19	16	11	14	10	41	81	79	65	83	73	62	78	28	61	23	15	15	14	28	83.2
27-Oct	16	7	7	5	5	5	8	14	10	9	10	12	14	14	12	10	13	17	6	22	19	16	10	13	21.7
28-Oct	35	93	71	68	84	14	8	53	53	55	15	50	41	10	19	15	9	26	22	29	24	24	65	50	93.5
29-Oct	18	25	21	21	30	35	71	25	19	14	21	13	11	14	23	22	46	23	8	10	20	10	17	9	70.8
30-Oct	33	38	76	27	32	14	13	28	10	8	12	16	14	13	14	13	11	11	7	9	18	13	10	22	75.6
31-Oct	9	16	27	41	21	20	74	18	49	79	54	85	40	38	20	31	26	28	14	11	37	26	17	10	85.4
85.4 93.5 78.0 83.9 86.0 92.4 73.9 89.8 76.9 79.3 81.7 86.7 76.9 86.4 86.2 62.1 96.4 91.2 61.2 91.5 67.1 84.1 88.9 93.0																									

PAZA

Donnelly Station

Monthly Summary Tables, Graphs and Roses

Hourly Averages

Sulphur Dioxide (SO₂) - ppb


Donnelly - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 1.5 ppb on Oct 17 11:00 Maximum Daily Average: 0.4 ppb on Oct 16																			Hours in Service: 744 Hours of Data: 709 Hours of Missing Data: 35 Hours of Calibration: 35 Percent Operational Time: 100.0							
Minimum Value: 0 ppb on Oct 4 03:00 Maximum Diurnal Average: 0.2 ppb at hour 11 Monthly Average: 0.11 ppb																			Minimum Daily Average: 0.0 ppb on Oct 12 Minimum Diurnal Average: 0.1 ppb at hour 22 Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.1 Q ₃ = 0.1 P ₉₀ = 0.3 P ₉₉ = 0.7							
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
3-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
4-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.5
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.0
10-Oct	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
11-Oct	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0.0	0.0
13-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0.0	0.0
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0.1	0.2
15-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0.1	0.1
16-Oct	0	0	0	0	0	0	0	0	0	1	1	1	1	0	0	1	1	0	0	A	0	0	0	1	0.4	0.8
17-Oct	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	A	0	0	0	0	0	0.2	1.5
18-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.0	0.1
19-Oct	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	A	0	0	0	0	0	0	0	0.2	0.5
20-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0.1	0.2
21-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0.2	0.3
22-Oct	0	0	0	0	0	0	0	0	0	0	0	1	0	A	0	0	0	0	0	0	0	0	0	0	0.2	0.5
23-Oct	0	0	0	0	0	0	0	0	1	0	1	1	A	0	0	1	0	0	0	0	0	0	0	0	0.3	0.6
24-Oct	0	0	0	0	0	0	0	0	0	0	1	A	1	C	C	C	1	0	0	0	0	0	0	0	0.3	0.7
25-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
26-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
27-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.1
28-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
29-Oct	0	0	0	0	0	0	A	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5
30-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.2
31-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
0.1 0.1 0.1 0.1 0.1 0.1 0.1 0.1 0.1 0.1 0.2 0.2 0.2 0.1 0.1 0.1 0.1 0.1 0.1 0.1 0.1 0.1 0.1 0.1																									Diurnal Average	
0.3 0.3 0.2 0.2 0.4 0.5 0.4 0.4 0.5 0.8 1.5 0.7 0.7 0.5 0.4 0.7 0.6 0.5 0.4 0.4 0.3 0.2 0.4 0.8																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 172 ppb 24-hr 48 ppb 30-day 11 ppb																										

Hourly Averages

Sulphur Dioxide (SO₂) - ppb

Donnelly - October 2018



Hourly Maximums

Sulphur Dioxide (SO₂) - ppb


Donnelly - October 2018

Maximum Value: 13.5 ppb on Oct 29 06:00																			Maximum Daily Average: 1.1 ppb on Oct 29						Hours in Service: 744	
Minimum Value: 0 ppb on Oct 12 05:00																			Minimum Daily Average: 0.3 ppb on Oct 12						Hours of Data: 709	
Maximum Diurnal Average: 0.9 ppb at hour 6																			Minimum Diurnal Average: 0.4 ppb at hour 3						Hours of Missing Data: 35	
Monthly Average: 0.52 ppb																			Percentiles: P ₁ = 0.2 P ₁₀ = 0.3 Q ₁ = 0.3 Median = 0.4 Q ₃ = 0.6 P ₉₀ = 0.7 P ₉₉ = 1.5						Hours of Calibration: 35	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	0	0	0	0	0	0	0	0	1	A	0	1	0	1	0	0	0	1	0	1	0	0	0.4	0.7
2-Oct	0	0	0	0	0	0	1	0	0	0	A	0	0	0	0	1	0	1	0	0	1	1	0	0	0.4	0.6
3-Oct	0	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	0	0	0	0	0	1	0	0.6	0.8
4-Oct	0	0	0	0	0	0	0	0	A	1	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0.5	1.1
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0.4	0.6
6-Oct	0	0	0	0	0	1	A	1	0	0	0	0	0	1	1	1	1	0	0	0	0	0	0	0	0.5	0.8
7-Oct	0	0	0	0	0	A	0	0	1	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0.4	0.6
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.5
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.4
10-Oct	0	0	A	0	0	0	0	1	0	0	0	1	0	1	1	1	1	0	1	1	0	0	0	0	0.5	0.7
11-Oct	1	A	0	0	0	1	1	1	1	1	1	1	1	0	1	1	1	0	0	0	0	0	0	0	0.5	0.7
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0.3	0.4
13-Oct	0	0	0	0	1	0	1	0	0	7	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0.6	6.6
14-Oct	0	0	0	0	0	0	0	0	0	0	0	1	1	0	1	1	0	0	1	0	0	A	0	0	0.4	0.6
15-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	A	0	0	0	0.4	0.7
16-Oct	1	0	0	1	1	1	1	1	1	2	2	1	1	1	1	1	1	1	1	A	1	1	1	1	0.9	1.8
17-Oct	1	1	0	0	0	0	0	1	1	1	9	3	1	1	0	0	1	0	A	0	0	0	0	0	1.0	9.3
18-Oct	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	A	0	0	0	0	1	1	0.4	0.7
19-Oct	1	0	0	0	1	1	1	1	0	1	1	1	1	1	1	0	A	1	1	1	0	1	0	0	0.6	1.1
20-Oct	1	1	1	1	1	1	1	0	0	0	0	0	0	1	0	A	0	0	0	0	0	1	1	1	0.5	0.8
21-Oct	1	0	0	1	1	1	1	1	1	1	1	1	1	0	A	0	1	0	1	0	0	0	0	0	0.5	0.8
22-Oct	0	0	0	1	0	0	1	1	1	1	1	1	1	A	1	1	1	1	1	0	1	0	1	1	0.6	1.0
23-Oct	1	1	1	1	1	0	0	1	1	1	1	1	A	1	1	1	1	1	1	0	1	1	1	1	0.7	1.3
24-Oct	1	1	0	0	1	0	0	0	0	1	2	A	2	C	C	C	1	1	1	1	1	1	1	1	0.7	1.7
25-Oct	1	1	0	1	0	0	0	0	0	0	A	1	1	0	0	0	0	0	0	0	0	0	0	0	0.4	0.8
26-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	1	1	1	0	0	0	0	0	0	0.4	0.8
27-Oct	0	1	0	0	0	0	1	0	A	0	0	0	0	0	1	0	0	0	0	0	0	1	0	0	0.4	0.5
28-Oct	1	1	1	0	0	0	0	A	0	0	0	0	0	0	0	0	1	1	1	1	1	0	0	1	0.5	0.9
29-Oct	0	1	1	1	0	13	A	1	0	1	1	1	1	1	1	1	1	0	0	0	0	0	0	0	1.1	13.5
30-Oct	1	0	0	0	0	A	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0.4	0.8
31-Oct	0	0	0	0	A	0	0	0	0	0	1	1	0	0	0	1	0	0	0	0	0	0	0	0	0.4	0.6
																								Diurnal Average		
																								Diurnal Maximum		
C - Calibration																										
A - Automated Daily Zero Span																										

Hourly Maximums

Sulphur Dioxide (SO₂) - ppb


Donnelly - October 2018




Pollutant Rose

Sulphur Dioxide (SO₂) - ppb

Donnelly - October 2018




Pollutant Classes (ppb)



Span Responses

Sulphur Dioxide (SO₂)
Donnelly - October 2018



Hourly Averages

Hydrogen Sulphide (H₂S) - ppb


Donnelly - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0																				Hours in Service: 744						
Maximum Value: 1.9 ppb on Oct 12 06:00																				Maximum Daily Average: 0.5 ppb on Oct 13		Hours of Data: 710				
Minimum Value: 0 ppb on Oct 7 16:00																				Minimum Daily Average: 0.0 ppb on Oct 8		Hours of Missing Data: 34				
Maximum Diurnal Average: 0.3 ppb at hour 19																				Minimum Diurnal Average: 0.1 ppb at hour 14		Hours of Calibration: 34				
Monthly Average: 0.18 ppb																				Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.1 Median = 0.1 Q ₃ = 0.2 P ₉₀ = 0.3 P ₉₉ = 0.9		Percent Operational Time: 100.0				
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	1	1	1	1	1	0.3	1.1
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.4
3-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
4-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
7-Oct	0	0	0	0	0	A	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.5
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2
10-Oct	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
11-Oct	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3
12-Oct	A	0	0	0	0	2	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	1	A	0.3	1.9
13-Oct	1	1	1	1	1	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0.5	1.0
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0	0	A	0	0	0.2	0.6
15-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0.3	0.5
16-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0.1	0.2
17-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0.0	0.2
18-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.1	0.2
19-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0.1	0.5
20-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0.1	0.5
21-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0.2	0.4
22-Oct	0	0	0	0	0	0	0	0	1	0	0	0	0	A	0	0	0	0	0	0	0	0	1	0	0.1	0.7
23-Oct	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	1	0	0	0	0	0	0.1	1.1
24-Oct	0	0	0	0	0	0	0	1	0	0	0	C	C	C	0	0	0	0	0	0	0	0	0	0	0.2	0.8
25-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
26-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
27-Oct	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3
28-Oct	0	0	0	0	0	0	0	A	0	0	0	0	1	0	0	0	0	0	0	0	0	0	1	0	0.3	0.8
29-Oct	1	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0.3	1.1
30-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.5
31-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0.2	0.8
0.2 0.1 0.1 0.1 0.2 0.3 0.2 0.2 0.2 0.2 0.2 0.1 0.1 0.1 0.1 0.1 0.1 0.2 0.3 0.2 0.2 0.2 0.2 0.2																									Diurnal Average	
0.8 0.6 0.6 0.6 1.0 1.9 0.8 1.0 0.9 0.6 0.5 0.3 0.5 0.4 0.4 0.5 0.5 0.4 1.1 0.9 0.9 1.0 1.1 0.7																									Diurnal Maximum	
C - Calibration A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 10 ppb 24-hr 3 ppb																										

Hourly Averages

Hydrogen Sulphide (H₂S) - ppb

Donnelly - October 2018



Hourly Maximums

Hydrogen Sulphide (H₂S) - ppb


Donnelly - October 2018

Maximum Value: 14.3 ppb on Oct 29 19:00																				Maximum Daily Average: 1.5 ppb on Oct 29										Hours in Service: 744									
Minimum Value: 0 ppb on Oct 7 17:00																				Minimum Daily Average: 0.6 ppb on Oct 17										Hours of Data: 710									
Maximum Diurnal Average: 1.6 ppb at hour 19																				Minimum Diurnal Average: 0.7 ppb at hour 14										Hours of Missing Data: 34									
Monthly Average: 0.85 ppb																				Percentiles: P ₁ = 0.4 P ₁₀ = 0.5 Q ₁ = 0.6 Median = 0.7 Q ₃ = 0.9 P ₉₀ = 1.0 P ₉₉ = 3.3										Hours of Calibration: 34									
																				Percent Operational Time: 100.0																			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum													
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24															
1-Oct	1	1	1	0	1	1	1	1	1	1	1	A	0	1	1	1	1	1	0	2	2	2	2	1	0.9	1.7													
2-Oct	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	0.7	1.0													
3-Oct	1	1	1	0	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.0													
4-Oct	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.1													
5-Oct	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.7	0.9													
6-Oct	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.7	1.1													
7-Oct	1	2	1	1	1	A	1	1	1	1	3	1	1	1	0	0	0	1	1	1	1	0	0	1	0.8	2.7													
8-Oct	0	1	1	1	A	0	0	1	0	1	1	0	1	0	1	1	1	1	1	1	1	1	1	1	0.6	1.2													
9-Oct	1	1	1	A	1	1	1	1	1	1	1	1	0	1	1	1	1	1	1	1	1	1	1	1	0.6	1.2													
10-Oct	0	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.0													
11-Oct	1	A	1	1	1	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	1	1	0.8	2.2													
12-Oct	A	1	1	1	1	4	2	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1	A	1.0	4.1													
13-Oct	1	2	1	1	2	2	1	1	2	2	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1.1	1.7													
14-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	0.8	1.2													
15-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	0.9	1.1													
16-Oct	1	1	1	1	1	1	1	0	1	1	1	0	1	0	1	0	1	1	1	A	1	0	1	0	0.6	0.9													
17-Oct	1	0	1	1	0	1	1	1	1	1	0	1	1	0	1	0	0	1	A	1	1	1	0	1	0.6	0.8													
18-Oct	1	1	1	1	1	1	1	1	1	1	0	1	1	0	1	1	0	A	1	0	1	1	1	1	0.7	1.0													
19-Oct	1	1	1	1	1	1	1	1	9	3	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1.2	9.4													
20-Oct	2	1	1	1	1	1	1	1	1	1	1	0	1	1	0	A	1	0	1	1	1	1	1	1	0.7	2.1													
21-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	0.8	1.1													
22-Oct	1	1	1	1	1	0	1	1	4	1	1	1	1	A	1	1	1	2	1	0	1	1	5	1	1.0	4.9													
23-Oct	1	1	1	1	1	1	1	1	1	1	1	0	A	1	1	1	1	1	11	1	1	1	1	1	1.1	10.9													
24-Oct	1	0	1	1	1	1	1	14	1	1	1	C	C	C	1	1	1	1	1	1	1	1	1	1	1.4	13.9													
25-Oct	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	0	1	1	1	1	1	0.7	1.0													
26-Oct	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.7	0.9													
27-Oct	1	1	1	1	1	1	1	1	A	1	1	1	0	1	1	1	1	1	1	1	1	1	1	1	0.7	0.9													
28-Oct	1	1	1	1	1	1	1	A	1	1	1	1	3	1	1	1	1	1	1	1	1	2	3	1	1.1	3.1													
29-Oct	1	1	1	1	1	2	A	1	1	1	1	1	1	1	1	1	1	1	14	1	1	1	1	1	1.5	14.3													
30-Oct	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.9	1.3													
31-Oct	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1	1	1	0.9	1.7													
0.8 0.8 0.7 0.7 0.8 0.9 0.8 1.2 1.2 0.9 0.8 0.7 0.8 0.7 0.7 0.7 0.7 0.8 1.6 0.8 0.8 0.8 1.0 0.7																									Diurnal Average														
2.1 2.0 1.2 1.5 1.6 4.1 1.8 13.9 9.4 3.4 2.7 1.1 2.6 1.3 1.2 1.2 1.3 2.1 14.3 1.7 1.6 1.9 4.9 1.3																									Diurnal Maximum														
C - Calibration																									A - Automated Daily Zero Span														

Hourly Maximums

Hydrogen Sulphide (H₂S) - ppb


Donnelly - October 2018



Pollutant Rose

Hydrogen Sulphide (H₂S) - ppb


Donnelly - October 2018



Span Responses

Hydrogen Sulphide (H_2S)


Donnelly - October 2018



Hourly Averages

External Temperature (ET) - °C


Donnelly - October 2018



Wind Rose

Wind Speed (WS) (km/h)

Donnelly - October 2018



Hourly Averages - Wind Speed (Scalar)

Wind Speed (km/h)

Donnelly - October 2018

Maximum Speed: 45 km/h on Oct 20 14:00																				Maximum Daily Speed Average: 25.3 km/h on Oct 20										Hours in Service: 744	
Minimum Speed: 1 km/h on Oct 28 23:00																				Minimum Daily Speed Average: 5.3 km/h on Oct 5										Hours of Data: 744	
Maximum Diurnal Speed Average: 13.3 km/h at hour 15																				Minimum Diurnal Speed Average: 9.0 km/h at hour 5										Hours of Missing Data: 0	
Monthly Average Speed: 11.06 km/h																				Percentiles: P ₁ = 2.2 P ₁₀ = 4.5 Q ₁ = 6.9 Median = 9.9 Q ₃ = 13.7 P ₉₀ = 18.1 P ₉₉ = 34.2										Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	5	3	3	2	2	1	2	2	5	5	7	10	10	12	14	15	17	19	16	14	13	13	14	12	9.1	19.2					
2-Oct	10	10	11	9	4	1	4	3	4	8	12	14	15	17	18	17	18	15	11	14	18	14	14	15	11.6	18.1					
3-Oct	12	10	11	7	8	7	5	5	5	2	3	4	4	4	5	5	3	3	4	5	8	11	10	12	6.4	11.8					
4-Oct	12	11	9	8	6	5	7	5	8	10	9	9	8	8	12	11	11	8	2	3	4	4	4	6	7.6	12.4					
5-Oct	5	5	4	5	4	4	3	4	6	6	4	4	7	6	5	4	5	4	4	4	4	8	9	10	5.3	10.0					
6-Oct	9	9	9	7	4	9	8	9	10	9	11	8	9	9	10	7	10	8	9	10	11	9	7	6	8.6	11.5					
7-Oct	9	10	8	6	10	11	12	10	11	14	13	14	15	13	16	15	18	17	14	14	13	10	14	12	12.5	18.4					
8-Oct	12	11	10	7	9	10	9	8	9	12	13	14	12	10	9	8	11	11	9	10	12	13	16	14	10.7	15.6					
9-Oct	11	9	8	9	9	9	10	10	11	9	8	9	9	10	9	9	10	10	9	10	10	9	9	9	9.5	11.1					
10-Oct	9	8	8	9	9	10	17	16	14	12	7	8	9	7	10	8	8	7	8	7	8	7	4	5	9.0	16.7					
11-Oct	6	5	6	8	6	7	9	11	10	14	14	13	14	14	13	12	13	10	10	7	7	11	11	9	10.0	14.5					
12-Oct	11	9	8	4	2	3	12	10	15	19	21	19	18	18	16	14	13	10	8	7	6	6	6	6	10.9	21.2					
13-Oct	6	6	6	4	4	6	9	5	5	4	4	5	7	9	10	11	12	7	8	10	13	14	14	14	8.1	14.5					
14-Oct	14	14	15	14	17	16	18	17	18	18	17	17	18	21	16	16	17	15	10	18	20	22	23	22	17.3	23.4					
15-Oct	19	14	13	10	16	15	12	12	11	11	12	10	11	12	7	5	4	9	8	5	4	6	7	8	10.1	19.0					
16-Oct	9	11	9	10	10	14	15	14	14	11	13	10	13	12	13	13	9	7	12	8	11	15	15	15	11.8	15.4					
17-Oct	15	14	13	11	12	11	15	18	20	18	21	19	24	26	27	28	20	15	15	19	23	21	23	25	18.8	28.2					
18-Oct	22	17	14	11	12	12	16	14	17	22	24	26	23	22	25	25	17	12	12	12	6	8	8	7	16.0	26.3					
19-Oct	10	14	9	9	8	8	7	8	5	9	2	4	8	11	11	13	10	6	9	12	12	9	9	9	8.9	13.9					
20-Oct	7	15	17	13	14	16	24	26	29	34	36	40	41	45	42	31	30	24	17	20	20	20	25	22	25.3	44.6					
21-Oct	18	17	18	13	12	9	7	7	8	8	11	13	14	13	14	15	10	7	10	9	8	5	9	5	10.7	18.1					
22-Oct	5	5	5	7	7	6	6	4	5	5	10	13	12	10	11	8	6	7	6	7	6	9	7	6	7.2	12.5					
23-Oct	5	5	5	5	5	7	5	10	11	10	9	7	5	3	5	6	4	4	4	5	4	3	5	8	5.9	11.5					
24-Oct	6	6	5	4	6	2	2	3	3	4	3	4	7	8	8	7	3	4	10	11	14	12	10	10	6.3	13.7					
25-Oct	10	8	7	6	5	2	5	9	10	10	7	5	8	10	11	8	8	16	34	24	14	15	16	15	11.0	34.0					
26-Oct	16	17	17	16	13	18	15	11	11	14	11	11	6	8	11	5	7	6	10	10	13	14	16	17	12.3	18.0					
27-Oct	24	25	26	27	27	25	27	26	29	37	37	34	33	29	27	23	17	14	12	12	12	10	12	11	23.2	36.9					
28-Oct	12	9	6	4	6	8	9	9	9	9	10	11	10	7	9	8	9	12	13	11	7	3	1	3	8.1	12.9					
29-Oct	5	8	10	8	6	7	10	10	4	4	7	7	8	8	9	7	4	4	5	5	11	12	12	13	7.6	12.5					
30-Oct	12	10	8	14	16	12	9	10	9	12	12	14	15	14	9	5	6	10	16	16	15	15	12	13	11.8	16.4					
31-Oct	13	14	14	13	12	10	10	12	12	12	13	13	12	11	10	15	9	6	6	6	9	12	16	16	11.5	16.4					
10.9 10.7 10.1 9.1 9.0 9.1 10.3 10.4 10.9 12.0 12.3 12.4 13.2 13.1 13.3 12.2 11.0 9.9 10.4 10.6 10.8 10.9 11.6 11.4																									Diurnal Average						
23.8 25.3 26.1 27.3 27.0 24.9 27.5 26.2 29.4 36.9 36.9 40.2 40.9 44.6 41.5 31.1 29.6 23.5 34.0 23.7 22.8 22.5 25.4 24.9																									Diurnal Maximum						
All monthly, daily, and diurnal averages have been calculated using scalar methods																															

Hourly Standard Deviations

Wind Direction (WD) - deg

Donnelly - October 2018

Maximum Value: 86.1 deg on Oct 29 00:00																	Hours in Service: 744									
Minimum Value: 1.9 deg on Oct 17 01:00																	Hours of Data: 744									
Percentiles: P ₁ = 2.2 P ₁₀ = 3.5 Q ₁ = 5.1 Median = 7.6 Q ₃ = 13.4 P ₉₀ = 25.2 P ₉₉ = 64.9																	Hours of Missing Data: 0									
																	Hours of Calibration: 0									
																	Percent Operational Time: 100.0									
Day	Hourly Period Ending At (MST)																									Daily Maximum
1-Oct	10	10	31	14	27	57	21	17	40	15	14	24	11	14	11	9	5	6	6	7	6	7	7	7	7	57.1
2-Oct	8	6	5	6	16	24	23	25	32	12	20	15	22	17	14	13	10	7	9	12	3	4	5	5	5	32.4
3-Oct	3	4	7	9	7	12	14	12	19	49	52	79	83	55	77	37	63	42	11	6	5	4	4	5	5	83.0
4-Oct	4	4	6	10	8	12	9	14	9	12	16	19	23	27	18	18	12	16	18	65	24	23	28	18	18	64.5
5-Oct	8	20	30	10	19	26	14	35	17	21	57	54	24	38	54	52	18	18	12	17	14	8	7	7	7	57.2
6-Oct	6	6	5	9	16	39	18	13	36	16	21	24	21	20	26	32	10	7	6	9	9	9	8	11	11	38.7
7-Oct	7	6	6	12	9	5	5	6	6	4	4	4	4	5	5	4	6	6	3	3	4	5	4	3	3	11.8
8-Oct	4	5	6	7	5	5	6	9	9	6	6	6	8	9	11	10	9	8	6	8	6	5	4	5	5	10.6
9-Oct	8	11	11	8	9	7	5	4	6	9	11	11	11	10	12	11	8	7	7	6	6	9	6	6	6	12.4
10-Oct	7	6	7	6	6	6	4	3	4	7	14	10	16	43	13	9	7	10	7	19	11	9	22	7	7	42.9
11-Oct	11	17	7	12	29	6	7	7	8	7	7	5	7	9	11	10	5	4	10	12	6	10	4	10	10	28.9
12-Oct	5	3	7	30	56	38	5	8	8	7	6	5	7	6	6	7	11	7	6	6	12	7	7	19	19	56.4
13-Oct	12	6	18	20	13	13	6	8	7	22	36	29	22	20	15	11	9	20	9	4	4	4	4	2	2	35.7
14-Oct	2	3	3	3	2	3	3	3	3	4	6	6	7	7	4	13	3	5	7	4	4	2	2	3	3	12.7
15-Oct	3	5	2	6	3	4	7	5	9	14	11	14	5	6	6	30	9	5	27	57	22	10	9	8	8	57.3
16-Oct	8	4	5	3	8	3	3	3	4	10	8	6	12	8	8	8	5	19	3	31	9	3	6	3	3	31.2
17-Oct	2	3	7	12	7	5	6	4	3	4	12	6	6	7	5	3	11	5	5	10	4	4	7	3	3	12.1
18-Oct	5	4	10	8	4	3	3	5	6	8	5	6	8	9	12	9	5	7	4	4	82	12	6	5	5	82.4
19-Oct	9	5	4	4	3	6	4	6	61	30	65	59	32	32	12	8	8	8	4	4	6	6	17	25	25	65.3
20-Oct	36	9	7	9	8	5	4	3	5	3	4	4	6	6	6	3	3	4	7	5	4	3	3	3	3	35.9
21-Oct	5	2	3	12	10	5	7	14	6	21	13	11	13	16	11	8	4	30	7	15	15	19	2	25	25	29.5
22-Oct	14	13	12	20	5	7	8	7	18	11	13	7	9	9	9	14	9	5	4	7	14	6	14	22	22	21.5
23-Oct	10	8	10	11	18	10	8	7	5	11	18	21	32	51	11	14	11	20	12	9	40	41	12	10	10	51.2
24-Oct	18	9	6	7	81	38	51	41	58	33	31	44	19	19	14	15	47	17	6	4	5	4	6	7	7	81.1
25-Oct	4	5	10	7	14	74	44	13	9	6	13	19	11	8	11	12	18	37	7	7	20	3	8	5	5	74.2
26-Oct	4	4	3	3	8	7	5	10	4	4	8	6	30	28	7	41	28	11	4	4	3	5	4	11	11	41.4
27-Oct	4	4	3	3	3	3	3	3	4	5	5	5	4	4	5	5	7	2	9	7	6	6	4	7	7	9.1
28-Oct	3	5	19	25	18	6	4	7	6	5	6	7	5	7	11	6	34	21	5	11	16	57	49	86	86	86.1
29-Oct	11	8	5	7	7	20	9	6	22	18	11	16	17	13	10	13	13	19	16	13	5	4	5	4	4	21.9
30-Oct	3	4	7	10	5	4	3	10	3	7	5	8	11	6	11	11	16	9	24	5	4	5	5	6	6	24.4
31-Oct	7	2	3	3	3	7	5	2	3	3	3	3	9	17	11	6	7	26	4	16	14	13	8	6	6	26.4
35.9 20.2 30.9 29.6 81.1 74.2 50.6 40.5 60.8 49.0 65.3 79.4 83.0 55.4 77.1 51.8 62.8 41.8 26.7 64.5 82.4 56.6 48.8 86.1																										

PAZA


Portable – Rycroft Station

Monthly Summary Tables, Graphs and
Roses

Number of Exceedences (AAAQO): Maximum Value: 1.8 ppb on Oct 4 11:00																			1-hr: 0 24-hr: 0 Maximum Daily Average: 0.5 ppb on Oct 25						Hours in Service: 635 Hours of Data: 601 Hours of Missing Data: 34 Hours of Calibration: 34 Percent Operational Time: 100.0						
Minimum Value: 0 ppb on Oct 1 01:00 Maximum Diurnal Average: 0.4 ppb at hour 15 Monthly Average: 0.20 ppb																			Minimum Daily Average: 0.0 ppb on Oct 12 Minimum Diurnal Average: 0.1 ppb at hour 9 Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.2 Q ₃ = 0.3 P ₉₀ = 0.4 P ₉₉ = 1.1												
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1					
2-Oct	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2					
3-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.5					
4-Oct	0	0	0	0	0	0	0	0	A	1	2	1	0	0	0	0	0	0	0	0	0	0	0	0	0.3	1.8					
5-Oct	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2					
6-Oct	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2					
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.0	0.1					
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2					
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.3					
10-Oct	0	0	A	0	0	0	0	0	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0.3	0.5					
11-Oct	0	A	0	0	0	0	0	0	0	0	0	1	0	0	1	1	1	0	0	0	0	0	0	0	0.2	1.0					
12-Oct	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0.0	0.1					
13-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0.1	0.4					
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0.2	0.4					
15-Oct	0	1	1	0	0	0	0	0	0	1	1	0	0	0	2	2	0	0	0	0	A	0	0	0	0.5	1.6					
16-Oct	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	1	0	0.3	0.5					
17-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	A	0	0	0	0	0	0.3	1.2					
18-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0.2	0.3					
19-Oct	0	1	1	1	0	0	0	0	0	0	1	1	1	1	0	0	A	0	0	0	0	0	0	0	0.4	1.1					
20-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	1	0	0	0	0	0.3	0.6					
21-Oct	0	0	0	0	0	1	0	0	0	0	0	0	0	0	A	0	0	0	0	0											

Hourly Averages

Sulphur Dioxide (SO₂) - ppb
Portable Rycroft - October 2018




Hourly Maximums

Sulphur Dioxide (SO₂) - ppb Portable Rycroft - October 2018

Maximum Value: 25.0 ppb on Oct 23 17:00																			Maximum Daily Average: 1.8 ppb on Oct 23										Hours in Service: 635	
Minimum Value: 0 ppb on Oct 23 18:00																			Minimum Daily Average: 0.5 ppb on Oct 12										Hours of Data: 601	
Maximum Diurnal Average: 2.0 ppb at hour 17																			Minimum Diurnal Average: 0.7 ppb at hour 5										Hours of Missing Data: 34	
Monthly Average: 0.89 ppb																			Percentiles: P ₁ = 0.2 P ₁₀ = 0.5 Q ₁ = 0.6 Median = 0.7 Q ₃ = 0.9 P ₉₀ = 1.3 P ₉₉ = 2.6										Hours of Calibration: 34	
																			Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum				
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24						
1-Oct	1	1	0	1	1	1	1	1	1	1	1	A	1	1	1	1	1	0	0	1	1	1	1	1	0.6	0.8				
2-Oct	0	1	1	1	1	1	1	1	1	1	A	1	1	1	2	1	2	2	1	1	2	1	1	1	0.9	2.2				
3-Oct	1	1	1	1	1	1	1	1	1	A	2	1	1	1	1	1	1	1	1	1	1	1	1	0	0.8	1.7				
4-Oct	1	1	1	1	0	1	1	1	A	2	4	2	1	1	3	1	1	1	1	1	1	1	1	1	1.1	4.4				
5-Oct	0	0	0	1	1	1	1	A	1	1	1	0	1	1	1	1	1	1	1	1	1	1	1	1	0.7	1.1				
6-Oct	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	1	1	0.7	1.0				
7-Oct	1	1	1	0	1	A	1	1	0	1	1	1	1	1	1	0	1	1	1	1	1	1	0	1	0.6	0.9				
8-Oct	0	1	0	0	A	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.6	0.8				
9-Oct	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.1				
10-Oct	1	1	A	1	1	1	1	1	1	1	3	2	1	1	1	1	1	1	1	1	1	1	1	1	1.0	2.6				
11-Oct	1	A	1	1	1	1	1	1	1	1	1	1	1	1	2	2	3	1	1	1	1	1	1	1	1.0	2.8				
12-Oct	A	1	1	1	0	1	1	0	0	1	0	0	0	1	0	1	0	1	0	1	0	0	0	A	0.5	0.9				
13-Oct	1	1	1	0	1	1	1	0	1	1	1	1	1	2	1	2	1	1	2	1	0	1	A	1	0.8	2.1				
14-Oct	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	0.8	1.4				
15-Oct	1	1	2	1	1	1	1	1	1	1	2	1	2	1	3	2	2	1	1	1	A	1	1	1	1.3	3.4				
16-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	0.9	1.3				
17-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	4	A	1	1	2	1	1	1.0	3.7				
18-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	A	1	1	1	1	1	0.8	1.4				
19-Oct	1	1	2	2	1	1	1	1	1	1	1	2	1	1	1	1	A	1	1	1	1	1	1	1	1.0	1.7				
20-Oct	1	1	1	1	1	1	1	2	1	1	1	1	0	1	1	A	1	1	1	2	1	1	1	1	0.9	1.7				
21-Oct	1	1	1	1	1	2	1	1	1	1	1	1	1	1	A	1	1	1	1	0	1	1	1	1	0.8	1.7				
22-Oct	1	1	1	0	0	0	1	0	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	0.9	1.4				
23-Oct	1	1	0	1	1	1	1	0	1	1	1	2	A	C	C	C	25	0	0	0	0	0	0	0	1.8	25.0				
24-Oct	0	0	0	0	0	0	1	0	0	1	1	A	2	1	1	1	1	1	1	1	1	1	1	1	0.8	1.9				
25-Oct	1	1	1	1	2	2	1	1	1	1	A	2	2	2	3	1	1	2	1	1	1	1	2	1	1.3	2.6				
26-Oct	1	1	1	1	1	1	1	1	1	A	1	1	1	1	2	1	1	1	1	1	1	1	0	0.8	1.7					
27-Oct	1	0	1	0	0	1	1	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	0.6					
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--					
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--					
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--					
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--					
0.8 0.8 0.7 0.7 0.7 0.8 0.7 0.7 0.7 0.9 1.2 1.0 0.9 1.0 1.3 1.0 2.0 1.0 0.8 0.8 0.8 0.7 0.8 0.8																									Diurnal Average					
1.4 1.4 1.7 1.6 1.6 1.7 1.3 1.7 1.4 1.8 4.4 2.0 1.9 2.5 3.4 2.3 25.0 3.7 1.6 1.7 2.0 1.8 2.0 1.4																									Diurnal Maximum					
C - Calibration				NS - Not in service							A - Automated Daily Zero Span																			


Hourly Maximums

Sulphur Dioxide (SO₂) - ppb
Portable Rycroft - October 2018




Pollutant Rose

Sulphur Dioxide (SO₂) - ppb
Portable Rycroft - October 2018



Span Responses


Sulphur Dioxide (SO₂)
Portable Rycroft - October 2018



233 of 390

Hourly Averages

Total Reduced Sulphur (TRS) - ppb
Portable Rycroft - October 2018




Maximum Value: 8.4 ppb on Oct 4 02:00																			Maximum Daily Average: 1.6 ppb on Oct 13										Hours in Service: 635	
Minimum Value: 0 ppb on Oct 9 22:00																			Minimum Daily Average: 0.5 ppb on Oct 3										Hours of Data: 600	
Maximum Diurnal Average: 0.9 ppb at hour 2																			Minimum Diurnal Average: 0.6 ppb at hour 13										Hours of Missing Data: 35	
Monthly Average: 0.72 ppb																			Percentiles: P ₁ = 0.3 P ₁₀ = 0.4 Q ₁ = 0.5 Median = 0.6 Q ₃ = 0.8 P ₉₀ = 1.0 P ₉₉ = 2.7										Hours of Calibration: 35	
																			Percent Operational Time: 100.0											
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum				
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24						
1-Oct	1	1	1	1	1	1	1	1	1	1	1	A		1	0	0	1	1	0	1	0	0	0	0	0.6	1.0				
2-Oct	0	1	0	1	1	1	1	1	1	0	A	1	0	1	1	0	1	1	1	0	0	0	1	1	0.6	1.1				
3-Oct	1	0	0	0	0	0	1	1	0	A	1	0	1	1	1	0	0	0	1	0	0	0	1	1	0.5	1.0				
4-Oct	5	8	1	1	1	1	1	1	A	1	1	0	1	0	0	0	1	1	1	1	0	1	0	1	1.1	8.4				
5-Oct	1	1	0	1	1	1	1	A	1	1	0	0	0	0	1	0	1	1	1	1	1	1	1	1	0.6	1.0				
6-Oct	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0.7	1.1				
7-Oct	1	1	1	1	1	A	1	0	1	1	0	1	1	1	1	1	1	1	1	1	1	0	1	0	0.6	1.1				
8-Oct	1	1	0	1	A	1	1	0	1	1	1	0	0	1	0	1	1	1	0	1	1	1	0	0	0.6	1.1				
9-Oct	1	1	1	A	1	0	1	0	1	0	0	1	1	0	0	0	1	1	1	0	0	1	0	0	0.5	1.0				
10-Oct	0	0	A	0	0	1	1	1	0	1	0	1	1	0	3	1	1	1	1	1	1	0	1	0	0.6	3.4				
11-Oct	0	A	1	1	0	1	1	0	0	1	0	0	0	0	0	1	1	1	1	1	1	1	1	1	0.6	1.1				
12-Oct	A	1	1	2	2	3	1	1	1	1	0	0	1	1	0	1	1	1	1	1	2	3	1	2	A	1.2	2.7			
13-Oct	2	1	2	2	2	2	2	2	2	2	1	1	1	1	1	1	1	3	3	3	2	1	A	1	1.6	3.4				
14-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	0.7	1.0				
15-Oct	1	0	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	0.7	1.2				
16-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	0.8	1.3				
17-Oct	1	1	1	1	1	1	1	1	0	C	C	C	C	0	0	0	1	1	A	1	0	1	1	0	0.6	0.8				
18-Oct	1	0	1	1	1	0	1	1	1	1	1	0	0	1	0	1	1	A	1	0	1	1	1	1	0.6	1.2				
19-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	0	A	1	1	0	0	1	1	1	0.7	1.0				
20-Oct	1	1	1	1	0	0	1	0	0	0	0	0	0	0	0	A	1	1	0	1	1	0	1	1	0.5	0.9				
21-Oct	0	1	1	1	0	1	1	1	1	0	1	1	0	1	A	1	0	1	1	1	1	1	1	1	0.6	1.1				
22-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	A	1	0	1	1	1	1	1	1	1	1	0.7	1.0				
23-Oct	1	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	0.7	0.9				
24-Oct	1	1	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	0.8	1.1				
25-Oct	1	1	1	1	1	1	1	1	1	1	A	1	0	1	1	0	1	1	1	1	1	0	1	1	0.6	0.9				
26-Oct	1	1	1	1	1	1	1	1	1	A	1	1	1	1	1	1	1	1	0	1	1	1	1	1	0.7	1.0				
27-Oct	1	0	1	1	0	1	1	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	0.7				
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--				
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--				
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--				
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--				
																									Diurnal Average	Diurnal Maximum				
																									0.9	4.6				
																									0.9	8.4				
																									0.7	1.5				
																									0.7	1.6				
																									0.8	2.4				
																									0.8	2.7				
																									0.7	1.6				
																									0.7	1.8				
																									0.7	2.1				
																									0.7	2.0				
																									0.6	1.5				
																									0.6	1.2				
																									0.6	0.8				
																									0.6	0.9				
																									0.7	3.4				
																									0.6	1.1				
																									0.8	1.2				
																									0.9	2.5				
																									0.8	3.3				
																									0.8	3.4				
																									0.8	2.7				
																									0.7	1.4				
																									0.7	1.6				
																									0.6	1.2				
																									</					

Hourly Maximums

Total Reduced Sulphur (TRS) - ppb


Portable Rycroft - October 2018



Pollutant Rose


Total Reduced Sulphur (TRS) - ppb

Portable Rycroft - October 2018



Span Responses

Total Reduced Sulphur (TRS)
Portable Rycroft - October 2018




Hourly Averages

Nitrogen Dioxide (NO₂) - ppb Portable Rycroft - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 14.4 ppb on Oct 17 21:00 Maximum Daily Average: 8.0 ppb on Oct 23																					Hours in Service: 635 Hours of Data: 599					
Minimum Value: 0 ppb on Oct 2 10:00 Maximum Diurnal Average: 3.8 ppb at hour 21 Monthly Average: 2.51 ppb																					Minimum Daily Average: 0.7 ppb on Oct 8 Minimum Diurnal Average: 1.7 ppb at hour 14 Percentiles: P ₁ = 0.4 P ₁₀ = 0.7 Q ₁ = 1.1 Median = 1.8 Q ₃ = 3.0 P ₉₀ = 5.5 P ₉₉ = 11.5		Hours of Missing Data: 36 Hours of Calibration: 36 Percent Operational Time: 100.0			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	8	5	4	5	3	3	4	3	3	2	1	A	1	1	1	2	2	2	2	1	1	1	1	0	2.4	7.8
2-Oct	0	1	0	0	1	2	2	1	1	0	A	1	1	1	1	1	2	2	2	3	4	2	2	2	1.3	3.7
3-Oct	2	2	3	2	2	2	2	3	3	A	3	2	1	1	1	1	2	2	3	3	3	1	2	4	2.2	4.0
4-Oct	6	6	2	2	2	3	4	6	A	5	4	3	2	2	2	2	1	2	8	5	4	6	3	1	3.6	7.7
5-Oct	2	1	1	1	3	1	1	A	3	1	1	1	1	1	1	1	1	1	2	3	3	3	4	4	1.8	3.7
6-Oct	4	4	3	2	2	4	A	3	3	2	2	2	2	2	2	2	2	2	3	1	1	1	1	1	2.3	4.4
7-Oct	1	1	1	1	1	A	2	1	1	1	2	2	2	2	2	2	2	2	2	2	1	1	1	1	1.4	2.2
8-Oct	1	2	2	2	A	2	1	1	1	0	0	0	0	0	1	0	0	0	0	0	1	0	0	0	0.7	1.7
9-Oct	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	2	3	1.0	2.6
10-Oct	3	3	A	3	2	2	2	2	2	2	3	4	5	5	3	3	3	4	4	4	4	3	2	2	3.1	5.4
11-Oct	2	A	2	2	2	3	5	6	4	4	3	2	2	2	3	4	3	7	5	6	6	4	3	3.5	6.5	
12-Oct	A	2	2	2	3	2	1	1	1	1	0	0	0	0	1	1	1	1	2	2	1	2	1	A	1.3	3.3
13-Oct	2	1	1	1	3	3	5	2	3	2	1	1	1	1	1	1	2	1	2	2	2	1	A	2	1.9	5.2
14-Oct	1	1	1	1	1	1	1	1	1	1	1	1	1	1	1	0	1	1	1	4	A	3	1	1.3	4.0	
15-Oct	1	1	1	1	1	2	3	3	2	2	6	3	2	2	2	5	1	1	3	3	A	3	2	2	2.3	5.7
16-Oct	1	1	2	2	2	2	2	2	2	1	1	3	1	1	1	2	2	4	2	A	2	1	1	1	1.8	4.3
17-Oct	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	2	2	2	A	3	14	6	1	1	2.0	14.4
18-Oct	1	1	2	2	1	1	2	3	6	8	3	2	1	1	1	1	3	A	5	8	3	3	3	3	2.7	8.5
19-Oct	3	2	2	2	3	2	3	2	6	5	3	2	2	2	2	2	A	3	3	3	7	9	3	3	3.2	8.9
20-Oct	3	8	4	2	1	1	1	1	1	0	0	0	0	1	1	A	1	1	1	2	1	1	2	2	1.5	7.7
21-Oct	1	1	1	1	1	1	2	1	2	3	2	1	1	0	A	1	1	3	2	2	2	2	1	1	1.5	2.8
22-Oct	1	2	2	2	2	2	2	6	4	4	5	5	4	A	5	5	6	6	7	6	6	9	10	12	4.8	11.5
23-Oct	8	6	5	5	5	7	6	6	4	4	5	7	A	C	C	C	C	C	13	11	14	14	12	11	8.0	14.2
24-Oct	10	10	7	8	6	6	7	7	7	7	9	A	10	10	9	7	9	9	9	6	6	5	3	3	7.3	9.7
25-Oct	4	4	3	3	2	2	3	3	3	3	A	3	1	1	1	1	1	1	1	1	1	1	1	1	1.9	4.2
26-Oct	1	1	1	1	1	1	1	1	1	A	2	2	2	3	4	4	4	4	4	3	3	3	1	1	2.1	3.9
27-Oct	1	1	1	0	1	0	1	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	0.7
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
2.7 2.6 2.1 2.1 2.0 2.2 2.5 2.7 2.6 2.5 2.5 2.1 1.9 1.7 1.8 2.0 2.2 2.5 3.5 3.2 3.8 3.5 2.7 2.6																								Diurnal Average		
9.7 9.7 7.1 7.8 6.0 6.7 6.9 7.1 7.5 8.5 9.0 7.2 9.7 9.7 8.8 6.8 9.1 9.0 13.3 10.5 14.4 14.2 11.9 11.5																								Diurnal Maximum		
C - Calibration NS - Not in service A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 159 ppb 24-hr 106 ppb																										

Hourly Averages

Nitrogen Dioxide (NO₂) - ppb
Portable Rycroft - October 2018




Hourly Maximums

Nitrogen Dioxide (NO₂) - ppb Portable Rycroft - October 2018

Maximum Value: 68.9 ppb on Oct 18 09:00																								Hours in Service: 635		
Maximum Daily Average: 10.4 ppb on Oct 23																								Hours of Data: 599		
Minimum Value: 0 ppb on Oct 2 10:00																								Hours of Missing Data: 36		
Maximum Diurnal Average: 9.7 ppb at hour 21																								Hours of Calibration: 36		
Monthly Average: 4.76 ppb																								Percent Operational Time: 100.0		
Percentiles: P ₁ = 0.6 P ₁₀ = 1.2 Q ₁ = 1.7 Median = 2.9 Q ₃ = 5.6 P ₉₀ = 10.1 P ₉₉ = 26.8																										
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	10	6	7	7	3	4	5	5	4	2	2	A	2	2	3	10	5	8	10	4	3	6	4	1	4.9	10.0
2-Oct	1	1	1	1	3	3	4	2	2	0	A	1	3	1	2	1	4	6	2	6	13	3	2	2	2.8	12.6
3-Oct	2	3	3	2	3	3	3	6	5	A	5	2	2	1	1	1	3	3	8	8	10	2	6	15	4.3	14.9
4-Oct	9	12	3	3	4	5	13	13	A	7	5	5	4	5	6	4	4	4	24	10	9	11	5	6	7.3	23.9
5-Oct	2	2	2	5	7	1	2	A	5	2	1	1	1	1	1	2	2	3	5	5	7	4	4	5	2.9	6.9
6-Oct	10	5	4	2	3	6	A	5	6	3	3	3	3	2	3	2	3	4	8	11	7	2	1	1	4.2	10.8
7-Oct	1	1	1	1	1	A	2	1	1	2	2	3	3	2	2	2	2	2	2	2	1	2	1	2	1.9	3.2
8-Oct	2	2	2	2	A	3	1	1	1	1	1	1	1	3	1	1	1	1	1	1	1	1	1	1	1.1	2.6
9-Oct	1	1	1	A	2	1	1	1	2	2	1	1	2	1	1	1	3	4	3	1	2	2	2	3	1.7	3.7
10-Oct	3	3	A	4	3	3	2	3	2	6	3	7	8	7	4	4	4	4	4	6	6	5	3	3	4.2	7.6
11-Oct	3	A	3	3	2	4	6	7	6	4	4	3	2	2	2	6	7	5	33	12	25	18	16	4	7.8	32.9
12-Oct	A	4	2	2	6	3	3	2	2	2	1	2	2	1	1	1	1	3	6	3	2	2	2	A	2.4	6.3
13-Oct	2	2	1	2	5	4	15	4	3	3	2	2	3	3	4	3	13	2	4	3	3	2	A	3	3.9	15.4
14-Oct	2	2	1	1	1	2	1	2	2	1	2	2	3	1	1	3	1	1	2	3	7	A	5	2	2.1	6.7
15-Oct	4	3	4	3	1	7	8	4	9	2	10	8	4	4	9	12	2	1	4	6	A	10	2	2	5.2	11.8
16-Oct	2	2	2	2	2	2	3	2	3	4	2	6	5	12	5	6	3	18	6	A	3	2	4	2	4.3	17.9
17-Oct	6	2	1	2	1	1	1	1	1	2	2	6	1	2	2	22	2	20	A	24	60	50	2	2	9.3	60.2
18-Oct	1	6	9	5	1	2	3	6	69	18	6	4	4	1	1	5	8	A	5	15	27	5	3	3	9.0	68.9
19-Oct	3	3	2	3	3	3	3	3	9	9	5	3	2	5	3	4	A	5	6	8	35	32	9	3	6.9	35.2
20-Oct	3	21	7	3	2	2	5	5	2	1	1	1	1	3	2	A	4	3	2	3	3	2	3	3	3.4	20.9
21-Oct	2	2	1	1	2	2	3	2	11	13	4	2	1	1	A	2	3	17	2	2	2	2	2	2	3.6	17.3
22-Oct	2	3	2	2	2	2	2	17	8	5	6	6	6	A	7	6	6	13	14	8	7	10	12	14	6.9	16.7
23-Oct	10	6	6	6	6	9	8	9	5	14	8	9	A	C	C	C	C	C	15	12	17	17	13	14	10.4	17.4
24-Oct	13	11	10	10	8	7	8	8	11	14	11	A	10	11	11	8	17	15	13	9	9	7	4	4	9.9	17.4
25-Oct	5	4	4	3	3	3	3	3	3	4	A	13	2	1	1	1	1	2	2	1	1	1	1	1	2.8	12.6
26-Oct	1	1	1	1	1	2	2	2	2	A	3	3	2	3	6	4	4	4	5	4	4	4	1	1	2.8	6.2
27-Oct	1	1	1	1	1	1	1	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	1.0
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
3.8 4.2 3.1 3.0 2.9 3.2 4.2 4.6 6.9 5.0 3.7 3.9 3.0 3.2 3.3 4.6 4.5 6.2 7.9 7.2 9.7 8.1 4.4 4.0																										Diurnal Average
12.6 20.9 9.5 9.6 7.7 9.4 15.4 16.7 68.9 18.0 11.0 12.6 10.2 12.3 10.9 21.7 17.4 20.0 32.9 26.7 60.2 49.6 16.4 14.9																										Diurnal Maximum
C - Calibration NS - Not in service A - Automated Daily Zero Span																										


Hourly Maximums

Nitrogen Dioxide (NO₂) - ppb
Portable Rycroft - October 2018




Pollutant Rose

Nitrogen Dioxide (NO₂) - ppb
Portable Rycroft - October 2018



Span Responses

Nitrogen Dioxide (NO₂)
Portable Rycroft - October 2018




Hourly Averages

Nitrogen Oxide (NO) - ppb Portable Rycroft - October 2018

Maximum Value: 22.7 ppb on Oct 17 21:00		Maximum Daily Average: 2.1 ppb on Oct 23		Hours in Service: 635																							
Minimum Value: 0 ppb on Oct 14 04:00		Minimum Daily Average: 0.1 ppb on Oct 8		Hours of Data: 599																							
Maximum Diurnal Average: 1.9 ppb at hour 21		Minimum Diurnal Average: 0.1 ppb at hour 5		Hours of Missing Data: 36																							
Monthly Average: 0.51 ppb		Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.0 Median = 0.1 Q ₃ = 0.3 P ₉₀ = 1.0 P ₉₉ = 6.5		Hours of Calibration: 36																							
				Percent Operational Time: 100.0																							
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
1-Oct	0	0	0	1	0	0	0	0	0	0	0	0	A	0	0	0	0	0	1	0	0	0	0	0	0.2	0.7	
2-Oct	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	1	0	0	0	0	0	0	0.1	0.5	
3-Oct	0	0	0	0	0	0	0	0	1	A	1	1	0	0	0	0	1	0	0	0	0	0	0	1	0.3	0.9	
4-Oct	0	0	0	0	0	0	2	3	A	3	2	1	1	1	1	1	0	0	1	0	0	0	0	0	0.8	3.0	
5-Oct	0	0	0	0	0	0	0	A	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.2	0.7	
6-Oct	0	0	0	0	0	0	A	0	0	1	1	1	1	1	0	0	0	0	1	0	1	0	0	0	0.4	1.0	
7-Oct	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.4	
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.2	
9-Oct	0	0	0	A	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0.1	0.3	
10-Oct	0	0	A	0	0	0	0	0	0	1	1	1	2	1	0	0	0	0	0	0	0	0	0	0	0.3	1.7	
11-Oct	0	A	0	0	0	0	0	0	1	1	1	1	0	0	0	1	1	0	2	0	5	3	1	0	0.8	5.0	
12-Oct	A	0	0	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0.1	0.6	
13-Oct	0	0	0	0	0	0	4	0	1	2	1	1	1	1	0	1	1	0	0	0	0	0	A	0	0.7	3.8	
14-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	A	1	0	0	0.1	0.9	
15-Oct	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	1	0	0	0	0	A	0	0	0	0.3	1.3	
16-Oct	0	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	0	1	0	A	0	0	0	0	0.1	0.7	
17-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	2	0	6	A	4	23	10	0	0	2.0	22.7	
18-Oct	0	0	0	0	0	0	0	1	6	5	1	1	1	0	0	0	1	A	A	0	3	0	0	0	0	0.8	6.5
19-Oct	0	0	0	0	0	0	0	0	1	3	2	1	1	1	1	1	A	0	0	0	5	4	0	0	0	0.9	4.7
20-Oct	0	1	0	0	0	0	0	0	0	0	0	0	0	0	0	A	0	0	0	0	0	0	0	0	0.1	1.5	
21-Oct	0	0	0	0	0	0	0	0	0	1	0	0	0	0	A	0	0	0	0	0	0	0	0	0	0.1	0.7	
22-Oct	0	0	0	0	0	0	0	1	1	2	2	3	3	A	2	2	1	1	0	0	0	0	0	0	0.7	2.6	
23-Oct	0	0	0	0	0	1	0	0	1	4	5	6	A	C	C	C	C	C	1	0	12	7	1	1	2.1	11.6	
24-Oct	1	0	0	0	0	0	0	0	3	6	10	A	7	6	4	2	2	0	0	0	0	0	0	0	1.9	10.2	
25-Oct	0	0	0	0	0	0	0	0	0	0	A	2	0	0	0	0	0	0	0	0	0	0	0	0	0.2	1.6	
26-Oct	0	0	0	0	0	0	0	0	0	A	0	0	0	1	1	0	0	0	0	0	0	0	0	0	0.2	0.5	
27-Oct	0	0	0	0	0	0	0	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	0.1	
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
0.1		0.1	0.1	0.1	0.1	0.1	0.3	0.3	0.8	1.2	1.3	0.9	0.8	0.6	0.5	0.5	0.4	0.5	0.3	0.4	1.9	1.0	0.2	0.1	Diurnal Average		
0.6		1.5	0.4	0.6	0.6	0.7	3.8	2.9	6.5	5.6	10.2	6.0	7.0	5.6	4.2	2.0	1.5	5.6	2.3	3.9	22.7	10.0	1.3	1.5	Diurnal Maximum		
C - Calibration		NS - Not in service							A - Automated Daily Zero Span																		

Hourly Averages

Nitrogen Oxide (NO) - ppb
Portable Rycroft - October 2018



Hourly Maximums


Nitrogen Oxide (NO) - ppb

Portable Rycroft - October 2018

Maximum Value: 236.4 ppb on Oct 17 18:00																								Hours in Service: 635		
Maximum Daily Average: 27.1 ppb on Oct 17																								Hours of Data: 599		
Minimum Value: 0 ppb on Oct 16 02:00																								Hours of Missing Data: 36		
Maximum Diurnal Average: 12.2 ppb at hour 18																								Hours of Calibration: 36		
Monthly Average: 3.56 ppb																								Percent Operational Time: 100.0		
Minimum Daily Average: 0.4 ppb on Oct 8																								Daily Average		
Minimum Diurnal Average: 0.3 ppb at hour 4																										
Percentiles: P ₁ = 0.1 P ₁₀ = 0.1 Q ₁ = 0.2 Median = 0.4 Q ₃ = 1.5 P ₉₀ = 4.5 P ₉₉ = 60.9																								Daily Maximum		
Day	Hourly Period Ending At (MST)																									
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	1	0	2	1	0	0	0	1	1	1	1	A	0	1	1	2	1	3	6	1	1	2	1	0	1.1	5.5
2-Oct	0	0	0	0	0	0	0	0	1	0	A	0	1	1	1	1	2	2	0	0	0	0	0	0	0.5	2.4
3-Oct	0	0	0	0	0	0	0	2	2	A	2	1	1	0	0	1	1	1	4	1	2	0	1	11	1.3	10.5
4-Oct	0	1	0	0	0	0	29	19	A	4	3	2	2	3	3	2	2	1	11	1	3	3	0	5	4.1	29.2
5-Oct	0	0	0	0	0	0	1	A	1	1	0	0	0	0	0	1	2	2	1	2	3	1	0	2	0.9	3.5
6-Oct	9	1	0	0	0	2	A	0	1	1	1	1	1	1	1	1	2	2	9	8	25	2	0	0	3.0	25.1
7-Oct	0	0	0	0	0	A	0	1	1	1	1	1	1	0	1	1	1	0	0	0	1	0	0	2	0.6	1.8
8-Oct	0	0	0	0	A	0	0	0	0	0	0	0	1	2	1	0	0	0	1	1	0	0	0	0	0.4	1.6
9-Oct	0	0	0	A	0	0	0	1	1	2	1	0	2	0	1	2	2	3	1	0	0	0	0	0	0.8	2.9
10-Oct	0	0	A	0	0	0	0	1	0	14	1	3	3	2	1	0	1	0	0	0	0	0	0	0	1.3	14.3
11-Oct	0	A	0	0	0	0	0	1	1	1	1	5	1	1	1	2	2	1	60	7	100	38	26	0	10.8	100.3
12-Oct	A	0	0	0	3	1	1	0	1	1	0	1	1	1	0	1	0	0	1	0	0	0	0	A	0.6	2.5
13-Oct	0	0	0	0	0	0	36	1	2	3	6	2	3	3	3	2	7	0	3	0	0	0	A	0	3.2	36.3
14-Oct	0	0	0	0	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	1	2	A	2	1	0.4	1.8
15-Oct	1	1	1	1	0	1	2	0	0	0	3	3	1	1	2	3	0	0	0	0	A	0	0	0	1.0	3.3
16-Oct	0	0	0	0	0	0	0	0	0	1	1	2	2	5	1	1	0	16	0	A	1	0	0	0	1.4	16.4
17-Oct	0	0	0	0	0	0	0	0	0	0	1	0	0	0	0	96	0	236	A	50	96	140	0	0	27.1	236.4
18-Oct	0	1	1	0	0	0	0	23	108	15	3	3	2	0	0	2	2	A	7	51	0	0	0	0	9.5	108.0
19-Oct	0	0	0	0	0	0	0	1	2	22	23	3	2	3	2	3	A	3	1	3	43	30	4	0	6.3	42.7
20-Oct	0	16	0	0	0	0	5	10	0	0	0	0	0	2	1	A	1	1	0	0	0	0	0	0	1.7	16.3
21-Oct	0	0	0	0	0	0	0	1	7	7	1	1	0	0	A	0	0	6	0	0	0	0	0	0	1.1	7.1
22-Oct	0	0	0	0	0	0	0	6	2	18	3	3	4	A	3	2	1	13	6	0	0	0	0	0	2.7	17.6
23-Oct	0	0	0	0	0	5	0	2	14	32	7	8	A	C	C	C	C	C	4	1	18	17	1	3	6.2	31.6
24-Oct	2	0	1	0	0	0	0	1	51	21	19	A	8	7	6	2	4	2	0	0	0	0	0	0	5.5	51.3
25-Oct	0	0	0	0	0	0	0	0	0	1	A	35	1	0	0	0	0	0	0	0	0	0	0	0	1.9	35.1
26-Oct	0	0	0	0	0	0	0	0	0	A	0	1	1	1	3	1	0	1	0	0	0	0	0	0	0.5	3.3
27-Oct	0	0	0	0	0	0	0	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	0.2
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
0.7 0.9 0.3 0.3 0.3 0.5 3.0 2.9 7.9 6.1 3.3 3.2 1.5 1.5 1.3 5.2 1.3 12.2 4.7 5.2 11.9 9.5 1.5 1.1																									Diurnal Average	
9.4 16.3 1.9 1.5 2.5 5.0 36.3 22.9 108.0 31.6 23.1 35.1 8.1 6.9 6.2 96.1 7.2 236.4 60.2 50.7 100.3 140.3 25.7 10.5																										Diurnal Maximum
C - Calibration NS - Not in service A - Automated Daily Zero Span																										

Hourly Maximums

Nitrogen Oxide (NO) - ppb
Portable Rycroft - October 2018



Hourly Averages


Oxides of Nitrogen (NO_x) - ppb

Portable Rycroft - October 2018

Maximum Value: 37.2 ppb on Oct 17 21:00																				Maximum Daily Average: 10.2 ppb on Oct 23					Hours in Service: 635	
Minimum Value: 0 ppb on Oct 2 10:00																				Minimum Daily Average: 0.8 ppb on Oct 8					Hours of Data: 599	
Maximum Diurnal Average: 5.8 ppb at hour 21																				Minimum Diurnal Average: 2.1 ppb at hour 5					Hours of Missing Data: 36	
Monthly Average: 3.04 ppb																				Percentiles: P ₁ = 0.4 P ₁₀ = 0.8 Q ₁ = 1.2 Median = 2.0 Q ₃ = 3.4 P ₉₀ = 6.5 P ₉₉ = 15.5					Hours of Calibration: 36	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	8	5	5	6	3	3	4	3	3	2	2	A	2	1	1	2	2	3	3	1	1	2	1	0	2.7	8.2
2-Oct	0	1	0	1	1	2	2	1	1	0	A	1	1	1	1	1	2	2	2	3	4	2	2	2	1.5	3.8
3-Oct	2	2	3	2	2	2	2	3	4	A	4	3	2	1	1	1	3	3	3	3	3	1	2	5	2.5	4.6
4-Oct	6	7	2	2	2	3	6	9	A	8	7	4	3	3	3	3	2	3	9	6	4	6	4	2	4.4	9.4
5-Oct	2	1	1	1	3	1	1	A	4	1	1	1	1	1	1	1	2	2	2	4	4	3	4	4	2.0	3.7
6-Oct	5	4	3	2	2	4	A	4	3	3	3	3	3	2	2	2	2	2	4	2	2	1	1	1	2.6	4.8
7-Oct	1	1	1	1	1	A	2	1	1	1	2	3	3	2	2	2	2	2	2	2	1	1	1	1	1.5	2.6
8-Oct	1	2	2	2	A	2	1	1	1	1	1	1	1	1	1	1	1	0	0	0	1	0	0	0	0.8	1.7
9-Oct	1	1	1	A	1	1	1	1	1	1	1	1	1	1	1	1	1	2	1	1	1	2	2	3	1.1	2.7
10-Oct	3	3	A	3	3	2	2	2	2	3	3	5	7	6	3	3	4	4	4	4	4	3	2	2	3.4	7.2
11-Oct	2	A	2	2	2	3	5	7	5	4	4	2	2	3	2	4	5	3	9	5	11	9	6	3	4.4	10.8
12-Oct	A	2	2	2	4	2	1	1	1	1	0	1	1	1	1	1	1	1	2	2	1	2	1	A	1.4	3.9
13-Oct	2	1	1	1	3	3	9	3	4	5	3	3	3	2	1	1	4	1	2	2	2	2	A	2	2.6	9.1
14-Oct	2	1	1	1	1	1	1	1	2	1	1	1	1	1	1	1	1	1	1	1	5	A	3	2	1.4	4.9
15-Oct	2	1	2	2	1	3	4	3	2	2	7	4	3	2	3	6	1	1	3	3	A	3	2	2	2.6	7.0
16-Oct	1	1	2	2	2	1	2	2	2	1	2	3	1	2	2	3	2	5	2	A	2	1	1	1	1.9	4.9
17-Oct	2	1	1	1	1	1	1	1	1	1	1	1	1	1	1	4	2	6	A	7	37	16	1	1	3.9	37.2
18-Oct	1	1	2	2	1	1	2	3	13	13	4	3	2	1	1	1	3	A	5	11	3	3	3	3	3.6	13.4
19-Oct	3	2	2	2	2	2	2	2	7	8	5	3	2	3	3	3	A	3	3	3	12	14	3	3	4.1	13.5
20-Oct	3	9	4	2	1	1	1	2	1	0	0	0	0	1	1	A	2	1	1	2	1	1	1	2	1.6	9.3
21-Oct	1	1	1	1	1	1	1	1	2	4	2	2	1	1	A	1	1	3	2	2	2	2	2	1	1.6	3.5
22-Oct	1	2	2	2	2	2	2	6	5	6	7	8	7	A	8	7	7	7	7	6	6	9	10	12	5.6	11.8
23-Oct	8	6	5	5	5	7	6	6	6	8	10	13	A	C	C	C	C	C	14	11	26	21	13	13	10.2	25.9
24-Oct	10	10	8	8	6	6	7	7	11	12	19	A	17	15	13	9	11	10	9	6	6	5	3	3	9.3	19.3
25-Oct	4	4	3	3	2	2	3	3	3	3	A	4	2	1	1	1	1	1	2	1	1	1	1	1	2.1	4.4
26-Oct	1	1	1	1	1	1	1	1	2	A	3	3	2	3	4	4	4	4	4	4	3	3	1	1	2.3	4.4
27-Oct	1	1	1	1	1	1	1	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	0.7
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
2.8 2.8 2.2 2.2 2.1 2.3 2.8 3.0 3.4 3.8 3.8 3.1 2.7 2.3 2.4 2.6 2.6 2.9 3.9 3.6 5.8 4.6 2.8 2.8																									Diurnal Average	
10.5 10.0 7.5 8.1 6.2 7.4 9.1 9.4 12.6 13.4 19.3 13.4 16.9 15.5 13.1 8.9 10.8 9.6 14.1 11.0 37.2 21.5 12.5 13.1																									Diurnal Maximum	
C - Calibration NS - Not in service A - Automated Daily Zero Span																										

Hourly Averages

Oxides of Nitrogen (NO_x) - ppb
Portable Rycroft - October 2018



Hourly Maximums


Oxides of Nitrogen (NO_x) - ppb

Portable Rycroft - October 2018

Maximum Value: 188.0 ppb on Oct 17 22:00																				Maximum Daily Average: 26.3 ppb on Oct 17					Hours in Service: 635	
Minimum Value: 1 ppb on Oct 20 13:00																				Minimum Daily Average: 1.3 ppb on Oct 8					Hours of Data: 599	
Maximum Diurnal Average: 19.4 ppb at hour 21																				Minimum Diurnal Average: 3.0 ppb at hour 5					Hours of Missing Data: 36	
Monthly Average: 7.74 ppb																				Percentiles: P ₁ = 0.7 P ₁₀ = 1.3 Q ₁ = 2.0 Median = 3.3 Q ₃ = 7.0 P ₉₀ = 14.4 P ₉₉ = 92.4					Hours of Calibration: 36	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	10	7	9	9	3	4	5	7	4	3	2	A	2	2	4	11	5	11	16	4	3	8	4	1	5.9	15.7
2-Oct	1	1	1	1	3	3	4	2	2	1	A	2	4	2	2	2	6	9	2	6	13	3	2	2	3.2	12.8
3-Oct	2	3	3	2	3	3	3	8	7	A	6	3	2	1	1	2	4	3	13	10	12	2	7	26	5.6	25.6
4-Oct	9	12	3	3	4	5	42	30	A	11	8	7	5	9	8	5	6	5	35	11	12	14	5	11	11.3	41.7
5-Oct	2	2	2	5	7	2	2	A	6	3	1	1	1	1	1	2	3	4	6	6	11	6	4	8	3.7	10.5
6-Oct	19	6	4	2	3	7	A	5	6	4	4	4	4	3	3	2	4	7	17	19	33	4	1	1	7.1	32.5
7-Oct	2	2	1	1	1	A	2	2	2	2	3	3	4	3	3	3	3	3	2	2	2	2	1	4	2.3	4.0
8-Oct	2	2	2	2	A	3	1	1	1	1	1	1	1	4	1	1	1	1	1	2	1	1	1	1	1.3	3.9
9-Oct	1	1	1	A	2	1	1	2	3	4	2	1	3	1	2	3	5	7	4	1	2	2	2	3	2.3	6.6
10-Oct	3	3	A	4	3	3	2	4	3	19	4	10	11	9	5	4	4	4	4	6	6	5	3	3	5.3	18.8
11-Oct	3	A	3	3	2	4	6	7	7	5	5	7	3	3	3	8	9	6	92	19	123	53	42	4	18.1	122.6
12-Oct	A	4	2	2	8	3	3	2	3	2	1	3	2	1	2	1	2	3	8	3	2	3	2	A	2.9	8.1
13-Oct	2	2	1	3	5	4	51	4	4	6	8	4	6	6	6	5	20	3	7	4	3	2	A	3	6.9	50.9
14-Oct	2	2	1	1	1	2	1	2	2	2	2	3	3	2	1	3	1	2	2	3	9	A	7	3	2.4	8.6
15-Oct	5	3	6	4	1	9	10	4	9	2	13	11	6	6	10	15	3	1	4	6	A	10	2	2	6.2	15.2
16-Oct	2	2	2	2	2	2	3	2	3	6	3	8	7	18	7	8	4	34	6	A	4	2	4	2	5.6	33.5
17-Oct	6	2	1	1	1	1	1	1	1	2	3	6	1	2	2	103	2	115	A	51	109	188	2	2	26.3	188.0
18-Oct	1	7	10	5	2	2	3	18	177	33	9	7	6	2	2	6	11	A	22	78	5	3	3	3	18.1	177.4
19-Oct	3	3	2	3	3	3	3	3	10	30	22	6	4	7	4	6	A	7	7	11	78	60	13	3	12.6	78.1
20-Oct	3	37	7	3	1	2	9	15	2	1	1	1	1	4	3	A	5	5	2	3	3	2	3	3	5.0	36.6
21-Oct	2	2	1	1	1	1	3	3	18	20	5	3	1	2	A	2	3	23	3	2	2	2	2	2	4.6	23.0
22-Oct	2	3	2	2	2	2	2	23	9	23	8	9	9	A	10	7	7	26	20	8	7	10	12	14	9.4	25.8
23-Oct	10	7	6	7	6	13	8	11	18	46	15	17	A	C	C	C	C	C	20	13	33	32	14	15	16.2	46.2
24-Oct	14	12	10	10	8	7	8	8	57	34	30	A	18	18	17	11	21	17	14	9	9	7	5	4	15.1	56.6
25-Oct	5	5	4	3	3	3	3	3	3	5	A	48	3	1	2	2	1	2	2	1	1	1	1	1	4.5	47.7
26-Oct	1	1	1	1	1	2	2	2	2	A	3	3	3	4	10	5	4	5	5	4	4	4	1	1	3.1	9.5
27-Oct	1	1	1	1	1	1	1	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	1.0
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
4.4 4.9 3.3 3.1 3.0 3.5 7.0 6.8 14.4 10.9 6.7 6.9 4.4 4.6 4.6 9.1 5.6 12.5 12.5 11.4 19.4 17.1 5.8 4.9																									Diurnal Average	
19.3 36.6 10.0 9.9 8.1 13.2 50.9 29.8 177.4 46.2 30.1 47.7 18.3 18.0 17.2 102.8 21.2 114.6 92.1 77.5 122.6 188.0 42.2 25.6																									Diurnal Maximum	
C - Calibration NS - Not in service A - Automated Daily Zero Span																										


Hourly Maximums

Oxides of Nitrogen (NO_x) - ppb
Portable Rycroft - October 2018



Span Responses

Oxides of Nitrogen (NO_x)
Portable Rycroft - October 2018



Hourly Averages


Ozone (O₃) - ppb

Portable Rycroft - October 2018

Number of Exceedences (AAAQO): 1-hr: 0 24-hr: 0 Maximum Value: 48.0 ppb on Oct 16 14:00 Maximum Daily Average: 41.4 ppb on Oct 17																				Hours in Service: 635 Hours of Data: 601 Hours of Missing Data: 34 Hours of Calibration: 34 Percent Operational Time: 100.0						
Minimum Value: 1 ppb on Oct 23 21:00 Maximum Diurnal Average: 29.2 ppb at hour 14 Monthly Average: 23.24 ppb										Minimum Daily Average: 10.0 ppb on Oct 23 Minimum Diurnal Average: 18.9 ppb at hour 8 Percentiles: P ₁ = 3.5 P ₁₀ = 11.1 Q ₁ = 16.3 Median = 22.4 Q ₃ = 30.0 P ₉₀ = 36.6 P ₉₉ = 45.7																
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	5	11	9	7	12	12	11	15	17	21	25	A	27	30	28	28	29	29	28	26	26	25	27	27	20.7	29.6
2-Oct	27	25	24	21	21	20	20	24	28	33	A	32	32	32	33	33	33	33	30	25	23	25	25	24	27.0	33.4
3-Oct	22	21	21	21	18	17	16	16	17	A	25	28	32	35	35	33	30	31	29	27	28	31	28	21	25.3	35.2
4-Oct	17	12	21	20	12	12	8	5	A	16	25	30	33	33	33	32	32	31	21	21	20	15	16	20	21.0	33.0
5-Oct	19	18	19	10	11	11	8	A	13	19	21	22	24	24	25	26	26	26	22	18	17	17	16	15	18.5	26.2
6-Oct	11	11	10	14	10	6	A	8	8	11	11	13	15	17	18	20	20	20	16	19	19	18	16	15	14.1	20.5
7-Oct	13	13	13	12	12	A	12	13	15	16	16	15	15	16	16	17	17	17	17	18	19	20	20	19	15.7	20.2
8-Oct	17	17	16	16	A	18	19	19	19	20	20	20	19	20	20	20	20	21	22	23	22	23	22	21	19.7	22.7
9-Oct	21	20	20	A	20	21	20	20	20	21	22	23	24	24	24	24	24	23	24	23	26	25	24	21	22.4	26.1
10-Oct	19	18	A	17	17	16	14	13	14	15	14	13	11	12	16	16	16	13	12	11	15	16	16	15	14.7	19.1
11-Oct	15	A	9	11	10	9	8	6	12	14	17	25	28	27	28	27	28	26	20	18	17	17	18	19	17.8	28.1
12-Oct	A	13	10	9	4	11	16	14	13	16	22	20	18	18	17	16	16	15	13	13	10	10	11	A	13.9	21.9
13-Oct	12	14	15	11	7	8	7	10	10	13	15	16	16	21	24	25	23	23	17	15	13	14	A	14	14.9	25.2
14-Oct	16	15	17	18	19	19	19	18	20	24	27	37	39	42	43	43	43	41	40	37	33	A	32	33	29.4	43.2
15-Oct	33	32	30	29	29	28	30	30	30	32	34	40	41	40	39	36	38	37	33	30	A	31	28	26	32.9	40.5
16-Oct	26	27	26	29	30	30	32	35	35	36	37	40	45	48	47	42	38	37	34	A	31	31	32	34	34.8	48.0
17-Oct	35	35	36	38	44	45	45	46	45	46	47	44	46	46	43	43	41	40	A	40	33	37	39	39	41.4	47.0
18-Oct	38	36	31	32	38	35	30	29	26	25	34	35	36	38	37	37	34	A	26	20	23	24	23	23	30.8	38.3
19-Oct	22	23	23	21	20	18	17	15	10	12	20	26	29	29	30	32	A	31	30	30	28	25	30	29	23.8	31.6
20-Oct	26	14	21	25	33	31	29	30	34	38	38	38	38	38	38	A	36	35	33	29	29	29	26	28	31.1	38.1
21-Oct	28	26	27	27	27	28	28	27	31	31	35	35	39	40	A	41	40	35	32	31	28	28	27	27	31.1	40.6
22-Oct	27	24	25	23	21	24	20	14	15	17	20	23	22	A	26	26	26	26	24	21	20	15	11	6	20.7	26.9
23-Oct	12	14	11	9	11	7	9	10	9	10	12	16	A	20	19	17	C	C	C	9	1	1	3	1	10.0	20.2
24-Oct	2	4	3	3	6	7	5	6	6	8	9	A	15	19	20	21	18	19	20	22	21	22	23	21	13.1	23.4
25-Oct	20	21	21	22	24	21	19	20	23	22	A	26	37	39	38	37	35	34	33	34	34	34	32	30	28.5	39.2
26-Oct	29	29	28	26	29	29	29	30	26	A	25	26	23	22	26	25	23	21	21	21	19	25	30	31	25.8	31.5
27-Oct	31	30	31	34	33	33	33	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	33.6
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
20.8 20.1 19.8 19.5 20.0 19.8 19.3 18.9 19.8 21.5 23.7 26.7 28.2 29.2 28.9 28.7 28.6 27.6 24.8 23.2 22.3 22.2 23.1 22.3																									Diurnal Average	
38.3 35.8 36.2 37.8 44.4 44.7 44.6 46.4 44.7 46.1 47.0 43.9 45.9 48.0 47.1 43.4 43.0 41.4 39.8 39.6 34.1 37.2 39.1 38.5																									Diurnal Maximum	
C - Calibration NS - Not in service A - Automated Daily Zero Span																										
Alberta Ambient Air Quality Objectives (AAAQO): 1-hr 82 ppb 24-hr na																										


Hourly Averages

Ozone (O₃) - ppb
Portable Rycroft - October 2018




Hourly Maximums

Ozone (O₃) - ppb
Portable Rycroft - October 2018



Pollutant Rose

Ozone (O₃) - ppb
Portable Rycroft - October 2018



Eight Hour Running Averages


Ozone (O₃) - ppb

Portable Rycroft - October 2018

Maximum Value: 45.5 ppb on Oct 17 14:00																								Hours in Service: 642	
Minimum Value: 2.2 ppb on Oct 24 04:00																								Hours of Data: 626	
Percentiles: P ₁ = 4.0 P ₁₀ = 12.2 Q ₁ = 16.4 Median = 22.3 Q ₃ = 29.8 P ₉₀ = 35.6 P ₉₉ = 44.3																								Hours of Missing Data: 16	
																								Hours of Calibration: 16	
																								Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																							Daily Maximum	
1-Oct	18	15	13	11	10	9	9	10	12	13	15	16	18	21	23	25	27	28	28	28	27	27	27	28.4	
2-Oct	27	26	26	25	25	24	23	23	23	24	24	25	27	29	30	32	33	33	32	31	30	29	28	27	32.5
3-Oct	26	24	23	23	22	21	20	19	18	18	19	20	22	24	27	29	31	31	32	32	31	30	30	28	31.7
4-Oct	26	24	23	22	20	18	15	13	13	13	14	15	18	21	25	29	29	31	31	29	28	26	24	22	31.0
5-Oct	21	19	19	17	16	15	14	14	13	13	13	15	17	19	21	22	24	24	24	24	23	22	21	20	24.4
6-Oct	18	16	14	14	13	12	11	10	9	9	10	10	10	12	13	14	16	17	17	18	19	19	19	18	18.7
7-Oct	17	16	16	15	14	13	13	13	13	13	14	14	15	15	15	16	16	16	16	17	17	18	18	18	18.3
8-Oct	18	18	18	18	18	18	17	18	18	18	19	19	19	20	20	20	20	20	20	21	21	21	22	22	21.7
9-Oct	22	22	21	21	21	21	20	20	20	20	21	21	21	22	22	23	23	24	24	24	24	24	24	24	24.3
10-Oct	23	23	22	21	20	19	18	16	16	15	15	15	14	13	13	14	14	14	13	13	14	14	14	14	23.2
11-Oct	14	14	14	14	13	12	11	10	9	10	11	13	15	17	20	22	24	26	26	25	24	23	21	20	26.2
12-Oct	19	18	16	15	13	12	12	11	11	12	13	15	16	17	17	18	18	18	17	16	15	14	13	12	19.3
13-Oct	12	12	12	12	11	11	11	11	10	10	10	11	12	14	16	18	19	20	21	20	20	19	18	17	20.6
14-Oct	16	15	15	15	16	17	17	17	18	19	20	23	25	28	31	34	37	40	41	41	40	40	39	37	41.1
15-Oct	36	34	33	32	31	31	31	30	30	30	30	32	33	35	36	36	37	38	38	37	36	35	33	32	38.1
16-Oct	30	29	28	28	28	28	28	29	30	31	33	34	36	38	40	41	42	42	41	42	40	37	35	34	41.9
17-Oct	33	33	33	34	36	37	39	40	42	43	44	45	45	46	45	45	45	44	43	43	41	40	39	38	45.5
18-Oct	38	37	37	36	36	36	35	34	32	31	31	31	31	31	32	33	34	36	35	33	31	29	27	25	38.0
19-Oct	23	23	23	23	22	22	21	20	18	17	17	17	18	20	21	23	25	28	29	30	30	29	29	29	29.9
20-Oct	29	27	25	25	25	26	26	26	27	30	32	34	34	35	36	37	38	37	36	35	34	33	31	31	37.6
21-Oct	30	29	28	28	27	27	27	27	28	28	29	30	32	33	34	36	37	38	37	37	35	33	33	31	37.7
22-Oct	29	28	27	26	25	25	24	22	21	20	19	20	20	19	20	21	23	24	25	24	24	23	21	19	29.2
23-Oct	17	15	14	12	11	10	10	10	10	9	10	10	10	12	14	15	16	N	N	N	N	N	N	N	16.9
24-Oct	3	3	3	2	3	4	4	5	5	5	6	7	8	10	12	14	16	17	19	19	20	21	21	21	21.0
25-Oct	21	21	21	22	22	22	21	21	21	21	22	22	24	27	29	32	33	35	35	36	36	35	34	33	35.9
26-Oct	32	32	31	30	30	29	29	29	28	28	28	28	27	26	25	25	24	24	23	23	22	23	23	24	32.5
27-Oct	25	26	27	29	31	32	32	32	32	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	32.4
28-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--
29-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--
30-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--
31-Oct	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	N	--
Diurnal Maximums																									
38.0 37.4 36.5 35.6 36.2 37.4 38.9 40.5 41.7 43.1 44.5 45.2 45.4 45.5 45.4 45.0 44.5 43.8 43.3 42.7 40.9 40.1 39.0 38.3																									
N - Not Valid																									

Span Responses

Ozone (O₃)
Portable Rycroft - October 2018



Hourly Averages


Total Hydrocarbons (THC) - ppm

Portable Rycroft - October 2018

Maximum Value: 2.39 ppm on Oct 1 07:00		Maximum Daily Average: 2.14 ppm on Oct 24		Hours in Service: 635																							
Minimum Value: 1.8 ppm on Oct 12 11:00		Minimum Daily Average: 1.84 ppm on Oct 12		Hours of Data: 584																							
Maximum Diurnal Average: 1.94 ppm at hour 7		Minimum Diurnal Average: 1.91 ppm at hour 14		Hours of Missing Data: 51																							
Monthly Average: 1.921 ppm		Percentiles: P ₁ = 1.81 P ₁₀ = 1.84 Q ₁ = 1.87 Median = 1.90 Q ₃ = 1.96 P ₉₀ = 2.01 P ₉₉ = 2.19		Hours of Calibration: 38																							
				Percent Operational Time: 98.0																							
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
1-Oct	1.9	1.9	1.9	2.0	2.1	2.2	2.4	2.1	1.9	1.9	1.9	A	1.9	1.8	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.92	2.39	
2-Oct	1.9	1.9	1.8	1.9	1.9	1.9	1.9	1.8	1.8	1.8	A	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.87	1.91		
3-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.89	1.91		
4-Oct	1.8	1.9	1.8	1.8	1.8	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	1.87	1.92		
5-Oct	1.9	1.9	1.9	1.9	1.9	1.8	1.8	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.87	1.89		
6-Oct	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.90	1.92		
7-Oct	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.93		
8-Oct	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.91		
9-Oct	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.88	1.91		
10-Oct	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.90	1.94		
11-Oct	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.87	1.93		
12-Oct	A	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.9	1.8	1.9	1.8	1.9	A	1.84	1.90		
13-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.8	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.88	1.97		
14-Oct	1.9	1.8	1.8	1.9	1.8	1.9	1.8	1.9	1.8	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.9	A	1.9	1.9	1.85	1.88		
15-Oct	1.9	1.9	1.9	1.9	1.8	1.9	1.9	1.8	1.9	1.9	1.9	1.9	1.9	1.8	1.9	1.9	1.9	1.8	1.9	A	1.9	1.9	1.9	1.86	1.87		
16-Oct	1.9	1.9	1.9	1.9	1.9	N	N	N	N	N	N	N	N	1.8	1.8	1.8	1.8	1.8	1.8	A	1.8	1.8	1.8	--	1.90		
17-Oct	1.8	1.8	1.8	1.8	1.9	1.9	1.9	1.9	1.9	C	C	C	C	C	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.88	1.93		
18-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.9	1.9	1.9	1.9	1.9	1.9	A	2.0	2.0	2.0	2.0	2.0	1.95	1.99		
19-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	1.99	2.02		
20-Oct	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.0	2.0	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.96	2.02		
21-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.9	1.9	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.94	1.98		
22-Oct	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	2.1	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.1	2.00	2.07		
23-Oct	2.0	2.0	2.0	2.0	2.0	2.1	2.1	N	N	N	N	N	C	C	C	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	--	2.19		
24-Oct	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.1	2.1	2.1	2.3	A	2.3	2.2	2.2	2.2	2.2	2.2	2.1	2.0	2.0	2.0	2.0	2.14	2.27		
25-Oct	2.0	2.0	2.1	2.1	2.0	2.0	2.0	2.1	2.1	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.01	2.15		
26-Oct	2.0	1.9	1.9	1.9	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.97	2.01		
27-Oct	1.9	2.0	2.0	2.0	2.0	2.0	2.0	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	1.98		
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--		
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--		
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--		
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--		
		1.92	1.92	1.92	1.92	1.92	1.93	1.94	1.92	1.92	1.91	1.92	1.91	1.92	1.93	1.92	1.92	1.92	1.92	1.93	1.92	1.92	1.92	1.92	Diurnal Average		
		2.18	2.17	2.18	2.20	2.17	2.18	2.39	2.15	2.15	2.11	2.25	2.06	2.27	2.22	2.20	2.19	2.15	2.17	2.17	2.17	2.18	2.17	2.17	2.18	Diurnal Maximum	
C - Calibration		N - Not Valid						NS - Not in service						A - Automated Daily Zero Span													

Hourly Averages

Total Hydrocarbons (THC) - ppm
Portable Rycroft - October 2018



Hourly Maximums


Total Hydrocarbons (THC) - ppm

Portable Rycroft - October 2018

Maximum Value: 2.50 ppm on Oct 1 07:00																								Hours in Service: 635		
Maximum Daily Average: 2.17 ppm on Oct 24																								Hours of Data: 584		
Minimum Value: 1.8 ppm on Oct 12 11:00																								Hours of Missing Data: 51		
Maximum Diurnal Average: 1.97 ppm at hour 6																								Hours of Calibration: 38		
Monthly Average: 1.949 ppm																								Percent Operational Time: 98.0		
Percentiles: P ₁ = 1.84 P ₁₀ = 1.87 Q ₁ = 1.89 Median = 1.92 Q ₃ = 1.99 P ₉₀ = 2.05 P ₉₉ = 2.25																										
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	1.9	1.9	1.9	2.0	2.2	2.4	2.5	2.2	2.0	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.97	2.50
2-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.94
3-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.91	1.94
4-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	A	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.9	1.91	1.97
5-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.91
6-Oct	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.92	1.95
7-Oct	1.9	2.0	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.92	1.95
8-Oct	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.91	1.93
9-Oct	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.90	1.93
10-Oct	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.1	1.9	1.9	2.3	1.9	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.94	2.28
11-Oct	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.9	1.8	1.9	1.9	1.9	1.89	1.95
12-Oct	A	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.87	1.95
13-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.91	2.01
14-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.88	1.90
15-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.88	1.90
16-Oct	1.9	1.9	1.9	1.9	1.9	N	N	N	N	N	N	N	N	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.8	--	1.93
17-Oct	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	1.9	C	C	C	C	C	2.0	2.0	1.9	2.0	A	1.9	1.9	2.0	1.9	1.9	1.91	2.01
18-Oct	1.9	1.9	2.0	1.9	2.0	2.0	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	1.97	2.01
19-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.01	2.05
20-Oct	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.98	2.08
21-Oct	1.9	1.9	2.0	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.97	2.01
22-Oct	2.0	2.0	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.1	2.1	2.1	A	2.1	2.1	2.0	2.0	2.1	2.1	2.1	2.1	2.1	2.1	2.02	2.10
23-Oct	2.1	2.0	2.0	2.0	2.0	2.1	2.1	N	N	N	N	N	C	C	C	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	--	2.22
24-Oct	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.1	2.3	A	2.3	2.3	2.2	2.2	2.2	2.2	2.2	2.0	2.0	2.0	2.0	2.1	2.17	2.33
25-Oct	2.0	2.1	2.2	2.2	2.1	2.0	2.0	2.1	2.2	2.1	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.04	2.19
26-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.00	2.04
27-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	2.01
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
1.95 1.95 1.94 1.95 1.96 1.97 1.96 1.95 1.95 1.94 1.95 1.94 1.95 1.93 1.96 1.95 1.94 1.94 1.95 1.95 1.95 1.95 1.95 1.95																									Diurnal Average	
2.20 2.20 2.21 2.24 2.20 2.39 2.50 2.22 2.19 2.14 2.30 2.09 2.33 2.27 2.28 2.22 2.17 2.19 2.20 2.18 2.21 2.21 2.18 2.22																									Diurnal Maximum	
C - Calibration N - Not Valid NS - Not in service A - Automated Daily Zero Span																										


Hourly Maximums

Total Hydrocarbons (THC) - ppm
Portable Rycroft - October 2018




Pollutant Rose

Total Hydrocarbons (THC) - ppm
Portable Rycroft - October 2018



Span Responses

Total Hydrocarbons (THC)
Portable Rycroft - October 2018



Hourly Averages


Methane (CH₄) - ppm

Portable Rycroft - October 2018

Maximum Value: 2.38 ppm on Oct 1 07:00																				Maximum Daily Average: 2.13 ppm on Oct 24										Hours in Service: 635									
Minimum Value: 1.8 ppm on Oct 12 11:00																				Minimum Daily Average: 1.84 ppm on Oct 12										Hours of Data: 584									
Maximum Diurnal Average: 1.93 ppm at hour 7																				Minimum Diurnal Average: 1.90 ppm at hour 14										Hours of Missing Data: 51									
Monthly Average: 1.915 ppm																				Percentiles: P ₁ = 1.81 P ₁₀ = 1.84 Q ₁ = 1.86 Median = 1.89 Q ₃ = 1.95 P ₉₀ = 2.01 P ₉₉ = 2.18										Hours of Calibration: 38									
																				Percent Operational Time: 98.0																			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum													
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24															
1-Oct	1.9	1.9	1.9	1.9	2.1	2.2	2.4	2.1	1.9	1.9	1.9	A	1.9	1.8	1.8	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.92	2.38													
2-Oct	1.8	1.9	1.8	1.9	1.9	1.9	1.9	1.8	1.8	1.8	A	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.86	1.91													
3-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.88	1.91													
4-Oct	1.8	1.9	1.8	1.8	1.8	1.8	1.9	1.9	A	1.8	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	1.86	1.92													
5-Oct	1.9	1.9	1.9	1.8	1.9	1.8	1.8	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.86	1.88													
6-Oct	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.92													
7-Oct	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.92													
8-Oct	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.88	1.90													
9-Oct	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.88	1.91													
10-Oct	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.94													
11-Oct	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.86	1.93													
12-Oct	A	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.9	1.8	1.8	1.9	A	1.84	1.90													
13-Oct	1.9	1.9	1.8	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.9	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.88	1.96													
14-Oct	1.9	1.8	1.8	1.8	1.8	1.9	1.8	1.8	1.8	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	A	1.9	1.9	1.84	1.88													
15-Oct	1.8	1.8	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.9	1.9	1.9	1.9	1.8	1.9	1.9	1.8	1.8	1.8	1.9	A	1.9	1.9	1.9	1.85	1.87													
16-Oct	1.9	1.9	1.9	1.9	1.9	N	N	N	N	N	N	N	N	1.8	1.8	1.8	1.8	1.8	1.8	A	1.8	1.8	1.8	1.8	--	1.89													
17-Oct	1.8	1.8	1.8	1.8	1.9	1.9	1.8	1.8	1.9	C	C	C	C	C	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.88	1.92													
18-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	2.0	2.0	2.0	2.0	2.0	2.0	1.94	1.99													
19-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.99	2.02													
20-Oct	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.0	2.0	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.95	2.02													
21-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	1.9	1.9	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.9	2.0	1.93	1.98													
22-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	2.1	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.1	2.1	1.99	2.07													
23-Oct	2.0	2.0	2.0	2.0	2.0	2.1	2.1	N	N	N	N	N	C	C	C	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	--	2.18													
24-Oct	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.1	2.1	2.1	2.2	A	2.3	2.2	2.2	2.2	2.1	2.2	2.1	2.0	2.0	2.0	2.0	2.0	2.13	2.26													
25-Oct	2.0	2.0	2.1	2.1	2.0	2.0	2.0	2.1	2.1	2.0	A	2.0	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.00	2.14													
26-Oct	2.0	1.9	1.9	1.9	2.0	2.0	2.0	1.9	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.97	2.01													
27-Oct	1.9	2.0	1.9	2.0	2.0	2.0	2.0	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	1.98													
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--													
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--													
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--													
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--													
																									Diurnal Average														
																									Diurnal Maximum														
C - Calibration																									N - Not Valid					NS - Not in service					A - Automated Daily Zero Span				

Hourly Averages

Methane (CH₄) - ppm
Portable Rycroft - October 2018



Hourly Maximums


Methane (CH₄) - ppm

Portable Rycroft - October 2018

Maximum Value: 2.49 ppm on Oct 1 07:00																				Maximum Daily Average: 2.16 ppm on Oct 24										Hours in Service: 635									
Minimum Value: 1.8 ppm on Oct 12 11:00																				Minimum Daily Average: 1.86 ppm on Oct 12										Hours of Data: 584									
Maximum Diurnal Average: 1.96 ppm at hour 6																				Minimum Diurnal Average: 1.92 ppm at hour 14										Hours of Missing Data: 51									
Monthly Average: 1.943 ppm																				Percentiles: P ₁ = 1.83 P ₁₀ = 1.86 Q ₁ = 1.88 Median = 1.91 Q ₃ = 1.98 P ₉₀ = 2.04 P ₉₉ = 2.24										Hours of Calibration: 38									
																				Percent Operational Time: 98.0																			
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum													
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24															
1-Oct	1.9	1.9	1.9	2.0	2.2	2.4	2.5	2.2	2.0	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.96	2.49													
2-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.8	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.94													
3-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.90	1.93													
4-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.90	1.96													
5-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.89	1.90													
6-Oct	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.91	1.94													
7-Oct	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.91	1.94													
8-Oct	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.90	1.92													
9-Oct	1.9	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.90	1.93													
10-Oct	1.9	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.1	1.9	1.9	2.3	1.9	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.94	2.29													
11-Oct	1.9	A	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.9	1.89	1.95													
12-Oct	A	1.9	1.9	1.9	1.9	1.9	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.86	1.94													
13-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.91	2.01													
14-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.8	1.9	1.8	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.87	1.90													
15-Oct	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	1.9	A	1.9	1.9	1.9	1.88	1.90													
16-Oct	1.9	1.9	1.9	1.9	1.9	N	N	N	N	N	N	N	N	1.8	1.8	1.8	1.9	1.9	1.9	A	1.8	1.8	1.8	1.8	--	1.92													
17-Oct	1.8	1.8	1.8	1.9	1.9	1.9	1.9	1.9	1.9	C	C	C	C	C	2.0	1.9	1.9	2.0	A	1.9	1.9	1.9	1.9	1.9	1.90	2.01													
18-Oct	1.9	1.9	1.9	1.9	2.0	2.0	1.9	2.0	2.0	1.9	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	1.96	2.00													
19-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.01	2.04													
20-Oct	2.0	2.1	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	1.9	2.0	2.0	2.0	2.0	2.0	2.0	1.9	1.98	2.08													
21-Oct	1.9	1.9	2.0	1.9	1.9	1.9	1.9	1.9	1.9	1.9	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.96	2.00													
22-Oct	2.0	2.0	2.0	2.0	2.0	2.0	1.9	2.0	2.0	2.0	2.1	2.1	2.1	A	2.1	2.1	2.0	2.0	2.1	2.1	2.1	2.1	2.1	2.1	2.02	2.10													
23-Oct	2.1	2.0	2.0	2.0	2.0	2.1	2.1	N	N	N	N	N	C	C	C	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	--	2.21													
24-Oct	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.2	2.1	2.3	A	2.3	2.3	2.2	2.2	2.2	2.2	2.2	2.0	2.0	2.0	2.0	2.1	2.16	2.32													
25-Oct	2.0	2.1	2.1	2.1	2.1	2.0	2.0	2.1	2.2	2.1	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.04	2.17													
26-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	A	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	2.0	1.99	2.03													
27-Oct	2.0	2.0	2.0	2.0	2.0	2.0	2.0	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	2.00													
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--													
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--													
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--													
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--													
																									Diurnal Average														
																									Diurnal Maximum														
C - Calibration																									N - Not Valid					NS - Not in service					A - Automated Daily Zero Span				


Hourly Maximums

Methane (CH₄) - ppm
Portable Rycroft - October 2018




Pollutant Rose

Methane (CH₄) - ppm
Portable Rycroft - October 2018



Span Responses


Methane (CH₄)
Portable Rycroft - October 2018



Hourly Averages

Non Methane Hydrocarbon (NMHC) - ppm


Portable Rycroft - October 2018



Hourly Maximums

Non Methane Hydrocarbon (NMHC) - ppm


Portable Rycroft - October 2018




Pollutant Rose

Non Methane Hydrocarbon (NMHC) - ppm

Portable Rycroft - October 2018





Hourly Averages


PM2.5 (PM_{2.5}) - µg/m³

Portable Rycroft - October 2018

Number of Exceedences: 1-hr: 0 24-hr: 0 Maximum Value: 19.4 µg/m³ on Oct 23 21:00 Maximum Daily Average: 7.8 µg/m³ on Oct 24																					Hours in Service: 635 Hours of Data: 629 Hours of Missing Data: 6 Hours of Calibration: 6 Percent Operational Time: 100.0					
Minimum Value: 0 µg/m³ on Oct 2 00:00 Maximum Diurnal Average: 4.1 µg/m³ at hour 19 Monthly Average: 2.79 µg/m³										Minimum Daily Average: 0.4 µg/m³ on Oct 8 Minimum Diurnal Average: 1.7 µg/m³ at hour 13 Percentiles: P ₁ = 0.0 P ₁₀ = 0.0 Q ₁ = 0.8 Median = 2.0 Q ₃ = 3.9 P ₉₀ = 6.5 P ₉₉ = 12.5																
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1-Oct	11	11	6	3	4	2	2	1	2	1	2	2	2	1	2	1	2	4	3	1	1	1	0	0	2.8	11.5
2-Oct	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	2	2	3	4	2	1	1	0.7	4.4
3-Oct	0	1	1	0	1	1	1	1	3	3	5	2	1	0	1	1	3	2	7	7	2	1	1	5	2.1	7.1
4-Oct	7	19	2	2	3	2	2	4	4	7	6	3	2	3	3	3	4	5	14	11	6	7	6	9	5.6	18.6
5-Oct	3	3	2	2	2	2	4	2	2	1	1	2	1	1	1	5	6	11	9	7	10	4	4	9	3.8	11.3
6-Oct	7	8	7	4	5	5	5	5	5	4	3	4	3	3	3	3	3	4	5	3	2	2	2	2	4.0	7.9
7-Oct	3	3	2	2	2	2	2	3	3	2	2	3	3	3	3	3	2	3	2	2	2	2	1	1	2.4	3.1
8-Oct	2	1	1	1	1	1	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	1	0.4	1.7
9-Oct	1	1	1	1	1	1	1	1	2	3	2	1	1	1	1	1	2	4	3	0	1	2	3	3	1.6	4.3
10-Oct	4	4	4	5	5	5	5	5	5	5	5	5	5	4	3	3	4	4	3	4	3	2	3	2	4.0	5.3
11-Oct	3	2	3	2	2	4	5	4	2	1	1	0	1	1	1	1	3	2	1	1	3	1	1	1	1.9	4.5
12-Oct	0	0	2	4	4	4	3	3	3	3	3	2	2	3	1	2	1	1	0	1	0	1	1	0	1.9	4.3
13-Oct	1	0	1	1	2	2	0	0	2	2	1	0	1	0	0	0	0	0	0	0	0	0	0	0	0.5	2.1
14-Oct	0	0	0	0	0	0	0	1	1	0	0	0	0	0	0	0	0	1	1	2	1	2	2	2	0.6	2.1
15-Oct	2	2	2	2	2	1	0	2	2	3	1	2	2	2	2	3	2	1	2	3	1	2	2	3	1.9	3.3
16-Oct	2	3	2	1	1	2	1	0	1	0	0	4	2	1	3	7	9	10	5	2	1	2	1	2	2.6	10.0
17-Oct	3	1	1	2	1	3	3	3	4	3	3	2	1	4	3	4	5	4	5	3	4	4	6	8	3.3	8.4
18-Oct	9	8	8	8	7	6	7	5	5	5	7	5	2	0	1	1	4	6	5	12	4	2	2	2	5.1	12.0
19-Oct	2	2	1	1	1	1	1	1	2	1	0	0	0	0	0	0	0	0	0	0	0	12	4	4	1.5	11.8
20-Oct	5	5	4	4	1	0	2	0	0	0	0	0	0	6	4	0	2	1	2	5	2	2	2	2	2.0	6.5
21-Oct	2	1	2	1	1	1	1	1	1	1	0	1	0	0	0	0	2	2	2	2	2	3	3	2	1.3	3.0
22-Oct	1	1	1	1	0	1	1	3	3	3	2	2	2	2	2	3	3	3	5	7	8	7	7	6	3.2	8.1
23-Oct	5	7	3	3	4	6	2	3	5	6	7	7	5	4	5	C	C	18	14	12	19	14	11	10	7.8	19.4
24-Oct	9	7	7	7	5	5	5	5	6	6	9	9	7	8	10	9	12	8	11	13	9	8	7	6	7.8	12.6
25-Oct	6	5	5	5	5	5	7	6	5	4	6	4	0	0	0	0	0	0	1	1	1	0	0	1	2.8	6.7
26-Oct	1	0	0	0	0	3	1	0	1	2	1	1	3	3	2	2	2	3	3	3	2	1	0	0	1.5	3.4
27-Oct	0	1	1	2	4	4	4	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	4.0
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
3.3 3.6 2.5 2.4 2.4 2.6 2.4 2.3 2.6 2.6 2.6 2.4 1.7 1.9 2.0 2.1 2.9 3.9 4.1 4.1 3.5 3.3 2.6 3.1																									Diurnal Average	
11.3 18.6 8.5 7.5 7.2 6.2 6.7 6.1 6.2 7.2 8.8 8.8 6.8 7.8 9.9 9.3 12.2 18.4 14.5 12.6 19.4 14.4 11.0 10.2																									Diurnal Maximum	
C - Calibration NS - Not in service Alberta Ambient Air Quality Guideline (AAAQG): 1-hr 80 µg/m³ Alberta Ambient Air Quality Objective (AAAQO): 24-hr 30 µg/m³																										

Hourly Averages

PM_{2.5} (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Portable Rycroft - October 2018



Hourly Maximums


PM2.5 (PM_{2.5}) - µg/m³

Portable Rycroft - October 2018

Maximum Value: 44.8 μg/m³ on Oct 19 22:00																		Maximum Daily Average: 10.9 μg/m³ on Oct 23								Hours in Service: 635	
Minimum Value: 0 μg/m³ on Oct 2 03:00																		Minimum Daily Average: 1.5 μg/m³ on Oct 2								Hours of Data: 629	
Maximum Diurnal Average: 7.8 μg/m³ at hour 19																		Minimum Diurnal Average: 3.4 μg/m³ at hour 13								Hours of Missing Data: 6	
Monthly Average: 4.96 μg/m³																		Percentiles: P ₁ = 0.0 P ₁₀ = 1.3 Q ₁ = 2.4 Median = 3.6 Q ₃ = 6.0 P ₉₀ = 9.3 P ₉₉ = 27.8								Hours of Calibration: 6	
																										Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	18	17	13	5	8	3	3	3	4	2	4	3	3	2	4	4	4	7	4	2	3	2	1	1	5.1	18.1	
2-Oct	1	0	0	1	0	1	0	0	0	0	0	0	0	0	0	0	4	5	3	5	7	3	3	2	1.5	7.1	
3-Oct	2	2	2	1	3	3	3	2	5	5	8	6	3	1	3	3	7	7	18	14	3	3	6	7	4.8	17.5	
4-Oct	18	30	5	3	4	5	4	5	6	9	8	5	4	5	5	5	7	7	40	28	8	10	7	27	10.7	40.1	
5-Oct	4	4	4	4	4	3	7	3	4	2	3	8	3	2	2	17	11	28	14	13	16	7	6	22	8.0	28.2	
6-Oct	15	17	8	6	7	6	6	6	7	6	5	5	5	4	4	3	6	6	7	5	4	4	3	3	6.2	17.1	
7-Oct	4	6	4	4	3	4	4	4	4	4	4	5	4	4	4	4	4	4	4	3	3	3	2	3	3.9	6.1	
8-Oct	3	3	2	2	3	2	1	1	1	1	1	1	1	1	0	1	1	1	1	1	1	2	2	2	1.5	3.1	
9-Oct	2	2	2	2	2	2	1	2	4	4	3	2	2	3	2	2	3	7	4	2	3	4	5	5	3.0	6.6	
10-Oct	5	5	5	6	6	6	6	6	6	7	6	6	6	5	4	4	5	5	6	6	6	4	4	4	5.4	6.6	
11-Oct	6	4	5	4	4	7	6	6	5	2	2	2	3	3	2	3	5	3	2	3	8	3	2	2	3.8	7.8	
12-Oct	1	2	5	6	6	7	6	4	4	4	4	3	3	4	3	3	2	2	2	2	1	3	2	1	3.5	6.6	
13-Oct	2	2	2	3	4	3	2	2	3	4	3	2	3	1	1	1	1	1	2	2	1	1	1	1	2.0	4.4	
14-Oct	0	1	0	1	1	1	3	2	2	2	1	1	3	0	1	4	3	3	2	3	3	4	3	3	2.0	3.6	
15-Oct	3	3	4	3	4	3	1	4	4	4	3	4	3	3	4	4	4	3	3	5	3	3	3	3	3.5	5.4	
16-Oct	4	4	4	2	2	3	2	2	2	2	2	9	5	3	6	12	11	12	11	4	2	2	2	3	4.7	12.4	
17-Oct	9	3	2	3	3	4	5	5	5	5	4	2	9	4	18	10	7	8	5	6	5	8	10	6.1	18.2		
18-Oct	10	9	10	10	9	8	8	6	7	9	12	7	4	1	2	2	8	12	10	24	6	3	4	3	7.6	23.9	
19-Oct	3	3	3	2	3	3	3	4	4	3	2	2	1	1	1	1	2	2	2	1	2	45	6	5	4.3	44.8	
20-Oct	6	8	6	6	6	3	3	2	0	4	1	2	2	29	9	2	3	3	3	9	4	4	3	5	5.2	29.3	
21-Oct	3	3	2	2	2	3	3	3	3	3	3	2	1	1	1	2	4	4	4	4	4	4	4	3	2.8	4.4	
22-Oct	2	3	2	2	3	2	3	7	6	6	4	5	5	7	5	6	6	4	7	14	14	9	8	8	5.7	14.5	
23-Oct	8	9	6	5	6	13	4	4	7	8	9	9	6	6	6	C	C	32	22	14	24	18	13	12	10.9	32.1	
24-Oct	11	9	9	10	7	6	6	7	9	8	10	17	8	10	11	13	19	11	16	16	10	10	8	7	10.3	19.3	
25-Oct	7	6	6	7	6	7	8	7	6	7	7	7	3	1	1	2	2	2	4	2	3	1	1	2	4.3	7.9	
26-Oct	2	1	1	1	9	10	2	2	3	3	3	3	5	6	4	5	4	5	5	4	3	2	0	1	3.6	9.9	
27-Oct	2	2	2	4	5	5	5	C	C	C	C	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	5.2	
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
5.6 5.9 4.2 3.8 4.4 4.5 4.0 3.9 4.3 4.4 4.4 4.7 3.4 4.3 3.5 4.9 5.4 7.0 7.8 7.4 5.7 6.2 4.2 5.6																									Diurnal Average		
18.1 30.2 12.5 9.5 9.5 12.7 8.3 7.5 8.5 9.0 11.6 17.5 8.2 29.3 11.0 18.2 19.3 32.1 40.1 28.4 23.8 44.8 12.5 26.7																									Diurnal Maximum		
C - Calibration NS - Not in service																											


Hourly Maximums

PM_{2.5} (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Portable Rycroft - October 2018



Pollutant Rose

PM_{2.5} (PM_{2.5}) - $\mu\text{g}/\text{m}^3$
Portable Rycroft - October 2018



Hourly Averages

External Temperature (ET) - °C


Portable Rycroft - October 2018

Maximum Value: 21.2 °C on Oct 16 15:00																				Maximum Daily Average: 14.5 °C on Oct 17					Hours in Service: 631	
Minimum Value: -8 °C on Oct 11 03:00																				Minimum Daily Average: -1.8 °C on Oct 5					Hours of Data: 631	
Maximum Diurnal Average: 9.0 °C at hour 15																				Minimum Diurnal Average: -0.2 °C at hour 8					Hours of Missing Data: 0	
Monthly Average: 3.59 °C																				Percentiles: P ₁ = -5.9 P ₁₀ = -3.1 Q ₁ = -1.3 Median = 1.9 Q ₃ = 8.2 P ₉₀ = 12.9 P ₉₉ = 19.3					Hours of Calibration: 0	
																									Percent Operational Time: 100.0	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
1-Oct	-2	-3	-3	-3	-3	-3	-3	-2	-2	-1	1	1	2	4	1	1	1	0	0	-1	-1	-1	-2	-2	-0.9	3.6
2-Oct	-3	-3	-3	-4	-4	-4	-5	-5	-1	1	1	1	2	3	3	3	3	2	-1	-3	-4	-4	-4	-4	-1.4	3.3
3-Oct	-4	-5	-5	-5	-6	-6	-4	-4	-2	1	3	5	7	8	7	7	6	5	3	0	0	1	0	-3	0.3	7.6
4-Oct	-3	-3	-4	-4	-5	-5	-6	-6	-3	2	7	8	9	9	9	9	8	6	2	1	0	-2	-3	-3	1.1	9.4
5-Oct	-3	-4	-4	-5	-6	-6	-5	-5	-4	-3	-2	0	3	4	3	3	3	1	0	-2	-3	-3	-3	-4	-1.8	3.9
6-Oct	-5	-4	-3	-2	-1	-1	-2	-2	-1	0	1	2	3	3	4	4	5	4	3	2	0	0	0	0	0.4	4.5
7-Oct	-1	-1	-1	-1	-1	-2	-2	-2	-2	-2	-2	-2	-1	-1	-1	-1	-1	-1	-2	-2	-2	-2	-2	-2	-1.4	-0.5
8-Oct	-2	-2	-2	-2	-3	-3	-3	-3	-3	-2	-2	-1	-1	-1	0	0	0	-1	-1	-1	-1	-2	-2	-3	-1.7	-0.2
9-Oct	-3	-3	-3	-3	-3	-3	-3	-3	-3	-3	-2	-1	-1	0	1	2	1	0	0	0	0	0	0	-1	-1.2	1.5
10-Oct	-1	-1	-1	-1	-1	-2	-2	-2	-1	-1	0	0	0	0	0	0	0	0	-1	-1	-1	-1	-3	-4	-1.0	0.2
11-Oct	-4	-6	-8	-7	-6	-6	-3	-3	-1	0	2	5	7	7	9	9	9	8	7	5	5	4	4	4	1.7	9.1
12-Oct	4	4	3	3	3	3	1	0	0	0	0	0	0	1	1	1	1	0	0	0	0	0	0	0	1.1	4.2
13-Oct	0	0	0	0	0	0	0	-1	1	2	5	7	8	8	9	9	8	6	3	1	0	0	0	0	2.7	9.1
14-Oct	0	-1	0	0	1	1	1	2	4	7	10	12	12	13	14	14	13	12	12	11	11	12	12	12	7.7	13.9
15-Oct	12	11	11	11	10	10	10	9	9	12	15	15	16	16	15	15	14	13	10	8	8	8	6	6	11.1	15.6
16-Oct	5	6	6	5	5	4	5	6	7	11	15	18	19	21	21	20	17	16	12	12	12	11	12	12	11.6	21.2
17-Oct	12	11	9	8	11	11	10	12	12	17	19	20	21	20	19	19	17	17	15	14	12	13	14	15	14.5	20.7
18-Oct	14	14	11	10	10	9	7	6	6	10	12	13	14	14	14	14	13	10	6	5	3	2	2	2	9.2	14.4
19-Oct	1	1	1	0	-1	-1	-1	-1	0	3	6	8	11	13	13	13	12	11	9	8	9	8	7	6	5.8	13.5
20-Oct	7	4	5	7	11	10	8	8	9	12	13	13	14	14	14	13	13	11	9	5	5	4	2	3	8.9	13.9
21-Oct	3	1	1	1	1	1	1	1	4	6	9	11	13	14	14	14	11	8	6	4	3	2	2	1	5.4	14.0
22-Oct	1	0	0	-2	-2	-2	-2	-4	-3	3	7	10	13	14	14	14	13	10	8	4	3	2	0	-2	4.0	14.0
23-Oct	-2	-1	-3	-4	-3	-5	-5	-5	-4	1	6	9	12	13	13	12	10	6	4	3	-1	-1	-1	-3	2.1	13.1
24-Oct	-4	-3	-4	-4	-4	-4	-4	-4	-3	1	4	8	12	15	17	16	12	11	8	6	5	5	5	3	3.9	16.5
25-Oct	2	3	2	2	2	1	1	3	5	6	7	12	15	15	16	14	12	10	9	9	9	7	6	4	7.2	15.7
26-Oct	3	3	2	0	1	2	1	0	0	1	3	4	5	6	7	7	7	6	4	2	2	5	6	6	3.4	7.0
27-Oct	6	6	6	7	6	6	6	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	6.6
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--
1.3 0.8 0.4 0.3 0.4 0.2 0.1 -0.2 0.9 3.2 5.3 6.9 8.2 9.0 9.0 8.9 8.0 6.6 4.7 3.5 2.8 2.6 2.1 1.6																									Diurnal Average	
14.1 13.6 11.4 10.5 11.0 10.9 10.4 12.0 12.2 16.6 19.3 20.5 20.7 20.8 21.2 20.2 17.4 17.2 14.5 14.0 12.4 13.3 13.7 14.5																									Diurnal Maximum	
NS - Not in service																										

Hourly Averages

External Temperature (ET) - °C

Portable Rycroft - October 2018



Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Portable Rycroft - October 2018

Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24		
1 Spd	5	1	3	4	4	6	6	6	3	7	9	8	11	8	8	10	11	15	15	16	15	16	14	13	7.1	16.4
Dir	323	271	231	244	244	262	258	242	295	269	304	340	324	343	28	360	17	356	353	344	339	346	345	353	333.7	346.1
2 Spd	8	5	5	5	6	4	7	8	8	9	10	11	12	15	14	13	14	14	7	4	7	9	7	7	7.0	14.8
Dir	346	316	289	269	263	240	283	293	297	298	283	266	267	267	248	252	250	243	228	215	186	186	187	187	257.5	267.4
3 Spd	7	7	7	6	6	6	3	4	2	2	2	2	3	2	8	11	9	7	4	3	3	7	3	3	2.2	11.0
Dir	181	176	171	164	178	166	164	176	156	307	268	83	41	353	24	51	69	75	76	60	77	81	94	336	107.2	51.0
4 Spd	1	3	2	2	1	2	3	3	3	3	3	7	14	18	16	18	19	15	11	13	13	6	6	5	6.4	19.4
Dir	292	317	33	109	241	186	207	230	194	177	249	273	259	261	254	254	268	274	260	241	241	276	4	44	259.4	267.9
5 Spd	3	3	2	1	2	1	1	0	3	7	6	6	5	6	5	5	10	10	9	8	7	8	9	4	4.9	10.3
Dir	84	82	101	56	75	90	61	67	35	55	60	41	18	35	76	62	47	61	61	59	63	66	68	56	58.3	47.0
6 Spd	6	4	1	6	1	6	4	2	5	4	3	2	3	3	2	4	3	3	8	16	15	15	12	14	3.9	15.9
Dir	62	61	206	144	101	266	232	217	314	338	324	337	8	0	348	19	32	61	47	46	49	49	53	55	40.6	46.5
7 Spd	12	12	14	13	14	14	12	13	12	14	15	13	15	13	14	13	10	12	12	12	11	13	14	17	12.8	16.5
Dir	50	52	51	49	49	47	44	44	36	24	29	28	24	25	28	19	22	25	27	20	32	34	39	48	35.5	47.5
8 Spd	16	12	11	13	14	13	13	12	15	16	16	16	15	17	18	18	19	20	19	17	18	18	18	16	15.7	19.8
Dir	48	49	50	51	53	50	49	50	51	58	56	55	56	49	52	55	56	56	57	52	52	50	48	45	52.3	55.9
9 Spd	15	13	14	14	13	12	11	13	12	13	14	14	17	13	14	15	14	13	14	12	9	8	6	6	12.2	16.6
Dir	52	49	47	49	49	47	43	47	50	56	52	54	54	55	52	63	62	58	59	63	90	87	97	107	56.9	54.0
10 Spd	4	5	5	5	4	2	1	2	2	1	3	4	8	7	7	6	3	2	3	3	2	3	5	6	1.0	7.6
Dir	112	88	76	72	78	75	84	84	82	185	304	250	253	261	267	279	250	183	171	183	200	163	149	143	180.6	253.0
11 Spd	4	4	5	7	5	6	6	6	9	8	7	11	10	6	6	6	12	6	3	6	3	2	4	5	4.7	11.7
Dir	142	160	184	148	162	175	159	150	159	156	168	188	193	203	217	245	253	292	284	251	184	122	159	160	186.9	252.7
12 Spd	4	4	4	0	2	9	N	N	N	19	22	20	17	14	17	13	12	7	6	3	2	2	3	2	6.9	22.0
Dir	161	182	156	92	329	2	N	N	N	297	305	290	300	293	278	282	279	276	251	225	195	195	216	209	285.3	305.4
13 Spd	3	4	2	3	4	5	5	3	3	2	3	2	2	10	14	13	11	5	4	4	7	10	10	10	3.9	14.4
Dir	154	143	141	186	227	235	246	215	203	48	47	16	153	264	262	260	254	226	193	193	173	170	170	175	213.8	261.7
14 Spd	11	10	10	12	10	11	10	10	10	10	8	17	14	16	13	22	21	14	19	22	24	28	30	30	14.1	29.9
Dir	170	168	173	181	179	182	186	184	184	186	194	238	236	232	226	238	234	233	238	241	248	245	246	246	223.6	246.0
15 Spd	28	28	23	20	20	16	16	6	3	4	10	29	30	28	22	19	17	15	8	10	9	3	7	7	14.5	29.6
Dir	245	247	245	245	246	248	246	265	148	213	249	253	253	256	258	244	234	234	218	299	301	165	167	175	246.2	252.9
16 Spd	7	7	7	5	6	8	8	6	6	6	3	13	19	19	20	17	4	5	9	9	9	9	11	12	7.9	19.8
Dir	160	161	179	192	179	177	172	173	164	149	182	244	240	236	240	244	211	180	168	183	200	199	194	191	203.8	239.9
17 Spd	10	8	8	4	6	9	9	12	10	10	11	14	21	17	8	10	6	10	10	10	10	10	13	20	10.3	21.2
Dir	191	192	181	193	201	192	187	185	190	204	206	221	225	223	214	221	212	220	215	208	224	214	222	235	210.6	224.8
18 Spd	20	17	11	14	14	13	12	7	6	17	33	32	33	31	28	25	19	12	8	7	8	10	8	9	14.7	33.0
Dir	233	257	265	243	230	229	251	247	197	248	253	252	257	269	271	263	246	230	213	207	189	168	172	170	244.7	252.9
19 Spd	8	8	7	7	7	8	5	6	5	2	2	7	8	7	7	8	8	5	6	5	5	7	9	6	5.0	9.4
Dir	173	171	178	186	167	176	163	182	177	180	78	61	61	111	103	113	124	108	72	92	125	150	148	143	137.8	148.5
20 Spd	8	7	8	9	14	15	12	10	11	26	30	32	35	40	43	40	34	27	15	8	10	8	8	10	18.1	43.0
Dir	148	231	179	198	216	203	204	207	212	226	228	228	229	241	242	239	241	242	232	190	199	197	198	192	225.4	242.2
21 Spd	7	6	7	7	6	7	6	8	9	11	15	11	5	4	4	2	3	4	6	8	7	7	6	7	4.9	14.9
Dir	199	199	191	191	194	214	190	175	219	236	269	275	268	269	274	259	124	145	143	148	149	157	170	171	203.2	268.6
22 Spd	7	6	4	3	4	6	6	2	3	3	7	4	6	8	7	7	6	5	5	6	9	7	7	6	3.0	9.3
Dir	173	164	173	172	166	176	135	3	191	184	135	121	5	14	42	50	74	88	146	161	154	180	195	204	142.3	154.3


Hourly Averages

Wind Speed (km/h)
Wind Direction (deg)
Portable Rycroft - October 2018

Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum	
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
23 Spd	7	6	6	6	5	5	3	6	6	3	1	2	2	4	6	8	8	8	5	4	6	6	6	5	1.7	8.4	
Dir	179	174	200	166	184	237	212	176	172	173	294	286	315	12	26	27	34	43	61	121	195	182	173	193	160.6	34.2	
24 Spd	6	6	4	6	7	7	8	8	7	4	4	3	2	2	4	4	3	4	5	8	7	9	8	7	4.0	8.5	
Dir	189	166	202	177	179	176	186	185	191	208	257	220	246	345	341	24	24	98	188	183	177	173	165	181	183.0	173.5	
25 Spd	8	8	10	10	8	8	11	11	10	8	7	6	8	11	11	13	17	16	13	12	16	15	13	12	10.4	17.3	
Dir	176	164	167	149	150	169	159	161	158	161	169	166	197	194	205	204	189	186	188	191	195	192	187	192	180.7	188.6	
26 Spd	11	12	12	9	8	14	10	7	6	6	8	7	5	4	4	2	3	3	5	5	5	8	11	13	7.1	14.4	
Dir	196	193	187	172	174	183	166	153	171	172	164	159	171	165	163	147	143	130	134	149	172	186	194	195	175.0	183.3	
27 Spd	11	11	11	20	17	17	18	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	19.6	
Dir	205	202	207	225	222	221	224	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	224.9	
28 Spd	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
Dir	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
29 Spd	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
Dir	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
30 Spd	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
Dir	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
31 Spd	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
Dir	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--	
Spd	3.6	3.4	3.4	3.7	3.6	4.0	3.9	2.9	2.5	2.3	3.4	5.1	5.9	6.4	5.3	4.8	3.3	1.6	1.4	1.4	2.6	3.0	3.4	3.5	Diurnal Average		
Dir	179.7	185.7	182.4	183.7	191.7	198.3	196.1	176.1	168.3	223.9	254.7	254.1	258.3	263.1	264.5	261.2	248.7	235.9	178.1	184.0	185.2	167.0	171.2	181.2	Diurnal Maximum		
Spd	28.2	28.4	22.5	20.5	20.1	16.6	18.5	13.1	14.8	26.0	33.0	32.3	34.7	39.9	43.0	39.8	34.4	27.0	18.8	21.8	24.0	27.9	29.6	29.9	Diurnal Maximum		
Dir	245.3	247.2	244.6	244.7	245.7	220.6	223.8	44.1	50.5	226.2	252.9	252.1	228.6	241.4	242.2	239.5	241.0	242.3	237.6	241.0	247.9	245.0	246.0	246.0			
Maximum Speed Value: 43 km/h on Oct 20 15:00																		Minimum Speed Value: 0 km/h on Oct 5 08:00									
Maximum Daily Speed Average: 18.1 km/h on Oct 20																		Minimum Daily Speed Average: 1.0 km/h on Oct 10									
Maximum Diurnal Speed Average: 6.4 km/h at hour 14																		Minimum Diurnal Speed Average: 1.4 km/h at hour 19									
Monthly Average Velocity: 2.81 km/h 214.80 deg																		Speed Percentiles: P ₁ = 1.1 P ₁₀ = 2.8 Q ₁ = 5.0 Median = 7.7 Q ₃ = 12.6 P ₉₀ = 16.7 P ₉₉ = 32.6									
All monthly, daily, and diurnal averages have been calculated using vector methods																		Percent Operational Time: 99.5									
N - Not Valid NS - Not in service																											
Frequency Distribution																											
		Speed Range (km/h)																									
Direction		0 to 5	5 to 11	11 to 19	19 to 28	28 to 38	> 38																			Total	
North		11	11	9	0	0	0																			31	
NorthEast		16	26	68	1	0	0																			111	
East		24	15	0	0	0	0																			39	
SouthEast		20	23	0	0	0	0																			43	
South		40	132	20	0	0	0																			192	
SouthWest		24	30	33	16	8	3																			114	
West		12	28	21	5	8	0																			74	
NorthWest		12	8	2	2	0	0																			24	
Total		159	273	153	24	16	3																			628	

Wind Rose

Wind Speed (WS) (km/h)
Portable Rycroft - October 2018



Hourly Averages - Wind Speed (Scalar)

Wind Speed (km/h)

Portable Rycroft - October 2018

Maximum Speed: 43 km/h on Oct 20 15:00																				Maximum Daily Speed Average: 19.3 km/h on Oct 20										Hours in Service: 631	
Minimum Speed: 1 km/h on Oct 5 08:00																				Minimum Daily Speed Average: 4.2 km/h on Oct 10										Hours of Data: 628	
Maximum Diurnal Speed Average: 12.9 km/h at hour 14																				Minimum Diurnal Speed Average: 6.9 km/h at hour 9										Hours of Missing Data: 3	
Monthly Average Speed: 9.61 km/h																				Percentiles: P ₁ = 1.7 P ₁₀ = 3.2 Q ₁ = 5.2 Median = 7.8 Q ₃ = 12.8 P ₉₀ = 16.9 P ₉₉ = 32.8										Percent Operational Time: 99.5	
Day	Hourly Period Ending At (MST)																								Daily Average	Daily Maximum					
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24							
1-Oct	5	2	3	4	4	6	6	6	4	8	9	8	11	9	9	10	11	15	15	16	15	17	14	13	9.3	16.6					
2-Oct	8	5	6	6	6	4	8	8	8	10	10	12	13	15	14	13	14	15	7	4	7	9	7	7	9.0	15.5					
3-Oct	7	7	7	6	6	6	3	4	2	3	3	3	5	5	8	11	9	7	4	3	3	7	4	4	5.4	11.2					
4-Oct	2	4	2	2	3	3	4	3	3	3	4	8	14	19	16	19	20	15	11	13	13	7	8	5	8.3	19.8					
5-Oct	3	3	3	2	2	1	1	1	4	7	6	6	6	7	6	6	11	10	9	8	7	8	9	4	5.4	10.5					
6-Oct	6	4	2	7	3	6	4	3	5	4	3	3	4	4	3	4	4	4	8	16	15	15	12	14	6.3	16.1					
7-Oct	12	12	14	13	14	14	12	13	13	14	15	14	15	14	14	13	10	12	12	12	11	13	14	17	13.1	16.6					
8-Oct	16	12	11	13	14	13	13	12	15	16	16	17	15	17	18	18	19	20	19	17	18	18	18	16	15.9	19.9					
9-Oct	15	13	14	14	13	12	11	13	12	13	14	15	17	14	15	15	14	14	14	12	9	8	6	7	12.7	16.9					
10-Oct	5	5	5	5	4	3	1	2	3	2	3	4	8	8	7	6	3	2	3	3	3	4	5	6	4.2	7.7					
11-Oct	4	4	5	7	5	6	6	6	9	8	7	11	10	6	7	7	12	6	4	7	3	2	4	5	6.3	11.9					
12-Oct	4	4	4	2	3	10	N	N	N	21	22	20	17	14	17	13	12	7	6	3	2	2	3	3	9.0	22.2					
13-Oct	3	4	3	3	4	5	6	3	3	2	3	3	3	10	14	13	11	5	4	5	7	10	10	10	6.0	14.5					
14-Oct	11	10	10	12	10	11	10	10	10	10	8	17	14	16	14	22	21	14	19	22	24	28	30	30	16.0	29.9					
15-Oct	28	28	23	20	20	17	16	6	4	4	10	29	30	28	22	19	17	15	8	11	9	4	7	7	15.9	29.7					
16-Oct	7	7	7	5	7	8	8	6	6	6	4	13	19	19	20	17	4	5	9	9	9	9	11	12	9.4	19.9					
17-Oct	10	8	8	5	6	9	9	12	10	10	11	14	21	17	8	10	6	10	10	10	10	10	13	20	10.7	21.3					
18-Oct	20	18	11	14	14	13	12	8	6	17	33	32	33	31	28	25	20	12	8	7	8	10	8	9	16.5	33.0					
19-Oct	8	8	7	7	7	8	5	6	6	3	3	7	9	8	8	9	8	5	6	5	6	8	9	7	6.8	9.5					
20-Oct	8	8	8	9	14	15	12	10	11	26	31	32	35	40	43	40	34	27	16	8	10	9	8	10	19.3	43.1					
21-Oct	7	6	8	7	7	7	7	8	10	12	15	11	6	5	5	3	3	4	7	8	7	7	7	7	7.2	15.2					
22-Oct	7	6	5	3	5	6	6	3	4	4	7	5	6	8	8	8	6	5	5	6	9	7	7	6	5.9	9.4					
23-Oct	7	6	6	6	5	5	4	7	6	3	1	3	3	4	7	8	8	8	5	5	6	6	6	5	5.4	8.4					
24-Oct	7	6	4	6	7	7	8	8	7	4	4	3	2	2	4	4	3	4	5	8	7	9	8	7	5.6	8.5					
25-Oct	8	8	10	10	8	8	11	11	10	8	7	6	9	11	11	14	17	16	13	12	16	15	13	12	10.9	17.4					
26-Oct	12	12	12	9	9	15	10	8	6	6	8	7	5	5	4	3	3	4	5	5	6	8	11	13	7.6	14.5					
27-Oct	11	11	11	20	17	17	19	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	19.6					
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--					
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--					
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--					
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--	--					
9.0 8.3 7.7 8.0 8.1 8.7 8.2 7.1 6.9 8.6 9.9 11.7 12.7 12.9 12.7 12.6 11.6 10.0 8.9 9.1 9.3 9.6 9.6 9.8																									Diurnal Average						
28.2 28.4 22.5 20.5 20.1 16.6 18.5 13.2 14.9 26.1 33.0 32.5 34.9 40.2 43.1 39.9 34.5 27.0 18.9 21.8 24.0 27.9 29.6 29.9																									Diurnal Maximum						
N - Not Valid NS - Not in service																															
All monthly, daily, and diurnal averages have been calculated using scalar methods																															

Hourly Standard Deviations

Wind Direction (WD) - deg


Portable Rycroft - October 2018

Maximum Value: 91.5 deg on Oct 23 11:00																		Hours in Service: 631									
Minimum Value: 2.2 deg on Oct 15 04:00																		Hours of Data: 628									
																		Hours of Missing Data: 3									
Percentiles: P ₁ = 2.6 P ₁₀ = 4.1 Q ₁ = 5.8 Median = 9.3 Q ₃ = 18.4 P ₉₀ = 37.8 P ₉₉ = 78.6																		Hours of Calibration: 0									
																		Percent Operational Time: 99.5									
Day	Hourly Period Ending At (MST)																								Daily Maximum		
	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24			
1-Oct	7	65	21	17	15	17	6	20	40	10	21	12	14	23	25	12	8	7	10	5	4	8	5	6	65.3		
2-Oct	9	44	6	6	6	16	19	6	9	14	16	22	25	20	13	18	12	7	7	11	5	3	4	6	43.8		
3-Oct	7	5	5	9	5	8	31	8	42	80	44	79	62	78	21	13	11	9	16	57	14	10	71	47	79.5		
4-Oct	71	51	38	25	88	44	29	26	20	39	52	33	15	13	15	13	12	4	9	3	3	40	37	12	87.8		
5-Oct	30	18	28	55	31	41	41	74	14	18	25	21	35	43	29	45	11	6	4	5	4	5	3	27	73.8		
6-Oct	11	12	77	9	79	23	25	38	18	40	30	80	42	30	55	42	48	86	7	9	5	5	7	6	85.9		
7-Oct	7	12	6	8	5	5	8	8	8	5	7	9	6	7	6	9	9	5	4	3	8	3	3	4	12.3		
8-Oct	4	5	5	6	7	6	5	5	6	7	7	8	9	9	7	7	7	6	7	5	5	5	5	5	9.2		
9-Oct	6	6	4	5	5	8	4	5	8	7	8	9	9	11	12	11	8	6	7	13	13	13	14	19	18.8		
10-Oct	19	15	9	12	18	41	48	29	58	69	31	17	12	9	11	12	20	11	13	10	32	25	13	5	68.7		
11-Oct	9	20	6	12	8	8	12	7	6	7	8	9	9	9	14	22	9	14	49	20	21	32	13	10	49.2		
12-Oct	11	5	21	77	19	20	N	N	N	28	6	5	7	7	5	7	4	14	11	15	16	12	20	14	77.4		
13-Oct	16	6	34	25	7	12	20	9	22	34	12	46	58	16	6	7	11	10	7	26	15	3	3	4	58.4		
14-Oct	2	3	6	3	4	4	5	5	4	6	7	14	7	10	9	5	4	5	4	3	3	3	4	3	13.7		
15-Oct	3	3	3	2	2	3	6	38	28	19	14	4	4	5	6	5	4	4	5	33	19	26	12	12	37.9		
16-Oct	15	4	10	10	14	6	4	9	12	6	36	26	8	7	6	7	18	12	9	18	5	3	9	9	35.9		
17-Oct	6	7	6	18	5	9	4	3	6	5	5	7	5	6	6	8	7	9	3	6	6	7	10	3	17.8		
18-Oct	4	15	16	7	3	4	15	16	12	7	3	5	7	9	8	13	7	4	9	5	7	4	4	3	16.1		
19-Oct	8	3	7	5	5	7	29	9	35	44	39	14	17	19	17	14	12	22	12	15	22	20	8	56	55.6		
20-Oct	10	32	17	11	8	6	4	5	7	6	4	5	5	7	5	4	4	3	19	10	13	7	7	6	31.8		
21-Oct	6	8	7	6	10	13	25	7	19	29	13	12	31	52	62	68	34	14	12	9	8	7	8	8	68.2		
22-Oct	7	8	36	17	17	15	20	55	41	41	17	51	20	13	23	17	20	16	27	22	8	18	6	7	54.9		
23-Oct	18	13	10	16	13	15	34	12	12	20	91	38	39	26	14	8	5	3	11	43	6	7	8	10	91.5		
24-Oct	11	16	9	19	8	10	7	8	8	23	15	30	50	64	24	19	19	30	10	9	4	3	7	5	63.7		
25-Oct	6	5	4	5	6	8	10	6	7	10	9	17	18	15	10	7	4	6	6	12	6	4	5	6	18.0		
26-Oct	3	3	6	3	5	9	8	13	10	15	11	10	20	46	32	40	18	22	18	56	12	12	6	4	56.3		
27-Oct	5	4	8	4	4	4	4	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	7.8		
28-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--		
29-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--		
30-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--		
31-Oct	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	NS	--		
	70.9	65.3	76.7	77.4	87.8	44.1	47.7	73.8	58.0	79.5	91.5	80.0	62.3	77.7	61.9	68.2	48.2	85.9	49.2	56.8	32.4	40.1	71.4	55.6			
N - Not Valid NS - Not in service																											

PAZA

Monthly Passive Data Summary

Location of PAZA Passive Monitoring Stations



PAZA Passive Results for October 2018

Station Number	Station Name	SO2 ppb	O3 ppb	NO2 ppb	H2S ppb	Site Legal
Duplicates						
2a	Bay Tree		50.9			
2b	Bay Tree		31.5			
19a	Wanham			1.2		
19b	Wanham			1.2		
21a	Eaglesham	0.1				
21b	Eaglesham	0.1				
46a	Little Smoky	0.1		3.9		
46b	Little Smoky	0.1		3.9		
63a	Girouxville 3				Damaged	
63b	Girouxville 3				Damaged	
2	Bay Tree	0.2	41.2	1.1		13-16-078-13 W6M
3	Fourth Creek	0.3		0.3		04-13-082-07 W6M
9	Spirit River	0.2		0.2		08-12-079-07 W6M
12	Hythe	0.2		0.2		14-36-072-11 W6M
14	Sylvester	0.1		0.7		08-06-069-12 W6M
16	Beaverlodge	0.2		2.1		15-36-071-10 W6M
18	Saddle Hills	0.2		1.0		04-25-074-07 W6M
19	Wanham	0.3		1.2		16-22-077-03 W6M
21	Eaglesham	0.1		1.3		16-21-079-25 W5M
24	Wembley	0.1		2.5		12-31-070-08 W6M
25	Pinto Creek	0.1		1.0		04-24-069-11 W6M
27	Grande Prairie I	0.2		6.3		08-15-071-06 W6M
29	Smoky Heights	0.3		2.1		04-06-075-02 W6M
30	Fitzsimmons	0.2		2.0		15-36-072-03 W6M
32	Gold Creek	0.4		4.2		06-33-067-05 W6M
35	Jean Cote	0.1		1.1		12-35-079-21 W5M
36	Guy	0.1		1.1	0.1	03-04-076-22 W5M
37	Crooked Creek	0.3	25.6	2.3		16-01-071-26 W5M
39	Clouston Creek	0.2		1.1		12-01-073-22 W5M
40	McLennan	0.2		1.1		03-29-077-19 W5M
42	Sunset House	0.2		0.9		05-32-070-19 W5M
44	Peavine	0.1		0.6		03-05-079-15 W5M
46	Little Smoky	0.1		3.9		12-01-065-21 W5M
47	Kinuso	0.1	26.0	0.9		12-10-073-10 W5M
49	Grande Prairie HP	0.2		8.3		17-26-071-06 W6M
50	East Prairie	0.1		0.6		13-02-072-15 W5M
63	Girouxville 3				Damaged	14-02-077-23 W5M
64	Girouxville 4				0.2	4-08-077-22 W5M

*BDL = Below Detection Level

*N/S - No sample

Passive Summary for October 2018


Stats	Sulphur Dioxide SO ₂	Ozone O ₃	Nitrogen Dioxide NO ₂	Hydrogen Sulphide H ₂ S
	ppb	ppb	ppb	ppb
Passive Summary for October 2018 (PAZA Zone)				
Mean	0.2	30.9	1.9	0.2
Standard Deviation	0.1	8.9	1.9	0.1
Minimum	0.1	25.6	0.2	0.1
Minimum At	Sylvester (#14)	Crooked Creek (#37)	Hythe (#12)	Guy (#36)
Maximum	0.4	41.2	8.3	0.2
Maximum At	Gold Creek (#32)	Bay Tree (#2)	Grande Prairie HP	Girouxville 4 (#64)

Comparison between Continuous and Passive monitoring at Beaverlodge (passive #16 Beaverlodge)


	SO ₂	NO ₂
PAZA Beaverlodge station	0.1	4.0
PAZA Beaverlodge passive	0.2	2.1


Comparison between Continuous and Passive monitoring at Henry Pirker (passive #49 Grande Prairie HP)

	SO ₂	NO ₂
PAZA Henry Pirker station	0.3	10.9
PAZA Grande Prairie passive	0.2	6.3




Alberta Ambient Air Quality Objective - 30-day Objective is 11 ppb







Alberta Ambient Air Quality Objective - No Annual O₃ Objective



Alberta Ambient Air Quality Objective - Annual NO₂ Objective is 24 ppb





October 2018 Calibration Reports

Calibration Report



Parameter SO2

Air Monitoring Network PAZA

Station Information

Calibration Date	October 14, 2018	Previous Calibration	September 11, 2018
Station Number	4	Station Location	Beaverlodge
Reason:	Routine	☐ Install	☐ Removal
		☐ Other:	
Start Time (MST)	13:15	End Time (MST)	15:40
Barometric Pressure	0.921 atm	Station Temperature	20.0 Deg C
Calibrator	EnviroNics 6100	Serial Number	3016
Cal Gas Concentratio	10.1 ppm	Cal Gas Expiry Date	21/02/2021
Gas Cert Reference	EX0012231		
DACS make	CR3000	DACS serial No.	5237
DACS voltage range	0 - 5 volt	DACS channel #	5
	<u>Before</u>		<u>After</u>
DACS Scale High	100	DACS slope	100
DACS Scale Low	0	DACS intercept	0
Calculated slope	1.000983	Calculated slope	0.997848
Calculated intercept	0.197878	Calculated intercept	0.184743
Analyzer make	TEI Model 43i-TLE	Analyzer serial #	713021137

	before		after	
Concentration range	0 - 100	ppb	0 - 100	ppb
Background	2.74		2.8	
Coefficient	0.936		0.962	
PMT	-768.2	V	-768.2	V
UV Lamp Voltage	1121	V	1121	V
Chamber Temp	45	Deg C	45	Deg C
Pressure	667.3	mm Hg	667.00	mm Hg
Sample Flow	0.530	LPM	0.53	LPM
Lamp Intensity	94	%	94	%

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
4995	0.00	0.0	0.1	N/A
4995	39.94	80.1	80.2	0.9993
4995	19.97	40.2	40.2	1.0014
4995	9.97	20.1	19.6	1.0268
4995	0.00	0.0	0.1	As found zero
4995	39.94	80.1	78.1	As found span
Average Correction Factor				1.0092

Calculated value of As Found Response: 78.245 ppm Percent Change of As Found: 2.3%

	before calibration		after calibration	
Auto zero	0.1	ppb	0.3	ppb
Auto span	60.0	ppb	60.2	ppb

Notes: Slight span adjustment made.

Calibration Performed By: Grover Christiansen

Calibration Summary

Parameter SO2

Air Monitoring Network PAZA




Station Information

Calibration Date	October 14, 2018	Previous Calibration	September 11, 2018
Station Number	4	Station Location	Beaverlodge
Start Time (MST)	13:15	End Time (MST)	15:40
Analyzer make/model	TEI Model 43i-TLE	Analyzer serial #	713021137


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.1	N/A	Correlation Coefficient	0.999935
80.1	80.2	0.9993		
40.2	40.2	1.0014		
20.1	19.6	1.0268	Slope	0.997848
			Intercept	0.184743

SO2 Calibration Curve



SO2 Calibration



October 14, 2018

Parameter

Air Monitoring Network

NO_x-NO-NO₂
PAZA



Station Information

Calibration Date	October 13, 2018		Previous Calibration	September 6, 2018	
Station Number	4		Station Location	Beaverlodge	
Reason:	Routine Installation Removal		Other:		
Start Time (MST)	10:20		End Time (MST)	14:25	
Barometric Pressure	0.921	Atm	Station Temperature	20.0	Deg C
Calibrator	EnviroNics 6100		Serial Number	3016	
NO Cal Gas Conc	50.4	ppm	Cal Gas Expiry Date	10/24/2020	
NOx Cal Gas Conc	50.5	ppm	Cal Gas Serial #	LL107945	

DACS Information

DACS make CR3000 DACS serial No. 5237

Parameter		NO ₂	NO _x	NO
Before	Data Slope	1.000422	0.986002	0.987135
	Data Offset	-0.108666	1.555934	1.750910
After	Data Slope	1.000685	0.997107	0.997429
	Data Offset	0.305393	1.635322	1.471479
Channel #		8	6	7
Voltage Range		0 - 5 VDC	0 - 5 VDC	0 - 5 VDC

Analyzer Information

Analyzer make/model	TEI 42i	Analyzer serial #	906535068	
Test Point	before		after	
Concentration range	0-500	ppb	0-500	ppb
NO offset	2.1	mV	2.1	mV
NOx bkgnd	2.9	mV	2.9	mV
NO coefficient	0.993		0.993	
NOx coefficient	1.001		1.001	
NO2 conv temp	322.4	Deg C	322.4	Deg C
PMT Temp	-3.0	Deg C	-3.0	Deg C
PMT Volt	-729.6	mV	-729.6	mV
R Cell Press	175.9	in Hg	175.9	in Hg
Sample Flow	0.802	LPM	0.803	LPM

Notes: No adjustment made

Calibration Report

Parameter **NOX-NO-NO2**
Air Monitoring Network **PAZA**



Station Information

Calibration Date: October 13, 2018 Station Location: Beaverlodge

Calibration Data

	Dilution flow rate (ccm)	Source gas flow rate (ccm)	Calculated NOx conc (ppb)	Calculated NO conc (ppb)	Calculated NO2 conc (ppb)	Indicated NOx conc (ppb)	Indicated NO conc (ppb)	Indicated NO2 conc (ppb)	NOx Correction factor	NO Correction factor
zero	4995	0.00	0.0	0.0	0.0	-0.3	-0.1	-0.3	N/A	N/A
1	4995	39.94	400.6	399.8	0.8	400.7	400.1	-0.2	0.9998	0.9992
2	4995	19.97	201.1	200.7	0.4	199.7	198.8	0.1	1.0070	1.0096
3	4995	9.97	100.6	100.4	0.2	97.7	98.0	-0.4	1.0296	1.0246
AFZ	4995	0.00	0.0	0.0	0.0	-0.3	-0.1	-0.3	0.0000	0.0000
AFS	4995	39.94	400.6	399.8	0.8	401.4	401.0	-0.3	0.9979	0.9970
								Average Correction Factor	1.0121	1.0111

As Found Concentrations: NO_x= 397.7 NO= 397.7 As Found Percent Change NO_x= -0.7% NO= -0.5%

GPT Calibration Data

Dilution Flow 4995 ccm Source Gas Flow 39.94 ccm

O3 Setpoint (ppb)	Indicated NOx high point (ppb)	Indicated NO drop conc (ppb)	Calculated NO2 conc (ppb)	Indicated NOx conc (ppb)	Indicated NO conc (ppb)	Indicated NO2 conc (ppb)	NOx Correction factor	NO Correction factor	NO2 Correction factor	Converter Efficiency
0	-0.1	-0.1	0.0	-0.3	-0.1	-0.3	N/A	N/A	N/A	N/A
NO point	399.6	399.6	0.0	399.8	399.6	-0.6	0.9997	1.0000	N/A	N/A
300	399.6	101.8	297.8	399.8	101.8	297.4	0.9996	1.0000	1.0014	99.9%
200	399.6	194.3	205.3	399.7	194.3	204.7	0.9999	1.0000	1.0033	99.7%
100	399.6	288.8	110.9	399.9	288.8	110.6	0.9994	1.0000	1.0021	99.8%
Average Correction Factor							0.9996	1.0000	1.0023	99.8%

AIC Data

	Previous calibration				Current calibration			
Parameter	NOx	NO2	NO		NOx	NO2	NO	
Auto zero	-0.3	-0.3	-0.1	ppb	0.1	0.0	0.0	ppb
Auto span	289.4	286.0	2.5	ppb	307.3	304.6	2.0	ppb

Calibration Performed By: Grover Christiansen

Calibration Summary

Parameter NO₂

Air Monitoring Network PAZA




Station Information

Calibration Date	October 13, 2018	Previous Calibration	September 6, 2018
Station Number	4	Station Location	Beaverlodge
Start Time (MST)	10:20	End Time (MST)	14:25
Analyzer make	TEI 42i	Analyzer serial #	906535068

Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	-0.3	N/A	Correlation Coefficient	0.999998
297.8	297.4	1.0014		
205.3	204.7	1.0033		
110.9	110.6	1.0021	Slope	1.000685
			Intercept	0.305393

NO₂ Calibration Curve



Calibration Summary

Parameter NO_x

Air Monitoring Network PAZA




Station Information

Calibration Date	October 13, 2018	Previous Calibration	September 6, 2018
Station Number	4	Station Location	Beaverlodge
Start Time (MST)	10:20	End Time (MST)	14:25
Analyzer make	TEI 42i	Analyzer serial #	906535068

Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	-0.3	N/A	Correlation Coefficient	0.999948
400.6	400.7	0.9998		
201.1	199.7	1.0070		
100.6	97.7	1.0296	Slope	0.997107
			Intercept	1.635322

NO_x Calibration Curve



Calibration Summary

Parameter NO

Air Monitoring Network PAZA




Station Information

Calibration Date	October 13, 2018	Previous Calibration	September 6, 2018
Station Number	4	Station Location	Beaverlodge
Start Time (MST)	10:20	End Time (MST)	14:25
Analyzer make	TEI 42i	Analyzer serial #	906535068


Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	-0.1	N/A		
399.8	400.1	0.9992	Correlation Coefficient	0.999945
200.7	198.8	1.0096		
100.4	98.0	1.0246		
			Slope	0.997429
			Intercept	1.471479

NO Calibration Curve



NO_x Calibration



October 13, 2018

Calibration Report

Parameter 03

Air Monitoring Network PAZA



Station Information

Calibration Date	October 13, 2018	Previous Calibration	September 6, 2018
Station Number	4	Station Location	Beaverlodge
Reason:	☒ Routine	☐ Install	☐ Removal
		☐ Other:	
Start Time (MST)	13:25:00 PM	End Time (MST)	15:50
Barometric Pressure	0.921 atm	Station Temperature	20.0 Deg C
Calibrator	EnviroNics 6103	Serial Number	6586
Cal Gas Concentration	NA	Cal Gas Expiry Date	NA
DACS make	CR3000	DACS serial No.	5237
DACS voltage range	0 - 5 volt	DACS channel #	9
	Before		After
Calculated slope	0.965556	Calculated slope	1.000903
Calculated intercept	0.250823	Calculated intercept	0.811066

Analyzer make Teco 49i Analyzer serial # 1136451236,AMU 1879

	before		after	
Concentration range	0 - 500	ppb	0 - 500	ppb
offset	0.20	ppb	0.10	ppb
slope	1.021		0.991	
Lamp temp	67.5	mV	67.5	mV
Lamp Intensity A/B	63679/77398	mV	63492/77011	mV
Pressure	692	mm Hg	692	mm Hg
Flow A	0.784	LPM	0.790	LPM
Flow B	0.742	LPM	0.748	LPM

Calibration Data

Dilution air flow rate (cc/min)	Calibrator Setting	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
5032	0.00	0.0	0.1	N/A
5032	0.30	297.8	297.8	0.9999
5032	0.20	205.3	203.2	1.0103
5032	0.10	110.9	109.0	1.0172
5032				
5032	0.00	0.0	0.1	As found zero
5032	0.30	297.8	306.4	As found span
Average Correction Factor				1.0091

Calculated value of As Found Response: 295.9 ppm Percent Change of As Found: -0.6%

	before calibration		after calibration	
Auto zero	0.1	ppb	0.2	ppb
Auto span	274.3	ppb	270.0	ppb

Notes: Slight span adjustment made

Calibration Performed By: Grover Christiansen

Calibration Summary

Parameter 03

Air Monitoring Network PAZA




Station Information

Calibration Date	October 13, 2018	Previous Calibration	September 6, 2018
Station Number	4	Station Location	Beaverlodge
Start Time (MST)	13:25:00 PM	End Time (MST)	15:50
Analyzer make/model	Teco 49i	Analyzer serial #	1136451236,AMU 1879


Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.1	NA	Correlation Coefficient	0.999913
297.8	297.8	0.9999		
205.3	203.2	1.0103		
110.9	109.0	1.0172	Slope	1.000903
			Intercept	0.811066

O3 Calibration Curve



O3 Calibration



October 13, 2018

Calibration Report



Parameter SO₂
Air Monitoring Network PAZA

Station Information

Calibration Date	October 19, 2018	Previous Calibration	September 27, 2018
Station Number	2	Station Location	Evergreen Park
Reason:	☒ Routine	☐ Install	☐ Removal
		☐ Other:	

Start Time (MST)	13:00:00 PM	End Time (MST)	14:57:00 PM
Barometric Pressure	0.936 ATM	Station Temperature	22.0 Deg C
Calibrator	EnviroNics 6100	Serial Number	3474
Cal Gas Concentratio	49.6 ppm	Cal Gas Expiry Date	04/05/2019
Correction factor	0.031602	Cal Gas Cylinder #	LL107945
DACS make	CR3000	DACS serial No.	5236
DACS voltage range	0 - 5 volt	DACS channel #	6
	Before		After
Calculated slope	1.000773	Calculated slope	1.000878
Calculated intercept	-0.273893	Calculated intercept	-0.654359

Analyzer make	Teco 43i	Analyzer serial #	701120008
---------------	----------	-------------------	-----------

	before		after	
Concentration range	0 - 500	ppb	0 - 500	ppb
Background	12.1		12	
coefficient	1.072		1.072	
Lamp Voltage	842	volts	845	volts
Chamber Temp	45	Deg C	45	Deg C
Perm Gas Temp	45	Deg C	45	Deg C
Pressure	680.2	mm Hg	678.6	mm Hg
Sample Flow	0.436	ccm	0.439	ccm
Lamp Intensity	92	%	92	%

Dilution air flow rate (cc/min)	Corrected gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
4997	0.0	0.00	0.8	N/A
4997	39.97	393.6	393.9	0.9993
4997	19.99	197.6	198.4	0.9961
4997	9.98	98.9	99.0	0.9984
4997	0.0	0.0	0.8	As Found Zero
4997	39.98	393.7	393.9	As Found Span
Average Correction Factor				0.9979

Calculated value of As Found Response: 393.157 ppm Percent Change of As Found: 0.1%

	before calibration		after calibration	
Auto zero	0.8	ppm	0.9	ppm
Auto span	259.3	ppm	260.0	ppm

Notes: No adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter SO2
Air Monitoring Network PAZA




Station Information

Calibration Date	October 19, 2018	Previous Calibration	September 27, 2018
Station Number	2	Station Location	Evergreen Park
Start Time (MST)	13:00:00 PM	End Time (MST)	14:57:00 PM
Analyzer make/model	Teco 43i	Analyzer serial #	701120008


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.8	N/A	Correlation Coefficient	0.999997
393.6	393.9	0.9993		
197.6	198.4	0.9961		
98.9	99.0	0.9984	Slope	1.000878
			Intercept	-0.654359

SO2 Calibration Curve



SO2 Calibration



October 19, 2018

Calibration Report



Parameter TRS

Air Monitoring Network PAZA

Station Information

Calibration Date	October 19, 2018	Previous Calibration	September 27, 2018
Station Number	2	Station Location	Evergreen Park
Reason:	Routine	Install	Removal
Start Time (MST)	11:10	End Time (MST)	14:10:00 PM
Barometric Pressure	0.936 ATM	Station Temperature	22.0 Deg C
Calibrator	Evironics 6100	Serial Number	3474
Cal Gas Conc	9.78 ppm	Cal Gas Expiry Date	23/02/2019
Correction factor	0.031602	Cal Gas Cylinder #	EY0000380
DACS make	CR3000	DACS serial No.	5236
DACS voltage range	0 - 5 volt	DACS channel #	5
	<u>Before</u>		<u>After</u>
Calculated slope	1.000771	Calculated slope	0.994402
Calculated intercept	0.255830	Calculated intercept	-0.009313

Analyzer make TEI Model 43I

Analyzer serial #

	before		after	
Concentration range	0 - 100	ppb	0 - 100	ppb
Background	27.2	ppb	26.5	ppb
coefficient	0.880		0.880	
Lamp Voltage	888	volts	890	volts
Chamber Temp	44.6	Deg C	44.6	Deg C
Perm Gas Temp	44.9	Deg C	45	Deg C
Pressure	664.9	mm Hg	662.2	mm Hg
Sample Flow	0.466	ccm	0.464	ccm
Lamp Intensity	30,458	mv	30,722	mv

Dilution air flow rate (cc/min)	Corrected gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
4996	0.00	0.00	0.2	N/A
4996	39.97	77.62	78.2	0.9930
4996	19.96	38.92	39.0	0.9984
6999	9.97	13.91	13.9	1.0018
4996	10.00	0.00	0.4	Sox
4996	0.00	0.00	-0.6	As Found Zero
4996	39.97	77.62	76.9	As Found Span
Average Correction Factor				0.9977

Calculated value of As Found Response: 77.79 ppm

Percent Change of As Found: **-0.2%**

	before calibration		after calibration	
Auto zero	-0.2	ppm	0.3	ppm
Auto span	94.4	ppm	96.5	ppm

Notes: Zero adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter TRS

Air Monitoring Network PAZA




Station Information

Calibration Date	October 19, 2018	Previous Calibration	September 27, 2018
Station Number	2	Station Location	Evergreen Park
Start Time (MST)	11:10	End Time (MST)	14:10:00 PM
Analyzer make/model	TEI Model 43I	Analyzer serial #	0


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.2	N/A	Correlation Coefficient	0.999977
77.6	78.2	0.9930		
38.9	39.0	0.9984		
13.9	13.9	1.0018	Slope	0.994402
			Intercept	-0.009313

TRS Calibration Curve



TRS Calibration



Calibration Report

Parameter H2S
Air Monitoring Network

PAZA



Station Information

Calibration Date	October 9, 2018	Previous Calibration	September 12, 2018
Station Number	1	Station Location	Henry Pirker
Reason:	☒ Routine	☐ Install	☐ Removal
		☐ Other:	
Start Time (MST)	12:05	End Time (MST)	15:00:00 PM
Barometric Pressure	0.921 mBar	Station Temperature	22.0 Deg C
Calibrator	EnviroNics 6103	Serial Number	6586
Cal Gas Conc	9.78 ppb	Cal Gas Expiry Date	02/28/2020
		Cal Gas Cylinder #	LL119574
DACS make	CR3000	DACS serial No.	5408
DACS voltage range	0 - 1 volt	DACS channel #	9
	Before		After
Calculated slope	0.991497	Calculated slope	0.991809
Calculated intercept	0.601916	Calculated intercept	0.163916
Analyzer make	TEI 450i	Analyzer serial #	1170050143(AMU 2198)

	before		after	
Concentration range	0 - 100	ppb	0 - 100	ppb
Coefficient	0.875		0.831	
Background	13.7		12.7	
Pressure	558.0	mm Hg	555.6	mm Hg
Flow	0.968	ccm	0.964	ccm
Lamp Voltage	777	V	778	V

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
4997	0.00	0.00	0.32	N/A
4997	39.98	77.63	78.26	0.9919
4998	19.97	38.92	38.91	1.0002
6997	9.98	13.93	13.40	1.0395
4997	0.00	0.00	-0.27	As Found Zero
4997	41.20	79.98	81.40	As Found Span
Average Correction Factor				1.0105

Calculated value of As Found Response: 81.6 ppb Percent Change of As Found: -2.0%

	before calibration		after calibration	
Auto zero	-0.31	ppb	0.35	ppb
Auto span	30.06	ppb	29.43	ppb

Notes: Zero & span adjustments made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter H2S

Air Monitoring Network PAZA




Station Information

Calibration Date	October 9, 2018	Previous Calibration	September 12, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	12:05	End Time (MST)	15:00:00 PM
Analyzer make/model	TEI 450i	Analyzer serial #	1170050143(AMU 2198)


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.316	N/A	Correlation Coefficient	0.999855
77.627	78.264	0.9919		
38.921	38.914	1.0002		
13.930	13.401	1.0395	Slope	0.991809
			Intercept	0.163916

TRS Calibration Curve



TRS Calibration



Calibration Report

Parameter SO2

Air Monitoring Network PAZA



Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 21, 2018
Station Number	1	Station Location	Henry Pirker
Reason:	☒ Routine	☐ Install	☐ Removal
		☐ Other:	

Start Time (MST)	8:35	End Time (MST)	11:35
Barometric Pressure	923.0 mbar	Station Temperature	20.0 Deg C
Calibrator	EnviroNics	Serial Number	6586
Cal Gas Conc	50.1 ppm	Cal Gas Cert Date	10/24/2020
		Cal Gas Cylinder #	LL107945
DACS make	CR3000	DACS serial No.	5408
DACS voltage range	0 - 1 volt	DACS channel #	10
	Before		After
Calculated slope	1.000385	Calculated slope	0.996997
Calculated intercept	1.564898	Calculated intercept	2.404827

Analyzer make TEI 43I-TLE Analyzer serial # 1507864682 (AMU 2006)

	before		after	
Concentration range	0 - 500	ppb	0 - 500	ppb
Background	1.40		1.43	
Coefficient	0.905		0.926	
Pressure	671.0	mm Hg	669.8	mm Hg
Flow	0.443	lpm	0.443	lpm
Lamp intensity	91	Hz	91	Hz

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4995	0.00	0.0	0.2	N/A
4995	39.99	397.9	397.9	1.0000
4995	19.97	199.5	196.5	1.0151
4995	9.98	99.9	95.2	1.0496
4995	0.00	0.0	0.2	As Found Zero
4995	39.97	397.7	385.8	As Found Span
Average Correction Factor				1.0216

Calculated value of As Found Response: 387.4 ppb Percent Change of As Found: 2.6%

	before calibration		after calibration	
Auto zero	0.2	ppb	0.2	ppb
Auto span	276.3	ppb	279.4	ppb

Notes: Span adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter SO2
Air Monitoring Network PAZA




Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 21, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	8:35	End Time (MST)	11:35
Analyzer make/model	TEI 43I-TLE	Analyzer serial #	1507864682 (AMU 2006)


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.2	N/A	Correlation Coefficient	0.999814
397.9	397.9	1.0000		
199.5	196.5	1.0151		
99.9	95.2	1.0496	Slope	0.996997
			Intercept	2.404827

SO2 Calibration Curve



SO2 Calibration



Calibration Report

Parameter
Air Monitoring Network

NO_x-NO-NO₂
PAZA



Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 11, 2018
Station Number	1	Station Location	Henry Pirker
Reason:	Routine	Installation	Removal
Start Time (MST)	8:45	End Time (MST)	13:30:00 PM
Barometric Pressure	0.923 mbar	Station Temperature	20.0 Deg C
Calibrator	EnviroNics 6103	Serial Number	6586
NO Cal Gas Conc	50.7 ppm	Cal Gas Expiry Date	10/24/2020
NO _x Cal Gas Conc	50.8 ppm	Cal Gas Serial #	LL107945

DACS Information

DACS make	CR3000	DACS serial No.	5408	
	Parameter	NO ₂	NO _x	NO
Before	Data Slope	1.007421	0.999633	0.995852
	Data Offset	-0.741001	0.757281	1.161816
After	Data Slope	1.003940	1.000158	0.996899
	Data Offset	-0.995607	3.143073	3.506504
	Channel #	8	6	7
	Voltage Range	0 - 5 VDC	0 - 5 VDC	0 - 5 VDC

Analyzer Information

Analyzer make/model	42i	Analyzer serial #	906535087	
Test Point	before		after	
Concentration range	0 - 500	ppb	0 - 500	ppb
NO offset	8.3	mV	8.5	mV
NO _x bkgnd	8.5	mV	8.7	mV
NO coefficient	1.198		1.230	
NO _x coefficient	0.999		0.999	
NO ₂ conv temp	326.8	Deg C	326.8	Deg C
Cooler Temp	-2.9	Deg C	-2.8	Deg C
PMT Volt	-834.4	mV	-834.4	mV
R Cell Press	205.6	in Hg	203.8	in Hg
Sample Flow	0.511	LPM	0.514	LPM

Notes: Span adjustment made

Calibration Report

Parameter **NOX-NO-NO2**
Air Monitoring Network **PAZA**



Station Information

Calibration Date: October 5, 2018 Station Location: Henry Pirker

Calibration Data

	Dilution flow rate (ccm)	Source gas flow rate (ccm)	Calculated NOx conc (ppb)	Calculated NO conc (ppb)	Calculated NO2 conc (ppb)	Indicated NOx conc (ppb)	Indicated NO conc (ppb)	Indicated NO2 conc (ppb)	NOx Correction factor	NO Correction factor
zero	4996	0.00	0.0	0.0	0.0	0.2	0.2	0.6	N/A	N/A
1	4996	39.97	403.2	402.4	0.8	401.7	402.2	-0.1	1.0036	1.0006
2	4996	19.98	202.4	202.0	0.4	197.0	196.4	0.3	1.0269	1.0283
3	4996	9.97	101.2	101.0	0.2	95.1	94.8	0.4	1.0643	1.0657
AFZ	4996	0.00	0.0	0.0	0.0	0.2	0.2	0.6	0.0000	0.0000
AFS	4996	39.97	403.2	402.4	0.8	391.2	391.7	-0.1	1.0307	1.0274
Average Correction Factor									1.0316	1.0315

As Found Concentrations: NO_x= 391.6 NO= 391.0 As Found Percent Change NO_x= -2.9% NO= -2.8%

GPT Calibration Data

Dilution Flow 4995 ccm Source Gas Flow 39.93 ccm

O3 Setpoint (ppb)	Indicated NO high point (ppb)	Calculated NO drop conc (ppb)	Calculated NO2 conc (ppb)	Indicated NOx conc (ppb)	Indicated NO conc (ppb)	Indicated NO2 conc (ppb)	NOx Correction factor	NO Correction factor	NO2 Correction factor	Converter Efficiency
0	0.2	0.2	0.0	0.2	0.2	0.6	N/A	N/A	N/A	N/A
NO point	398.9	398.9	0.0	398.5	398.9	-0.2	1.0010	1.0000	N/A	N/A
300	398.9	98.8	300.0	398.4	98.8	299.5	1.0012	1.0000	1.0017	99.8%
200	398.9	193.6	205.3	400.0	193.6	205.6	0.9971	1.0000	0.9983	100.2%
100	398.9	290.3	108.5	400.9	290.3	109.7	0.9948	1.0000	0.9896	101.1%
Average Correction Factor							0.9977	1.0000	0.9965	100.4%

AIC Data

	Previous calibration				Current calibration			
Parameter	NOx	NO2	NO		NOx	NO2	NO	
Auto zero	0.0	0.2	0.1	ppb	0.0	0.2	0.2	ppb
Auto span	291.6	288.9	2.4	ppb	290.9	288.3	2.3	ppb

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter NO₂

Air Monitoring Network PAZA




Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 11, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	8:45	End Time (MST)	13:30:00 PM
Analyzer make	42i	Analyzer serial #	906535087

Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.6	N/A	Correlation Coefficient	0.999987
300.0	299.5	1.0017		
205.3	205.6	0.9983		
108.5	109.7	0.9896	Slope	1.003940
			Intercept	-0.995607

NO₂ Calibration Curve



Calibration Summary

Parameter NO_x

Air Monitoring Network PAZA




Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 11, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	8:45	End Time (MST)	13:30:00 PM
Analyzer make	42i	Analyzer serial #	906535087

Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.2	N/A	Correlation Coefficient	0.999690
403.2	401.7	1.0036		
202.4	197.0	1.0269		
101.2	95.1	1.0643	Slope	1.000158
			Intercept	3.143073

NO_x Calibration Curve



Calibration Summary

Parameter NO

Air Monitoring Network PAZA




Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 11, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	8:45	End Time (MST)	13:30:00 PM
Analyzer make	42i	Analyzer serial #	906535087


Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.2	N/A		
402.4	402.2	1.0006	Correlation Coefficient	0.999620
202.0	196.4	1.0283		
101.0	94.8	1.0657		
			Slope	0.996899
			Intercept	3.506504

NO Calibration Curve



NO_x Calibration



Calibration Report

Parameter 03

Air Monitoring Network PAZA



Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 11, 2018
Station Number	1	Station Location	Henry Pirker
Reason:	Routine	Install	Removal
		Other:	

Start Time (MST)	12:00	End Time (MST)	14:20:00 PM
Barometric Pressure	920.0 mbar	Station Temperature	20.0 Deg C
Calibrator	Envionics	Serial Number	6586
Cal Gas Concentration	NA	Cal Gas Expiry Date	NA

DACS make	CR3000	DACS serial No.	5237
DACS voltage range	0 - 5 volt	DACS channel #	9
	Before		After
Calculated slope	1.005155	Calculated slope	0.983351
Calculated intercept	-0.321566	Calculated intercept	-0.115935

Analyzer make	Teco 49I	Analyzer serial #	1507964699 (AMU:2015)
---------------	----------	-------------------	-----------------------

	before		after	
Concentration range	0 - 500	ppb	0 - 500	ppb
offset	-0.30	ppb	-0.30	ppb
slope	1.013		1.013	
Lamp temp	53.6	mV	53.6	mV
Lamp Intensity A/B	62969/68213	mV	62812/68134	mV
Pressure	692.8	mm Hg	692.9	mm Hg
Flow A	0.728	ccm	0.728	ccm
Flow B	0.732	ccm	0.732	ccm

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
4996	0.0	0.0	-0.1	N/A
4998	0.3	300.0	305.5	0.9820
4998	0.2	205.3	208.1	0.9865
4998	0.1	108.5	111.1	0.9764
4996	0.0	0.0	-0.1	As found zero
4996	0.3	300.0	305.5	As found span
Average Correction Factor				0.9816

Calculated value of As Found Response: 306.8 ppm Percent Change of As Found: 2.3%

	before calibration		after calibration	
Auto zero	0.3	ppb	0.0	ppb
Auto span	374.5	ppb	380.9	ppb

Notes: No adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter 03
Air Monitoring Network PAZA




Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 11, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	12:00	End Time (MST)	14:20:00 PM
Analyzer make/model	Teco 49I	Analyzer serial #	1507964699 (AMU:2015)


Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	-0.1	NA	Correlation Coefficient	0.999977
300.0	305.5	0.9820		
205.3	208.1	0.9865		
108.5	111.1	0.9764	Slope	0.983351
			Intercept	-0.115935

O3 Calibration Curve



O3 Calibration



October 5, 2018

Calibration Report

Parameter CO
Air Monitoring Network PAZA



Station Information

Calibration Date	October 9, 2018	Previous Calibration	September 12, 2018
Station Number	1	Station Location	Henry Pirker
Reason:	Routine	Install	Removal
Start Time (MST)	13:40:00 PM	End Time (MST)	16:35:00 PM
Barometric Pressure	0.921 mBar	Station Temperature	23.0 Deg C
Calibrator	EnviroNics 6100	Serial Number	6586
Cal Gas Conc	2906 ppm	Cal Gas Expiry Date	07/07/2023
		Cal Gas Cylinder #	LL109096
DACS make	CR3000	DACS serial No.	5408
DACS voltage range	0 - 5 volt	DACS channel #	9
	Before		After
Calculated slope	1.003742	Calculated slope	1.000999
Calculated intercept	-0.298527	Calculated intercept	-0.170489
Analyzer make	Model 48I-TLE	Analyzer serial #	1408761378

	before		after	
Concentration range	0 - 50	ppm	0 - 50	ppm
CO zero setting	10.771		10.814	
CO span setting	1.087		1.087	
Sample pressure	704.9	mm Hg	708.2	mm Hg
Sample Flow	0.450	LPM	0.452	LPM

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
4997	0.00	0.00	0.00	N/A
4998	69.96	40.12	40.12	0.9998
4998	34.98	20.20	20.53	0.9836
4998	17.99	10.42	10.69	0.9752
4998	0.00	0.00	0.25	As Found Zero
4998	69.96	40.12	40.90	As Found Span
Average Correction Factor				0.9862

Calculated value of As Found Response: 40.503 ppm Percent Change of As Found: -1.0%

	before calibration		after calibration	
Auto zero	0.05	ppm	0.15	ppm
Auto span	20.72	ppm	20.56	ppm

Notes: Zero & span adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter CO
Air Monitoring Network PAZA




Station Information

Calibration Date	October 9, 2018	Previous Calibration	September 12, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	13:40:00 PM	End Time (MST)	16:35:00 PM
Analyzer make/model	Model 48I-TLE	Analyzer serial #	1408761378


Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.003	N/A	Correlation Coefficient	0.999899
40.116	40.123	0.9998		
20.197	20.533	0.9836		
10.422	10.687	0.9752	Slope	1.000999
			Intercept	-0.170489

CO Calibration Curve



CO Calibration



October 9, 2018

Calibration Report



Parameter CH4 / NMHC / THC

Air Monitoring Network PAZA

Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 15, 2018
Station Number	1	Station Location	Henry Pirker
Reason:	Routine	Install	Removal
			Other: Mair

Start Time (MST)	13:10:00 PM	End Time (MST)	16:25:00 PM
Barometric Pressure	0.921 mBar	Station Temperature	23.0 Deg C
Calibrator	EnviroNics 6103	Serial Number	6586
Cal Gas CH4 Conc	995.6 ppm CH4	Cal Gas Expiry Date	05/07/2015
Cal Gas C3H8 Conc	370.5 1018.875 ppm CH4	Cal Gas Cylinder #	LL34318
DACS make	CR3000	DACS serial No.	5408
DACS voltage range	0 - 5 volt	DACS channel #	SE 11,12,13

Analyzer make TEI 55I Analyzer serial # 1134650658

	before		after	
Concentration range	0-20 (CH4, NMHC); 0-40 (THC)	ppm	0-20 (CH4, NMHC); 0-40 (THC)	ppm
Air pressure	27.8	PSI	27.8	PSI
Fuel pressure	42.1	PSI	42.1	PSI
Carrier pressure	30.3	PSI	30.3	PSI
CH4 cal factor	5.22		5.22	E ⁻⁴
NMHC cal factor	1.51		1.51	E ⁻⁴
Rt	12.31	Sec	12.31	Sec
Pk Index	23.43		23.43	

CH4 Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
3495	0.00	0.00	0.02	N/A
3495	54.94	15.41	15.23	1.0116
3494	34.97	9.87	9.68	1.0192
3495	14.96	4.24	4.13	1.0286
3495	0.00	0.00	0.02	As Found Zero
3495	54.94	15.41	15.24	As Found Span
Average Correction Factor				1.0198

Calculated value of As Found Response: 15.242 ppm Percent Change of As Found: 1.1%

	Before		After
Calculated slope	0.999910	Calculated slope	1.012353
Calculated intercept	0.022820	Calculated intercept	0.025704

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.05	ppm	0.03	ppm
Auto span	9.11	ppm	9.13	ppm

NMHC Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
3495	0.00	0.00	0.02	N/A
3495	54.94	15.77	15.60	1.0111
3494	34.97	10.10	9.96	1.0141
3495	14.96	4.34	4.22	1.0298
3495	0.00	0.00	0.02	As Found Zero
3495	54.94	15.77	15.57	As Found Span
Average Correction Factor				1.0183

Calculated value of As Found Response: 15.562 ppm Percent Change of As Found: 1.3%

	Before		After
Calculated slope	0.997279	Calculated slope	1.011124
Calculated intercept	0.055328	Calculated intercept	0.021033

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.02	ppm	0.02	ppm
Auto span	12.21	ppm	12.27	ppm

THC Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
3495	0.00	0.00	0.04	N/A
3495	54.94	31.18	30.80	1.0122
3494	34.97	19.96	19.63	1.0171
3495	14.96	8.59	8.32	1.0324
3495	0.00	0.00	0.03	As Found Zero
3495	54.94	31.18	30.80	As Found Span
Average Correction Factor				1.0205

Calculated value of As Found Response: 30.823 ppm Percent Change of As Found: 1.1%

	Before		After
Calculated slope	0.999091	Calculated slope	1.012078
Calculated intercept	0.089062	Calculated intercept	0.058339

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.04	ppm	0.02	ppm
Auto span	21.32	ppm	21.39	ppm

Notes: Hydrogen cylinder replaced
No adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter CH4

Air Monitoring Network PAZA




Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 15, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	13:10:00 PM	End Time (MST)	16:25:00 PM
Analyzer make/model	TEI 55I	Analyzer serial #	1134650658

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.019	N/A	Correlation Coefficient	0.999950
15.408	15.231	1.0116		
9.866	9.680	1.0192	Slope	1.012353
4.243	4.125	1.0286		
			Intercept	0.025704

CH4 Calibration Data



Calibration Summary

Parameter NMHC

Air Monitoring Network PAZA




Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 15, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	13:10:00 PM	End Time (MST)	16:25:00 PM
Analyzer make/model	TEI 55I	Analyzer serial #	1134650658

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.024	N/A	Correlation Coefficient	0.999958
15.768	15.595	1.0111		
10.096	9.956	1.0141		
4.343	4.217	1.0298	Slope	1.011124
			Intercept	0.021033

NMHC Calibration Data



Calibration Summary

Parameter THC

Air Monitoring Network PAZA




Station Information

Calibration Date	October 5, 2018	Previous Calibration	September 15, 2018
Station Number	1	Station Location	Henry Pirker
Start Time (MST)	13:10:00 PM	End Time (MST)	16:25:00 PM
Analyzer make/model	TEI 55I	Analyzer serial #	1134650658


Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.037	N/A	Correlation Coefficient	0.999953
31.177	30.802	1.0122		
19.962	19.626	1.0171	Slope	1.012078
8.586	8.317	1.0324		
			Intercept	0.058339

THC Calibration Data



THC/CH₄/NMHC Calibration



Calibration Report

Parameter SO₂

Air Monitoring Network PAZA



Station Information

Calibration Date	October 19, 2018	Previous Calibration	September 26, 2018
Station Number	3	Station Location	Smokey Heights
Reason:	Routine	Install	Removal
		Other:	
Start Time (MST)	8:50	End Time (MST)	11:40
Barometric Pressure	0.924 ATM	Station Temperature	19.0 Deg C
Calibrator	EnviroNics 6100	Serial Number	3016
Cal Gas Concentration	50.1 ppm	Cal Gas Cert Date	10/10/2020
Correction factor	0.031517	Cal Gas Cylinder #	LL104185
DACS make	CR3000	DACS serial No.	5238
DACS voltage range	0 - 5 volt	DACS channel #	6
	<u>Before</u>		<u>After</u>
Calculated slope	1.003839	Calculated slope	0.997695
Calculated intercept	0.896348	Calculated intercept	0.953357
Analyzer make	Teco 43i	Analyzer serial #	701120009

	before		after	
Concentration range	0 - 500	ppb	0 - 500	ppb
Background	15.7		14.3	
coefficient	0.951		0.951	
Lamp Voltage	944	volts	944	volts
Chamber Temp	45	Deg C	45	Deg C
Perm Gas Temp	45	Deg C	45	Deg C
Pressure	676.5	mm Hg	670.9	mm Hg
Sample Flow	0.452	lpm	0.450	lpm
Lamp Intensity	88	%	88	%

Dilution air flow rate (cc/min)	Corrected gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4995	0.0	0.00	0.2	N/A
4995	39.94	397.42	398.0	0.9985
4995	19.97	199.50	198.3	1.0062
4995	9.99	100.00	98.2	1.0186
4995	0.0	0.00	0.2	As Found Zero
4995	39.94	397.42	393.0	As Found Span
Average Correction Factor				1.0078

Calculated value of As Found Response: 395.174 ppb Percent Change of As Found: 0.6%

	before calibration		after calibration	
Auto zero	1.0	ppb	0.2	ppb
Auto span	213.2	ppb	205.3	ppb

Notes: Slight span adjustment made

Calibration Performed By: Grover Christiansen

Calibration Summary

Parameter SO2
Air Monitoring Network PAZA




Station Information

Calibration Date	October 19, 2018	Previous Calibration	September 26, 2018
Station Number	3	Station Location	Smokey Heights
Start Time (MST)	8:50	End Time (MST)	11:40
Analyzer make/model	Teco 43i	Analyzer serial #	701120009


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.2	N/A	Correlation Coefficient	0.999958
397.4	398.0	0.9985		
199.5	198.3	1.0062		
100.0	98.2	1.0186	Slope	0.997695
			Intercept	0.953357

SO2 Calibration Curve



Smokey Heights SO₂ Calibration



Calibration Report

Parameter TRS

Air Monitoring Network PAZA



Station Information

Calibration Date	October 19, 2018		Previous Calibration	September 26, 2018	
Station Number	3		Station Location	Smokey Heights	
Reason:	Routine	Install	Removal	Other:	
Start Time (MST)	10:55		End Time (MST)	13:22	
Barometric Pressure	0.923	ATM	Station Temperature	19.0	Deg C
Calibrator	EnviroNics 6103		Serial Number	6586	
Cal Gas Conc	13.4	ppm	Cal Gas Expiry Date	02/28/2020	
Correction factor	0.031483		Cal Gas Cylinder #	LL119574	
DACS make	CR3000		DACS serial No.	5238	
DACS voltage range	0 - 5 volt		DACS channel #	5	
	<u>Before</u>			<u>After</u>	
Calculated slope	1.370301		Calculated slope	0.997059	
Calculated intercept	0.854545		Calculated intercept	0.540201	
Analyzer make	TEI Model 43I APSAA		Analyzer serial #	1153630151	

	before		after	
Concentration range	100	ppb	100	ppb
Background	14.9	ppb	15.1	ppb
coefficient	1.075		1.075	
Lamp Voltage	804	volts	804	volts
Chamber Temp	45	Deg C	45	Deg C
Perm Gas Temp	45	Deg C	45	Deg C
Pressure	680.2	mm Hg	676.1	mm Hg
Sample Flow	0.422	lpm	0.420	lpm
Lamp Intensity	91	mv	91	mv

Dilution air flow rate (cc/min)	Corrected gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4995	0.0	0.00	-0.1	N/A
6995	41.94	79.86	79.8	1.0004
6995	20.97	40.05	39.2	1.0214
7995	9.97	16.69	15.9	1.0519
4995	9.97		-0.1	Sox test
4995	0.0	0.00	-0.1	As Found Zero
6995	41.94	79.86	74.6	As Found Span
Average Correction Factor				1.0246

Calculated value of As Found Response: 103.17 ppb Percent Change of As Found: -29.2%

	before calibration		after calibration	
Auto zero	0.3	ppb	-0.1	ppb
Auto span	70.3	ppb	70.9	ppb

Notes: Slight span adjustment made

Calibration Performed By: Grover Christiansen

Calibration Summary

Parameter TRS

Air Monitoring Network PAZA




Station Information

Calibration Date	October 19, 2018	Previous Calibration	September 26, 2018
Station Number	3	Station Location	Smokey Heights
Start Time (MST)	10:55	End Time (MST)	13:22
Analyzer make/model	TEI Model 43I APSAA	Analyzer serial #	1153630151


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	-0.1	N/A	Correlation Coefficient	0.999838
79.9	79.8	1.0004		
40.1	39.2	1.0214		
16.7	15.9	1.0519	Slope	0.997059
			Intercept	0.540201

TRS Calibration Curve



Smokey Heights TRS Calibration



October 19, 2018

Calibration Report

Parameter SO2
Air Monitoring Network PAZA



Station Information

Calibration Date	October 25, 2018	Previous Calibration	September 24, 2018
Station Number	6	Station Location	Valleyview
Reason:	Routine	Install	Removal
Start Time (MST)	10:50	End Time (MST)	12:45
Barometric Pressure	0.929 atm	Station Temperature	22.0 Deg C
Calibrator	Envionics	Serial Number	3016
Cal Gas Concentratio	49.4 ppm	Cal Gas Exp Date	February 21, 2020
Gas Cylinder Num.	EY0000751		
DACS make	CR3000	DACS serial No.	5409
DACS voltage range	0 - 5 volt	DACS channel #	4
	Before		After
DACS Scale High	500	DACS slope	500
DACS Scale Low	0	DACS intercept	0
Calculated slope	1.004279	Calculated slope	0.997555
Calculated intercept	0.287931	Calculated intercept	-0.557414
Analyzer make	TEI Model 43i - APSCB	Analyzer serial #	701120010

	before		after	
Concentration range	0-500	ppb	0-500	ppb
Background	7.1		7.1	
Coefficient	0.831		0.831	
UV Lamp Voltage	794	LPM	794	LPM
Chamber Temp	45	V	45	V
Perm Gas Temp	45	C	45	C
Pressure	673	in Hg	6733	in Hg
Sample Flow	0.454	LPM	0.455	LPM
Lamp Intensity	91	Hz	91	%

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4995	0.00	0.0	0.8	N/A
4995	39.94	391.9	393.4	0.9960
4995	19.97	196.7	197.9	0.9942
4995	9.97	98.4	98.8	0.9959
4995	0.00	0.0	0.8	As found zero
4995	39.94	391.9	393.4	As found span
Average Correction Factor				0.9954

Calculated value of As Found Response: 394.6 ppb Percent Change of As Found: -0.7%

	before calibration		after calibration	
Auto zero	0.6	ppb	0.8	ppb
Auto span	264.7	ppb	265.8	ppb

Notes: No adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter SO2
Air Monitoring Network PAZA




Station Information

Calibration Date	October 25, 2018	Previous Calibration	September 24, 2018
Station Number	6	Station Location	Valleyview
Start Time (MST)	10:50	End Time (MST)	12:45
Analyzer make/model	TEI Model 43i - APSCB	Analyzer serial #	701120010


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.8	N/A	Correlation Coefficient	0.999997
391.9	393.4	0.9960		
196.7	197.9	0.9942		
98.4	98.8	0.9959	Slope	0.997555
			Intercept	-0.557414

SO2 Calibration Curve



SO2 Calibration



Calibration Report

Parameter H2S

Air Monitoring Network PAZA



Station Information

Calibration Date	October 25, 2018	Previous Calibration	September 24, 2018
Station Number	6	Station Location	Valleyview
Reason:	Routine	Install	Removal
		Other:	
Start Time (MST)	9:00	End Time (MST)	11:55
Barometric Pressure	0.929 atm	Station Temperature	22.0 Deg C
Calibrator	EnviroNics 6100	Serial Number	3016
Cal Gas Concentration	9.78 ppm	Cal Gas Expiry Date	19/03/2021
Gas Cert Reference	EY0000943		
DACS make	CR3000	DACS serial No.	5409
DACS voltage range	0 - 5 volt	DACS channel #	9
	Before		After
DACS Scale High	100	DACS slope	100
DACS Scale Low	0	DACS intercept	0
Calculated slope	0.991407	Calculated slope	1.000454
Calculated intercept	0.645592	Calculated intercept	0.023966
Analyzer make	TEI Model 450i-APHAA	Analyzer serial #	1170050144

	before		after	
Concentration range	0 - 100	ppb	0 - 100	ppb
Back Ground	14.7	ppb	14.8	ppb
Coefficient	0.947		00.965	
Lamp Voltage	771	V	769	V
Chamber Temp	45.1	c	45.1	c
Convertor	323.9	c	324.1	c
Pressure	484.20	mm Hg	483.60	mm Hg
Sample Flow	0.804	ccm	0.810	lpm
Lamp Intensity	93	%	90	%

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4995	0.00	0.0	-0.1	N/A
4995	39.94	77.6	77.6	1.0001
4995	19.97	38.9	38.8	1.0050
6995	9.97	13.9	14.0	0.9913
4995	0.00	0.0	-0.1	As found zero
4995	39.94	77.6	75.5	As found span
Average Correction Factor				0.9988

Calculated value of As Found Response: 75.52 ppb Percent Change of As Found: 2.7%

	before calibration		after calibration	
Auto zero	0.1	ppb	0.0	ppb
Auto span	78.3	ppb	79.7	ppb

Notes: Slight span adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter H2S

Air Monitoring Network PAZA




Station Information

Calibration Date	October 25, 2018	Previous Calibration	September 24, 2018
Station Number	6	Station Location	Valleyview
Start Time (MST)	9:00	End Time (MST)	11:55
Analyzer make/model	TEI Model 450i-APHAA	Analyzer serial #	1170050144


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	-0.1	N/A	Correlation Coefficient	0.999985
77.6	77.6	1.0001		
38.9	38.8	1.0050		
13.9	14.0	0.9913	Slope	1.000454
			Intercept	0.023966

H2S Calibration Curve



H2S Calibration



Calibration Report

Parameter SO₂

Air Monitoring Network PAZA



Station Information

Calibration Date	October 24, 2018	Previous Calibration	September 25, 2018
Station Number	1	Station Location	Donnelly
Reason:	Routine	Install	Removal
		Other:	
Start Time (MST)	13:10:00 PM	End Time (MST)	15:30:00 PM
Barometric Pressure	0.934 atm	Station Temperature	21.0 Deg C
Calibrator	EnviroNics 6103	Serial Number	6586
Cal Gas Concentration	49.6 ppm	Cal Gas Expiry Date	October 24, 2020
Correction factor	0.031641	Cal Gas Cylinder #	LL107945
DACS make	CR1000	DACS serial No.	3980
DACS voltage range	0 - 5 volt	DACS channel #	6
	<u>Before</u>		<u>After</u>
Calculated slope	0.984949	Calculated slope	0.995190
Calculated intercept	1.342359	Calculated intercept	-0.994771
Analyzer make	Teco 43i	Analyzer serial #	1207452008

	before		after	
Concentration range	0 - 500	ppb	0 - 500	ppb
Background	6.9		6.8	
coefficient	1.029		1.016	
Lamp Voltage	849	volts	847	volts
Chamber Temp	45	Deg C	45	Deg C
Perm Gas Temp	45	Deg C	45	Deg C
Pressure	678.2	mm Hg	674	mm Hg
Sample Flow	0.432	ccm	0.43	ccm
Lamp Intensity	96	%	97	%

Dilution air flow rate (cc/min)	Corrected gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4995	0.0	0.00	0.8	N/A
4995	39.89	393.0	395.3	0.9941
4995	19.78	195.6	199.0	0.9832
4995	9.90	98.1	99.0	0.9913
4995	0.0	0.0	0.8	As Found Zero
4995	39.99	393.9	399.4	As Found Span
Average Correction Factor				0.9895

Calculated value of As Found Response: 394.021 ppb Percent Change of As Found: 0.0%

	before calibration		after calibration	
Auto zero	0.6	ppb	0.7	ppb
Auto span	245.0	ppb	246.5	ppb

Notes: Span adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter SO2
Air Monitoring Network PAZA




Station Information

Calibration Date	October 24, 2018	Previous Calibration	September 25, 2018
Station Number	1	Station Location	Donnely
Start Time (MST)	13:10:00 PM	End Time (MST)	15:30:00 PM
Analyzer make/model	Teco 43i	Analyzer serial #	1207452008


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.8	N/A	Correlation Coefficient	0.999968
393.0	395.3	0.9941		
195.6	199.0	0.9832		
98.1	99.0	0.9913	Slope	0.995190
			Intercept	-0.994771

SO2 Calibration Curve



SO2 Calibration



October 24, 2018

Calibration Report

Parameter H2S
Air Monitoring Network PAZA



Station Information

Calibration Date	24/10/2018	Previous Calibration	25/09/2018
Station Number	1	Station Location	Donnelly
Reason:	Routine	Install	Removal
Start Time (MST)	11:45	End Time (MST)	14:20:00 PM
Barometric Pressure	0.926 atm	Station Temperature	20.0 Deg C
Calibrator	EnviroNics 6103	Serial Number	6586
Cal Gas Conc	9.78 ppm	Cal Gas Expiry Date	March 19, 2021
Correction factor	0.031477	Cal Gas Cylinder #	LL119574
DACS make	CR1000	DACS serial No.	3980
DACS voltage range	0 - 5 volt	DACS channel #	5
	<u>Before</u>		<u>After</u>
Calculated slope	0.978232	Calculated slope	0.999119
Calculated intercept	-0.012988	Calculated intercept	-0.821425
Analyzer make	Thermo 450i	Analyzer serial #	1207452006

	before		after	
Concentration range	0 - 100	ppb	0 - 100	ppb
Background	10.1	ppb	10.1	ppb
coefficient	1.091		1.091	
Lamp Voltage	767	volts	767	volts
Chamber Temp	45	Deg C	45	Deg C
Perm Gas Temp	45	Deg C	45	Deg C
Pressure	463	mm Hg	469.3	mm Hg
Sample Flow	0.685	lpm	0.650	lpm
Lamp Intensity	91	mv	91	mv

Calibration Data

Dilution air flow rate (cc/min)	Corrected gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4992	0.00	0.00	1.1	N/A
4991	39.92	77.60	78.5	0.9885
4991	19.84	38.72	39.7	0.9760
6970	9.93	13.91	14.4	0.9659
4992	0.00	0.00	1.1	As Found Zero
4991	39.92	77.60	78.5	As Found Span
Average Correction Factor				0.9768

Calculated value of As Found Response: 75.75 ppb Percent Change of As Found: **2.4%**

	before calibration		after calibration	
Auto zero	0.6	ppb	0.8	ppb
Auto span	76.4	ppb	78.0	ppb

Notes: No adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter H2S

Air Monitoring Network PAZA




Station Information

Calibration Date	24/10/2018	Previous Calibration	25/09/2018
Station Number	1	Station Location	Donnely
Start Time (MST)	11:45	End Time (MST)	14:20:00 PM
Analyzer make/model	Thermo 450i	Analyzer serial #	1207452006


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	1.1	N/A	Correlation Coefficient	0.999948
77.6	78.5	0.9885		
38.7	39.7	0.9760		
13.9	14.4	0.9659	Slope	0.999119
			Intercept	-0.821425

TRS Calibration Curve



H2S Calibration



24/10/2018

Calibration Report

Parameter SO2

Air Monitoring Network PAZA



Station Information

Calibration Date	October 23, 2018	Previous Calibration	September 19, 2018
Station Number	10	Station Location	Rycroft
Reason:	Routine	☐ Install	☐ Removal
		☐ Other:	
Start Time (MST)	13:10:00 PM	End Time (MST)	16:20:00 PM
Barometric Pressure	0.919 Atm	Station Temperature	20.0 Deg C
Calibrator	Envionics 6103	Serial Number	6586
Cal Gas Concentratio	49.6 ppm	Cal Gas Expiry Date	10/10/2020
Gas Cert Reference	LL104185		
DACS make	CR3000	DACS serial No.	5407
DACS voltage range	0 - 5 Volt	DACS channel #	2
	<u>Before</u>		<u>After</u>
DACS Scale High	500	DACS slope	500
DACS Scale Low	0	DACS intercept	0
Calculated slope	1.000406	Calculated slope	0.991056
Calculated intercept	0.350038	Calculated intercept	-0.355742
Analyzer make	TEI 43I	Analyzer serial #	1173410005

	before		after	
Concentration range	0-500	ppb	0-500	ppb
Background	15.9		17.3	
Coefficient	0.971		0.997	
UV Lamp Voltage	756	V	756	V
Chamber Temp	45	C	45	C
Perm Gas Temp	45	C	45	C
Pressure	670.1	mm Hg	668.3	mm Hg
Sample Flow	0.444	LPM	0.441	LPM
Lamp Intesity	96	Hz	96	Hz

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4996	0.00	0.0	0.2	N/A
4996	39.96	393.6	397.2	0.9909
4996	19.99	197.7	200.5	0.9860
4996	9.98	98.9	100.0	0.9891
4996	0.00	0.0	1.2	As found zero
4996	39.96	393.6	384.5	As found span
Average Correction Factor				0.9887

Calculated value of As Found Response: 383.790 ppb Percent Change of As Found: 2.5%

	before calibration		after calibration	
Auto zero	0.3	ppb	0.3	ppb
Auto span	195.3	ppb	199.7	ppb

Notes: Zero/span adjustmen made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter SO2
Air Monitoring Network PAZA




Station Information

Calibration Date	October 23, 2018	Previous Calibration	September 19, 2018
Station Number	10	Station Location	Rycroft
Start Time (MST)	13:10:00 PM	End Time (MST)	16:20:00 PM
Analyzer make/model	TEI 43I	Analyzer serial #	1173410005


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.2	N/A	Correlation Coefficient	0.999993
393.6	397.2	0.9909		
197.7	200.5	0.9860		
98.9	100.0	0.9891	Slope	0.991056
			Intercept	-0.355742

SO2 Calibration Curve



SO2 Calibration



Calibration Report



Parameter TRS
Air Monitoring Network PAZA

Station Information

Calibration Date October 17, 2018 Previous Calibration September 20, 2018
Station Name PAZA Rover Station Location Rycroft
Reason: Routine Install Removal Other:

Start Time (MST) 9:50 AM End Time (MST) 12:30
Barometric Pressure 0.919 mbar Station Temperature 21.0 Deg C
Calibrator EnviroNics 6103 Serial Number 6586
Cal Gas Concentratio 9.78 ppm Cal Gas Expiry Date 02/02/2020
Gas Cert Reference LL119574
DACS make CR3000 DACS serial No. 5407
DACS voltage range 0 - 5 Volt DACS channel # 1
Before After
DACS Scale High 100 DACS slope 100
DACS Scale Low 0 DACS intercept 0
Calculated slope 0.987933 Calculated slope 1.010660
Calculated intercept 0.734319 Calculated intercept -0.009781

Analyzer make TEI 43C Analyzer serial #

	before		after	
Concentration range	0 - 100	ppb	0 - 100	ppb
Background	11.5	ppb	11.5	ppb
Coefficient	0.964		0.964	
Lamp Voltage	774	V	775	V
Chamber Temp	45	C	45	C
Perm gas Temp	45	C	45	C
Pressure	673.9	mmHg	673.3	mmHg
Sample Flow	0.385	lpm	0.385	lpm
Lamp Intensity	85	Hz	87	Hz

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)
4996	0.0	0.0	0.2	N/A
4996	39.96	77.6	76.9	1.0098
4998	19.98	38.9	38.5	1.0113
6996	9.97	13.9	13.6	1.0269
			0.1	Sox Test
4996	0.00	0.0	0.2	As found zero
4996	39.96	77.6	76.9	As found span
Average Correction Factor				1.0160

Calculated value of As Found Response: 76.45 ppb Percent Change of As Found: 1.5%

	before calibration		after calibration	
Auto zero	0.2	ppb	0.1	ppb
Auto span	52.2	ppb	53.3	ppb

Notes: No adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter TRS
Air Monitoring Network PAZA




Station Information

Calibration Date	October 17, 2018	Previous Calibration	September 20, 2018
Station Number	PAZA Rover	Station Location	Rycroft
Start Time (MST)	9:50	End Time (MST)	12:30
Analyzer make/model	TEI 43C	Analyzer serial #	0


Calibration Data

Calculated concentration (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.2	N/A	Correlation Coefficient	0.999971
77.6	76.9	1.0098		
38.9	38.5	1.0113		
13.9	13.6	1.0269	Slope	1.010660
			Intercept	-0.009781

TRS Calibration Curve



TRS Calibration



Parameter

Air Monitoring Network

NO_x-NO-NO₂
PAZA



Station Information

Calibration Date	October 23, 2018		Previous Calibration	September 19, 2018	
Station Number	10		Station Location	Rycroft	
Reason:	Routine	Installation	Removal	Other:	
Start Time (MST)	13:05:00 PM		End Time (MST)	17:45:00 PM	
Barometric Pressure	0.919	Atm	Station Temperature	20.0	Deg C
Calibrator	Enviroicsn 6103		Serial Number	6586	
NO Cal Gas Conc	51.9	ppm	Cal Gas Expiry Date	10/24/2020	
NOx Cal Gas Conc	52	ppm	Cal Gas Serial #	LL107945	

DACS Information

DACS make CR3000 DACS serial No. 5407

Parameter		NO ₂	NO _x	NO
Before	Data Slope	1.006542	1.002179	0.999409
	Data Offset	0.452592	1.790533	1.897272
After	Data Slope	1.003572	1.000955	1.000631
	Data Offset	0.502992	0.208167	-0.148580
Channel #		8	6	7
Voltage Range		0 - 5 VDC	0 - 5 VDC	0 - 5 VDC

Analyzer Information

Analyzer make/model	TEI 42i	Analyzer serial #	0701120011	
Test Point	before		after	
Concentration range	0-500	ppb	0-500	ppb
NO offset	6.8	mV	6.7	mV
NOx bkgnd	7.0	mV	7.2	mV
NO coefficient	1.044		1.053	
NOx coefficient	0.995		0.995	
NO2 conv temp	326.8	Deg C	326.7	Deg C
Cooler	-2.7	Deg C	-2.7	Deg C
PMT Volt	-840.6	mV	-840.6	mV
R Cell Press	175.4	in Hg	176.9	in Hg
Sample Flow	0.729	LPM	0.729	LPM

Notes: Zero/span adjustment made

Calibration Report

Parameter **NOX-NO-NO2**
Air Monitoring Network **PAZA**



Station Information

Calibration Date: October 23, 2018 Station Location: Rycroft

Calibration Data

	Dilution flow rate (ccm)	Source gas flow rate (ccm)	Calculated NOx conc (ppb)	Calculated NO conc (ppb)	Calculated NO2 conc (ppb)	Indicated NOx conc (ppb)	Indicated NO conc (ppb)	Indicated NO2 conc (ppb)	NOx Correction factor	NO Correction factor
zero	4996	0.00	0.0	0.0	0.0	-0.1	0.0	0.0	N/A	N/A
1	4996	39.96	412.6	411.8	0.8	412.0	411.6	-1.2	1.0015	1.0006
2	4996	19.99	207.2	206.8	0.4	207.0	207.2	-0.9	1.0010	0.9983
3	4996	9.98	103.7	103.5	0.2	103.1	103.5	-0.2	1.0056	0.9999
AFZ	4996	0.00	0.0	0.0	0.0	0.5	-0.1	0.6	0.0000	0.0000
AFS	4996	39.96	412.6	411.8	0.8	409.1	408.9	-1.3	1.0085	1.0073
Average Correction Factor									1.0027	0.9996

As Found Concentrations: NO_x= 411.4 NO= 410.6 As Found Percent Change NO_x= -0.3% NO= -0.3%

GPT Calibration Data

Dilution Flow 4995 ccm Source Gas Flow 39.94 ccm

O3 Setpoint (ppb)	Indicated NOX high point (ppb)	Indicated NO drop conc (ppb)	Calculated NO2 conc (ppb)	Indicated NOx conc (ppb)	Indicated NO conc (ppb)	Indicated NO2 conc (ppb)	NOx Correction factor	NO Correction factor	NO2 Correction factor	Converter Efficiency
0	-0.1	-0.1	0.0	-0.1	0.0	0.0	N/A	N/A	N/A	N/A
NO point	411.9	411.9	0.0	411.9	411.9	-2.0	0.9999	1.0000	N/A	N/A
300	411.9	136.5	275.4	412.2	136.5	274.2	0.9993	1.0000	1.0043	99.6%
200	411.9	242.5	169.4	412.6	242.5	168.1	0.9983	1.0000	1.0079	99.2%
100	411.9	340.3	71.6	412.3	340.3	70.3	0.9989	1.0000	1.0191	98.1%
Average Correction Factor							0.9988	1.0000	1.0104	99.0%

AIC Data

	Previous calibration				Current calibration			
Parameter	NOx	NO2	NO		NOx	NO2	NO	
Auto zero	0.3	0.4	0.0	ppb	-0.2	-0.3	0.2	ppb
Auto span	192.3	189.9	1.4	ppb	183.3	181.0	1.3	ppb

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter NO₂

Air Monitoring Network PAZA




Station Information

Calibration Date	October 23, 2018	Previous Calibration	September 19, 2018
Station Number	10	Station Location	Rycroft
Start Time (MST)	13:05:00 PM	End Time (MST)	17:45:00 PM
Analyzer make	TEI 42i	Analyzer serial #	0701120011

Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.0	N/A	Correlation Coefficient	0.999982
275.4	274.2	1.0043		
169.4	168.1	1.0079		
71.6	70.3	1.0191	Slope	1.003572
			Intercept	0.502992

NO₂ Calibration Curve



Calibration Summary

Parameter NO_x

Air Monitoring Network PAZA




Station Information

Calibration Date	October 23, 2018	Previous Calibration	September 19, 2018
Station Number	10	Station Location	Rycroft
Start Time (MST)	13:05:00 PM	End Time (MST)	17:45:00 PM
Analyzer make	TEI 42i	Analyzer serial #	0701120011

Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	-0.1	N/A	Correlation Coefficient	0.999999
412.6	412.0	1.0015		
207.2	207.0	1.0010		
103.7	103.1	1.0056	Slope	1.000955
			Intercept	0.208167

NO_x Calibration Curve



Calibration Summary

Parameter NO

Air Monitoring Network PAZA




Station Information

Calibration Date	October 23, 2018	Previous Calibration	September 19, 2018
Station Number	10	Station Location	Rycroft
Start Time (MST)	13:05:00 PM	End Time (MST)	17:45:00 PM
Analyzer make	TEI 42i	Analyzer serial #	0701120011


Calibration Data

Calculated conc (ppb) (Cc)	Indicated concentration (ppb) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	0.0	N/A		
411.8	411.6	1.0006	Correlation Coefficient	0.999998
206.8	207.2	0.9983		
103.5	103.5	0.9999		
			Slope	1.000631
			Intercept	-0.148580

NO Calibration Curve



NO_x Calibration



Calibration Report



Parameter 03
Air Monitoring Network PAZA

Station Information

Calibration Date	October 23 2018	Previous Calibration	September 19 2018
Station Number	10	Station Location	Rycroft
Reason:	☒ Routine	☐ Install	☐ Removal
		☐ Other:	
Start Time (MST)	16:40:00 PM	End Time (MST)	19:10:00 PM
Barometric Pressure	0.919 atm	Station Temperature	20.0 Deg C
Calibrator	Envionics 6103	Serial Number	6586
DACS make	CR3000	DACS serial No.	5407
DACS voltage range	0 - 5 Volts	DACS channel #	6
	Before		After
Calculated slope	1.000937	Calculated slope	1.005291
Calculated intercept	0.273305	Calculated intercept	-0.257758
Analyzer make	TEI Model 49I	Analyzer serial #	49I-115363156

	before		after	
Concentration range	0-500	ppb	0-500	ppb
Offset	-0.8	ppb	-0.8	ppb
Span	0.991		1.010	
Cell A intensity	91706	Hz	91609	Hz
Cell B intensity	79100	Hz	79323	Hz
Pressure	693.80	in Hg	695.60	in Hg
CellA Flow	0.682	ccm	0.683	ccm
Cell B Flow	0.682	cmm	0.683	cmm

Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
4996	0.00	0.0	-0.3	N/A
4996	0.30	275.4	273.8	1.0057
4996	0.20	169.4	169.1	1.0017
4996	0.10	71.6	72.1	0.9934
4996	0.00	0.0	-0.3	As found zero
4996	0.30	275.4	269.6	As found span
Average Correction Factor				1.0003

Calculated value of As Found Response: 270.4 ppm Percent Change of As Found: -1.8%

	before calibration		after calibration	
Auto zero	0.8	ppb	0.9	ppb
Auto span	371.2	ppb	388.3	ppb

Notes: Slight span adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter 03
Air Monitoring Network PAZA




Station Information


Calibration Date	October 23 2018	Previous Calibration	September 19 2018
Station Number	10	Station Location	Rycroft
Start Time (MST)	16:40:00 PM	End Time (MST)	19:10:00 PM
Analyzer make/model	TEI Model 49I	Analyzer serial #	49I-115363156

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.0	-0.3	NA	Correlation Coefficient	0.999978
275.4	273.8	1.0057		
169.4	169.1	1.0017		
71.6	72.1	0.9934	Slope	1.005291
			Intercept	-0.257758

O3 Calibration Curve





October 23 2018

Calibration Report



Parameter CH4 / NMHC / THC

Air Monitoring Network PAZA

Station Information

Calibration Date	October 17, 2018	Previous Calibration	September 18, 2018
Station Number	1	Station Location	Rycroft
Reason:	Routine	Install	Removal Other:

Start Time (MST)	11:40	End Time (MST)	14:15:00 PM
Barometric Pressure	0.930 atm	Station Temperature	21.0 Deg C
Calibrator	EnviroNics 6103	Serial Number	6586
Cal Gas CH4 Conc	404 ppm CH4	Cal Gas Expiry Date	28/03/2014
Cal Gas C3H8 Conc	201 552.75 ppm CH4	Cal Gas Cylinder #	LL34988
DACS make	CR3000	DACS serial No.	5408
DACS voltage range	0 - 5 volt	DACS channel #	SE 12,13,14

Analyzer make TEI 55I-A3PHAA Analyzer serial # 1151980005

	before		after	
Concentration range	0-20 (CH4, NMHC); 0-40 (THC)	ppm	0-20 (CH4, NMHC); 0-40 (THC)	ppm
Air pressure	34.7	PSI	34.5	PSI
Fuel pressure	49.4	PSI	49.5	PSI
Carrier pressure	40.1	PSI	42.2	PSI
CH4 cal factor	7.06582160		7.48918500	E ⁻⁴
NMHC cal factor	1.8934355		1.9680489	E ⁻⁴
CH4 Rt	12.60	Sec	12.60	Sec
CH4 Pk HT	18009.00		17155.00	

CH4 Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
2000	0.00	0.00	0.00	N/A
2000	64.96	12.71	12.77	0.9954
2000	34.97	6.94	6.95	0.9991
2000	15.96	3.20	3.19	1.0013
2000	0.00	0.00	0.00	As Found Zero
2000	64.96	12.71	12.11	As Found Span
Average Correction Factor				0.9986

Calculated value of As Found Response: 12.066 ppm Percent Change of As Found: 5.1%

	Before		After
Calculated slope	0.991757	Calculated slope	0.995263
Calculated intercept	0.055769	Calculated intercept	0.011542

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.02	ppm	0.00	ppm
Auto span	7.91	ppm	8.18	ppm

NMHC Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
2000	0.00	0.00	0.00	N/A
2002	64.96	17.37	17.35	1.0012
2000	34.97	9.50	9.43	1.0076
2000	15.96	4.38	4.24	1.0328
2000	0.00	0.00	0.00	As Found Zero
2000	64.96	17.39	16.65	As Found Span
Average Correction Factor				1.0139

Calculated value of As Found Response: 16.704 ppm Percent Change of As Found: 3.9%

	Before		After
Calculated slope	0.996274	Calculated slope	0.999053
Calculated intercept	0.122504	Calculated intercept	0.065092

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.00	ppm	0.00	ppm
Auto span	10.73	ppm	10.99	ppm

THC Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
2000	0.00	0.00	0.00	N/A
2000	64.96	30.10	30.16	0.9981
2000	34.97	16.44	16.39	1.0034
2000	15.96	7.57	7.46	1.0155
2000	0.00	0.00	0.00	As Found Zero
2000	64.96	30.10	28.79	As Found Span
Average Correction Factor				1.0057

Calculated value of As Found Response: 28.791 ppm Percent Change of As Found: 4.3%

	Before		After
Calculated slope	0.993982	Calculated slope	0.997022
Calculated intercept	0.172336	Calculated intercept	0.068739

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.02	ppm	0.00	ppm
Auto span	18.64	ppm	19.17	ppm

Notes: Span adjustment made
Span & hydrogen cylinders replaced on Oct 16 2018

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter CH4
Air Monitoring Network PAZA




Station Information

Calibration Date	October 17, 2018	Previous Calibration	September 18, 2018
Station Number	1	Station Location	Rycroft
Start Time (MST)	11:40	End Time (MST)	14:15:00 PM
Analyzer make/model	TEI 55I-A3PHAA	Analyzer serial #	1151980005

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.001	N/A	Correlation Coefficient	0.999994
12.709	12.768	0.9954		
6.943	6.949	0.9991		
3.198	3.194	1.0013	Slope	0.995263
			Intercept	0.011542

CH4 Calibration Data



Calibration Summary

Parameter NMHC
Air Monitoring Network PAZA




Station Information

Calibration Date	October 17, 2018	Previous Calibration	September 18, 2018
Station Number	1	Station Location	Rycroft
Start Time (MST)	11:40	End Time (MST)	14:15:00 PM
Analyzer make/model	TEI 55I-A3PHAA	Analyzer serial #	1151980005

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.001	N/A	Correlation Coefficient	0.999931
17.372	17.351	1.0012		
9.499	9.427	1.0076		
4.376	4.237	1.0328	Slope	0.999053
			Intercept	0.065092

NMHC Calibration Data



Calibration Summary

Parameter THC
Air Monitoring Network PAZA




Station Information

Calibration Date	October 17, 2018	Previous Calibration	September 18, 2018
Station Number	1	Station Location	Rycroft
Start Time (MST)	11:40	End Time (MST)	14:15:00 PM
Analyzer make/model	TEI 55I-A3PHAA	Analyzer serial #	1151980005


Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.000	N/A	Correlation Coefficient	0.999976
30.098	30.156	0.9981		
16.441	16.385	1.0034		
7.574	7.459	1.0155	Slope	0.997022
			Intercept	0.068739

THC Calibration Data



THC/CH₄/NMHC Calibration



Calibration Report



Parameter CH4 / NMHC / THC

Air Monitoring Network PAZA

Station Information

Calibration Date	October 23, 2018	Previous Calibration	October 17, 2018
Station Number	1	Station Location	Rycroft
Reason:	Routine	Install	Removal Other:

Start Time (MST)	11:25	End Time (MST)	14:25:00 PM
Barometric Pressure	0.930 atm	Station Temperature	21.0 Deg C
Calibrator	EnviroNics 6103	Serial Number	6586
Cal Gas CH4 Conc	404 ppm CH4	Cal Gas Expiry Date	28/03/2014
Cal Gas C3H8 Conc	201 552.75 ppm CH4	Cal Gas Cylinder #	LL34988
DACS make	CR3000	DACS serial No.	5408
DACS voltage range	0 - 5 volt	DACS channel #	SE 12,13,14

Analyzer make TEI 55I-A3PHAA Analyzer serial # 1151980005

	before		after	
Concentration range	0-20 (CH4, NMHC); 0-40 (THC)	ppm	0-20 (CH4, NMHC); 0-40 (THC)	ppm
Air pressure	34.5	PSI	34.6	PSI
Fuel pressure	49.5	PSI	49.5	PSI
Carrier pressure	42.2	PSI	42.2	PSI
CH4 cal factor	7.48918500		7.59986650	E ⁻⁴
NMHC cal factor	1.9680489		2.0225895	E ⁻⁴
CH4 Rt	12.60	Sec	12.60	Sec
CH4 Pk HT	17155		16724	

CH4 Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
2000	0.00	0.00	0.02	N/A
2000	64.96	12.71	12.83	0.9907
2000	34.97	6.94	6.89	1.0075
2000	15.96	3.20	3.23	0.9912
2000	0.00	0.00	0.02	As Found Zero
2000	64.93	12.70	12.29	As Found Span
Average Correction Factor				0.9965

Calculated value of As Found Response: 12.227 ppm Percent Change of As Found: 3.7%

	Before		After
Calculated slope	0.991757	Calculated slope	0.993464
Calculated intercept	0.055769	Calculated intercept	0.008284

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.02	ppm	0.01	ppm
Auto span	7.91	ppm	8.29	ppm

NMHC Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
2000	0.00	0.00	0.00	N/A
2002	64.96	17.37	17.41	0.9981
2000	34.97	9.50	9.30	1.0213
2000	15.96	4.38	4.25	1.0286
2000	0.00	0.00	0.00	As Found Zero
2000	64.93	17.38	16.87	As Found Span
Average Correction Factor				1.0160

Calculated value of As Found Response: 16.926 ppm Percent Change of As Found: 2.6%

	Before		After
Calculated slope	0.996274	Calculated slope	0.997410
Calculated intercept	0.122504	Calculated intercept	0.091253

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.00	ppm	0.00	ppm
Auto span	10.73	ppm	11.04	ppm

THC Calibration Data

Dilution air flow rate (cc/min)	Source gas flow rate (cc/min)	Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)
2000	0.00	0.00	0.02	N/A
2000	64.96	30.10	30.27	0.9943
2000	34.97	16.44	16.20	1.0149
2000	15.96	7.57	7.50	1.0095
2000	0.00	0.00	0.02	As Found Zero
2000	64.93	30.08	29.20	As Found Span
Average Correction Factor				1.0062

Calculated value of As Found Response: 29.176 ppm Percent Change of As Found: 3.0%

	Before		After
Calculated slope	0.993982	Calculated slope	0.995255
Calculated intercept	0.172336	Calculated intercept	0.094392

Final Zero/Span Data

	before calibration		after calibration	
Auto zero	0.02	ppm	0.01	ppm
Auto span	18.64	ppm	19.33	ppm

Notes: Hydrogen & nitrogen cylinders replaced
Span adjustment made

Calibration Performed By: Dmytro Dolotii

Calibration Summary

Parameter CH4
Air Monitoring Network PAZA




Station Information

Calibration Date	October 23, 2018	Previous Calibration	October 17, 2018
Station Number	1	Station Location	Rycroft
Start Time (MST)	11:25	End Time (MST)	14:25:00 PM
Analyzer make/model	TEI 55I-A3PHAA	Analyzer serial #	1151980005

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.021	N/A	Correlation Coefficient	0.999880
12.709	12.828	0.9907		
6.943	6.891	1.0075		
3.198	3.227	0.9912	Slope	0.993464
			Intercept	0.008284

CH4 Calibration Data



Calibration Summary

Parameter NMHC
Air Monitoring Network PAZA




Station Information

Calibration Date	October 23, 2018	Previous Calibration	October 17, 2018
Station Number	1	Station Location	Rycroft
Start Time (MST)	11:25	End Time (MST)	14:25:00 PM
Analyzer make/model	TEI 55I-A3PHAA	Analyzer serial #	1151980005

Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.001	N/A	Correlation Coefficient	0.999798
17.372	17.406	0.9981		
9.499	9.301	1.0213		
4.376	4.254	1.0286	Slope	0.997410
			Intercept	0.091253

NMHC Calibration Data



Calibration Summary

Parameter THC
Air Monitoring Network PAZA




Station Information

Calibration Date	October 23, 2018	Previous Calibration	October 17, 2018
Station Number	1	Station Location	Rycroft
Start Time (MST)	11:25	End Time (MST)	14:25:00 PM
Analyzer make/model	TEI 55I-A3PHAA	Analyzer serial #	1151980005


Calibration Data

Calculated concentration (ppm) (Cc)	Indicated concentration (ppm) (Ic)	Correction factor (Cc/Ic)	Statistical Evaluation	
0.000	0.019	N/A	Correlation Coefficient	0.999845
30.098	30.269	0.9943		
16.441	16.201	1.0149		
7.574	7.503	1.0095	Slope	0.995255
			Intercept	0.094392

THC Calibration Data



THC/CH₄/NMHC Calibration



AB TEOM PM2.5 Calibration



STATION: Rover-Rycroft
 LOCATION: PAZA - Grande Prairie

OPERATOR: Dmytro Dolotii
 DATE: 23-Oct-18

MONITOR INFO / PARAMETER VALUES:

Make/Model	TEOM AB
Configuration	PM2.5
Serial Number	140AB245710304
Site Number	10
Inlet Type	PM 10 / SCC
FAdj. Main Setting	1.000
FAdj. Aux. Setting	0.990
T-Case Indicated / Set Point	40/40
T-Air Indicated / Set Point	40/40
T-Cap Indicated / Set Point	40/40
Splitter Assembly Alignment (cm)	15.5
(vs. specified depth of 15.5 cm from top of flow tube to top of concentric 1/2 in. tube)	

RECENT CALIBRATION AND AUDIT HISTORY

Previous Audit	
Previous Calibration	14-Aug-18

PUMP CAPACITY CHECK *	PASS
-----------------------	------

* capacity test or pump on timed test utilized to verify pump integrity

"FAIL" indicates that pump requires service.

LEAK CHECK	Indicated Flow (lpm)	
	Main	Auxiliary
PUMP ON	0.130	0.200
PUMP OFF	0.100	0.160
NET	0.030	0.040
LIMITS	<0.15	<0.60

	Ambient Temp. (° C)	Ambient Pres. (atm)	Ko *	Bypass flow (lpm)	Sample flow (lpm)
SET POINT (S)	na	na	12247	13.67	3.000
INDICATED (I)	9.8	0.923		13.69	3.000
MEASURED (AF)	9.2	0.923		13.41	3.280
MEASURED (M)	9.2	0.923	12276	13.68	3.140
DIFFERENCE (M-I)	-0.6	0.000	0.2%	0.01	0.14
LIMITS	± 2 ° C	± 0.005 atm	± 2.5 %	± 1.0 L/min	± 0.2 L/min

As Found Data
 Adjusted Data

Ko Audit Filter data Weight: 0.09687 Serial #: CVK 5445

COMMENTS: Pass
 Full audit was performed.
 No adjustment required
 Cleaned the heads

Sample Head Inspection/Cleaning: Large In Line Filter Inspection & Or Cleaning: